

LABORATORY REPORT

If you have any questions concerning this report, please do not hesitate to call us at (800) 332-4345 or (574) 233-4777.

This report may not be reproduced, except in full, without written approval from EEA.

STATE CERTIFICATION LIST

State	Certification	State	Certification
Alabama	40700	Missouri	880
Alaska	IN00035	Montana	CERT0026
Arizona	AZ0432	Nebraska	NE-OS-05-04
Arkansas	IN00035	Nevada	IN00035
California	2920	New Hampshire*	2124
Colorado	IN00035	New Jersey*	IN598
Colorado Radiochemistry	IN00035	New Mexico	IN00035
Connecticut	PH-0132	New York*	11398
Delaware	IN035	North Carolina	18700
Florida*	E87775	North Dakota	R-035
Georgia	929	Ohio	87775
Hawaii	IN035	Oklahoma	D9508
Idaho	IN00035	Oregon (Primary AB)*	4074
Illinois*	200001	Pennsylvania*	68-00466
Illinois Microbiology	17767	Puerto Rico	IN00035
Illinois Radiochemistry	IN00035	Rhode Island	LAO00343
Indiana Chemistry	C-71-01	South Carolina	95005
Indiana Microbiology	M-76-07	South Dakota	IN00035
Iowa	098	Tennessee	TN02973
Kansas*	E-10233	Texas*	T104704187-18-12
Kentucky	90056	Texas/TCEQ	TX207
Louisiana*	LA014	Utah*	IN00035
Maine	IN00035	Vermont	VT-8775
Maryland	209	Virginia*	460275
Massachusetts	M-IN035	Washington	C837
Michigan	9926	West Virginia	9927 C
Minnesota*	018-999-338	Wisconsin	999766900
Mississippi	IN035	Wyoming	IN035
EPA	IN00035		

*NELAP/TNI Recognized Accreditation Bodies



LABORATORY CASE NARRATIVE

Client: Charleston Water System

Report #: 451915QC

All method QC was within acceptance limits.

Note: This report may not be reproduced, except in full, without written approval from EEA.

		06/22/2019
Authorized Signature	Title	Date

Page 1 of 1

110 South Hill Street
 South Bend, IN 46617
 Tel: (574) 233-4777
 Fax: (574) 233-8207
 1 800 332 4345

Laboratory Report

Client: Charleston Water System
 Attn: Becky Thames
 1104 Hanahan Road
 Hanahan, SC 29410

Report: 451915
 Priority: Standard Written
 Status: Final
 PWS ID: SC1010001
 Lab ID #: 95005

Sample Information					
EEA ID #	Client ID	Method	Collected Date / Time	Collected By:	Received Date / Time
4291756	Finished	522	05/13/19 11:59	Client	05/14/19 10:00
4291757	Finished	515.3	05/13/19 12:15	Client	05/14/19 10:00
4291758	Finished	521	05/13/19 12:04	Client	05/14/19 10:00
4291759	Finished	525.2	05/13/19 11:55	Client	05/14/19 10:00
4291760	Finished	549.2	05/13/19 12:01	Client	05/14/19 10:00
4291761	Finished	1613	05/13/19 12:09	Client	05/14/19 10:00
4291762	Finished	SRL 524M-TCP	05/13/19 12:24	Client	05/14/19 10:00
4291763	Finished	524.2	05/13/19 12:27	Client	05/14/19 10:00
4291764	Finished	L306	05/13/19 12:13	Client	05/14/19 10:00
4291765	Finished	531.2 LL	05/13/19 12:19	Client	05/14/19 10:00
4291766	Finished	LC/MS/MS	05/13/19 12:17	Client	05/14/19 10:00
4291767	Finished	L520	05/13/19 12:20	Client	05/14/19 10:00
4291768	Finished	556	05/13/19 12:21	Client	05/14/19 10:00
4291769	Finished	L402	05/13/19 12:32	Client	05/14/19 10:00
4291770	Finished	200.8	05/13/19 12:11	Client	05/14/19 10:00
4291771	Finished	331.0	05/13/19 12:34	Client	05/14/19 10:00
4291772	Finished	218.6	05/13/19 12:52	Client	05/14/19 10:00
4292046	Finished	8330A	05/13/19 11:17	Client	05/14/19 10:00
4292047	Finished	527	05/13/19 11:13	Client	05/14/19 10:00
4292048	Finished	545	05/13/19 11:05	Client	05/14/19 10:00
4292049	Finished	544	05/13/19 11:09	Client	05/14/19 10:00

Report Summary

Note: Method 1613 analysis was performed by Pace Analytical Services, Minneapolis, MN.

Note: Method 8330A analysis was performed by Eurofins Lancaster Laboratories, Lancaster, PA.

Note: Method EDC/PPCP/Hormone analysis was performed by Eurofins Eaton Analytical, Monrovia, CA.

Note: In the Method L402 analysis, PFMOPrA is not reportable in the sample submitted due to matrix interference.

Client Name: Charleston Water System

Report #: 451915

Detailed quantitative results are presented on the following pages. The results presented relate only to the samples provided for analysis.

We appreciate the opportunity to provide you with this analysis. If you have any questions concerning this report, please do not hesitate to call Pat Muff at (574) 233-4777.

Note: This report may not be reproduced, except in full, without written approval from EEA.



ASM

Authorized Signature

Title

06/22/2019

Date

Client Name: Charleston Water System

Report #: 451915

Sampling Point: Finished

PWS ID: SC1010001

General Chemistry									
Analyte ID #	Analyte	Method	Reg Limit	MRL†	Result	Units	Preparation Date	Analyzed Date	EEA ID #
1080	Chromium, Hexavalent	218.6	---	0.02	0.06	ug/L	---	05/23/19 22:46	4291772
1039	Perchlorate	331.0	---	0.05	0.12	ug/L	---	05/16/19 03:03	4291771

Metals									
Analyte ID #	Analyte	Method	Reg Limit	MRL†	Result	Units	Preparation Date	Analyzed Date	EEA ID #
1002	Aluminum	200.8	50 ^	2.0	35	ug/L	---	05/21/19 11:55	4291770
1074	Antimony	200.8	6 *	1.0	< 1.0	ug/L	---	05/21/19 11:55	4291770
1005	Arsenic	200.8	10 *	1.0	< 1.0	ug/L	---	05/21/19 11:55	4291770
1010	Barium	200.8	2000 *	2.0	17	ug/L	---	05/21/19 11:55	4291770
1075	Beryllium	200.8	4 *	0.3	< 0.3	ug/L	---	05/21/19 11:55	4291770
1079	Boron	200.8	---	5.0	22	ug/L	---	05/21/19 11:55	4291770
1015	Cadmium	200.8	5 *	1.0	< 1.0	ug/L	---	05/21/19 11:55	4291770
1020	Chromium	200.8	100 *	0.9	< 0.9	ug/L	---	05/21/19 11:55	4291770
1081	Cobalt	200.8	---	2.0	< 2.0	ug/L	---	05/21/19 11:55	4291770
1022	Copper	200.8	1300 !	1.0	< 1.0	ug/L	---	05/21/19 11:55	4291770
1030	Lead	200.8	15 !	1.0	< 1.0	ug/L	---	05/21/19 11:55	4291770
1083	Lithium	200.8	---	2.0	< 2.0	ug/L	---	05/21/19 11:55	4291770
1032	Manganese	200.8	50 ^	2.0	9.6	ug/L	---	05/21/19 11:55	4291770
1084	Molybdenum	200.8	---	2.0	< 2.0	ug/L	---	05/21/19 11:55	4291770
1036	Nickel	200.8	---	1.0	< 1.0	ug/L	---	05/21/19 11:55	4291770
1045	Selenium	200.8	50 *	2.0	< 2.0	ug/L	---	05/21/19 11:55	4291770
1050	Silver	200.8	100 ^	2.0	< 2.0	ug/L	---	05/21/19 11:55	4291770
1051	Strontium	200.8	---	2.0	53	ug/L	---	05/21/19 11:55	4291770
1085	Thallium	200.8	2 *	0.3	< 0.3	ug/L	---	05/21/19 11:55	4291770
7440-29-1	Thorium	200.8	---	5.0	< 5.0	ug/L	---	05/21/19 11:55	4291770
1086	Tin	200.8	---	2.0	< 2.0	ug/L	---	05/21/19 11:55	4291770
1087	Titanium	200.8	---	5.0	< 5.0	ug/L	---	05/21/19 11:55	4291770
4006	Uranium	200.8	30 *	1.0	< 1.0	ug/L	---	05/21/19 11:55	4291770
1088	Vanadium	200.8	---	2.0	< 2.0	ug/L	---	05/21/19 11:55	4291770
1095	Zinc	200.8	5000 ^	5.0	< 5.0	ug/L	---	05/21/19 11:55	4291770

Semi-volatile Organic Chemicals									
Analyte ID #	Analyte	Method	Reg Limit	MRL†	Result	Units	Preparation Date	Analyzed	EEA ID #
2108	Acifluorfen	515.3	---	1.0	< 1.0	ug/L	05/15/19 08:00	05/21/19 04:49	4291757
2625	Bentazon	515.3	---	0.5	< 0.5	ug/L	05/15/19 08:00	05/21/19 04:49	4291757
2205	Chloramben	515.3	---	2.0	< 2.0	ug/L	05/15/19 08:00	05/21/19 04:49	4291757
2105	2,4-D	515.3	70 *	0.1	< 0.1	ug/L	05/15/19 08:00	05/21/19 04:49	4291757
2031	Dalapon	515.3	200 *	1.0	< 1.0	ug/L	05/15/19 08:00	05/21/19 04:49	4291757
2106	2,4-DB	515.3	---	2.0	< 2.0	ug/L	05/15/19 08:00	05/21/19 04:49	4291757
2108	DCPA acid metabolites	515.3	---	0.5	< 0.5	ug/L	05/15/19 08:00	05/21/19 04:49	4291757
2440	Dicamba	515.3	---	0.1	< 0.1	ug/L	05/15/19 08:00	05/21/19 04:49	4291757
51-36-5	3,5-Dichlorobenzoic acid	515.3	---	0.5	< 0.5	ug/L	05/15/19 08:00	05/21/19 04:49	4291757
2206	Dichlorprop	515.3	---	2.0	< 2.0	ug/L	05/15/19 08:00	05/21/19 04:49	4291757
2041	Dinoseb	515.3	7 *	0.1	< 0.1	ug/L	05/15/19 08:00	05/21/19 04:49	4291757

2113	MCPA	515.3	---	0.5	< 0.5	ug/L	05/15/19 08:00	05/21/19 19:55	4291757
2112	Mecoprop	515.3	---	0.5	< 0.5	ug/L	05/15/19 08:00	05/21/19 19:55	4291757
2326	Pentachlorophenol	515.3	1 *	0.04	< 0.04	ug/L	05/15/19 08:00	05/21/19 04:49	4291757
2040	Picloram	515.3	500 *	0.1	< 0.1	ug/L	05/15/19 08:00	05/21/19 04:49	4291757
2111	2,4,5-T	515.3	---	0.5	< 0.5	ug/L	05/15/19 08:00	05/21/19 04:49	4291757
2110	2,4,5-TP (Silvex)	515.3	50 *	0.1	< 0.1	ug/L	05/15/19 08:00	05/21/19 04:49	4291757
2107	Triclopyr	515.3	---	0.5	< 0.5	ug/L	05/15/19 08:00	05/21/19 04:49	4291757
924-16-3	N-Nitrosodi-N-butylamine (NDBA)	521	---	2.0	< 2.0	ng/L	05/15/19 09:30	05/22/19 17:16	4291758
2316	N-Nitrosodi-N-propylamine (NDPA)	521	---	2.0	< 2.0	ng/L	05/15/19 09:30	05/22/19 17:16	4291758
55-18-5	N-Nitrosodiethylamine (NDEA)	521	---	2.0	< 2.0	ng/L	05/15/19 09:30	05/22/19 17:16	4291758
2314	N-Nitrosodimethylamine (NDMA)	521	---	2.0	5.1	ng/L	05/15/19 09:30	05/22/19 17:16	4291758
86-30-6	N-Nitrosodiphenylamine (NDPhA)	521	---	20	< 20	ng/L	05/15/19 09:30	05/22/19 17:16	4291758
10595-95-6	N-Nitrosomethylethylamine (NMEA)	521	---	2.0	< 2.0	ng/L	05/15/19 09:30	05/22/19 17:16	4291758
59-89-2	N-Nitrosomorpholine (NMOR)	521	---	2.0	< 2.0	ng/L	05/15/19 09:30	05/22/19 17:16	4291758
100-75-4	N-Nitrosopiperidine (NPIP)	521	---	2.0	< 2.0	ng/L	05/15/19 09:30	05/22/19 17:16	4291758
930-55-2	N-Nitrosopyrrolidine (NPYR)	521	---	2.0	< 2.0	ng/L	05/15/19 09:30	05/22/19 17:16	4291758
2261	Acenaphthene	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2260	Acenaphthylene	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2027	Acetochlor	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2051	Alachlor	525.2	2 *	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2356	Aldrin	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
834-12-8	Ametryn	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
101-05-3	Anilazine	525.2	---	1.0	< 1.0	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2280	Anthracene	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
3244-90-4	Aspon	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
1610-17-9	Atraton	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2050	Atrazine	525.2	3 *	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2642-71-9	Azinphos-ethyl	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2059	Azinphos-methyl	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
22781-23-3	Bendiocarb	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2207	Benfluralin	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2300	Benzo(a)anthracene	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2302	Benzo(b)fluoranthene	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2304	Benzo(k)fluoranthene	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2312	Benzo(g,h,i)perylene	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2306	Benzo(a)pyrene	525.2	0.2 *	0.02	< 0.02	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2348	alpha-Hexachlorocyclohexane	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2350	beta-BHC	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2354	delta-BHC	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2010	gamma-BHC (Lindane)	525.2	0.2 *	0.02	< 0.02	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2635	Bolstar	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2098	Bromacil	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2076	Butachlor	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2053	Butylate	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2294	Butylbenzylphthalate	525.2	---	1.0	< 1.0	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2061	Carbophenothion	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
5234-68-4	Carboxin	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
5103-71-9	alpha-Chlordane	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
5103-74-2	gamma-Chlordane	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
470-90-6	Chlorfenvinphos	525.2	---	5.0	< 5.0	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
510-15-6	Chlorobenzilate	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2051-60-7	2-Chlorobiphenyl	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2675-77-6	Chloroneb	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759

5836-10-2	Chloropropylate	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2213	Chlorothalonil	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
101-21-3	Chlorpropham	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2057	Chlorpyrifos	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
5598-13-0	Chlorpyrifos methyl	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2296	Chrysene	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
81777-89-1	Clomazone	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
1702-17-6	Clopyralid	525.2	---	10	< 10	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
56-72-4	Coumaphos	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
7700-17-6	Crotoxyphos	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2054	Cyanazine	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
1134-23-2	Cycloate	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2099	DCPA	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2071	4,4'-DDD	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2009	4,4'-DDE	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2075	4,4'-DDT	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
298-03-3	Demeton O	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
126-75-0	Demeton S	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2006	Desethylatrazine	525.2	---	1.0	< 1.0	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2007	Desisopropylatrazine	525.2	---	1.0	< 1.0	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2056	Diazinon	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2310	Dibenzo(a,h)anthracene	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2290	Di-n-butylphthalate	525.2	---	2.0	< 2.0	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
1194-65-6	Dichlobenil	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
97-17-6	Dichlofenthion	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2933	Dichloran	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
16605-91-7	2,3-Dichlorobiphenyl	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
62-73-7	Dichlorvos	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
141-66-2	Dicrotophos	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2070	Dieldrin	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2035	Di(2-ethylhexyl)adipate	525.2	400 *	0.6	< 0.6	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2039	Di(2-ethylhexyl)phthalate	525.2	6 *	0.6	< 0.6	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2284	Diethylphthalate	525.2	---	1.0	< 1.0	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2221	Dimethoate	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2282	Dimethylphthalate	525.2	---	1.0	< 1.0	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2270	2,4-Dinitrotoluene	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2266	2,6-Dinitrotoluene	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2223	Di-n-octylphthalate	525.2	---	2.0	< 2.0	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
78-34-2	Dioxathion A	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
78-34-2	Dioxathion B	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
957-51-7	Diphenamid	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2102	Disulfoton	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2497-06-5	Disulfoton sulfone	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2497-07-6	Disulfoton sulfoxide	525.2	---	10	< 10	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2068	Endosulfan I	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2072	Endosulfan II	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2074	Endosulfan sulfate	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2005	Endrin	525.2	2 *	0.01	< 0.01	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2372	Endrin aldehyde	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2104-64-5	EPN	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2052	EPTC	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
112-84-5	Erucylamide	525.2	---	5.0	< 5.0	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
66230-04-4	Esfenvalerate	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759

2252	Ethalfuralin	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2062	Ethion	525.2	---	5.0	< 5.0	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
26225-79-6	Ethofumesate	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2267	Ethoprop	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2593-15-9	Etridiazole	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
52-85-7	Famphur	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
22224-92-6	Fenamiphos	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
60168-88-9	Fenarimol	525.2	---	1.0	< 1.0	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
122-14-5	Fenitrothion	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
66441-23-4	Fenoxaprop-ethyl	525.2	---	1.0	< 1.0	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
115-90-2	Fensulfothion	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
55-38-9	Fenthion	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
69806-50-4	Fluazifop-butyl	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
33245-39-5	Fluchloralin	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2164-17-2	Fluometuron	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2286	Fluoranthene	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2264	Fluorene	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
59756-60-4	Fluridone	525.2	---	1.0	< 1.0	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2104	Fonofos	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2065	Heptachlor	525.2	0.4 *	0.04	< 0.04	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2067	Heptachlor epoxide	525.2	0.2 *	0.02	< 0.02	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
52663-71-5	2,2',3,3',4,4',6-Heptachlorobiphenyl	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2274	Hexachlorobenzene	525.2	1 *	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
60145-22-4	2,2',4,4',5,6'-Hexachlorobiphenyl	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2042	Hexachlorocyclopentadiene	525.2	50 *	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
51235-04-2	Hexazinone	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2308	Indeno(1,2,3-cd)pyrene	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
36734-19-7	Iprodione	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2060	Isofenphos	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2262	Isophorone	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
21609-90-5	Leptophos	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2058	Malathion	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
57837-19-1	Metalaxyl	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2015	Methoxychlor	525.2	40 *	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
90-12-0	1-Methylnaphthalene	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2241	2-Methylnaphthalene	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
950-35-6	Methyl paraoxon	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
298-00-0	Methyl parathion	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2045	Metolachlor	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2595	Metribuzin	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
74223-64-6	Metsulfuron-methyl	525.2	---	10	< 10	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2073	Mevinphos	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
113-48-4	MGK 264 isomer a	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
113-48-4	MGK 264 isomer b	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
136-45-8	MGK 326	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2932	Mirex	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2626	Molinate	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
6923-22-4	Monocrotophos	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
300-76-5	Naled	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2248	Naphthalene	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
15299-99-7	Napropamide	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
5103-73-1	cis-Nonachlor	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2273	trans-Nonachlor	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759

27314-13-2	Norflurazon	525.2	---	1.0	< 1.0	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
40186-71-8	2,2',3,3',4,5',6,6'-Octachlorobiphenyl	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
19044-88-3	Oryzalin	525.2	---	10	< 10	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
19666-30-9	Oxadiazon	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
27304-13-8	Oxychlor dane	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
42874-03-3	Oxyfluorfen	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2064	Parathion	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
1114-71-2	Pebulate	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2255	Pendimethalin	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
608-93-5	Pentachlorobenzene	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
60233-25-2	2,2',3',4,6-Pentachlorobiphenyl	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2934	Pentachloronitrobenzene	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2326	Pentachlorophenol	525.2	---	1.0	< 1.0	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2114	Permethrin, cis & trans	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2278	Phenanthrene	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2605	Phorate	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
732-11-6	Phosmet	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
297-99-4	E-Phosphamidon	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
23783-98-4	Z-Phosphamidon	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
26399-36-0	Profluralin	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2029	Prometon	525.2	---	1.0	< 1.0	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
7287-19-6	Prometryn	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
23950-58-5	Pronamide	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2077	Propachlor	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2289	Propanil	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2256	Propazine	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
60207-90-1	Propiconazole isomer a	525.2	---	5.0	< 5.0	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
60207-90-1	Propiconazole isomer b	525.2	---	5.0	< 5.0	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
34643-46-4	Prothiofos	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2288	Pyrene	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2037	Simazine	525.2	4 *	0.07	< 0.07	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
1014-70-6	Simetryn	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
22248-79-9	Stirofos	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
3689-24-5	Sulfotep	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2271	Tebuthiuron	525.2	---	10	< 10	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
107-49-3	TEPP	525.2	---	1.0	< 1.0	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2272	Terbacil	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2545	Terbufos	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
886-50-0	Terbutryn	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
95-94-3	1,2,4,5-Tetrachlorobenzene	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2437-79-8	2,2',4,4'-Tetrachlorobiphenyl	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
148-79-8	Thiabendazole	525.2	---	10	< 10	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2627	Thiobencarb	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
297-97-2	Thionazin	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
43121-43-3	Triadimefon	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
78-48-8	Tribufos	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
15862-07-4	2,4,5-Trichlorobiphenyl	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
327-98-0	Trichloronate	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
41814-78-2	Tricyclazole	525.2	---	1.0	< 1.0	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2055	Trifluralin	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
1929-77-7	Vernolate	525.2	---	0.1	< 0.1	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
50471-44-8	Vinclozolin	525.2	---	0.5	< 0.5	ug/L	05/21/19 08:45	05/22/19 22:58	4291759
2050	Atrazine	527	---	0.5	< 0.5	ug/L	05/21/19 08:55	05/23/19 15:23	4292047

82657-04-3	Bifenthrin	527	---	0.5	< 0.5	ug/L	05/21/19 08:55	05/23/19 15:23	4292047
2098	Bromacil	527	---	0.5	< 0.5	ug/L	05/21/19 08:55	05/23/19 15:23	4292047
2057	Chlorpyrifos	527	---	0.5	< 0.5	ug/L	05/21/19 08:55	05/23/19 15:23	4292047
2221	Dimethoate	527	---	0.7	< 0.7	ug/L	05/21/19 08:55	05/23/19 15:23	4292047
28434-00-6	Esbiol	527	---	0.5	< 0.5	ug/L	05/21/19 08:55	05/23/19 15:23	4292047
66230-04-4	Esfenvalerate	527	---	0.5	< 0.5	ug/L	05/21/19 08:55	05/23/19 15:23	4292047
51630-58-1	Fenvalerate	527	---	0.5	< 0.5	ug/L	05/21/19 08:55	05/23/19 15:23	4292047
59080-40-9	2,2',4,4',5,5'-Hexabromobiphenyl (HBB)	527	---	0.7	< 0.7	ug/L	05/21/19 08:55	05/23/19 15:23	4292047
68631-49-2	2,2',4,4',5,5'-Hexabromodiphenyl ether (BDE-153)	527	---	0.8	< 0.8	ug/L	05/21/19 08:55	05/23/19 15:23	4292047
51235-04-2	Hexazinone	527	---	0.5	< 0.5	ug/L	05/21/19 08:55	05/23/19 15:23	4292047
2930	Kepone	527	---	0.5	< 0.5	ug/L	05/21/19 08:55	05/23/19 15:23	4292047
2058	Malathion	527	---	0.5	< 0.5	ug/L	05/21/19 08:55	05/23/19 15:23	4292047
2932	Mirex	527	---	0.5	< 0.5	ug/L	05/21/19 08:55	05/23/19 15:23	4292047
1836-75-5	Nitrofen	527	---	0.5	< 0.5	ug/L	05/21/19 08:55	05/23/19 15:23	4292047
27314-13-2	Norflurazon	527	---	0.5	< 0.5	ug/L	05/21/19 08:55	05/23/19 15:23	4292047
27304-13-8	Oxychlorodane	527	---	0.5	< 0.5	ug/L	05/21/19 08:55	05/23/19 15:23	4292047
2064	Parathion	527	---	0.5	< 0.5	ug/L	05/21/19 08:55	05/23/19 15:23	4292047
60348-60-9	2,2',4,4',5-Pentabromodiphenyl ether (BDE-99)	527	---	0.9	< 0.9	ug/L	05/21/19 08:55	05/23/19 15:23	4292047
189084-64-8	2,2',4,4',6-Pentabromodiphenyl ether (BDE-100)	527	---	0.5	< 0.5	ug/L	05/21/19 08:55	05/23/19 15:23	4292047
7287-19-6	Prometryn	527	---	0.5	< 0.5	ug/L	05/21/19 08:55	05/23/19 15:23	4292047
2256	Propazine	527	---	0.5	< 0.5	ug/L	05/21/19 08:55	05/23/19 15:23	4292047
56070-16-7	Terbufos-sulfone	527	---	0.4	< 0.4	ug/L	05/21/19 08:55	05/23/19 15:23	4292047
5436-43-1	2,2',4,4'-Tetrabromodiphenyl ether (BDE-47)	527	---	0.3	< 0.3	ug/L	05/21/19 08:55	05/23/19 15:23	4292047
2627	Thiobencarb	527	---	0.5	< 0.5	ug/L	05/21/19 08:55	05/23/19 15:23	4292047
50471-44-8	Vinclozolin	527	---	0.5	< 0.5	ug/L	05/21/19 08:55	05/23/19 15:23	4292047
2047	Aldicarb	531.2 LL	---	0.1	< 0.1	ug/L	05/16/19 08:20	05/21/19 02:05	4291765
2044	Aldicarb sulfone	531.2 LL	---	0.1	< 0.1	ug/L	05/16/19 08:20	05/21/19 02:05	4291765
2043	Aldicarb sulfoxide	531.2 LL	---	0.1	< 0.1	ug/L	05/16/19 08:20	05/21/19 02:05	4291765
2023	Baygon	531.2 LL	---	0.1	< 0.1	ug/L	05/16/19 08:20	05/21/19 02:05	4291765
2021	Carbaryl	531.2 LL	---	0.1	< 0.1	ug/L	05/16/19 08:20	05/21/19 02:05	4291765
2046	Carbofuran	531.2 LL	---	0.1	< 0.1	ug/L	05/16/19 08:20	05/21/19 02:05	4291765
2066	3-Hydroxycarbofuran	531.2 LL	---	0.1	< 0.1	ug/L	05/16/19 08:20	05/21/19 02:05	4291765
2024	Methiocarb	531.2 LL	---	1.0	< 1.0	ug/L	05/16/19 08:20	05/21/19 02:05	4291765
2022	Methomyl	531.2 LL	---	0.1	< 0.1	ug/L	05/16/19 08:20	05/21/19 02:05	4291765
2259	1-Naphthol	531.2 LL	---	1.0	< 1.0	ug/L	05/16/19 08:20	05/21/19 02:05	4291765
2036	Oxamyl	531.2 LL	---	0.1	< 0.1	ug/L	05/16/19 08:20	05/21/19 02:05	4291765
96180-79-9	Microcystin-LA	544	---	0.008	< 0.008	ug/L	06/03/19 08:30	06/03/19 20:57	4292049
154037-70-4	Microcystin-LF	544	---	0.006	< 0.006	ug/L	06/03/19 08:30	06/03/19 20:57	4292049
101043-37-2	Microcystin-LR	544	---	0.020	< 0.020	ug/L	06/03/19 08:30	06/03/19 20:57	4292049
123304-10-9	Microcystin-LY	544	---	0.009	< 0.009	ug/L	06/03/19 08:30	06/03/19 20:57	4292049
111755-37-4	Microcystin-RR	544	---	0.006	< 0.006	ug/L	06/03/19 08:30	06/03/19 20:57	4292049
101064-48-6	Microcystin-YR	544	---	0.020	< 0.020	ug/L	06/03/19 08:30	06/03/19 20:57	4292049
118399-22-7	Nodularin	544	---	0.005	< 0.005	ug/L	06/03/19 08:30	06/03/19 20:57	4292049
64285-06-9	Anatoxin-a	545	---	0.03	< 0.03	ug/L	05/16/19 13:50	05/16/19 20:36	4292048
143545-90-8	Cylindrospermopsin	545	---	0.09	< 0.09	ug/L	05/16/19 13:50	05/16/19 20:36	4292048
2032	Diquat	549.2	20 *	0.4	< 0.4	ug/L	05/16/19 08:20	05/16/19 17:51	4291760
2028	Paraquat	549.2	---	0.4	< 0.4	ug/L	05/16/19 08:20	05/16/19 17:51	4291760
2470	Acetaldehyde	556	---	5.0	< 5.0	ug/L	05/17/19 07:30	05/19/19 02:08	4291768
2479	Benzaldehyde	556	---	5.0	< 5.0	ug/L	05/17/19 07:30	05/19/19 02:08	4291768
2472	Butanal	556	---	5.0	< 5.0	ug/L	05/17/19 07:30	05/19/19 02:08	4291768
123-73-9	Crotonaldehyde	556	---	5.0	< 5.0	ug/L	05/17/19 07:30	05/19/19 02:08	4291768
108-94-1	Cyclohexanone	556	---	5.0	< 5.0	ug/L	05/17/19 07:30	05/19/19 02:08	4291768
2481	Decanal	556	---	5.0	< 5.0	ug/L	05/17/19 07:30	05/19/19 02:08	4291768

2961	Formaldehyde	556	---	5.0	7.1	ug/L	05/17/19 07:30	05/19/19 02:08	4291768
2474	Glyoxal	556	---	5.0	< 5.0	ug/L	05/17/19 07:30	05/19/19 02:08	4291768
2477	Heptanal	556	---	5.0	< 5.0	ug/L	05/17/19 07:30	05/19/19 02:08	4291768
2476	Hexanal	556	---	5.0	< 5.0	ug/L	05/17/19 07:30	05/19/19 02:08	4291768
2475	Methyl glyoxal	556	---	5.0	< 5.0	ug/L	05/17/19 07:30	05/19/19 02:08	4291768
2480	Nonanal	556	---	5.0	< 5.0	ug/L	05/17/19 07:30	05/19/19 02:08	4291768
2478	Octanal	556	---	5.0	< 5.0	ug/L	05/17/19 07:30	05/19/19 02:08	4291768
2473	Pentanal	556	---	5.0	< 5.0	ug/L	05/17/19 07:30	05/19/19 02:08	4291768
2471	Propanal	556	---	5.0	< 5.0	ug/L	05/17/19 07:30	05/19/19 02:08	4291768

Volatile Organic Chemicals									
Analyte ID #	Analyte	Method	Reg Limit	MRL†	Result	Units	Preparation Date	Analyzed Date	EEA ID #
2243	Acetone	524.2	---	5.0	< 5.0	ug/L	---	05/15/19 19:19	4291763
2240	Acrylonitrile	524.2	---	1.0	< 1.0	ug/L	---	05/16/19 18:24	4291763
2402	Allyl chloride	524.2	---	5.0	< 5.0	ug/L	---	05/16/19 18:24	4291763
994-05-8	tert-Amyl Methyl ether	524.2	---	3.0	< 3.0	ug/L	---	05/15/19 19:19	4291763
2990	Benzene	524.2	5 *	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
100-44-7	Benzyl chloride	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2993	Bromobenzene	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2430	Bromochloromethane	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2943	Bromodichloromethane	524.2	---	0.5	2.9	ug/L	---	05/15/19 19:19	4291763
2942	Bromoform	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2214	Bromomethane	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2486	1,3-Butadiene	524.2	---	5.0	< 5.0	ug/L	---	05/15/19 19:19	4291763
2247	2-Butanone (MEK)	524.2	---	5.0	< 5.0	ug/L	---	05/15/19 19:19	4291763
141-32-2	n-Butyl acrylate	524.2	---	1.0	< 1.0	ug/L	---	05/15/19 19:19	4291763
2422	n-Butylbenzene	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2428	sec-Butylbenzene	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2426	tert-Butylbenzene	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
1902	Carbon disulfide	524.2	---	5.0	< 5.0	ug/L	---	05/15/19 19:19	4291763
2982	Carbon tetrachloride	524.2	5 *	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2466	Chloroacetonitrile	524.2	---	5.0	< 5.0	ug/L	---	05/16/19 18:24	4291763
2989	Chlorobenzene	524.2	100 *	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2086	1-Chlorobutane	524.2	---	5.0	< 5.0	ug/L	---	05/16/19 18:24	4291763
2216	Chloroethane	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2941	Chloroform	524.2	---	0.5	3.2	ug/L	---	05/15/19 19:19	4291763
2210	Chloromethane	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
126-99-8	Chloroprene	524.2	---	5.0	< 5.0	ug/L	---	05/15/19 19:19	4291763
2965	2-Chlorotoluene	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2966	4-Chlorotoluene	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
108-94-1	Cyclohexanone	524.2	---	5.0	< 5.0	ug/L	---	05/15/19 19:19	4291763
2944	Dibromochloromethane	524.2	---	0.5	1.5	ug/L	---	05/15/19 19:19	4291763
2950	Total Trihalomethanes	524.2	80 *	0.5	7.6	ug/L	---	05/15/19 19:19	4291763
2931	1,2-Dibromo-3-chloropropane (DBCP)	524.2	---	0.2	< 0.2	ug/L	---	05/15/19 19:19	4291763
2946	1,2-Dibromoethane (EDB)	524.2	---	0.2	< 0.2	ug/L	---	05/15/19 19:19	4291763
2408	Dibromomethane	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2968	1,2-Dichlorobenzene	524.2	600 *	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2967	1,3-Dichlorobenzene	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2969	1,4-Dichlorobenzene	524.2	75 *	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2401	Total Dichlorobenzene	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2970	trans-1,4-Dichloro-2-butylene	524.2	---	5.0	< 5.0	ug/L	---	05/16/19 18:24	4291763

2212	Dichlorodifluoromethane	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2978	1,1-Dichloroethane	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2980	1,2-Dichloroethane	524.2	5 *	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2977	1,1-Dichloroethylene	524.2	7 *	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2380	cis-1,2-Dichloroethylene	524.2	70 *	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2979	trans-1,2-Dichloroethylene	524.2	100 *	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2222	Di(2-chloroethyl)ether	524.2	---	2.0	< 2.0	ug/L	---	05/15/19 19:19	4291763
2964	Dichloromethane	524.2	5 *	0.5	< 0.5	ug/L	---	05/16/19 18:24	4291763
2983	1,2-Dichloropropane	524.2	5 *	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2412	1,3-Dichloropropane	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2463	1,1-Dichloropropanone	524.2	---	5.0	< 5.0	ug/L	---	05/16/19 18:24	4291763
2416	2,2-Dichloropropane	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2410	1,1-Dichloropropylene	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2228	cis-1,3-Dichloropropylene	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2224	trans-1,3-Dichloropropylene	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2413	1,3-Dichloropropylene, cis & trans	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
108-20-3	Diisopropyl ether	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2049	1,4-Dioxane	524.2	---	5.0	< 5.0	ug/L	---	05/15/19 19:19	4291763
2257	Epichlorohydrin	524.2	---	1.0	< 1.0	ug/L	---	05/15/19 19:19	4291763
140-88-5	Ethyl acrylate	524.2	---	1.0	< 1.0	ug/L	---	05/15/19 19:19	4291763
2090	Ethyl Ether	524.2	---	2.0	< 2.0	ug/L	---	05/16/19 18:24	4291763
2293	Ethyl methacrylate	524.2	---	1.0	< 1.0	ug/L	---	05/16/19 18:24	4291763
2992	Ethylbenzene	524.2	700 *	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
637-92-3	Ethyl tert-butyl ether	524.2	---	3.0	< 3.0	ug/L	---	05/15/19 19:19	4291763
2246	Hexachlorobutadiene	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2225	Hexachloroethane	524.2	---	2.0	< 2.0	ug/L	---	05/16/19 18:24	4291763
2269	2-Hexanone	524.2	---	5.0	< 5.0	ug/L	---	05/15/19 19:19	4291763
2994	Isopropylbenzene	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2030	4-Isopropyltoluene	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2297	Methylacrylate	524.2	---	1.0	< 1.0	ug/L	---	05/16/19 18:24	4291763
2458	Methyl iodide	524.2	---	2.0	< 2.0	ug/L	---	05/16/19 18:24	4291763
2295	Methylmethacrylate	524.2	---	1.0	< 1.0	ug/L	---	05/16/19 18:24	4291763
2249	4-Methyl-2-pentanone (MIBK)	524.2	---	2.0	< 2.0	ug/L	---	05/15/19 19:19	4291763
75-65-0	tert-Butyl alcohol	524.2	---	2.0	< 2.0	ug/L	---	05/15/19 19:19	4291763
2251	Methyl-t-butyl ether (MTBE)	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2248	Naphthalene	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2254	Nitrobenzene	524.2	---	5.0	< 5.0	ug/L	---	05/16/19 18:24	4291763
2469	2-Nitropropane	524.2	---	2.0	< 2.0	ug/L	---	05/16/19 18:24	4291763
2468	Propionitrile	524.2	---	5.0	< 5.0	ug/L	---	05/16/19 18:24	4291763
2998	n-Propylbenzene	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2996	Styrene	524.2	100 *	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2986	1,1,1,2-Tetrachloroethane	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2988	1,1,2,2-Tetrachloroethane	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2987	Tetrachloroethylene	524.2	5 *	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2263	Tetrahydrofuran	524.2	---	5.0	6.1	ug/L	---	05/15/19 19:19	4291763
2991	Toluene	524.2	1000 *	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2420	1,2,3-Trichlorobenzene	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2378	1,2,4-Trichlorobenzene	524.2	70 *	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2981	1,1,1-Trichloroethane	524.2	200 *	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2985	1,1,2-Trichloroethane	524.2	5 *	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2984	Trichloroethylene	524.2	5 *	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2218	Trichlorofluoromethane	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763

2414	1,2,3-Trichloropropane	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2904	1,1,2-Trichloro-1,2,2-trifluoroethane	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2419	1,2,3-Trimethylbenzene	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2418	1,2,4-Trimethylbenzene	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2424	1,3,5-Trimethylbenzene	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2447	Vinyl acetate	524.2	---	5.0	< 5.0	ug/L	---	05/15/19 19:19	4291763
2976	Vinyl chloride	524.2	2 *	0.2	< 0.2	ug/L	---	05/15/19 19:19	4291763
2997	1,2-Xylene	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2963	1,3 + 1,4-Xylene	524.2	---	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2955	Xylenes, Total	524.2	10000 *	0.5	< 0.5	ug/L	---	05/15/19 19:19	4291763
2049	1,4-Dioxane	522	---	0.07	0.33	ug/L	05/29/19 08:10	05/30/19 23:39	4291756
2414	1,2,3-Trichloropropane	SRL 524M-TCP	---	5.0	< 5.0	ng/L	---	05/17/19 17:03	4291762

Compliance monitoring for 1,2-Dibromo-3-chloropropane (DBCP) must be done using EPA method 504.1.

Compliance monitoring for 1,2-Dibromoethane (EDB) must be done using EPA method 504.1.

EEA Methods									
Analyte ID #	Analyte	Method	Reg Limit	MRL†	Result	Units	Preparation Date	Analyzed Date	EEA ID #
2044	Aldicarb sulfone	L306	---	0.5	< 0.5	ug/L	---	05/21/19 20:27	4291764
2043	Aldicarb sulfoxide	L306	---	0.5	< 0.5	ug/L	---	05/21/19 20:27	4291764
131860-33-8	Azoxystrobin	L306	---	0.5	< 0.5	ug/L	---	05/21/19 20:27	4291764
2023	Baygon	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
22781-23-3	Bendiocarb	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
741-58-2	Bensulide	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
2021	Carbaryl	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
2046	Carbofuran	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
81777-89-1	Clomazone	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
2291	Difluzenuron	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
2103	Diuron	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
101-42-8	Fenuron	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
2164-17-2	Fluometuron	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
112226-61-6	Halofenozide	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
100784-20-1	Halosulfuron methyl	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
138261-41-3	Imidacloprid	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
2283	Linuron	L306	---	0.5	< 0.5	ug/L	---	05/21/19 20:27	4291764
---	Metalaxyl/Metalaxyl-M (Mefenoxam)	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
2024	Methiocarb	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
2022	Methomyl	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
MONU	Monuron	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
555-37-3	Neburon	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
19044-88-3	Oryzalin	L306	---	0.5	< 0.5	ug/L	---	05/21/19 20:27	4291764
19666-30-9	Oxadiazon	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
2036	Oxamyl	L306	---	0.5	< 0.5	ug/L	---	05/21/19 20:27	4291764
76738-62-0	Paclobutrazol	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
2289	Propanil	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
2312-35-8	Propargite	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
2301	Siduron, Total	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
2271	Tebuthiuron	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
2303	Thidiazuron	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
43121-43-3	Triadimefon	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
55219-65-3	Triadimenol	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
2066	3-Hydroxycarbofuran	L306	---	0.05	< 0.05	ug/L	---	05/21/19 20:27	4291764
2265	Acrylamide	L520	---	0.5	< 0.5	ug/L	---	05/17/19 18:25	4291767

62-53-3	Aniline	L520	---	0.5	< 0.5	ug/L	---	05/17/19 18:25	4291767
51-79-6	Urethane	L520	---	1.0	< 1.0	ug/L	---	05/17/19 18:25	4291767
120226-60-0	10:2 Fluorotelomer sulfonic acid (10:2 FTS)	L402	---	2.0	< 2.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
757124-22-4	4:2 Fluorotelomer sulfonic acid (4:2 FTS)	L402	---	2.0	< 2.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
27619-97-2	6:2 Fluorotelomer sulfonic acid (6:2 FTS)	L402	---	2.0	< 2.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
39108-34-4	8:2 Fluorotelomer sulfonic acid (8:2 FTS)	L402	---	2.0	< 2.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
958445-44-8	ADONA	L402	---	2.0	< 2.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
73606-19-6	9CI-PF3ONS/F-53B Major	L402	---	2.0	< 2.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
83329-89-9	11CI-PF3OUdS/F-53B Minor	L402	---	2.0	< 2.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
13252-13-6	HFPO-DA/GenX	L402	---	5.0	< 5.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
4151-50-2	N-ethylperfluorooctane sulfonamide (NEtFOSA)	L402	---	2.0	< 2.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
1691-99-2	N-ethylperfluorooctane sulfonamidoethanol	L402	---	2.0	< 2.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
31506-32-8	N-methylperfluorooctane sulfonamide (NMeFOSA)	L402	---	2.0	< 2.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
24448-09-7	N-methylperfluorooctane sulfonamidoethanol	L402	---	2.0	< 2.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
2801	Perfluorobutanesulfonic acid (PFBS)	L402	---	2.0	3.5	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
375-22-4	Perfluorobutanoic acid (PFBA)	L402	---	5.0	< 5.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
335-76-2	Perfluorodecanoic acid (PFDA)	L402	---	2.0	< 2.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
2802	Perfluoroheptanoic acid (PFHpA)	L402	---	2.0	2.8	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
2803	Perfluorohexanesulfonic acid (PFHxS)	L402	---	2.0	2.2	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
307-24-4	Perfluorohexanoic acid (PFHxA)	L402	---	2.0	5.6	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
307-55-1	Perfluorododecanoic acid (PFDoA)	L402	---	2.0	< 2.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
2804	Perfluorononanoic acid (PFNA)	L402	---	2.0	< 2.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
6909	Perfluorooctanesulfonic acid (PFOS)	L402	---	2.0	7.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
2991-50-6	N-ethyl Perfluorooctanesulfonamidoacetic acid	L402	---	2.0	< 2.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
2355-31-9	N-methyl Perfluorooctanesulfonamidoacetic acid	L402	---	2.0	< 2.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
6912	Perfluorooctanoic acid (PFOA)	L402	---	2.0	5.3	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
72629-94-8	Perfluorotridecanoic acid (PFTrDA)	L402	---	2.0	< 2.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
2058-94-8	Perfluoroundecanoic acid (PFUnA)	L402	---	2.0	< 2.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
NA	Perfluorododecanesulfonic acid (PFDoS)	L402	---	2.0	< 2.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
335-77-3	Perfluorodecanesulfonic acid (PFDS)	L402	---	2.0	< 2.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
375-92-8	Perfluoroheptanesulfonic acid (PFHpS)	L402	---	2.0	< 2.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
67905-19-5	Perfluorohexadecanoic acid (PFHxDA)	L402	---	2.0	< 2.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
151772-58-6	Perfluoro-2-methoxyethoxyacetic acid	L402	---	5.0	< 5.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
801212-59-9	Perfluoro-4-isopropoxybutanoic acid	L402	---	5.0	< 5.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
863090-89-5	Perfluoro-4-methoxybutanoic acid (PFMOBA)	L402	---	5.0	< 5.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
68259-12-1	Perfluorononanesulfonic acid (PFNS)	L402	---	2.0	< 2.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
754-91-6	Perfluorooctane sulfonamide (PFOSA)	L402	---	2.0	< 2.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
2706-90-3	Perfluoropentanoic acid (PFPeA)	L402	---	2.0	5.8	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
2706-91-4	Perfluoropentanesulfonic acid (PFPeS)	L402	---	2.0	< 2.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769
376-06-7	Perfluorotetradecanoic acid (PFTeDA)	L402	---	2.0	< 2.0	ng/L	05/30/19 08:30	05/31/19 01:20	4291769

Reference Lab Tests

Analyte ID #	Analyte	Method	Reg Limit	MRL†	Result	Units	Preparation Date	Analyzed	EEA ID #
2063	Dioxin	1613	30 *	5.0	< 5.0	pg/L	05/22/19 00:00	05/24/19 16:11	4291761
55589-62-3	Acesulfame-K	LC/MS/MS	---	20	88	ng/L	---	05/23/19 21:00	4291766
103-90-2	Acetaminophen	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
18559-94-9	Albuterol	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
26787-78-0	Amoxicillin	LC/MS/MS	---	20	< 20	ng/L	---	05/23/19 21:00	4291766
2705	4-Androstene-3,17-dione	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
60-80-0	Antipyrine	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
29122-68-7	Atenolol	LC/MS/MS	---	5.0	5.8	ng/L	---	05/23/19 21:00	4291766
2050	Atrazine	LC/MS/MS	---	5.0	16	ng/L	---	05/23/19 21:00	4291766
83905-01-5	Azithromycin	LC/MS/MS	---	20	< 20	ng/L	---	05/23/19 21:00	4291766

74-48-3	Bendroflumethiazide	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
41859-67-0	Bezafibrate	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
80-05-7	Bisphenol A	LC/MS/MS	---	10	< 10	ng/L	---	05/23/19 21:00	4291766
2098	Bromacil	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
77-26-9	Butalbital	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
94-26-8	Butylparaben	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
58-08-2	Caffeine	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
6804-07-5	Carbadox	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
298-46-4	Carbamazepine	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
78-44-4	Carisoprodol	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
56-75-7	Chloramphenicol	LC/MS/MS	---	10	< 10	ng/L	---	05/23/19 21:00	4291766
1698-60-8	Chloridazon	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
15545-48-9	Chlorotoluron	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
51481-61-9	Cimetidine	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
882-09-7	Clofibrac acid	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
486-56-6	Cotinine	LC/MS/MS	---	10	< 10	ng/L	---	05/23/19 21:00	4291766
2054	Cyanazine	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
2105	2,4-D	LC/MS/MS	---	5.0	8.7	ng/L	---	05/23/19 21:00	4291766
134-62-3	DEET	LC/MS/MS	---	10	< 10	ng/L	---	05/23/19 21:00	4291766
67035-22-7	Dehydronifedipine	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
2006	Desethylatrazine	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
2007	Desisopropylatrazine	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
3397-62-4	Diaminochlorotriazine	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
439-14-5	Diazepam	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
15307-86-5	Diclofenac	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
57-41-0	Dilantin	LC/MS/MS	---	20	< 20	ng/L	---	05/23/19 21:00	4291766
42399-41-7	Diltiazem	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
611-59-6	1,7-Dimethylxanthine	LC/MS/MS	---	10	< 10	ng/L	---	05/23/19 21:00	4291766
2103	Diuron	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
114-07-8	Erythromycin	LC/MS/MS	---	10	< 10	ng/L	---	05/23/19 21:00	4291766
2701	17beta-Estradiol	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
2707	Estrone	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
120-47-8	Ethylparaben	LC/MS/MS	---	20	< 20	ng/L	---	05/23/19 21:00	4291766
2702	17alpha-Ethynyl estradiol	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
42835-25-6	Flumequine	LC/MS/MS	---	10	< 10	ng/L	---	05/23/19 21:00	4291766
59333-67-4	Fluoxetine (Prozac)	LC/MS/MS	---	10	< 10	ng/L	---	05/23/19 21:00	4291766
25812-30-0	Gemfibrozil	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
15687-27-1	Ibuprofen	LC/MS/MS	---	10	< 10	ng/L	---	05/23/19 21:00	4291766
66108-95-0	Iohexal	LC/MS/MS	---	10	51	ng/L	---	05/23/19 21:00	4291766
73334-07-3	Iopromide	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
4247-02-3	Isobutylparaben	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
34123-59-6	Isoproturon	LC/MS/MS	---	100	< 100	ng/L	---	05/23/19 21:00	4291766
22074-15-4	Ketoprofen	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
74103-06-3	Ketorolac	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
137-58-6	Lidocaine	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
154-21-2	Lincomycin	LC/MS/MS	---	10	< 10	ng/L	---	05/23/19 21:00	4291766
2283	Linuron	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
51384-51-1	Lopressor	LC/MS/MS	---	20	< 20	ng/L	---	05/23/19 21:00	4291766
644-62-2	Meclofenamic acid	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
57-53-4	Meprobamate	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
67129-08-2	Metazochlor	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
99-76-3	Methylparaben	LC/MS/MS	---	20	< 20	ng/L	---	05/23/19 21:00	4291766
2045	Metolachlor	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
22204-53-1	Naproxen	LC/MS/MS	---	10	< 10	ng/L	---	05/23/19 21:00	4291766

21829-25-4	Nifedipine	LC/MS/MS	---	20	< 20	ng/L	---	05/23/19 21:00	4291766
104-40-5	4-Nonylphenol	LC/MS/MS	---	100	< 100	ng/L	---	05/23/19 21:00	4291766
68-22-4	Norethisterone	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
140-66-9	4-tert-Octylphenol	LC/MS/MS	---	50	< 50	ng/L	---	05/23/19 21:00	4291766
14698-29-4	Oxolinic acid	LC/MS/MS	---	10	< 10	ng/L	---	05/23/19 21:00	4291766
6493-05-6	Pentoxifylline	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
125-33-7	Primidone	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
57-83-0	Progesterone	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
2256	Propazine	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
94-13-3	Propylparaben	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
91-22-5	Quinoline	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
2037	Simazine	LC/MS/MS	---	5.0	16	ng/L	---	05/23/19 21:00	4291766
56038-13-2	Sucralose	LC/MS/MS	---	100	580	ng/L	---	05/23/19 21:00	4291766
80-32-0	Sulfachloropyridazine	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
68-35-9	Sulfadiazine	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
122-11-2	Sulfadimethoxine	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
127-79-7	Sulfamerazine	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
57-68-1	Sulfamethazine	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
144-82-1	Sulfamethizole	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
723-46-6	Sulfamethoxazole	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
72-14-0	Sulfathiazole	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
2706	Testosterone	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
83-67-0	Theobromine	LC/MS/MS	---	10	< 10	ng/L	---	05/23/19 21:00	4291766
58-55-9	Theophylline	LC/MS/MS	---	20	< 20	ng/L	---	05/23/19 21:00	4291766
148-79-8	Thiabendazole	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
101-20-2	Triclocarban	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
3380-34-5	Triclosan	LC/MS/MS	---	10	< 10	ng/L	---	05/23/19 21:00	4291766
738-70-5	Trimethoprim	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
51805-45-9	Tris(2-carboxyethyl)phosphine f	LC/MS/MS	---	10	< 10	ng/L	---	05/23/19 21:00	4291766
13674-87-8	Tris(1,3-dichloro-2-propyl) phos	LC/MS/MS	---	100	< 100	ng/L	---	05/23/19 21:00	4291766
81-81-2	Warfarin	LC/MS/MS	---	5.0	< 5.0	ng/L	---	05/23/19 21:00	4291766
35572-78-2	2-Amino-4,6-dinitrotoluene	8330A	---	0.6	< 0.6	ug/L	---	05/24/19 08:37	4292046
35572-78-2	4-Amino-2,6-dinitrotoluene	8330A	---	0.6	< 0.6	ug/L	---	05/24/19 08:37	4292046
99-65-0	1,3-Dinitrobenzene	8330A	---	0.6	< 0.6	ug/L	---	05/24/19 08:37	4292046
2270	2,4-Dinitrotoluene	8330A	---	2.0	< 2.0	ug/L	---	05/24/19 08:37	4292046
2266	2,6-Dinitrotoluene	8330A	---	1.3	< 1.3	ug/L	---	05/24/19 08:37	4292046
2691-41-0	HMX (Octahydro-1,3,5,7-tetranitro	8330A	---	2.0	< 2.0	ug/L	---	05/24/19 08:37	4292046
2254	Nitrobenzene	8330A	---	0.6	< 0.6	ug/L	---	05/24/19 08:37	4292046
55-63-0	Nitroglycerin	8330A	---	15	< 15	ug/L	---	05/24/19 08:37	4292046
88-72-2	2-Nitrotoluene	8330A	---	0.75	< 0.75	ug/L	---	05/24/19 08:37	4292046
88-72-2	3-Nitrotoluene	8330A	---	1.2	< 1.2	ug/L	---	05/24/19 08:37	4292046
88-72-2	4-Nitrotoluene	8330A	---	1.2	< 1.2	ug/L	---	05/24/19 08:37	4292046
78-11-5	PETN (Pentaerythritol tetranitrat	8330A	---	18	< 18	ug/L	---	05/24/19 08:37	4292046
2096	RDX (Hexahydro-1,3,5-trinitro-1	8330A	---	0.5	< 0.5	ug/L	---	05/24/19 08:37	4292046
479-45-8	Tetryl (2,4,6-Trinitrophenylmethy	8330A	---	0.8	< 0.8	ug/L	---	05/24/19 08:37	4292046
99-35-4	1,3,5-Trinitrobenzene	8330A	---	0.6	< 0.6	ug/L	---	05/24/19 08:37	4292046
118-96-7	2,4,6-Trinitrotoluene	8330A	---	0.6	< 0.6	ug/L	---	05/24/19 08:37	4292046

† EEA has demonstrated it can achieve these report limits in reagent water, but can not document them in all sample matrices.

Reg Limit Type:	MCL	SMCL	AL
Symbol:	*	^	!

Lab Definitions

Continuing Calibration Check Standard (CCC) / Continuing Calibration Verification (CCV) / Initial Calibration Verification Standard (ICV) / Initial Performance Check (IPC) - is a standard containing one or more of the target analytes that is prepared from the same standards used to calibrate the instrument. This standard is used to verify the calibration curve at the beginning of each analytical sequence, and may also be analyzed throughout and at the end of the sequence. The concentration of continuing standards may be varied, when prescribed by the reference method, so that the range of the calibration curve is verified on a regular basis. CCL, CCM, and CCH are the CCC standards at low, mid, and high concentration levels, respectively.

Internal Standards (IS) - are pure compounds with properties similar to the analytes of interest, which are added to field samples or extracts, calibration standards, and quality control standards at a known concentration. They are used to measure the relative responses of the analytes of interest and surrogates in the sample, calibration standard or quality control standard.

Laboratory Duplicate (LD) - is a field sample aliquot taken from the same sample container in the laboratory and analyzed separately using identical procedures. Analysis of laboratory duplicates provides a measure of the precision of the laboratory procedures.

Laboratory Fortified Blank (LFB) / Laboratory Control Sample (LCS) - is an aliquot of reagent water to which known concentrations of the analytes of interest are added. The LFB is analyzed exactly the same as the field samples. LFBs are used to determine whether the method is in control. FBL, FBM, and FBH are the LFB samples at low, mid, and high concentration levels, respectively.

Laboratory Method Blank (LMB) / Laboratory Reagent Blank (LRB) - is a sample of reagent water included in the sample batch analyzed in the same way as the associated field samples. The LMB is used to determine if method analytes or other background contamination have been introduced during the preparation or analytical procedure. The LMB is analyzed exactly the same as the field samples.

Laboratory Trip Blank (LTB) / Field Reagent Blank (FRB) - is a sample of laboratory reagent water placed in a sample container in the laboratory and treated as a field sample, including storage, preservation, and all analytical procedures. The FRB/LTB container follows the collection bottles to and from the collection site, but the FRB/LTB is not opened at any time during the trip. The FRB/LTB is primarily a travel blank used to verify that the samples were not contaminated during shipment.

Matrix Spike Duplicate Sample (MSD) / Laboratory Fortified Sample Matrix Duplicate (LFSMD) - is a sample aliquot taken from the same field sample source as the Matrix Spike Sample to which known quantities of the analytes of interest are added in the laboratory. The MSD is analyzed exactly the same as the field samples. Analysis of the MSD provides a measure of the precision of the laboratory procedures in a specific matrix. SDL, SDM, and SDH / LFSMDL, LFSMDM, and LFSMDH are the MSD or LFSMD at low, mid, and high concentration levels, respectively.

Matrix Spike Sample (MS) / Laboratory Fortified Sample Matrix (LFSM) - is a sample aliquot taken from field sample source to which known quantities of the analytes of interest are added in the laboratory. The MS is analyzed exactly the same as the field samples. The purpose is to demonstrate recovery of the analytes from a sample matrix to determine if the specific matrix contributes bias to the analytical results. MSL, MSM, and MSH / LFSML, LFSMM, and LFSMH are the MS or LFSM at low, mid, and high concentration levels, respectively.

Quality Control Standard (QCS) / Second Source Calibration Verification (SSCV) - is a solution containing known concentrations of the analytes of interest prepared from a source different from the source of the calibration standards. The solution is obtained from a second manufacturer or lot if the lot can be demonstrated by the manufacturer as prepared independently from other lots. The QCS sample is analyzed using the same procedures as field samples. The QCS is used as a check on the calibration standards used in the method on a routine basis.

Reporting Limit Check (RLC) / Initial Calibration Check Standard (ICCS) - is a procedural standard that is analyzed each day to evaluate instrument performance at or below the minimum reporting limit (MRL).

Surrogate Standard (SS) / Surrogate Analyte (SUR) - is a pure compound with properties similar to the analytes of interest, which is highly unlikely to be found in any field sample, that is added to the field samples, calibration standards, blanks and quality control standards before sample preparation. The SS is used to evaluate the efficiency of the sample preparation process.



Eaton Analytical

110 S. Hill Street
South Bend, IN 46617
T: 1.800.332.4345
F: 1.574.233.8207

Order # 38860
Batch # 451915

www.EurofinsUS.com/Eaton

SOUTH CAROLINA CHAIN OF CUSTODY RECORD

Page 1 of 23

REPORT TO: Becky Thames
1104 Hanahan Rd, Hanahan, SC 29410

BILL TO: Charleston Water System
PO Box B, 103 St Philip St,
Charleston, SC 29403

SAMPLER (Signature): Becky Thames

STATE (if sample origin): South Carolina

PWS ID#: SC1010001

POPULATION SERVED: 400,000

SOURCE WATER: SC1010001

COMPLIANCE MONITORING: Yes No

SAMPLING SITE: PH-2-24-19

TEST NAME: PH-2-24-19

PRESERVATIVE CODE (or write preservative name): PH-2-24-19

CHLORINATED: YES NO

OF CONTAINERS: 2

MATRIX CODE: DW SW

TURNAROUND TIME: 23

LAB Number	COLLECTION		SAMPLING SITE	TEST NAME	PRESERVATIVE CODE (or write preservative name)	CHLORINATED		# OF CONTAINERS	MATRIX CODE	TURNAROUND TIME
	DATE	TIME				YES	NO			
1	5/13/19	1159	PH-2-24-19	1,4 Dioxane 1522	★	✓	✓	2	DW SW	23
2	5/13/19	1215	PH-2-24-19	Chlorinated Acids - Extended 1515.3	5	✓	✓	2	DW SW	23
3	5/13/19	1204	PH-2-24-19	Nitrosamines 1521	4	✓	✓	2	DW SW	23
4	5/13/19	1155	PH-2-24-19	Pesticides + Inert Chem Ext 1525.2	5+3	✓	✓	2	DW SW	23
5	5/13/19	1201	PH-2-24-19	Diquat / Paraquat 1549.2	4+2	✓	✓	1	DW SW	23
6	5/13/19	1209	PH-2-24-19	Dioxin 11613	4	✓	✓	3	DW SW	23
7	5/13/19	1224	PH-2-24-19	1,2,3 Trichloropropane 1524M-TP	6+3	✓	✓	3	DW SW	23
8	5/13/19	1227	PH-2-24-19	VOCs Extended 1524.2	6+3	✓	✓	4	DW SW	23
9	5/13/19	1213	PH-2-24-19	Turfgrass Pesticides 1L306	6	✓	✓	1	DW SW	23
10	5/13/19	1219	PH-2-24-19	Carbamates 1531.2L	★	✓	✓	2	DW SW	23
11	5/13/19	1217	PH-2-24-19	EDC/PPEP/Hormone/LC/MS/MS	★	✓	✓	2	DW SW	23
12	5/13/19	1220	PH-2-24-19	Acrylamide, Anilines, Urethane 1520	4	✓	✓	1	DW SW	23
13	5/13/19	1221	PH-2-24-19	Aldehydes 1556	★	✓	✓	3	DW SW	23
14	5/13/19	1232	PH-2-24-19	PFC - Guan + 1L402	1	✓	✓	2	DW SW	23

LAB COMMENTS: WPS

RECEIVED BY: (Signature) WPS

RECEIVED BY: (Signature) WPS

RECEIVED FOR LABORATORY BY: WPS

DATE: 5/13/19 TIME: 1300

DATE: 5/14/19 TIME: 1000

DATE: 5/14/19 TIME: 1000

CONDITIONS UPON RECEIPT (check one):
 Icep. Wet/Glue
 Ambient
 °C Upon Receipt 1.0
 N/A

TURN-AROUND TIME (TA) - SURCHARGES
 IV* = Immediate Verbal: (3 working days) 100%
 IW* = Immediate Written: (3 working days) 125%
 SP* = Weekend, Holiday CALL
 ST* = Less than 48 hours CALL

MATRIX CODES:
 DW=Drinking Water
 RW=Reagent Water
 GW=Ground Water
 EW=Exposure Water
 SW=Surface Water
 PW=Pool Water
 WW=Waste Water

PRESERVATIVE CODES:
 6=Ascorbic Acid
 7=MCAA Buffer
 8=NaOH/NaBH₄
 9=NH₄Cl
 10=EDA

SW = Standard Written: (15 working days) 0%
 RV* = Rush Verbal: (5 working days) 50%
 RW* = Rush Written: (5 working days) 75%

* Please call, expedited service not available for all testing

Sample analysis will be provided according to the standard EEA/Water Services Terms, which are available upon request. Any other terms proposed by Customer are deemed material alterations and are rejected unless expressly agreed to in writing by EEA.

★ Sodium sulfate + NaHSO₄
 ★ Potassium dihydrogen citrate + Sodium thiosulfate

*** Sodium amide + ascorbic acid
 *** CuSO₄·5H₂O + NH₄Cl

06-LO-F0435 Issue 5.0 Effective Date: 2018-03-01



Eaton Analytical

110 S. Hill Street
South Bend, IN 46617
T: 1.800.332.4345
F: 1.574.233.8207

Order # 208676
Batch # _____

www.EurofinsUS.com/Eaton

SOUTH CAROLINA CHAIN OF CUSTODY RECORD

Page 2 of 2

REPORT TO: Becky Thames
1104 Hanahan Rd, Hanahan, SC 29410

BILL TO: Charleston Water System
PO Box B, 103 St Philip St,
Charleston, SC 29403

SAMPLER (Signature): Stephanie Smith

STATE (if sample origin): South Carolina

PWS ID#: SC1010001

PROJECT NAME: _____

PO#: _____

CHLORINATED: YES NO

PRESERVATIVE CODE (or write preservative name): None

MATRIX CODE: _____

TURNAROUND TIME: _____

LAB NUMBER	COLLECTION		COMPLIANCE MONITORING	SAMPLING SITE	TEST NAME	PRESERVATIVE CODE (or write preservative name)	CHLORINATED	# OF CONTAINERS	MATRIX CODE	TURNAROUND TIME
	DATE	TIME								
1	5-13-19	12:11	X	Finished	ICP-MS Metals Scan 1200.8	None	✓	1	DW SW	
2	↓	12:34	X	Finished	Perchlorate 1331.0	None	✓	1	DW SW	
3	↓	12:52	X	Finished	Chromium VI 1218.6	None	✓	1	DW SW	
4										
5										
6	4-30-19	↓		LTP ↓	Ald			2		
7	774	↓			VOC			2		
8										
9										
10										
11										
12										
13										
14										

LAB RESERVES THE RIGHT TO RETURN UNUSED PORTIONS OF NON-AQUEOUS SAMPLES TO CLIENT

RELINQUISHED BY: (Signature) Stephanie Smith DATE 5-13-19 TIME 1300

RECEIVED BY: (Signature) _____ DATE _____ TIME _____

LAB COMMENTS: LTP VIALS CONTAIN ACCEPTABLE BUBBLES

RELINQUISHED BY: (Signature) _____ DATE _____ TIME _____

RECEIVED BY: (Signature) [Signature] DATE 5/14/19 TIME 1000

CONDITIONS UPON RECEIPT (check one):
 Ice/Wet/Blip Ambient 1.0 °C Upon Receipt N/A

TURN-AROUND TIME (TAT) - SURCHARGES
 IV* = Immediate Verbal: (3 working days) 100%
 IW* = Immediate Written: (3 working days) 125%
 RW* = Rush Verbal: (5 working days) 50%
 RW* = Rush Written: (5 working days) 75%
 SP* = Weekend, Holiday CALL
 ST* = Less than 48 hours CALL

* Please call, expedited service not available for all testing

06-LO-F0436 Issue 5.0 Effective Date: 2018-03-01

Samples received unannounced with less than 48 hours holding time remaining may be subject to additional charges.

Simple analysis will be provided according to the standard EEA/Water Services Terms, which are available upon request. Any other terms proposed by Customer are deemed material alterations and are rejected unless expressly agreed to in writing by EEA.

★ Please add Buffer solution upon receipt. No pH buffer soln included in Kit - SMS 5-13-19 per Jim with Eaton



Eaton Analytical

110 S. Hill Street
South Bend, IN 46617
T: 1.800.332.4345
F: 1.574.233.8207

Order # 20826
Batch # _____

www.EurofinsUS.com/Eaton

Shaded area for EEA use only

SOUTH CAROLINA CHAIN OF CUSTODY RECORD

REPORT TO: **Becky Thames**
1104 Hanahan Rd, Hanahan, SC 29410

BILL TO: **Charleston Water System**
PO Box B, 103 St Philip St,
Charleston, SC 29403

SAMPLER (Signature): *Alefumi Amell*

STATE (of sample origin): South Carolina

POPULATION SERVED: 400,000

PWS ID#: SC1010001

PROJECT NAME: _____

POF: _____

MATRIX CODE: _____

TURNAROUND TIME: _____

LAB Number	COLLECTION		SAMPLING SITE	TEST NAME	PRESERVATIVE CODE (or write preservative name)		CHLORINATED		# OF CONTAINERS	TURNAROUND TIME
	DATE	TIME			AM	PM	YES	NO		
1. <i>4292046</i>	5-13-19	1117	Finished	Nitroaromatics/Mitramines 18330A	None	✓		2	DW SW	
2. <i>047</i>	I	1113	Finished	Flame Retardants 1527	*	✓		2	DW SW	
3. <i>048</i>	I	1105	Finished	UICMR 4 545 1545 CIA	*	✓		1	DW SW	
4. <i>049</i>	I	1109	Finished	UICMR 4 544 1544	*	✓		3	DW SW	
5.										
6.										
7.										
8.										
9.										
10.										
11.										
12.										
13.										
14.										

RELINQUISHED BY: (Signature) *Alefumi Amell*

RECEIVED BY: (Signature) *WPS*

DATE: 5-13-19

TIME: 1130

AM | PM

LAB COMMENTS

LAB RESERVES THE RIGHT TO RETURN UNUSED PORTIONS OF NON-AQUEOUS SAMPLES TO CLIENT

RELINQUISHED BY: (Signature) _____

RECEIVED BY: (Signature) _____

DATE: _____

TIME: _____

AM | PM

RELINQUISHED BY: (Signature) _____

RECEIVED FOR LABORATORY BY: *SSS*

DATE: 5/14/19

TIME: 1000

AM | PM

CONDITIONS UPON RECEIPT (check one):

Iced Ambient Upon Receipt

0.6 °C

TURN-AROUND TIME (TAT) SURCHARGES

SW = Standard Written: (15 working days) 0% **IV* = Immediate Verbal: (3 working days) 100%**

RW* = Rush Written: (5 working days) 50% **IW* = Immediate Written: (3 working days) 125%**

RW* = Rush Written: (5 working days) 75% **SP* = Weekend, Holiday CALL**

STAT* = Less than 48 hours CALL

*Please call, expedited service not available for all testing

PRESERVATIVE CODES:

1= HNO₃ 6= Ascorbic Acid
2= H₂SO₄ 7= MCAA Buffer
3= HCl 8= NaOH/NaBH₄
4= Na₂S₂O₃ 9= NH₄Cl
5= Na₂SO₃ 10= EDA

MATRIX CODES:

DW=Drinking Water
RW=Reagent Water
GW=Ground Water
EW=Exposure Water
SW=Surface Water
WW=Waste Water

Sample analysis will be provided according to the standard EEA/Water Services Terms, which are available upon request. Any other terms proposed by Customer are deemed material alterations and are rejected unless expressly agree to in writing by EEA.

★ ★ Ascorbic acid, EDTA + potassium dihydrogen citrate ★ ★ ★ Trizms, 2 chloroacetamide, ascorbic acid, EDTA ★ ★ sodium bisulfate + ascorbic acid

Page 21 of 153

Effective Date: 2018-03-01

06-LO-F0436 Issue 5.0



Eaton Analytical

Eurofins Eaton Analytical

Run Log

Run ID: 260166 Method: 1613

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
FS	4291761	Finished	DW	R L	05/24/2019 16:11	Ref Lab Dioxin

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FS	2,3,7,8-Tetrachlorodibenzo-p-Dioxin	1613	5.0	Finished	<	5.0		pg/L	---	---	---	---	1.0	05/22/2019 00:00	05/24/2019 16:11	4291761

Eurofins Eaton Analytical

Run Log

Run ID: **259506** Method: **200.8**

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
QCS	4299389		RW	FQ	05/21/2019 11:37	
ICV	4299390		RW	FQ	05/21/2019 11:39	
ICB	4299391		RW	FQ	05/21/2019 11:42	
LRB	4299399		RW	FQ	05/21/2019 11:46	
LFB	4299400		RW	FQ	05/21/2019 11:51	
FS	4291770	Finished	DW	FQ	05/21/2019 11:55	
CCV	4299373		RW	FQ	05/21/2019 12:23	
CCB	4299374		RW	FQ	05/21/2019 12:25	
CCV	4299395		RW	FQ	05/21/2019 12:54	
CCB	4299396		RW	FQ	05/21/2019 12:57	

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
QCS	Aluminum	200.8	2.0	---		51.6050	50.0	ug/L	103	90 - 110	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	Antimony	200.8	1.0	---		51.6710	50.0	ug/L	103	90 - 110	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	Arsenic	200.8	1.0	---		50.6750	50.0	ug/L	101	90 - 110	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	Barium	200.8	2.0	---		50.2700	50.0	ug/L	101	90 - 110	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	Beryllium	200.8	0.3	---		49.7470	50.0	ug/L	99	90 - 110	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	IS-Bismuth	200.8	N/A	---		1.0189	1.0	N/A	102	60 - 125	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	Boron	200.8	5.0	---		49.9900	50.0	ug/L	100	90 - 110	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	Cadmium	200.8	1.0	---		50.2750	50.0	ug/L	101	90 - 110	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	Chromium	200.8	0.9	---		50.0030	50.0	ug/L	100	90 - 110	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	Cobalt	200.8	2.0	---		50.3490	50.0	ug/L	101	90 - 110	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	Copper	200.8	1.0	---		53.7240	50.0	ug/L	107	90 - 110	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	IS-Indium	200.8	N/A	---		1.0090	1.0	N/A	101	60 - 125	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	Lead	200.8	1.0	---		52.6070	50.0	ug/L	105	90 - 110	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	Lithium	200.8	2.0	---		52.9920	50.0	ug/L	106	90 - 110	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	Manganese	200.8	2.0	---		50.6730	50.0	ug/L	101	90 - 110	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	Molybdenum	200.8	2.0	---		50.1170	50.0	ug/L	100	90 - 110	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	Nickel	200.8	1.0	---		50.1070	50.0	ug/L	100	90 - 110	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	IS-Scandium	200.8	N/A	---		1.0040	1.0	N/A	100	60 - 125	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	Selenium	200.8	2.0	---		50.3450	50.0	ug/L	101	90 - 110	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	Silver	200.8	2.0	---		53.9810	50.0	ug/L	108	90 - 110	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	Strontium	200.8	2.0	---		50.6680	50.0	ug/L	101	90 - 110	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	IS-Terbium	200.8	N/A	---		1.0195	1.0	N/A	102	60 - 125	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	Thallium	200.8	0.3	---		49.6710	50.0	ug/L	99	90 - 110	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	Thorium	200.8	5.0	---		49.1050	50.0	ug/L	98	90 - 110	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	Tin	200.8	2.0	---		50.1670	50.0	ug/L	100	90 - 110	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	Titanium	200.8	5.0	---		50.6760	50.0	ug/L	101	90 - 110	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	Uranium	200.8	1.0	---		50.0070	50.0	ug/L	100	90 - 110	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	Vanadium	200.8	2.0	---		50.0350	50.0	ug/L	100	90 - 110	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	IS-Yttrium	200.8	N/A	---		1.0101	1.0	N/A	101	60 - 125	---	---	1.0	---	05/21/2019 11:37	4299389
QCS	Zinc	200.8	5.0	---		50.3150	50.0	ug/L	101	90 - 110	---	---	1.0	---	05/21/2019 11:37	4299389
ICV	Aluminum	200.8	2.0	---		51.3350	50.0	ug/L	103	90 - 110	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	Antimony	200.8	1.0	---		49.4880	50.0	ug/L	99	90 - 110	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	Arsenic	200.8	1.0	---		48.0230	50.0	ug/L	96	90 - 110	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	Barium	200.8	2.0	---		48.4680	50.0	ug/L	97	90 - 110	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	Beryllium	200.8	0.3	---		47.5650	50.0	ug/L	95	90 - 110	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	IS-Bismuth	200.8	N/A	---		1.0167	1.0	N/A	102	60 - 125	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	Boron	200.8	5.0	---		47.4740	50.0	ug/L	95	90 - 110	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	Cadmium	200.8	1.0	---		48.0640	50.0	ug/L	96	90 - 110	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	Chromium	200.8	0.9	---		47.6740	50.0	ug/L	95	90 - 110	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	Cobalt	200.8	2.0	---		47.5920	50.0	ug/L	95	90 - 110	---	---	1.0	---	05/21/2019 11:39	4299390

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
ICV	Copper	200.8	1.0	---		51.3390	50.0	ug/L	103	90 - 110	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	IS-Indium	200.8	N/A	---		1.0131	1.0	N/A	101	60 - 125	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	Lead	200.8	1.0	---		50.1040	50.0	ug/L	100	90 - 110	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	Lithium	200.8	2.0	---		50.5160	50.0	ug/L	101	90 - 110	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	Manganese	200.8	2.0	---		47.9580	50.0	ug/L	96	90 - 110	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	Molybdenum	200.8	2.0	---		48.4600	50.0	ug/L	97	90 - 110	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	Nickel	200.8	1.0	---		47.4390	50.0	ug/L	95	90 - 110	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	IS-Scandium	200.8	N/A	---		1.0056	1.0	N/A	101	60 - 125	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	Selenium	200.8	2.0	---		47.3720	50.0	ug/L	95	90 - 110	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	Silver	200.8	2.0	---		51.2720	50.0	ug/L	103	90 - 110	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	Strontium	200.8	2.0	---		48.3690	50.0	ug/L	97	90 - 110	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	IS-Terbium	200.8	N/A	---		1.0131	1.0	N/A	101	60 - 125	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	Thallium	200.8	0.3	---		47.5190	50.0	ug/L	95	90 - 110	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	Thorium	200.8	5.0	---		46.2880	50.0	ug/L	93	90 - 110	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	Tin	200.8	2.0	---		48.5640	50.0	ug/L	97	90 - 110	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	Titanium	200.8	5.0	---		48.2420	50.0	ug/L	96	90 - 110	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	Uranium	200.8	1.0	---		47.1750	50.0	ug/L	94	90 - 110	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	Vanadium	200.8	2.0	---		48.0390	50.0	ug/L	96	90 - 110	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	IS-Yttrium	200.8	N/A	---		1.0094	1.0	N/A	101	60 - 125	---	---	1.0	---	05/21/2019 11:39	4299390
ICV	Zinc	200.8	5.0	---		48.2130	50.0	ug/L	96	90 - 110	---	---	1.0	---	05/21/2019 11:39	4299390
ICB	Aluminum	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	Antimony	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	Arsenic	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	Barium	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	Beryllium	200.8	0.3	---	<	0.3		ug/L	---	---	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	IS-Bismuth	200.8	N/A	---		1.0250	1.0	N/A	102	60 - 125	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	Boron	200.8	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	Cadmium	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	Chromium	200.8	0.9	---	<	0.9		ug/L	---	---	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	Cobalt	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	Copper	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	IS-Indium	200.8	N/A	---		1.0168	1.0	N/A	102	60 - 125	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	Lead	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	Lithium	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	Manganese	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	Molybdenum	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	Nickel	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	IS-Scandium	200.8	N/A	---		0.9979	1.0	N/A	100	60 - 125	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	Selenium	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	Silver	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	Strontium	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:42	4299391

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
ICB	IS-Terbitum	200.8	N/A	---		1.0205	1.0	N/A	102	60 - 125	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	Thallium	200.8	0.3	---	<	0.3		ug/L	---	---	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	Thorium	200.8	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	Tin	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	Titanium	200.8	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	Uranium	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	Vanadium	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	IS-Yttrium	200.8	N/A	---		1.0140	1.0	N/A	101	60 - 125	---	---	1.0	---	05/21/2019 11:42	4299391
ICB	Zinc	200.8	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:42	4299391
LRB	Aluminum	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	Antimony	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	Arsenic	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	Barium	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	Beryllium	200.8	0.3	---	<	0.3		ug/L	---	---	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	IS-Bismuth	200.8	N/A	---		1.0089	1.0	N/A	101	60 - 125	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	Boron	200.8	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	Cadmium	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	Chromium	200.8	0.9	---	<	0.9		ug/L	---	---	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	Cobalt	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	Copper	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	IS-Indium	200.8	N/A	---		1.0084	1.0	N/A	101	60 - 125	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	Lead	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	Lithium	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	Manganese	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	Molybdenum	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	Nickel	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	IS-Scandium	200.8	N/A	---		0.9904	1.0	N/A	99	60 - 125	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	Selenium	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	Silver	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	Strontium	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	IS-Terbitum	200.8	N/A	---		1.0106	1.0	N/A	101	60 - 125	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	Thallium	200.8	0.3	---	<	0.3		ug/L	---	---	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	Thorium	200.8	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	Tin	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	Titanium	200.8	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	Uranium	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	Vanadium	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	IS-Yttrium	200.8	N/A	---		1.0057	1.0	N/A	101	60 - 125	---	---	1.0	---	05/21/2019 11:46	4299399
LRB	Zinc	200.8	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:46	4299399
LFB	Aluminum	200.8	2.0	---		49.5080	50.0	ug/L	99	85 - 115	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	Antimony	200.8	1.0	---		50.0610	50.0	ug/L	100	85 - 115	---	---	1.0	---	05/21/2019 11:51	4299400

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LFB	Arsenic	200.8	1.0	---		45.9270	50.0	ug/L	92	85 - 115	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	Barium	200.8	2.0	---		45.8850	50.0	ug/L	92	85 - 115	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	Beryllium	200.8	0.3	---		45.1060	50.0	ug/L	90	85 - 115	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	IS-Bismuth	200.8	N/A	---		0.9986	1.0	N/A	100	60 - 125	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	Boron	200.8	5.0	---		44.1710	50.0	ug/L	88	85 - 115	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	Cadmium	200.8	1.0	---		46.3530	50.0	ug/L	93	85 - 115	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	Chromium	200.8	0.9	---		45.7500	50.0	ug/L	92	85 - 115	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	Cobalt	200.8	2.0	---		46.0240	50.0	ug/L	92	85 - 115	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	Copper	200.8	1.0	---		49.1270	50.0	ug/L	98	85 - 115	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	IS-Indium	200.8	N/A	---		0.9994	1.0	N/A	100	60 - 125	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	Lead	200.8	1.0	---		48.0030	50.0	ug/L	96	85 - 115	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	Lithium	200.8	2.0	---		50.6540	50.0	ug/L	101	85 - 115	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	Manganese	200.8	2.0	---		46.5160	50.0	ug/L	93	85 - 115	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	Molybdenum	200.8	2.0	---		48.9340	50.0	ug/L	98	85 - 115	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	Nickel	200.8	1.0	---		45.7800	50.0	ug/L	92	85 - 115	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	IS-Scandium	200.8	N/A	---		0.9998	1.0	N/A	100	60 - 125	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	Selenium	200.8	2.0	---		45.4450	50.0	ug/L	91	85 - 115	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	Silver	200.8	2.0	---		49.5820	50.0	ug/L	99	85 - 115	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	Strontium	200.8	2.0	---		46.4500	50.0	ug/L	93	85 - 115	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	IS-Terbium	200.8	N/A	---		1.0108	1.0	N/A	101	60 - 125	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	Thallium	200.8	0.3	---		45.2410	50.0	ug/L	90	85 - 115	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	Thorium	200.8	5.0	---		42.9650	50.0	ug/L	86	85 - 115	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	Tin	200.8	2.0	---		48.6220	50.0	ug/L	97	85 - 115	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	Titanium	200.8	5.0	---		48.6990	50.0	ug/L	97	85 - 115	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	Uranium	200.8	1.0	---		45.3220	50.0	ug/L	91	85 - 115	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	Vanadium	200.8	2.0	---		46.3460	50.0	ug/L	93	85 - 115	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	IS-Yttrium	200.8	N/A	---		1.0046	1.0	N/A	100	60 - 125	---	---	1.0	---	05/21/2019 11:51	4299400
LFB	Zinc	200.8	5.0	---		46.3050	50.0	ug/L	93	85 - 115	---	---	1.0	---	05/21/2019 11:51	4299400
FS	Aluminum	200.8	2.0	Finished		35		ug/L	---	---	---	---	1.0	---	05/21/2019 11:55	4291770
FS	Antimony	200.8	1.0	Finished	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:55	4291770
FS	Arsenic	200.8	1.0	Finished	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:55	4291770
FS	Barium	200.8	2.0	Finished		17		ug/L	---	---	---	---	1.0	---	05/21/2019 11:55	4291770
FS	Beryllium	200.8	0.3	Finished	<	0.3		ug/L	---	---	---	---	1.0	---	05/21/2019 11:55	4291770
FS	IS-Bismuth	200.8	N/A	Finished		1.0282	1.0	N/A	103	60 - 125	---	---	1.0	---	05/21/2019 11:55	4291770
FS	Boron	200.8	5.0	Finished		22		ug/L	---	---	---	---	1.0	---	05/21/2019 11:55	4291770
FS	Cadmium	200.8	1.0	Finished	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:55	4291770
FS	Chromium	200.8	0.9	Finished	<	0.9		ug/L	---	---	---	---	1.0	---	05/21/2019 11:55	4291770
FS	Cobalt	200.8	2.0	Finished	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:55	4291770
FS	Copper	200.8	1.0	Finished	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:55	4291770
FS	IS-Indium	200.8	N/A	Finished		1.0374	1.0	N/A	104	60 - 125	---	---	1.0	---	05/21/2019 11:55	4291770
FS	Lead	200.8	1.0	Finished	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:55	4291770

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FS	Lithium	200.8	2.0	Finished	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:55	4291770
FS	Manganese	200.8	2.0	Finished		9.6		ug/L	---	---	---	---	1.0	---	05/21/2019 11:55	4291770
FS	Molybdenum	200.8	2.0	Finished	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:55	4291770
FS	Nickel	200.8	1.0	Finished	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:55	4291770
FS	IS-Scandium	200.8	N/A	Finished		1.0273	1.0	N/A	103	60 - 125	---	---	1.0	---	05/21/2019 11:55	4291770
FS	Selenium	200.8	2.0	Finished	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:55	4291770
FS	Silver	200.8	2.0	Finished	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:55	4291770
FS	Strontium	200.8	2.0	Finished		53		ug/L	---	---	---	---	1.0	---	05/21/2019 11:55	4291770
FS	IS-Terbium	200.8	N/A	Finished		1.0519	1.0	N/A	105	60 - 125	---	---	1.0	---	05/21/2019 11:55	4291770
FS	Thallium	200.8	0.3	Finished	<	0.3		ug/L	---	---	---	---	1.0	---	05/21/2019 11:55	4291770
FS	Thorium	200.8	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:55	4291770
FS	Tin	200.8	2.0	Finished	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:55	4291770
FS	Titanium	200.8	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:55	4291770
FS	Uranium	200.8	1.0	Finished	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:55	4291770
FS	Vanadium	200.8	2.0	Finished	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:55	4291770
FS	IS-Yttrium	200.8	N/A	Finished		1.0454	1.0	N/A	105	60 - 125	---	---	1.0	---	05/21/2019 11:55	4291770
FS	Zinc	200.8	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	---	05/21/2019 11:55	4291770
CCV	Aluminum	200.8	2.0	---		49.3780	50.0	ug/L	99	85 - 115	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	Antimony	200.8	1.0	---		48.6030	50.0	ug/L	97	85 - 115	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	Arsenic	200.8	1.0	---		47.6480	50.0	ug/L	95	85 - 115	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	Barium	200.8	2.0	---		48.1720	50.0	ug/L	96	85 - 115	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	Beryllium	200.8	0.3	---		47.5700	50.0	ug/L	95	85 - 115	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	IS-Bismuth	200.8	N/A	---		1.0101	1.0	N/A	101	60 - 125	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	Boron	200.8	5.0	---		48.7910	50.0	ug/L	98	85 - 115	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	Cadmium	200.8	1.0	---		47.4310	50.0	ug/L	95	85 - 115	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	Chromium	200.8	0.9	---		48.0480	50.0	ug/L	96	85 - 115	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	Cobalt	200.8	2.0	---		47.9220	50.0	ug/L	96	85 - 115	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	Copper	200.8	1.0	---		50.9590	50.0	ug/L	102	85 - 115	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	IS-Indium	200.8	N/A	---		1.0093	1.0	N/A	101	60 - 125	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	Lead	200.8	1.0	---		49.3450	50.0	ug/L	99	85 - 115	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	Lithium	200.8	2.0	---		50.9430	50.0	ug/L	102	85 - 115	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	Manganese	200.8	2.0	---		48.3320	50.0	ug/L	97	85 - 115	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	Molybdenum	200.8	2.0	---		48.2570	50.0	ug/L	97	85 - 115	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	Nickel	200.8	1.0	---		47.4980	50.0	ug/L	95	85 - 115	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	IS-Scandium	200.8	N/A	---		0.9986	1.0	N/A	100	60 - 125	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	Selenium	200.8	2.0	---		46.9190	50.0	ug/L	94	85 - 115	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	Silver	200.8	2.0	---		50.8150	50.0	ug/L	102	85 - 115	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	Strontium	200.8	2.0	---		48.3490	50.0	ug/L	97	85 - 115	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	IS-Terbium	200.8	N/A	---		1.0055	1.0	N/A	101	60 - 125	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	Thallium	200.8	0.3	---		46.8650	50.0	ug/L	94	85 - 115	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	Thorium	200.8	5.0	---		43.9010	50.0	ug/L	88	85 - 115	---	---	1.0	---	05/21/2019 12:23	4299373

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCV	Tin	200.8	2.0	---		48.0560	50.0	ug/L	96	85 - 115	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	Titanium	200.8	5.0	---		48.0880	50.0	ug/L	96	85 - 115	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	Uranium	200.8	1.0	---		46.8580	50.0	ug/L	94	85 - 115	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	Vanadium	200.8	2.0	---		48.2840	50.0	ug/L	97	85 - 115	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	IS-Yttrium	200.8	N/A	---		1.0068	1.0	N/A	101	60 - 125	---	---	1.0	---	05/21/2019 12:23	4299373
CCV	Zinc	200.8	5.0	---		47.8340	50.0	ug/L	96	85 - 115	---	---	1.0	---	05/21/2019 12:23	4299373
CCB	Aluminum	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	Antimony	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	Arsenic	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	Barium	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	Beryllium	200.8	0.3	---	<	0.3		ug/L	---	---	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	IS-Bismuth	200.8	N/A	---		1.0280	1.0	N/A	103	60 - 125	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	Boron	200.8	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	Cadmium	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	Chromium	200.8	0.9	---	<	0.9		ug/L	---	---	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	Cobalt	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	Copper	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	IS-Indium	200.8	N/A	---		1.0217	1.0	N/A	102	60 - 125	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	Lead	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	Lithium	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	Manganese	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	Molybdenum	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	Nickel	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	IS-Scandium	200.8	N/A	---		1.0020	1.0	N/A	100	60 - 125	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	Selenium	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	Silver	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	Strontium	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	IS-Terbium	200.8	N/A	---		1.0209	1.0	N/A	102	60 - 125	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	Thallium	200.8	0.3	---	<	0.3		ug/L	---	---	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	Thorium	200.8	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	Tin	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	Titanium	200.8	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	Uranium	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	Vanadium	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	IS-Yttrium	200.8	N/A	---		1.0211	1.0	N/A	102	60 - 125	---	---	1.0	---	05/21/2019 12:25	4299374
CCB	Zinc	200.8	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:25	4299374
CCV	Aluminum	200.8	2.0	---		48.5750	50.0	ug/L	97	85 - 115	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	Antimony	200.8	1.0	---		48.8950	50.0	ug/L	98	85 - 115	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	Arsenic	200.8	1.0	---		48.4920	50.0	ug/L	97	85 - 115	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	Barium	200.8	2.0	---		49.4150	50.0	ug/L	99	85 - 115	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	Beryllium	200.8	0.3	---		44.5860	50.0	ug/L	89	85 - 115	---	---	1.0	---	05/21/2019 12:54	4299395

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCV	IS-Bismuth	200.8	N/A	---		0.8785	1.0	N/A	88	60 - 125	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	Gadolinium	200.8	1.0	---		47.6030	50.0	ug/L	95	85 - 115	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	Chromium	200.8	0.9	---		47.3420	50.0	ug/L	95	85 - 115	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	Cobalt	200.8	2.0	---		47.5180	50.0	ug/L	95	85 - 115	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	Copper	200.8	1.0	---		53.0270	50.0	ug/L	106	85 - 115	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	IS-Indium	200.8	N/A	---		0.9309	1.0	N/A	93	60 - 125	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	Lead	200.8	1.0	---		49.8560	50.0	ug/L	100	85 - 115	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	Lithium	200.8	2.0	---		46.0750	50.0	ug/L	92	85 - 115	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	Manganese	200.8	2.0	---		47.5890	50.0	ug/L	95	85 - 115	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	Molybdenum	200.8	2.0	---		47.5020	50.0	ug/L	95	85 - 115	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	Nickel	200.8	1.0	---		46.8890	50.0	ug/L	94	85 - 115	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	IS-Scandium	200.8	N/A	---		1.0011	1.0	N/A	100	60 - 125	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	Selenium	200.8	2.0	---		47.3490	50.0	ug/L	95	85 - 115	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	Silver	200.8	2.0	---		51.5420	50.0	ug/L	103	85 - 115	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	Strontium	200.8	2.0	---		48.3320	50.0	ug/L	97	85 - 115	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	IS-Terbium	200.8	N/A	---		0.8996	1.0	N/A	90	60 - 125	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	Thallium	200.8	0.3	---		47.1390	50.0	ug/L	94	85 - 115	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	Thorium	200.8	5.0	---		44.3740	50.0	ug/L	89	85 - 115	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	Tin	200.8	2.0	---		47.5000	50.0	ug/L	95	85 - 115	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	Titanium	200.8	5.0	---		47.5240	50.0	ug/L	95	85 - 115	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	Uranium	200.8	1.0	---		46.6100	50.0	ug/L	93	85 - 115	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	Vanadium	200.8	2.0	---		47.6880	50.0	ug/L	95	85 - 115	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	IS-Yttrium	200.8	N/A	---		0.9593	1.0	N/A	96	60 - 125	---	---	1.0	---	05/21/2019 12:54	4299395
CCV	Zinc	200.8	5.0	---		49.0100	50.0	ug/L	98	85 - 115	---	---	1.0	---	05/21/2019 12:54	4299395
CCB	Aluminum	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	Antimony	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	Arsenic	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	Barium	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	Beryllium	200.8	0.3	---	<	0.3		ug/L	---	---	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	IS-Bismuth	200.8	N/A	---		0.8785	1.0	N/A	88	60 - 125	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	Gadolinium	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	Chromium	200.8	0.9	---	<	0.9		ug/L	---	---	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	Cobalt	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	Copper	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	IS-Indium	200.8	N/A	---		0.9326	1.0	N/A	93	60 - 125	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	Lead	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	Lithium	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	Manganese	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	Molybdenum	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	Nickel	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	IS-Scandium	200.8	N/A	---		0.9981	1.0	N/A	100	60 - 125	---	---	1.0	---	05/21/2019 12:57	4299396

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCB	Selenium	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	Silver	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	Strontium	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	IS-Terbium	200.8	N/A	---		0.9074	1.0	N/A	91	60 - 125	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	Thallium	200.8	0.3	---	<	0.3		ug/L	---	---	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	Thorium	200.8	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	Tin	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	Titanium	200.8	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	Uranium	200.8	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	Vanadium	200.8	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	IS-Yttrium	200.8	N/A	---		0.9533	1.0	N/A	95	60 - 125	---	---	1.0	---	05/21/2019 12:57	4299396
CCB	Zinc	200.8	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/21/2019 12:57	4299396



Eurofins Eaton Analytical

Run Log

Run ID: 259717 Method: 218.6

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
QCS	4267334		RW	DP	05/23/2019 21:02	
IPC	4267336		RW	DP	05/23/2019 21:28	
LRB	4267337		RW	DP	05/23/2019 21:41	
LFB	4267338		RW	DP	05/23/2019 21:54	
FS	4291772	Finished	DW	DP	05/23/2019 22:46	
IPC	4267339		RW	DP	05/24/2019 00:43	
CCB	4267340		RW	DP	05/24/2019 00:56	
CCB	4267372		RW	DP	05/24/2019 04:11	

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
QCS	Chromium, Hexavalent	218.6	0.02	---		1.0623	1.0	ug/L	106	90 - 110	---	---	1.0	---	05/23/2019 21:02	4267334
IPC	Chromium, Hexavalent	218.6	0.02	---		0.9868	1.0	ug/L	99	90 - 110	---	---	1.0	---	05/23/2019 21:28	4267336
LRB	Chromium, Hexavalent	218.6	0.02	---	<	0.02		ug/L	---	---	---	---	1.0	---	05/23/2019 21:41	4267337
LFB	Chromium, Hexavalent	218.6	0.02	---		0.9854	1.0	ug/L	99	90 - 110	---	---	1.0	---	05/23/2019 21:54	4267338
FS	Chromium, Hexavalent	218.6	0.02	Finished		0.06		ug/L	---	---	---	---	1.0	---	05/23/2019 22:46	4291772
IPC	Chromium, Hexavalent	218.6	0.02	---		1.0420	1.0	ug/L	104	90 - 110	---	---	1.0	---	05/24/2019 00:43	4267339
CCB	Chromium, Hexavalent	218.6	0.02	---	<	0.02		ug/L	---	---	---	---	1.0	---	05/24/2019 00:56	4267340
CCB	Chromium, Hexavalent	218.6	0.02	---	<	0.02		ug/L	---	---	---	---	1.0	---	05/24/2019 04:11	4267372

Eurofins Eaton Analytical

Run Log

Run ID: 259377 Method: 331.0

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
CCL	4292674		RW	PW6	05/15/2019 20:12	051519M331a.mdb
LMB	4292679		RW	PW6	05/15/2019 20:55	051519M331a.mdb
FBL	4292680		RW	PW6	05/15/2019 21:17	051519M331a.mdb
CCH	4292678		RW	PW6	05/16/2019 01:15	051519M331a.mdb
FS	4291771	Finished	DW	PW6	05/16/2019 03:03	051519M331a.mdb
MSL	4292681	Finished	DW	PW6	05/16/2019 03:24	051519M331a.mdb
SDL	4292690	Finished	DW	PW6	05/16/2019 03:46	051519M331a.mdb
CCL	4292675		RW	PW6	05/16/2019 05:34	051519M331a.mdb

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCL	Perchlorate	331.0	0.05	---		0.0487	0.05	ug/L	97	50 - 150	---	---	1.0	---	05/15/2019 20:12	4292674
LMB	Perchlorate	331.0	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/15/2019 20:55	4292679
FBL	Perchlorate	331.0	0.05	---		0.0452	0.05	ug/L	90	50 - 150	---	---	1.0	---	05/15/2019 21:17	4292680
CCH	Perchlorate	331.0	0.05	---		10.0277	10.0	ug/L	100	80 - 120	---	---	1.0	---	05/16/2019 01:15	4292678
FS	Perchlorate	331.0	0.05	Finished		0.12		ug/L	---	---	---	---	1.0	---	05/16/2019 03:03	4291771
MSL	Perchlorate	331.0	0.05	Finished		0.1941	0.169877	ug/L	148	50 - 150	---	---	1.0	---	05/16/2019 03:24	4292681
SDL	Perchlorate	331.0	0.05	Finished		0.1686	0.169877	ug/L	97	50 - 150	14	50	1.0	---	05/16/2019 03:46	4292690
CCL	Perchlorate	331.0	0.05	---		0.0519	0.05	ug/L	104	50 - 150	---	---	1.0	---	05/16/2019 05:34	4292675



Eurofins Eaton Analytical

Run Log

Run ID: 259065 Method: 515.3

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
CCC	4284778		RW	BS	05/10/2019 18:58	m 515.3 050619BS
CCC	4284779		RW	BS	05/11/2019 08:33	m 515.3 050619BS
CCC	4284780		RW	BS	05/11/2019 17:21	m 515.3 050619BS

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	IS-4,4'-Dibromodifluorobiphenyl	515.3	N/A	--		13738	13526	ug/L	102	70 - 130	---	---	1.0	05/08/2019 08:00	05/10/2019 18:58	4284778
CCC	SS-2,4-Dichlorophenylacetic acid	515.3	N/A	--		24.5458	25.0	ug/L	98	70 - 130	---	---	1.0	05/08/2019 08:00	05/10/2019 18:58	4284778
CCC	Bentazon	515.3	0.5	--		0.9525	1.0	ug/L	95	70 - 130	---	---	1.0	05/08/2019 08:00	05/10/2019 18:58	4284778
CCC	2,4-D	515.3	0.1	--		0.8766	1.0	ug/L	88	70 - 130	---	---	1.0	05/08/2019 08:00	05/10/2019 18:58	4284778
CCC	Dalapon	515.3	1.0	--		0.9316	1.0	ug/L	93	70 - 130	---	---	1.0	05/08/2019 08:00	05/10/2019 18:58	4284778
CCC	Dicamba	515.3	0.1	--		0.9558	1.0	ug/L	96	70 - 130	---	---	1.0	05/08/2019 08:00	05/10/2019 18:58	4284778
CCC	Dinoseb	515.3	0.1	--		0.9858	1.0	ug/L	99	70 - 130	---	---	1.0	05/08/2019 08:00	05/10/2019 18:58	4284778
CCC	Pentachlorophenol	515.3	0.04	--		0.2045	0.2	ug/L	102	70 - 130	---	---	1.0	05/08/2019 08:00	05/10/2019 18:58	4284778
CCC	Picloram	515.3	0.1	--		0.3575	0.5	ug/L	72	70 - 130	---	---	1.0	05/08/2019 08:00	05/10/2019 18:58	4284778
CCC	2,4,5-TP (Silvex)	515.3	0.1	--		0.4817	0.5	ug/L	96	70 - 130	---	---	1.0	05/08/2019 08:00	05/10/2019 18:58	4284778
CCC	IS-4,4'-Dibromodifluorobiphenyl	515.3	N/A	--		13936	13526	ug/L	103	70 - 130	---	---	1.0	05/08/2019 08:00	05/11/2019 08:33	4284779
CCC	SS-2,4-Dichlorophenylacetic acid	515.3	N/A	--		20.5466	25.0	ug/L	82	70 - 130	---	---	1.0	05/08/2019 08:00	05/11/2019 08:33	4284779
CCC	Bentazon	515.3	0.5	--		2.7397	3.0	ug/L	91	70 - 130	---	---	1.0	05/08/2019 08:00	05/11/2019 08:33	4284779
CCC	2,4-D	515.3	0.1	--		2.7408	3.0	ug/L	91	70 - 130	---	---	1.0	05/08/2019 08:00	05/11/2019 08:33	4284779
CCC	Dalapon	515.3	1.0	--		2.8621	3.0	ug/L	95	70 - 130	---	---	1.0	05/08/2019 08:00	05/11/2019 08:33	4284779
CCC	Dicamba	515.3	0.1	--		2.7382	3.0	ug/L	91	70 - 130	---	---	1.0	05/08/2019 08:00	05/11/2019 08:33	4284779
CCC	Dinoseb	515.3	0.1	--		2.7049	3.0	ug/L	90	70 - 130	---	---	1.0	05/08/2019 08:00	05/11/2019 08:33	4284779
CCC	Pentachlorophenol	515.3	0.04	--		0.5303	0.6	ug/L	88	70 - 130	---	---	1.0	05/08/2019 08:00	05/11/2019 08:33	4284779
CCC	Picloram	515.3	0.1	--		1.2976	1.5	ug/L	87	70 - 130	---	---	1.0	05/08/2019 08:00	05/11/2019 08:33	4284779
CCC	2,4,5-TP (Silvex)	515.3	0.1	--		1.3190	1.5	ug/L	88	70 - 130	---	---	1.0	05/08/2019 08:00	05/11/2019 08:33	4284779
CCC	IS-4,4'-Dibromodifluorobiphenyl	515.3	N/A	--		13928	13526	ug/L	103	70 - 130	---	---	1.0	05/08/2019 08:00	05/11/2019 17:21	4284780
CCC	SS-2,4-Dichlorophenylacetic acid	515.3	N/A	--		20.5035	25.0	ug/L	82	70 - 130	---	---	1.0	05/08/2019 08:00	05/11/2019 17:21	4284780
CCC	Bentazon	515.3	0.5	--		2.7161	3.0	ug/L	91	70 - 130	---	---	1.0	05/08/2019 08:00	05/11/2019 17:21	4284780
CCC	2,4-D	515.3	0.1	--		2.7189	3.0	ug/L	91	70 - 130	---	---	1.0	05/08/2019 08:00	05/11/2019 17:21	4284780
CCC	Dalapon	515.3	1.0	--		2.8946	3.0	ug/L	96	70 - 130	---	---	1.0	05/08/2019 08:00	05/11/2019 17:21	4284780
CCC	Dicamba	515.3	0.1	--		2.7352	3.0	ug/L	91	70 - 130	---	---	1.0	05/08/2019 08:00	05/11/2019 17:21	4284780
CCC	Dinoseb	515.3	0.1	--		2.7088	3.0	ug/L	90	70 - 130	---	---	1.0	05/08/2019 08:00	05/11/2019 17:21	4284780
CCC	Pentachlorophenol	515.3	0.04	--		0.5249	0.6	ug/L	87	70 - 130	---	---	1.0	05/08/2019 08:00	05/11/2019 17:21	4284780
CCC	Picloram	515.3	0.1	--		1.2776	1.5	ug/L	85	70 - 130	---	---	1.0	05/08/2019 08:00	05/11/2019 17:21	4284780
CCC	2,4,5-TP (Silvex)	515.3	0.1	--		1.3124	1.5	ug/L	87	70 - 130	---	---	1.0	05/08/2019 08:00	05/11/2019 17:21	4284780

Eurofins Eaton Analytical

Run Log

Run ID: 259507 Method: 515.3

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
CCC	4293309		RW	BS	05/20/2019 16:01	m 515.3 051319BS
LMB	4293308		RW	BS	05/20/2019 17:37	m 515.3 051319BS
FS	4291757	Finished	DW	BS	05/21/2019 04:49	m 515.3 051319BS
CCC	4293310		RW	BS	05/21/2019 09:37	m 515.3 051319BS

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	IS-4,4'-Dibromooctafluorobiphenyl	515.3	N/A	--		13738	12215	ug/L	112	70 - 130	---	---	1.0	05/15/2019 08:00	05/20/2019 16:01	4293309
CCC	SS-2,4-Dichlorophenylacetic acid	515.3	N/A	--		23.6451	25.0	ug/L	95	70 - 130	---	---	1.0	05/15/2019 08:00	05/20/2019 16:01	4293309
CCC	Acifluorfen	515.3	1.0	--		1.0280	1.0	ug/L	103	70 - 130	---	---	1.0	05/15/2019 08:00	05/20/2019 16:01	4293309
CCC	Bentazon	515.3	0.5	--		1.0562	1.0	ug/L	106	70 - 130	---	---	1.0	05/15/2019 08:00	05/20/2019 16:01	4293309
CCC	Chloramben	515.3	2.0	--		0.7598	1.0	ug/L	76	70 - 130	---	---	1.0	05/15/2019 08:00	05/20/2019 16:01	4293309
CCC	2,4-D	515.3	0.1	--		0.8113	1.0	ug/L	81	70 - 130	---	---	1.0	05/15/2019 08:00	05/20/2019 16:01	4293309
CCC	Dalapon	515.3	1.0	--		1.0042	1.0	ug/L	100	70 - 130	---	---	1.0	05/15/2019 08:00	05/20/2019 16:01	4293309
CCC	2,4-DB	515.3	2.0	--		0.7912	1.0	ug/L	79	70 - 130	---	---	1.0	05/15/2019 08:00	05/20/2019 16:01	4293309
CCC	DCPA acid metabolites	515.3	0.5	--		1.1128	1.0	ug/L	111	70 - 130	---	---	1.0	05/15/2019 08:00	05/20/2019 16:01	4293309
CCC	Dicamba	515.3	0.1	--		0.9765	1.0	ug/L	98	70 - 130	---	---	1.0	05/15/2019 08:00	05/20/2019 16:01	4293309
CCC	3,5-Dichlorobenzoic acid	515.3	0.5	--		0.8550	1.0	ug/L	86	70 - 130	---	---	1.0	05/15/2019 08:00	05/20/2019 16:01	4293309
CCC	Dichlorprop	515.3	2.0	--		0.9010	1.0	ug/L	90	70 - 130	---	---	1.0	05/15/2019 08:00	05/20/2019 16:01	4293309
CCC	Dinoseb	515.3	0.1	--		1.1099	1.0	ug/L	111	70 - 130	---	---	1.0	05/15/2019 08:00	05/20/2019 16:01	4293309
CCC	Pentachlorophenol	515.3	0.04	--		0.2034	0.2	ug/L	102	70 - 130	---	---	1.0	05/15/2019 08:00	05/20/2019 16:01	4293309
CCC	Picloram	515.3	0.1	--		0.4516	0.5	ug/L	90	70 - 130	---	---	1.0	05/15/2019 08:00	05/20/2019 16:01	4293309
CCC	2,4,5-T	515.3	0.5	--		0.8418	1.0	ug/L	84	70 - 130	---	---	1.0	05/15/2019 08:00	05/20/2019 16:01	4293309
CCC	2,4,5-TP (Silvex)	515.3	0.1	--		0.5145	0.5	ug/L	103	70 - 130	---	---	1.0	05/15/2019 08:00	05/20/2019 16:01	4293309
CCC	Triclopyr	515.3	0.5	--		0.9982	1.0	ug/L	100	69 - 141	---	---	1.0	05/15/2019 08:00	05/20/2019 16:01	4293309
LMB	IS-4,4'-Dibromooctafluorobiphenyl	515.3	N/A	--		12820	12215	ug/L	105	70 - 130	---	---	1.0	05/15/2019 08:00	05/20/2019 17:37	4293308
LMB	SS-2,4-Dichlorophenylacetic acid	515.3	N/A	--		21.0373	25.0	ug/L	84	70 - 130	---	---	1.0	05/15/2019 08:00	05/20/2019 17:37	4293308
LMB	Acifluorfen	515.3	1.0	--	<	1.0		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/20/2019 17:37	4293308
LMB	Bentazon	515.3	0.5	--	<	0.5		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/20/2019 17:37	4293308
LMB	Chloramben	515.3	2.0	--	<	2.0		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/20/2019 17:37	4293308
LMB	2,4-D	515.3	0.1	--	<	0.1		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/20/2019 17:37	4293308
LMB	Dalapon	515.3	1.0	--	<	1.0		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/20/2019 17:37	4293308
LMB	2,4-DB	515.3	2.0	--	<	2.0		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/20/2019 17:37	4293308
LMB	DCPA acid metabolites	515.3	0.5	--	<	0.5		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/20/2019 17:37	4293308
LMB	Dicamba	515.3	0.1	--	<	0.1		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/20/2019 17:37	4293308
LMB	3,5-Dichlorobenzoic acid	515.3	0.5	--	<	0.5		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/20/2019 17:37	4293308
LMB	Dichlorprop	515.3	2.0	--	<	2.0		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/20/2019 17:37	4293308
LMB	Dinoseb	515.3	0.1	--	<	0.1		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/20/2019 17:37	4293308
LMB	Pentachlorophenol	515.3	0.04	--	<	0.04		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/20/2019 17:37	4293308
LMB	Picloram	515.3	0.1	--	<	0.1		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/20/2019 17:37	4293308
LMB	2,4,5-T	515.3	0.5	--	<	0.5		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/20/2019 17:37	4293308
LMB	2,4,5-TP (Silvex)	515.3	0.1	--	<	0.1		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/20/2019 17:37	4293308
LMB	Triclopyr	515.3	0.5	--	<	0.5		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/20/2019 17:37	4293308
DF	IS-4,4'-Dibromooctafluorobiphenyl	515.3	N/A	Finished		12958	12215	ug/L	106	70 - 130	---	---	1.0	05/15/2019 08:00	05/21/2019 04:49	4291757
DF	SS-2,4-Dichlorophenylacetic acid	515.3	N/A	Finished		18.4651	25.0	ug/L	74	70 - 130	---	---	1.0	05/15/2019 08:00	05/21/2019 04:49	4291757
DF	Acifluorfen	515.3	1.0	Finished	<	1.0		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/21/2019 04:49	4291757
DF	Bentazon	515.3	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/21/2019 04:49	4291757

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FS	Chloramben	515.3	2.0	Finished	<	2.0		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/21/2019 04:49	4291757
FS	2,4-D	515.3	0.1	Finished	<	0.1		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/21/2019 04:49	4291757
FS	Dalapon	515.3	1.0	Finished	<	1.0		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/21/2019 04:49	4291757
FS	2,4-DB	515.3	2.0	Finished	<	2.0		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/21/2019 04:49	4291757
FS	DCPA acid metabolites	515.3	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/21/2019 04:49	4291757
FS	Dicamba	515.3	0.1	Finished	<	0.1		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/21/2019 04:49	4291757
FS	3,5-Dichlorobenzoic acid	515.3	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/21/2019 04:49	4291757
FS	Dichlorprop	515.3	2.0	Finished	<	2.0		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/21/2019 04:49	4291757
FS	Dinoseb	515.3	0.1	Finished	<	0.1		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/21/2019 04:49	4291757
FS	Pentachlorophenol	515.3	0.04	Finished	<	0.04		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/21/2019 04:49	4291757
FS	Picloram	515.3	0.1	Finished	<	0.1		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/21/2019 04:49	4291757
FS	2,4,5-T	515.3	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/21/2019 04:49	4291757
FS	2,4,5-TP (Silvex)	515.3	0.1	Finished	<	0.1		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/21/2019 04:49	4291757
FS	Triclopyr	515.3	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/21/2019 04:49	4291757
CCC	IS-4'-Dibromooctylfluorobiphenyl	515.3	N/A	---		13572	12215	ug/L	111	70 - 130	---	---	1.0	05/15/2019 08:00	05/21/2019 09:37	4293310
CCC	SS-2,4-Dichlorophenylacetic acid	515.3	N/A	---		21,0844	25.0	ug/L	84	70 - 130	---	---	1.0	05/15/2019 08:00	05/21/2019 09:37	4293310
CCC	Acifluorfen	515.3	1.0	---		2,9104	3.0	ug/L	97	70 - 130	---	---	1.0	05/15/2019 08:00	05/21/2019 09:37	4293310
CCC	Bentazon	515.3	0.5	---		2,8902	3.0	ug/L	96	70 - 130	---	---	1.0	05/15/2019 08:00	05/21/2019 09:37	4293310
CCC	Chloramben	515.3	2.0	---		2,2232	3.0	ug/L	74	70 - 130	---	---	1.0	05/15/2019 08:00	05/21/2019 09:37	4293310
CCC	2,4-D	515.3	0.1	---		2,6093	3.0	ug/L	87	70 - 130	---	---	1.0	05/15/2019 08:00	05/21/2019 09:37	4293310
CCC	Dalapon	515.3	1.0	---		2,5066	3.0	ug/L	84	70 - 130	---	---	1.0	05/15/2019 08:00	05/21/2019 09:37	4293310
CCC	2,4-DB	515.3	2.0	---		2,7538	3.0	ug/L	92	70 - 130	---	---	1.0	05/15/2019 08:00	05/21/2019 09:37	4293310
CCC	DCPA acid metabolites	515.3	0.5	---		2,6011	3.0	ug/L	87	70 - 130	---	---	1.0	05/15/2019 08:00	05/21/2019 09:37	4293310
CCC	Dicamba	515.3	0.1	---		2,5342	3.0	ug/L	84	70 - 130	---	---	1.0	05/15/2019 08:00	05/21/2019 09:37	4293310
CCC	3,5-Dichlorobenzoic acid	515.3	0.5	---		2,8722	3.0	ug/L	96	70 - 130	---	---	1.0	05/15/2019 08:00	05/21/2019 09:37	4293310
CCC	Dichlorprop	515.3	2.0	---		2,5286	3.0	ug/L	84	70 - 130	---	---	1.0	05/15/2019 08:00	05/21/2019 09:37	4293310
CCC	Dinoseb	515.3	0.1	---		2,9057	3.0	ug/L	97	70 - 130	---	---	1.0	05/15/2019 08:00	05/21/2019 09:37	4293310
CCC	Pentachlorophenol	515.3	0.04	---		0,5045	0.6	ug/L	84	70 - 130	---	---	1.0	05/15/2019 08:00	05/21/2019 09:37	4293310
CCC	Picloram	515.3	0.1	---		1,2916	1.5	ug/L	86	70 - 130	---	---	1.0	05/15/2019 08:00	05/21/2019 09:37	4293310
CCC	2,4,5-T	515.3	0.5	---		2,5208	3.0	ug/L	84	70 - 130	---	---	1.0	05/15/2019 08:00	05/21/2019 09:37	4293310
CCC	2,4,5-TP (Silvex)	515.3	0.1	---		1,2957	1.5	ug/L	86	70 - 130	---	---	1.0	05/15/2019 08:00	05/21/2019 09:37	4293310
CCC	Triclopyr	515.3	0.5	---		2,4789	3.0	ug/L	83	69 - 141	---	---	1.0	05/15/2019 08:00	05/21/2019 09:37	4293310

Eurofins Eaton Analytical

Run Log

Run ID: 259594 Method: 515.3

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
CCC	4293309		RW	CC	05/21/2019 16:17	515 3-042219CC-MS.mth
LMB	4293308		RW	CC	05/21/2019 16:41	515 3-042219CC-MS.mth
FS	4291757	Finished	DW	CC	05/21/2019 19:55	515 3-042219CC-MS.mth
CCC	4293310		RW	CC	05/21/2019 21:08	515 3-042219CC-MS.mth

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	MCPA	515.3	0.5	---		1.2790	1.0	ug/L	128	60 - 152	---	---	1.0	05/15/2019 08:00	05/21/2019 16:17	4293309
CCC	Mecoprop	515.3	0.5	---		1.2520	1.0	ug/L	125	62 - 154	---	---	1.0	05/15/2019 08:00	05/21/2019 16:17	4293309
LMB	IS-4,4'-Dibromooctafuorobiphenyl	515.3	N/A	---		28880	35367	ug/L	82	70 - 130	---	---	1.0	05/15/2019 08:00	05/21/2019 16:41	4293308
LMB	SS-2,4-Dichlorophenylacetic acid	515.3	N/A	---		27.4640	25.0	ug/L	110	70 - 130	---	---	1.0	05/15/2019 08:00	05/21/2019 16:41	4293308
LMB	MCPA	515.3	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/21/2019 16:41	4293308
LMB	Mecoprop	515.3	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/21/2019 16:41	4293308
FS	MCPA	515.3	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/21/2019 19:55	4291757
FS	Mecoprop	515.3	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	05/15/2019 08:00	05/21/2019 19:55	4291757
CCC	MCPA	515.3	0.5	---		3.2340	3.0	ug/L	108	60 - 152	---	---	1.0	05/15/2019 08:00	05/21/2019 21:08	4293310
CCC	Mecoprop	515.3	0.5	---		3.3030	3.0	ug/L	110	62 - 154	---	---	1.0	05/15/2019 08:00	05/21/2019 21:08	4293310



Eurofins Eaton Analytical

Run Log

Run ID: 259699 Method: 521

Type	Sample Id	Sample Site	Matrix	Instrument ID	Analysis Date	Calibration File
CCL	4293670		OS	DF	05/21/2019 22:44	521-031819-DF-up5.mth
LMB	4293658		RW	DF	05/22/2019 03:11	521-031819-DF-up5.mth
FBM	4293659		RW	DF	05/22/2019 03:50	521-031819-DF-up5.mth
CCM	4293671		OS	DF	05/22/2019 14:02	521-031819-DF-up5.mth
FS	4291758	Finished	DW	DF	05/22/2019 17:16	521-031819-DF-up5.mth
CCH	4293673		OS	DF	05/23/2019 00:54	521-031819-DF-up5.mth

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCL	IS-NDPA-d14	521	N/A	--		38265	51243	ng/L	75	50 - 150	---	---	1.0	05/15/2019 09:30	05/21/2019 22:44	42936670
CCL	SS-NDMA-d6	521	N/A	--		2.2200	2.0	ng/L	111	70 - 130	---	---	1.0	05/15/2019 09:30	05/21/2019 22:44	42936670
CCL	N-Nitrosopyrrolidine (NPYR)	521	2.0	--		1.3820	2.0	ng/L	69	50 - 150	---	---	1.0	05/15/2019 09:30	05/21/2019 22:44	42936670
CCL	N-Nitrosodi-N-butylamine (NDBA)	521	2.0	--		1.2710	2.0	ng/L	64	50 - 150	---	---	1.0	05/15/2019 09:30	05/21/2019 22:44	42936670
CCL	N-Nitrosodimethylamine (NDEA)	521	2.0	--		1.0380	2.0	ng/L	52	50 - 150	---	---	1.0	05/15/2019 09:30	05/21/2019 22:44	42936670
CCL	N-Nitrosodimethylamine (NDMA)	521	2.0	--		2.1410	2.0	ng/L	107	50 - 150	---	---	1.0	05/15/2019 09:30	05/21/2019 22:44	42936670
CCL	N-Nitrosodi-N-propylamine (NDPA)	521	2.0	--		2.3130	2.0	ng/L	116	50 - 150	---	---	1.0	05/15/2019 09:30	05/21/2019 22:44	42936670
CCL	N-Nitrosodiphenylamine (NDPhA)	521	20	--		12.5040	10.0	ng/L	125	3 - 250	---	---	1.0	05/15/2019 09:30	05/21/2019 22:44	42936670
CCL	N-Nitrosomethylamine (NMEA)	521	2.0	--		1.0040	2.0	ng/L	50	50 - 150	---	---	1.0	05/15/2019 09:30	05/21/2019 22:44	42936670
CCL	N-Nitrosomorpholine (NMOR)	521	2.0	--		1.9380	2.0	ng/L	97	50 - 150	---	---	1.0	05/15/2019 09:30	05/21/2019 22:44	42936670
CCL	N-Nitrosopiperidine (NPIP)	521	2.0	--		2.5850	2.0	ng/L	129	50 - 150	---	---	1.0	05/15/2019 09:30	05/21/2019 22:44	42936670
LMB	IS-NDPA-d14	521	N/A	--		40311	38265	ng/L	105	50 - 150	---	---	1.0	05/15/2019 09:30	05/22/2019 03:11	42936668
LMB	SS-NDMA-d6	521	N/A	--		16.4960	20.0	ng/L	82	70 - 130	---	---	1.0	05/15/2019 09:30	05/22/2019 03:11	42936668
LMB	N-Nitrosopyrrolidine (NPYR)	521	2.0	--		2.0		ng/L	---	---	---	---	1.0	05/15/2019 09:30	05/22/2019 03:11	42936668
LMB	N-Nitrosodi-N-butylamine (NDBA)	521	2.0	--		2.0		ng/L	---	---	---	---	1.0	05/15/2019 09:30	05/22/2019 03:11	42936668
LMB	N-Nitrosodimethylamine (NDEA)	521	2.0	--		2.0		ng/L	---	---	---	---	1.0	05/15/2019 09:30	05/22/2019 03:11	42936668
LMB	N-Nitrosodimethylamine (NDMA)	521	2.0	--		2.0		ng/L	---	---	---	---	1.0	05/15/2019 09:30	05/22/2019 03:11	42936668
LMB	N-Nitrosodi-N-propylamine (NDPA)	521	2.0	--		2.0		ng/L	---	---	---	---	1.0	05/15/2019 09:30	05/22/2019 03:11	42936668
LMB	N-Nitrosodiphenylamine (NDPhA)	521	20	--		20		ng/L	---	---	---	---	1.0	05/15/2019 09:30	05/22/2019 03:11	42936668
LMB	N-Nitrosomethylamine (NMEA)	521	2.0	--		2.0		ng/L	---	---	---	---	1.0	05/15/2019 09:30	05/22/2019 03:11	42936668
LMB	N-Nitrosomorpholine (NMOR)	521	2.0	--		2.0		ng/L	---	---	---	---	1.0	05/15/2019 09:30	05/22/2019 03:11	42936668
LMB	N-Nitrosopiperidine (NPIP)	521	2.0	--		2.0		ng/L	---	---	---	---	1.0	05/15/2019 09:30	05/22/2019 03:11	42936668
FBM	IS-NDPA-d14	521	N/A	--		38396	38265	ng/L	100	50 - 150	---	---	1.0	05/15/2019 09:30	05/22/2019 03:50	42936659
FBM	SS-NDMA-d6	521	N/A	--		18.2710	20.0	ng/L	91	70 - 130	---	---	1.0	05/15/2019 09:30	05/22/2019 03:50	42936659
FBM	N-Nitrosopyrrolidine (NPYR)	521	2.0	--		17.5830	20.0	ng/L	88	70 - 130	---	---	1.0	05/15/2019 09:30	05/22/2019 03:50	42936659
FBM	N-Nitrosodi-N-butylamine (NDBA)	521	2.0	--		17.2250	20.0	ng/L	86	70 - 130	---	---	1.0	05/15/2019 09:30	05/22/2019 03:50	42936659
FBM	N-Nitrosodimethylamine (NDEA)	521	2.0	--		17.6660	20.0	ng/L	88	70 - 130	---	---	1.0	05/15/2019 09:30	05/22/2019 03:50	42936659
FBM	N-Nitrosodimethylamine (NDMA)	521	2.0	--		20.0640	20.0	ng/L	100	70 - 130	---	---	1.0	05/15/2019 09:30	05/22/2019 03:50	42936659
FBM	N-Nitrosodi-N-propylamine (NDPA)	521	2.0	--		18.6760	20.0	ng/L	93	70 - 130	---	---	1.0	05/15/2019 09:30	05/22/2019 03:50	42936659
FBM	N-Nitrosodiphenylamine (NDPhA)	521	20	--		14.8450	100	ng/L	15	1 - 120	---	---	1.0	05/15/2019 09:30	05/22/2019 03:50	42936659
FBM	N-Nitrosomethylamine (NMEA)	521	2.0	--		18.7670	20.0	ng/L	94	70 - 130	---	---	1.0	05/15/2019 09:30	05/22/2019 03:50	42936659
FBM	N-Nitrosomorpholine (NMOR)	521	2.0	--		20.5320	20.0	ng/L	103	70 - 130	---	---	1.0	05/15/2019 09:30	05/22/2019 03:50	42936659
FBM	N-Nitrosopiperidine (NPIP)	521	2.0	--		17.1720	20.0	ng/L	86	70 - 130	---	---	1.0	05/15/2019 09:30	05/22/2019 03:50	42936659
CCM	IS-NDPA-d14	521	N/A	--		32313	32313	ng/L	100	50 - 150	---	---	1.0	05/15/2019 09:30	05/22/2019 14:02	42936671
CCM	SS-NDMA-d6	521	N/A	--		16.4030	20.0	ng/L	82	70 - 130	---	---	1.0	05/15/2019 09:30	05/22/2019 14:02	42936671
CCM	N-Nitrosopyrrolidine (NPYR)	521	2.0	--		20.0680	20.0	ng/L	100	70 - 130	---	---	1.0	05/15/2019 09:30	05/22/2019 14:02	42936671
CCM	N-Nitrosodi-N-butylamine (NDBA)	521	2.0	--		22.2280	20.0	ng/L	111	70 - 130	---	---	1.0	05/15/2019 09:30	05/22/2019 14:02	42936671
CCM	N-Nitrosodimethylamine (NDEA)	521	2.0	--		19.9870	20.0	ng/L	100	70 - 130	---	---	1.0	05/15/2019 09:30	05/22/2019 14:02	42936671
CCM	N-Nitrosodimethylamine (NDMA)	521	2.0	--		16.5500	20.0	ng/L	83	70 - 130	---	---	1.0	05/15/2019 09:30	05/22/2019 14:02	42936671
CCM	N-Nitrosodi-N-propylamine (NDPA)	521	2.0	--		17.6410	20.0	ng/L	88	70 - 130	---	---	1.0	05/15/2019 09:30	05/22/2019 14:02	42936671

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCM	N-Nitrosodiphenylamine (NDPhA)	521	20	---		135.2910	100	ng/L	135	20 - 250	---	---	1.0	05/15/2019 09:30	05/22/2019 14:02	4293671
CCM	N-Nitrosomethylethylamine (NMEA)	521	2.0	---		19.5780	20.0	ng/L	98	70 - 130	---	---	1.0	05/15/2019 09:30	05/22/2019 14:02	4293671
CCM	N-Nitrosomorpholine (NMOR)	521	2.0	---		20.2970	20.0	ng/L	101	70 - 130	---	---	1.0	05/15/2019 09:30	05/22/2019 14:02	4293671
CCM	N-Nitrosopiperidine (NPIP)	521	2.0	---		19.0270	20.0	ng/L	95	70 - 130	---	---	1.0	05/15/2019 09:30	05/22/2019 14:02	4293671
FS	IS-NDPA-d14	521	N/A	Finished		38795	32313	ng/L	120	50 - 150	---	---	1.0	05/15/2019 09:30	05/22/2019 17:16	4291758
FS	SS-NDMA-d6	521	N/A	Finished		16.2370	20.0	ng/L	81	70 - 130	---	---	1.0	05/15/2019 09:30	05/22/2019 17:16	4291758
FS	N-Nitrosopyrrolidine (NPYR)	521	2.0	Finished	<	2.0		ng/L	---	---	---	---	1.0	05/15/2019 09:30	05/22/2019 17:16	4291758
FS	N-Nitrosodi-N-butylamine (NDBA)	521	2.0	Finished	<	2.0		ng/L	---	---	---	---	1.0	05/15/2019 09:30	05/22/2019 17:16	4291758
FS	N-Nitrosodiethylamine (NDEA)	521	2.0	Finished	<	2.0		ng/L	---	---	---	---	1.0	05/15/2019 09:30	05/22/2019 17:16	4291758
FS	N-Nitrosodimethylamine (NDMA)	521	2.0	Finished		5.1		ng/L	---	---	---	---	1.0	05/15/2019 09:30	05/22/2019 17:16	4291758
FS	N-Nitrosodi-N-propylamine (NDPA)	521	2.0	Finished	<	2.0		ng/L	---	---	---	---	1.0	05/15/2019 09:30	05/22/2019 17:16	4291758
FS	N-Nitrosodiphenylamine (NDPhA)	521	20	Finished	<	20		ng/L	---	---	---	---	1.0	05/15/2019 09:30	05/22/2019 17:16	4291758
FS	N-Nitrosomethylethylamine (NMEA)	521	2.0	Finished	<	2.0		ng/L	---	---	---	---	1.0	05/15/2019 09:30	05/22/2019 17:16	4291758
FS	N-Nitrosomorpholine (NMOR)	521	2.0	Finished	<	2.0		ng/L	---	---	---	---	1.0	05/15/2019 09:30	05/22/2019 17:16	4291758
FS	N-Nitrosopiperidine (NPIP)	521	2.0	Finished	<	2.0		ng/L	---	---	---	---	1.0	05/15/2019 09:30	05/22/2019 17:16	4291758
CCH	IS-NDPA-d14	521	N/A	---		36307	36307	ng/L	100	50 - 150	---	---	1.0	05/15/2019 09:30	05/23/2019 00:54	4293673
CCH	SS-NDMA-d6	521	N/A	---		44.4730	40.0	ng/L	111	70 - 130	---	---	1.0	05/15/2019 09:30	05/23/2019 00:54	4293673
CCH	N-Nitrosopyrrolidine (NPYR)	521	2.0	---		43.9830	40.0	ng/L	110	70 - 130	---	---	1.0	05/15/2019 09:30	05/23/2019 00:54	4293673
CCH	N-Nitrosodi-N-butylamine (NDBA)	521	2.0	---		36.7420	40.0	ng/L	92	70 - 130	---	---	1.0	05/15/2019 09:30	05/23/2019 00:54	4293673
CCH	N-Nitrosodiethylamine (NDEA)	521	2.0	---		42.2590	40.0	ng/L	106	70 - 130	---	---	1.0	05/15/2019 09:30	05/23/2019 00:54	4293673
CCH	N-Nitrosodimethylamine (NDMA)	521	2.0	---		44.2060	40.0	ng/L	111	70 - 130	---	---	1.0	05/15/2019 09:30	05/23/2019 00:54	4293673
CCH	N-Nitrosodi-N-propylamine (NDPA)	521	2.0	---		35.9220	40.0	ng/L	90	70 - 130	---	---	1.0	05/15/2019 09:30	05/23/2019 00:54	4293673
CCH	N-Nitrosodiphenylamine (NDPhA)	521	20	---		228.1710	200	ng/L	114	20 - 250	---	---	1.0	05/15/2019 09:30	05/23/2019 00:54	4293673
CCH	N-Nitrosomethylethylamine (NMEA)	521	2.0	---		43.9730	40.0	ng/L	110	70 - 130	---	---	1.0	05/15/2019 09:30	05/23/2019 00:54	4293673
CCH	N-Nitrosomorpholine (NMOR)	521	2.0	---		41.5870	40.0	ng/L	104	70 - 130	---	---	1.0	05/15/2019 09:30	05/23/2019 00:54	4293673
CCH	N-Nitrosopiperidine (NPIP)	521	2.0	---		39.6890	40.0	ng/L	99	70 - 130	---	---	1.0	05/15/2019 09:30	05/23/2019 00:54	4293673



Eaton Analytical

Eurofins Eaton Analytical

Run Log

Run ID: 259867 Method: 521

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
CCL	4304731		OS	DF	05/23/2019 15:11	521-031819-DF-up5.mth
CCM	4304733		OS	DF	05/24/2019 01:21	521-031819-DF-up5.mth

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCL	IS-NDPA-d14	521	N/A	---		51922	51243	ng/L	101	50 - 150	---	---	1.0	05/15/2019 09:30	05/23/2019 15:11	4304731
CCL	SS-NDMA-d6	521	N/A	---		2.1340	2.0	ng/L	107	70 - 130	---	---	1.0	05/15/2019 09:30	05/23/2019 15:11	4304731
CCL	N-Nitrosopyrrolidine (NPYR)	521	2.0	---		1.3350	2.0	ng/L	67	50 - 150	---	---	1.0	05/15/2019 09:30	05/23/2019 15:11	4304731
CCL	N-Nitrosodi-N-butylamine (NDBA)	521	2.0	---		1.3420	2.0	ng/L	67	50 - 150	---	---	1.0	05/15/2019 09:30	05/23/2019 15:11	4304731
CCL	N-Nitrosodimethylamine (NDEA)	521	2.0	---		1.1490	2.0	ng/L	57	50 - 150	---	---	1.0	05/15/2019 09:30	05/23/2019 15:11	4304731
CCL	N-Nitrosodimethylamine (NDMA)	521	2.0	---		2.4240	2.0	ng/L	121	50 - 150	---	---	1.0	05/15/2019 09:30	05/23/2019 15:11	4304731
CCL	N-Nitrosodi-N-propylamine (NDPA)	521	2.0	---		2.4300	2.0	ng/L	122	50 - 150	---	---	1.0	05/15/2019 09:30	05/23/2019 15:11	4304731
CCL	N-Nitrosodiphenylamine (NDPhA)	521	20	---		16.7480	10.0	ng/L	167	3 - 250	---	---	1.0	05/15/2019 09:30	05/23/2019 15:11	4304731
CCL	N-Nitrosomethylethylamine (NMEA)	521	2.0	---		1.3430	2.0	ng/L	67	50 - 150	---	---	1.0	05/15/2019 09:30	05/23/2019 15:11	4304731
CCL	N-Nitrosomorpholine (NMOR)	521	2.0	---		2.3160	2.0	ng/L	116	50 - 150	---	---	1.0	05/15/2019 09:30	05/23/2019 15:11	4304731
CCL	N-Nitrosopiperidine (NPIP)	521	2.0	---		2.7190	2.0	ng/L	136	50 - 150	---	---	1.0	05/15/2019 09:30	05/23/2019 15:11	4304731
CCM	IS-NDPA-d14	521	N/A	---		47889	47889	ng/L	100	50 - 150	---	---	1.0	05/15/2019 09:30	05/24/2019 01:21	4304733
CCM	SS-NDMA-d6	521	N/A	---		21.9140	20.0	ng/L	110	70 - 130	---	---	1.0	05/15/2019 09:30	05/24/2019 01:21	4304733
CCM	N-Nitrosopyrrolidine (NPYR)	521	2.0	---		22.8150	20.0	ng/L	114	70 - 130	---	---	1.0	05/15/2019 09:30	05/24/2019 01:21	4304733
CCM	N-Nitrosodi-N-butylamine (NDBA)	521	2.0	---		17.7510	20.0	ng/L	89	70 - 130	---	---	1.0	05/15/2019 09:30	05/24/2019 01:21	4304733
CCM	N-Nitrosodimethylamine (NDEA)	521	2.0	---		23.5440	20.0	ng/L	118	70 - 130	---	---	1.0	05/15/2019 09:30	05/24/2019 01:21	4304733
CCM	N-Nitrosodimethylamine (NDMA)	521	2.0	---		21.6010	20.0	ng/L	108	70 - 130	---	---	1.0	05/15/2019 09:30	05/24/2019 01:21	4304733
CCM	N-Nitrosodi-N-propylamine (NDPA)	521	2.0	---		17.4700	20.0	ng/L	87	70 - 130	---	---	1.0	05/15/2019 09:30	05/24/2019 01:21	4304733
CCM	N-Nitrosodiphenylamine (NDPhA)	521	20	---		150.6530	100	ng/L	151	20 - 250	---	---	1.0	05/15/2019 09:30	05/24/2019 01:21	4304733
CCM	N-Nitrosomethylethylamine (NMEA)	521	2.0	---		25.6550	20.0	ng/L	128	70 - 130	---	---	1.0	05/15/2019 09:30	05/24/2019 01:21	4304733
CCM	N-Nitrosomorpholine (NMOR)	521	2.0	---		19.8840	20.0	ng/L	99	70 - 130	---	---	1.0	05/15/2019 09:30	05/24/2019 01:21	4304733
CCM	N-Nitrosopiperidine (NPIP)	521	2.0	---		18.4920	20.0	ng/L	92	70 - 130	---	---	1.0	05/15/2019 09:30	05/24/2019 01:21	4304733

Eurofins Eaton Analytical

Run Log

Run ID: 259987 Method: 522

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
CCL	4307878		OS	DM	05/30/2019 15:03	522-04082019-DM.M
LRB	4307870		RW	DM	05/30/2019 15:32	522-04082019-DM.M
FBL	4307873		RW	DM	05/30/2019 16:01	522-04082019-DM.M
FBH	4307876		RW	DM	05/30/2019 16:29	522-04082019-DM.M
CCM	4307881		OS	DM	05/30/2019 23:11	522-04082019-DM.M
FS	4291756	Finished	DW	DM	05/30/2019 23:39	522-04082019-DM.M
CCH	4307884		OS	DM	05/31/2019 04:24	522-04082019-DM.M

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCL	IS-Tetrahydrofuran-d8	522	N/A	---		22226	22226	ug/L	100	50 - 150	---	---	1.0	05/29/2019 09:49	05/30/2019 15:03	4307878
CCL	SS-1,4-Dioxane-d8	522	N/A	---		9.4400	10.0	ug/L	94	70 - 130	---	---	1.0	05/29/2019 09:49	05/30/2019 15:03	4307878
CCL	1,4-Dioxane	522	0.07	---		0.0700	0.07	ug/L	100	50 - 150	---	---	1.0	05/29/2019 09:49	05/30/2019 15:03	4307878
LRB	IS-Tetrahydrofuran-d8	522	N/A	---		21601	22226	ug/L	97	50 - 150	---	---	1.0	05/29/2019 08:10	05/30/2019 15:32	4307870
LRB	SS-1,4-Dioxane-d8	522	N/A	---		8.9900	10.0	ug/L	90	70 - 130	---	---	1.0	05/29/2019 08:10	05/30/2019 15:32	4307870
LRB	1,4-Dioxane	522	0.07	---	<	0.07		ug/L	---	---	---	---	1.0	05/29/2019 08:10	05/30/2019 15:32	4307870
FBL	IS-Tetrahydrofuran-d8	522	N/A	---		21678	22226	ug/L	98	50 - 150	---	---	1.0	05/29/2019 08:10	05/30/2019 16:01	4307873
FBL	SS-1,4-Dioxane-d8	522	N/A	---		8.9200	10.0	ug/L	89	70 - 130	---	---	1.0	05/29/2019 08:10	05/30/2019 16:01	4307873
FBL	1,4-Dioxane	522	0.07	---		0.0880	0.07	ug/L	126	50 - 150	---	---	1.0	05/29/2019 08:10	05/30/2019 16:01	4307873
FBH	IS-Tetrahydrofuran-d8	522	N/A	---		22144	22226	ug/L	100	50 - 150	---	---	1.0	05/29/2019 08:10	05/30/2019 16:29	4307876
FBH	SS-1,4-Dioxane-d8	522	N/A	---		8.9700	10.0	ug/L	90	70 - 130	---	---	1.0	05/29/2019 08:10	05/30/2019 16:29	4307876
FBH	1,4-Dioxane	522	0.07	---		9.3480	10.0	ug/L	93	70 - 130	---	---	1.0	05/29/2019 08:10	05/30/2019 16:29	4307876
CCM	IS-Tetrahydrofuran-d8	522	N/A	---		26135	26135	ug/L	100	50 - 150	---	---	1.0	05/29/2019 09:49	05/30/2019 23:11	4307881
CCM	SS-1,4-Dioxane-d8	522	N/A	---		8.8600	10.0	ug/L	89	70 - 130	---	---	1.0	05/29/2019 09:49	05/30/2019 23:11	4307881
CCM	1,4-Dioxane	522	0.07	---		0.9180	1.0	ug/L	92	70 - 130	---	---	1.0	05/29/2019 09:49	05/30/2019 23:11	4307881
FS	IS-Tetrahydrofuran-d8	522	N/A	Finished		23653	26135	ug/L	91	50 - 150	---	---	1.0	05/29/2019 08:10	05/30/2019 23:39	4291756
FS	SS-1,4-Dioxane-d8	522	N/A	Finished		9.1500	10.0	ug/L	92	70 - 130	---	---	1.0	05/29/2019 08:10	05/30/2019 23:39	4291756
FS	1,4-Dioxane	522	0.07	Finished		0.33		ug/L	---	---	---	---	1.0	05/29/2019 08:10	05/30/2019 23:39	4291756
CCH	IS-Tetrahydrofuran-d8	522	N/A	---		25743	25743	ug/L	100	50 - 150	---	---	1.0	05/29/2019 09:49	05/31/2019 04:24	4307884
CCH	SS-1,4-Dioxane-d8	522	N/A	---		8.7100	10.0	ug/L	87	70 - 130	---	---	1.0	05/29/2019 09:49	05/31/2019 04:24	4307884
CCH	1,4-Dioxane	522	0.07	---		9.2560	10.0	ug/L	93	70 - 130	---	---	1.0	05/29/2019 09:49	05/31/2019 04:24	4307884

Eurofins Eaton Analytical

Run Log

Run ID: **260543** Method: **522**

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
CCL	4307879		OS	DM	05/31/2019 14:52	522-04082019-DM.M
LRB	4307871		RW	DM	05/31/2019 15:20	522-04082019-DM.M
FBL	4307874		RW	DM	05/31/2019 15:48	522-04082019-DM.M
CCM	4307882		OS	DM	05/31/2019 21:35	522-04082019-DM.M
CCH	4307885		OS	DM	06/01/2019 03:48	522-04082019-DM.M

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCL	IS-Tetrahydrofuran-d8	522	N/A	--		25252	25252	ug/L	100	50 - 150	---	---	1.0	05/29/2019 09:49	05/31/2019 14:52	4307879
CCL	SS-1,4-Dioxane-d8	522	N/A	--		9.0000	10.0	ug/L	90	70 - 130	---	---	1.0	05/29/2019 09:49	05/31/2019 14:52	4307879
CCL	1,4-Dioxane	522	0.07	--		0.0610	0.07	ug/L	87	50 - 150	---	---	1.0	05/29/2019 09:49	05/31/2019 14:52	4307879
LRB	IS-Tetrahydrofuran-d8	522	N/A	--		23134	25252	ug/L	92	50 - 150	---	---	1.0	05/29/2019 08:10	05/31/2019 15:20	4307871
LRB	SS-1,4-Dioxane-d8	522	N/A	--		8.4800	10.0	ug/L	85	70 - 130	---	---	1.0	05/29/2019 08:10	05/31/2019 15:20	4307871
LRB	1,4-Dioxane	522	0.07	--	<	0.07		ug/L	---	---	---	---	1.0	05/29/2019 08:10	05/31/2019 15:20	4307871
FBL	IS-Tetrahydrofuran-d8	522	N/A	--		22562	25252	ug/L	89	50 - 150	---	---	1.0	05/29/2019 08:10	05/31/2019 15:48	4307874
FBL	SS-1,4-Dioxane-d8	522	N/A	--		8.5400	10.0	ug/L	85	70 - 130	---	---	1.0	05/29/2019 08:10	05/31/2019 15:48	4307874
FBL	1,4-Dioxane	522	0.07	--		0.0820	0.07	ug/L	117	50 - 150	---	---	1.0	05/29/2019 08:10	05/31/2019 15:48	4307874
CCM	IS-Tetrahydrofuran-d8	522	N/A	--		24595	24595	ug/L	100	50 - 150	---	---	1.0	05/29/2019 09:49	05/31/2019 21:35	4307882
CCM	SS-1,4-Dioxane-d8	522	N/A	--		8.7900	10.0	ug/L	88	70 - 130	---	---	1.0	05/29/2019 09:49	05/31/2019 21:35	4307882
CCM	1,4-Dioxane	522	0.07	--		0.8780	1.0	ug/L	88	70 - 130	---	---	1.0	05/29/2019 09:49	05/31/2019 21:35	4307882
CCH	IS-Tetrahydrofuran-d8	522	N/A	--		24594	24594	ug/L	100	50 - 150	---	---	1.0	05/29/2019 09:49	06/01/2019 03:48	4307885
CCH	SS-1,4-Dioxane-d8	522	N/A	--		8.8100	10.0	ug/L	88	70 - 130	---	---	1.0	05/29/2019 09:49	06/01/2019 03:48	4307885
CCH	1,4-Dioxane	522	0.07	--		9.1750	10.0	ug/L	92	70 - 130	---	---	1.0	05/29/2019 09:49	06/01/2019 03:48	4307885

Eurofins Eaton Analytical

Run Log

Run ID: **259404** Method: **524.2**

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
CCC	4293297		RW	CF	05/15/2019 11:04	524 2-032719CF-up1.mth
CCC	4293300		RW	CF	05/15/2019 11:46	524 2-032719CF-up1.mth
CCL	4293302		RW	CF	05/15/2019 12:27	524 2-032719CF-up1.mth
LMB	4293299		RW	CF	05/15/2019 13:35	524 2-032719CF-up1.mth
LTB	4291774	LTB-04/30/19	RW	CF	05/15/2019 18:48	524 2-032719CF-up1.mth
FS	4291763	Finished	DW	CF	05/15/2019 19:19	524 2-032719CF-up1.mth
MS	4293912	Finished	DW	CF	05/15/2019 19:50	524 2-032719CF-up1.mth
CCC	4293298		RW	CF	05/15/2019 20:21	524 2-032719CF-up1.mth
CCC	4293301		RW	CF	05/15/2019 20:52	524 2-032719CF-up1.mth
LMB	4293953		RW	CF	05/15/2019 21:55	524 2-032719CF-up1.mth

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	IS-1,4-Difluorobenzene	524.2	N/A	--		258253	258253	ug/L	100	50 - 150	---	---	1.0	---	05/15/2019 11:04	4293297
CCC	SS-Bromofluorobenzene	524.2	N/A	--		5.0010	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/15/2019 11:04	4293297
CCC	SS-1,2-Dichlorobenzene-d4	524.2	N/A	--		9.7110	10.0	ug/L	97	70 - 130	---	---	1.0	---	05/15/2019 11:04	4293297
CCC	SS-1,2-Dichloroethane-d4	524.2	N/A	--		9.7650	10.0	ug/L	98	70 - 130	---	---	1.0	---	05/15/2019 11:04	4293297
CCC	SS-Toluene-d8	524.2	N/A	--		9.9820	10.0	ug/L	100	70 - 130	---	---	1.0	---	05/15/2019 11:04	4293297
CCC	tert-Amyl Methyl ether	524.2	3.0	--		4.7470	5.0	ug/L	95	70 - 130	---	---	1.0	---	05/15/2019 11:04	4293297
CCC	Benzyl chloride	524.2	0.5	--		5.0760	5.0	ug/L	102	40 - 150	---	---	1.0	---	05/15/2019 11:04	4293297
CCC	1,3-Butadiene	524.2	5.0	--		5.2340	5.0	ug/L	105	70 - 130	---	---	1.0	---	05/15/2019 11:04	4293297
CCC	n-Butyl acrylate	524.2	1.0	--		4.4750	5.0	ug/L	90	50 - 150	---	---	1.0	---	05/15/2019 11:04	4293297
CCC	Chloroprene	524.2	5.0	--		4.9480	5.0	ug/L	99	70 - 130	---	---	1.0	---	05/15/2019 11:04	4293297
CCC	Cyclohexanone	524.2	5.0	--		5.2140	5.0	ug/L	104	31 - 149	---	---	1.0	---	05/15/2019 11:04	4293297
CCC	Diisopropyl ether	524.2	0.5	--		5.0350	5.0	ug/L	101	70 - 130	---	---	1.0	---	05/15/2019 11:04	4293297
CCC	1,4-Dioxane	524.2	5.0	--		6.3320	5.0	ug/L	127	50 - 150	---	---	1.0	---	05/15/2019 11:04	4293297
CCC	Epichlorohydrin	524.2	1.0	--		5.5020	5.0	ug/L	110	35 - 170	---	---	1.0	---	05/15/2019 11:04	4293297
CCC	Ethyl acrylate	524.2	1.0	--		4.1970	5.0	ug/L	84	70 - 130	---	---	1.0	---	05/15/2019 11:04	4293297
CCC	Ethyl tert-butyl ether	524.2	3.0	--		4.8010	5.0	ug/L	96	70 - 130	---	---	1.0	---	05/15/2019 11:04	4293297
CCC	tert-Butyl alcohol	524.2	2.0	--		4.2640	5.0	ug/L	85	60 - 130	---	---	1.0	---	05/15/2019 11:04	4293297
CCC	Vinyl acetate	524.2	5.0	--		5.3480	5.0	ug/L	107	45 - 160	---	---	1.0	---	05/15/2019 11:04	4293297
CCC	IS-1,4-Difluorobenzene	524.2	N/A	--		252266	252266	ug/L	100	50 - 150	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	SS-Bromofluorobenzene	524.2	N/A	--		5.2500	5.0	ug/L	105	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	SS-1,2-Dichlorobenzene-d4	524.2	N/A	--		10.2730	10.0	ug/L	103	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	SS-1,2-Dichloroethane-d4	524.2	N/A	--		9.6860	10.0	ug/L	97	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	SS-Toluene-d8	524.2	N/A	--		10.0130	10.0	ug/L	100	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Acetone	524.2	5.0	--		4.0640	5.0	ug/L	81	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	tert-Amyl Methyl ether	524.2	3.0	--		5.0790	5.0	ug/L	102	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Benzene	524.2	0.5	--		4.8180	5.0	ug/L	96	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Bromobenzene	524.2	0.5	--		5.0950	5.0	ug/L	102	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Bromochloromethane	524.2	0.5	--		5.1270	5.0	ug/L	103	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Bromodichloromethane	524.2	0.5	--		4.4630	5.0	ug/L	89	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Bromoform	524.2	0.5	--		4.5260	5.0	ug/L	91	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Bromomethane	524.2	0.5	--		4.0050	5.0	ug/L	80	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	2-Butanone (MEK)	524.2	5.0	--		5.4800	5.0	ug/L	110	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	n-Butylbenzene	524.2	0.5	--		5.3480	5.0	ug/L	107	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	sec-Butylbenzene	524.2	0.5	--		5.1770	5.0	ug/L	104	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	tert-Butylbenzene	524.2	0.5	--		4.9570	5.0	ug/L	99	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Carbon disulfide	524.2	5.0	--		4.8220	5.0	ug/L	96	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Carbon tetrachloride	524.2	0.5	--		4.6780	5.0	ug/L	94	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Chlorobenzene	524.2	0.5	--		5.1930	5.0	ug/L	104	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Chloroethane	524.2	0.5	--		4.8770	5.0	ug/L	98	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Chloroform	524.2	0.5	--		4.8930	5.0	ug/L	98	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	Chloromethane	524.2	0.5	---		4.5650	5.0	ug/L	91	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	2-Chlorotoluene	524.2	0.5	---		5.0500	5.0	ug/L	101	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	4-Chlorotoluene	524.2	0.5	---		5.0550	5.0	ug/L	101	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Dibromochloromethane	524.2	0.5	---		4.1770	5.0	ug/L	84	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	1,2-Dibromo-3-chloropropane (DBCP)	524.2	0.2	---		5.7580	5.0	ug/L	115	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	1,2-Dibromoethane (EDB)	524.2	0.2	---		4.9900	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Dibromomethane	524.2	0.5	---		4.9000	5.0	ug/L	98	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	1,2-Dichlorobenzene	524.2	0.5	---		5.1550	5.0	ug/L	103	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	1,3-Dichlorobenzene	524.2	0.5	---		5.1570	5.0	ug/L	103	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	1,4-Dichlorobenzene	524.2	0.5	---		5.0860	5.0	ug/L	102	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Dichlorodifluoromethane	524.2	0.5	---		4.5910	5.0	ug/L	92	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	1,1-Dichloroethane	524.2	0.5	---		4.9860	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	1,2-Dichloroethane	524.2	0.5	---		4.7910	5.0	ug/L	96	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	1,1-Dichloroethylene	524.2	0.5	---		4.9440	5.0	ug/L	99	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	cis-1,2-Dichloroethylene	524.2	0.5	---		4.7610	5.0	ug/L	95	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	trans-1,2-Dichloroethylene	524.2	0.5	---		4.9110	5.0	ug/L	98	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	1,2-Dichloropropane	524.2	0.5	---		5.1810	5.0	ug/L	104	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	1,3-Dichloropropane	524.2	0.5	---		4.9120	5.0	ug/L	98	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	2,2-Dichloropropane	524.2	0.5	---		5.1550	5.0	ug/L	103	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	1,1-Dichloropropylene	524.2	0.5	---		4.8490	5.0	ug/L	97	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	cis-1,3-Dichloropropylene	524.2	0.5	---		4.9010	5.0	ug/L	98	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	trans-1,3-Dichloropropylene	524.2	0.5	---		4.6810	5.0	ug/L	94	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Diisopropyl ether	524.2	0.5	---		4.9620	5.0	ug/L	99	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Ethylbenzene	524.2	0.5	---		5.0550	5.0	ug/L	101	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Ethyl tert-butyl ether	524.2	3.0	---		4.9620	5.0	ug/L	99	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Hexachlorobutadiene	524.2	0.5	---		5.3800	5.0	ug/L	108	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	2-Hexanone	524.2	5.0	---		5.2460	5.0	ug/L	105	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Isopropylbenzene	524.2	0.5	---		5.0610	5.0	ug/L	101	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	4-Isopropyltoluene	524.2	0.5	---		5.0470	5.0	ug/L	101	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	4-Methyl-2-pentanone (MIBK)	524.2	2.0	---		4.9640	5.0	ug/L	99	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Methyl-t-butyl ether (MTBE)	524.2	0.5	---		5.0330	5.0	ug/L	101	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Naphthalene	524.2	0.5	---		4.7060	5.0	ug/L	94	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	n-Propylbenzene	524.2	0.5	---		5.0420	5.0	ug/L	101	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Styrene	524.2	0.5	---		5.0460	5.0	ug/L	101	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	1,1,1,2-Tetrachloroethane	524.2	0.5	---		4.7270	5.0	ug/L	95	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	1,1,2,2-Tetrachloroethane	524.2	0.5	---		4.9140	5.0	ug/L	98	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Tetrachloroethylene	524.2	0.5	---		4.8970	5.0	ug/L	98	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Tetrahydrofuran	524.2	5.0	---		5.2200	5.0	ug/L	104	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Toluene	524.2	0.5	---		5.1130	5.0	ug/L	102	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	1,2,3-Trichlorobenzene	524.2	0.5	---		5.2270	5.0	ug/L	105	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	1,2,4-Trichlorobenzene	524.2	0.5	---		5.3070	5.0	ug/L	106	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	1,1,1-Trichloroethane	524.2	0.5	---		4.8770	5.0	ug/L	98	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	1,1,2-Trichloroethane	524.2	0.5	---		4.7370	5.0	ug/L	95	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Trichloroethylene	524.2	0.5	---		5.1730	5.0	ug/L	103	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Trichlorofluoromethane	524.2	0.5	---		4.6540	5.0	ug/L	93	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	1,1,2-Trichloro-1,2,2-trifluoroethane	524.2	0.5	---		4.6080	5.0	ug/L	92	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	1,2,3-Trimethylbenzene	524.2	0.5	---		5.0360	5.0	ug/L	101	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	1,2,4-Trimethylbenzene	524.2	0.5	---		5.0310	5.0	ug/L	101	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	1,3,5-Trimethylbenzene	524.2	0.5	---		5.0050	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	Vinyl chloride	524.2	0.2	---		4.6020	5.0	ug/L	92	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	1,2-Xylene	524.2	0.5	---		5.1320	5.0	ug/L	103	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCC	1,3 + 1,4-Xylene	524.2	0.5	---		9.9480	10.0	ug/L	99	70 - 130	---	---	1.0	---	05/15/2019 11:46	4293300
CCL	IS-1,4-Difluorobenzene	524.2	N/A	---		239048	239048	ug/L	100	50 - 150	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	SS-Bromofluorobenzene	524.2	N/A	---		5.2880	5.0	ug/L	106	70 - 130	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	SS-1,2-Dichlorobenzene-d4	524.2	N/A	---		10.2310	10.0	ug/L	102	70 - 130	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	SS-1,2-Dichloroethane-d4	524.2	N/A	---		10.1080	10.0	ug/L	101	70 - 130	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	SS-Toluene-d8	524.2	N/A	---		10.1180	10.0	ug/L	101	70 - 130	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	Benzene	524.2	0.5	---		0.4850	0.5	ug/L	97	50 - 150	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	Bromodichloromethane	524.2	0.5	---		0.4550	0.5	ug/L	91	50 - 150	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	Bromoform	524.2	0.5	---		0.5630	0.5	ug/L	113	50 - 150	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	Carbon tetrachloride	524.2	0.5	---		0.4780	0.5	ug/L	96	50 - 150	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	Chlorobenzene	524.2	0.5	---		0.4980	0.5	ug/L	100	50 - 150	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	Chloroform	524.2	0.5	---		0.4620	0.5	ug/L	92	50 - 150	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	Dibromochloromethane	524.2	0.5	---		0.4250	0.5	ug/L	85	50 - 150	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	1,2-Dichlorobenzene	524.2	0.5	---		0.5460	0.5	ug/L	109	50 - 150	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	1,4-Dichlorobenzene	524.2	0.5	---		0.5350	0.5	ug/L	107	50 - 150	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	1,2-Dichloroethane	524.2	0.5	---		0.4550	0.5	ug/L	91	50 - 150	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	1,1-Dichloroethylene	524.2	0.5	---		0.4860	0.5	ug/L	97	50 - 150	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	cis-1,2-Dichloroethylene	524.2	0.5	---		0.4310	0.5	ug/L	86	50 - 150	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	trans-1,2-Dichloroethylene	524.2	0.5	---		0.4800	0.5	ug/L	96	50 - 150	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	1,2-Dichloropropane	524.2	0.5	---		0.4530	0.5	ug/L	91	50 - 150	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	Ethylbenzene	524.2	0.5	---		0.4880	0.5	ug/L	98	50 - 150	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	Styrene	524.2	0.5	---		0.4830	0.5	ug/L	97	50 - 150	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	Tetrachloroethylene	524.2	0.5	---		0.4790	0.5	ug/L	96	50 - 150	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	Toluene	524.2	0.5	---		0.4970	0.5	ug/L	99	50 - 150	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	1,2,4-Trichlorobenzene	524.2	0.5	---		0.5050	0.5	ug/L	101	50 - 150	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	1,1,1-Trichloroethane	524.2	0.5	---		0.4740	0.5	ug/L	95	50 - 150	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	1,1,2-Trichloroethane	524.2	0.5	---		0.4540	0.5	ug/L	91	50 - 150	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	Trichloroethylene	524.2	0.5	---		0.4820	0.5	ug/L	96	50 - 150	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	Vinyl chloride	524.2	0.2	---		0.4810	0.5	ug/L	96	50 - 150	---	---	1.0	---	05/15/2019 12:27	4293302
CCL	1,2-Xylene	524.2	0.5	---		0.4720	0.5	ug/L	94	50 - 150	---	---	1.0	---	05/15/2019 12:27	4293302

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
COL	1,3 + 1,4-Xylene	524.2	0.5	---		0.9830	1.0	ug/L	98	50 - 150	---	---	1.0	---	05/15/2019 12:27	4293302
LMB	IS-1,4-Difluorobenzene	524.2	N/A	---		247180	239048	ug/L	103	70 - 130	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	SS-Bromofluorobenzene	524.2	N/A	---		5.0660	5.0	ug/L	101	70 - 130	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	SS-1,2-Dichlorobenzene-d4	524.2	N/A	---		9.8040	10.0	ug/L	98	70 - 130	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	SS-1,2-Dichloroethane-d4	524.2	N/A	---		9.4250	10.0	ug/L	94	70 - 130	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	SS-Toluene-d8	524.2	N/A	---		10.0080	10.0	ug/L	100	70 - 130	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Acetone	524.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	tert-Amyl Methyl ether	524.2	3.0	---	<	3.0		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Benzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Benzyl chloride	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Bromobenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Bromochloromethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Bromodichloromethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Bromoform	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Bromomethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	1,3-Butadiene	524.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	2-Butanone (MEK)	524.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	n-Butyl acrylate	524.2	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	n-Butylbenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	sec-Butylbenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	tert-Butylbenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Carbon disulfide	524.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Carbon tetrachloride	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Chlorobenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Chloroethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Chloroform	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Chloromethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Chloroprene	524.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	2-Chlorotoluene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	4-Chlorotoluene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Cyclohexanone	524.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Dibromochloromethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	1,2-Dibromo-3-chloropropane (DBCP)	524.2	0.2	---	<	0.2		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	1,2-Dibromoethane (EDB)	524.2	0.2	---	<	0.2		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Dibromomethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	1,2-Dichlorobenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	1,3-Dichlorobenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	1,4-Dichlorobenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Dichlorodifluoromethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	1,1-Dichloroethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	1,2-Dichloroethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LMB	1,1-Dichloroethylene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	cis-1,2-Dichloroethylene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	trans-1,2-Dichloroethylene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	1,2-Dichloropropane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	1,3-Dichloropropane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	2,2-Dichloropropane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	1,1-Dichloropropylene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	cis-1,3-Dichloropropylene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	trans-1,3-Dichloropropylene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Diisopropyl ether	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	1,4-Dioxane	524.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Epichlorohydrin	524.2	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Ethyl acrylate	524.2	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Ethylbenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Ethyl tert-butyl ether	524.2	3.0	---	<	3.0		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Hexachlorobutadiene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	2-Hexanone	524.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Isopropylbenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	4-Isopropyltoluene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	4-Methyl-2-pentanone (MIBK)	524.2	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	tert-Butyl alcohol	524.2	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Methyl-t-butyl ether (MTBE)	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Naphthalene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	n-Propylbenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Styrene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	1,1,1,2-Tetrachloroethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	1,1,2,2-Tetrachloroethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Tetrachloroethylene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Tetrahydrofuran	524.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Toluene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	1,2,3-Trichlorobenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	1,2,4-Trichlorobenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	1,1,1-Trichloroethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	1,1,2-Trichloroethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Trichloroethylene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Trichlorofluoromethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	1,2,3-Trichloropropane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	1,1,1-Trichloro-1,2,2-trifluoroethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	1,2,3-Trimethylbenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	1,2,4-Trimethylbenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	1,3,5-Trimethylbenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LMB	Vinyl acetate	524.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	Vinyl chloride	524.2	0.2	---	<	0.2		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	1,2-Xylene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LMB	1,3 + 1,4-Xylene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 13:35	4293299
LTB	IS-1,4-Difluorobenzene	524.2	N/A	LTB-04/30/19		240964	239048	ug/L	101	70 - 130	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	SS-Bromofluorobenzene	524.2	N/A	LTB-04/30/19	<	4.7080	5.0	ug/L	94	70 - 130	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	SS-1,2-Dichlorobenzene-d4	524.2	N/A	LTB-04/30/19		9.6900	10.0	ug/L	97	70 - 130	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	SS-1,2-Dichloroethane-d4	524.2	N/A	LTB-04/30/19		9.5870	10.0	ug/L	96	70 - 130	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	SS-Toluene-d8	524.2	N/A	LTB-04/30/19		9.9680	10.0	ug/L	100	70 - 130	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Acetone	524.2	5.0	LTB-04/30/19		9.8		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	tert-Amyl Methyl ether	524.2	3.0	LTB-04/30/19	<	3.0		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Benzene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Benzyl chloride	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Bromobenzene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Bromochloromethane	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Bromodichloromethane	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Bromoform	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Bromomethane	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,3-Butadiene	524.2	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	2-Butanone (MEK)	524.2	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	n-Butyl acrylate	524.2	1.0	LTB-04/30/19	<	1.0		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	n-Butylbenzene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	sec-Butylbenzene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	tert-Butylbenzene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Carbon disulfide	524.2	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Carbon tetrachloride	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Chlorobenzene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Chloroethane	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Chloroform	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Chloromethane	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Chloroprene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	2-Chlorotoluene	524.2	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	4-Chlorotoluene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Cyclohexanone	524.2	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Dibromochloromethane	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,2-Dibromo-3-chloropropane (DBCP)	524.2	0.2	LTB-04/30/19	<	0.2		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,2-Dibromoethane (EDB)	524.2	0.2	LTB-04/30/19	<	0.2		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Dibromomethane	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,2-Dichlorobenzene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,3-Dichlorobenzene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,4-Dichlorobenzene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,4-Dichlorobenzene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LTB	Total Dichlorobenzene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Dichlorodifluoromethane	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,1-Dichloroethane	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,2-Dichloroethane	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,1-Dichloroethylene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	cis-1,2-Dichloroethylene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	trans-1,2-Dichloroethylene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,2-Dichloropropane	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,3-Dichloropropane	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	2,2-Dichloropropane	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,1-Dichloropropylene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,3-Dichloropropylene, cis & trans	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	cis-1,3-Dichloropropylene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	trans-1,3-Dichloropropylene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Diisopropyl ether	524.2	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,4-Dioxane	524.2	1.0	LTB-04/30/19	<	1.0		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Epichlorohydrin	524.2	1.0	LTB-04/30/19	<	1.0		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Ethyl acrylate	524.2	1.0	LTB-04/30/19	<	1.0		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Ethylbenzene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Ethyl tert-butyl ether	524.2	3.0	LTB-04/30/19	<	3.0		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Hexachlorobutadiene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	2-Hexanone	524.2	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Isopropylbenzene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	4-Isopropyltoluene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	4-Methyl-2-pentanone (MIBK)	524.2	2.0	LTB-04/30/19	<	2.0		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	tert-Butyl alcohol	524.2	2.0	LTB-04/30/19	<	2.0		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Methyl-t-butyl ether (MTBE)	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Naphthalene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	n-Propylbenzene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Styrene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,1,1,2-Tetrachloroethane	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,1,2,2-Tetrachloroethane	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Tetrachloroethylene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Tetrahydrofuran	524.2	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Toluene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Total Trihalomethanes	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,2,3-Trichlorobenzene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,2,4-Trichlorobenzene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,1,1-Trichloroethane	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,1,2-Trichloroethane	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Trichloroethylene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LTB	Trichlorofluoromethane	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,2,3-Trichloropropane	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,1,2-Trichloro-1,2,2-trifluoroethane	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,2,3-Trimethylbenzene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,2,4-Trimethylbenzene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,3,5-Trimethylbenzene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Vinyl acetate	524.2	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Vinyl chloride	524.2	0.2	LTB-04/30/19	<	0.2		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,2-Xylene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	1,3 + 1,4-Xylene	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
LTB	Xylenes, Total	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 18:48	4291774
FS	IS-1,4-Difluorobenzene	524.2	N/A	Finished		233226	239048	ug/L	98	70 - 130	---	---	1.0	---	05/15/2019 19:19	4291763
FS	SS-Bromofluorobenzene	524.2	N/A	Finished		4.9950	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/15/2019 19:19	4291763
FS	SS-1,2-Dichlorobenzene-d4	524.2	N/A	Finished		9.5520	10.0	ug/L	96	70 - 130	---	---	1.0	---	05/15/2019 19:19	4291763
FS	SS-1,2-Dichloroethane-d4	524.2	N/A	Finished		9.7880	10.0	ug/L	98	70 - 130	---	---	1.0	---	05/15/2019 19:19	4291763
FS	SS-Toluene-d8	524.2	N/A	Finished		9.9870	10.0	ug/L	100	70 - 130	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Acetone	524.2	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	tert-Amyl Methyl ether	524.2	3.0	Finished	<	3.0		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Benzene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Benzyl chloride	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Bromobenzene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Bromochloromethane	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Bromodichloromethane	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Bromoform	524.2	0.5	Finished	<	2.9		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Bromomethane	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Bromotoluene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	1,3-Butadiene	524.2	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	2-Butanone (MEK)	524.2	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	n-Butyl acrylate	524.2	1.0	Finished	<	1.0		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	n-Butylbenzene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	sec-Butylbenzene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	tert-Butylbenzene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Carbon disulfide	524.2	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Carbon tetrachloride	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Chlorobenzene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Chloroethane	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Chloroform	524.2	0.5	Finished	<	3.2		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Chloromethane	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Chloroprene	524.2	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	2-Chlorotoluene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	4-Chlorotoluene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Cyclohexanone	524.2	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FS	Dibromochloromethane	524.2	0.5	Finished		1.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	1,2-Dibromo-3-chloropropane (DBCP)	524.2	0.2	Finished	<	0.2		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	1,2-Dibromoethane (EDB)	524.2	0.2	Finished	<	0.2		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Dibromomethane	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	1,2-Dichlorobenzene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	1,3-Dichlorobenzene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	1,4-Dichlorobenzene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Total Dichlorobenzene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Dichlorodifluoromethane	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	1,1-Dichloroethane	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	1,2-Dichloroethane	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	1,1-Dichloroethylene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	cis-1,2-Dichloroethylene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	trans-1,2-Dichloroethylene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	1,3-Dichloropropane	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	cis-1,3-Dichloropropylene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	trans-1,3-Dichloropropylene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Diisopropyl ether	524.2	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	1,4-Dioxane	524.2	1.0	Finished	<	1.0		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Epichlorohydrin	524.2	1.0	Finished	<	1.0		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Ethyl acrylate	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Ethylbenzene	524.2	3.0	Finished	<	3.0		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Ethyl tert-butyl ether	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Hexachlorobutadiene	524.2	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	2-Hexanone	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Isopropylbenzene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	4-Isopropyltoluene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	4-Methyl-2-pentanone (MIBK)	524.2	2.0	Finished	<	2.0		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	tert-Butyl alcohol	524.2	2.0	Finished	<	2.0		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Methyl-t-butyl ether (MTBE)	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Naphthalene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	n-Propylbenzene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Styrene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	1,1,1,2-Tetrachloroethane	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	1,1,2,2-Tetrachloroethane	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Tetrachloroethylene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Tetrahydrofuran	524.2	5.0	Finished		6.1		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FS	Toluene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Total Trihalomethanes	524.2	0.5	Finished		7.6		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	1,2,3-Trichlorobenzene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	1,2,4-Trichlorobenzene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	1,1,1-Trichloroethane	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	1,1,2-Trichloroethane	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Trichloroethylene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Trichlorofluoromethane	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	1,2,3-Trichloropropane	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	1,1,2-Trichloro-1,2,2-trifluoroethane	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	1,2,3-Trimethylbenzene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	1,2,4-Trimethylbenzene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	1,3,5-Trimethylbenzene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Vinyl acetate	524.2	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Vinyl chloride	524.2	0.2	Finished	<	0.2		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	1,2-Xylene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	1,3 + 1,4-Xylene	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
FS	Xylenes, Total	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 19:19	4291763
MS	IS-1,4-Difluorobenzene	524.2	N/A	Finished		238135	239048	ug/L	100	70 - 130	---	---	1.0	---	05/15/2019 19:50	4293912
MS	SS-Bromofluorobenzene	524.2	N/A	Finished		4.9340	5.0	ug/L	99	70 - 130	---	---	1.0	---	05/15/2019 19:50	4293912
MS	SS-1,2-Dichlorobenzene-d4	524.2	N/A	Finished		9.7710	10.0	ug/L	98	70 - 130	---	---	1.0	---	05/15/2019 19:50	4293912
MS	SS-1,2-Dichloroethane-d4	524.2	N/A	Finished		9.4300	10.0	ug/L	94	70 - 130	---	---	1.0	---	05/15/2019 19:50	4293912
MS	SS-Toluene-d8	524.2	N/A	Finished		9.9730	10.0	ug/L	100	70 - 130	---	---	1.0	---	05/15/2019 19:50	4293912
MS	tert-Amyl Methyl ether	524.2	3.0	Finished		7.4790	10.0	ug/L	75	81 - 124	---	---	1.0	---	05/15/2019 19:50	4293912
MS	Benzyl chloride	524.2	0.5	Finished		9.0070	10.0	ug/L	90	37 - 165	---	---	1.0	---	05/15/2019 19:50	4293912
MS	1,3-Butadiene	524.2	5.0	Finished		9.2410	10.0	ug/L	92	66 - 127	---	---	1.0	---	05/15/2019 19:50	4293912
MS	n-Butyl acrylate	524.2	1.0	Finished		7.2440	10.0	ug/L	72	54 - 123	---	---	1.0	---	05/15/2019 19:50	4293912
MS	Chloroprene	524.2	5.0	Finished		8.7190	10.0	ug/L	87	73 - 115	---	---	1.0	---	05/15/2019 19:50	4293912
MS	Cyclohexanone	524.2	5.0	Finished		7.9360	10.0	ug/L	79	31 - 149	---	---	1.0	---	05/15/2019 19:50	4293912
MS	Diisopropyl ether	524.2	0.5	Finished		8.2620	10.0	ug/L	83	70 - 130	---	---	1.0	---	05/15/2019 19:50	4293912
MS	1,4-Dioxane	524.2	5.0	Finished		6.7660	10.0	ug/L	68	20 - 194	---	---	1.0	---	05/15/2019 19:50	4293912
MS	Epichlorohydrin	524.2	1.0	Finished		6.9420	10.0	ug/L	69	1 - 179	---	---	1.0	---	05/15/2019 19:50	4293912
MS	Ethyl acrylate	524.2	1.0	Finished		6.4910	10.0	ug/L	65	73 - 119	---	---	1.0	---	05/15/2019 19:50	4293912
MS	Ethyl tert-butyl ether	524.2	3.0	Finished		7.5100	10.0	ug/L	75	79 - 125	---	---	1.0	---	05/15/2019 19:50	4293912
MS	tert-Butyl alcohol	524.2	2.0	Finished		5.4830	10.0	ug/L	55	47 - 140	---	---	1.0	---	05/15/2019 19:50	4293912
MS	Vinyl acetate	524.2	5.0	Finished		8.9360	10.0	ug/L	89	67 - 130	---	---	1.0	---	05/15/2019 19:50	4293912
CCC	IS-1,4-Difluorobenzene	524.2	N/A	---		241308	241308	ug/L	100	50 - 150	---	---	1.0	---	05/15/2019 20:21	4293298
CCC	SS-Bromofluorobenzene	524.2	N/A	---		4.9300	5.0	ug/L	99	70 - 130	---	---	1.0	---	05/15/2019 20:21	4293298
CCC	SS-1,2-Dichlorobenzene-d4	524.2	N/A	---		9.8590	10.0	ug/L	99	70 - 130	---	---	1.0	---	05/15/2019 20:21	4293298
CCC	SS-1,2-Dichloroethane-d4	524.2	N/A	---		9.7980	10.0	ug/L	98	70 - 130	---	---	1.0	---	05/15/2019 20:21	4293298
CCC	SS-Toluene-d8	524.2	N/A	---		10.2730	10.0	ug/L	103	70 - 130	---	---	1.0	---	05/15/2019 20:21	4293298

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	tert-Amyl Methyl ether	524.2	3.0	---		8.8020	10.0	ug/L	88	70 - 130	---	---	1.0	---	05/15/2019 20:21	4293298
CCC	Benzyl chloride	524.2	0.5	---		9.2700	10.0	ug/L	93	40 - 150	---	---	1.0	---	05/15/2019 20:21	4293298
CCC	1,3-Butadiene	524.2	5.0	---		10.0550	10.0	ug/L	101	70 - 130	---	---	1.0	---	05/15/2019 20:21	4293298
CCC	n-Butyl acrylate	524.2	1.0	---		7.5750	10.0	ug/L	76	50 - 150	---	---	1.0	---	05/15/2019 20:21	4293298
CCC	Chloroprene	524.2	5.0	---		10.2500	10.0	ug/L	102	70 - 130	---	---	1.0	---	05/15/2019 20:21	4293298
CCC	Cyclohexanone	524.2	5.0	---		8.9670	10.0	ug/L	90	31 - 149	---	---	1.0	---	05/15/2019 20:21	4293298
CCC	Diisopropyl ether	524.2	0.5	---		9.9180	10.0	ug/L	99	70 - 130	---	---	1.0	---	05/15/2019 20:21	4293298
CCC	1,4-Dioxane	524.2	5.0	---		8.8080	10.0	ug/L	88	50 - 150	---	---	1.0	---	05/15/2019 20:21	4293298
CCC	Epichlorohydrin	524.2	1.0	---		10.0090	10.0	ug/L	100	35 - 170	---	---	1.0	---	05/15/2019 20:21	4293298
CCC	Ethyl acrylate	524.2	1.0	---		7.7900	10.0	ug/L	78	70 - 130	---	---	1.0	---	05/15/2019 20:21	4293298
CCC	Ethyl tert-butyl ether	524.2	3.0	---		9.0410	10.0	ug/L	90	70 - 130	---	---	1.0	---	05/15/2019 20:21	4293298
CCC	tert-Butyl alcohol	524.2	2.0	---		6.7880	10.0	ug/L	68	60 - 130	---	---	1.0	---	05/15/2019 20:21	4293298
CCC	Vinyl acetate	524.2	5.0	---		8.6350	10.0	ug/L	86	45 - 160	---	---	1.0	---	05/15/2019 20:21	4293298
CCC	IS-1,4-Difluorobenzene	524.2	N/A	---		243271	243271	ug/L	100	50 - 150	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	SS-Bromofluorobenzene	524.2	N/A	---		5.0930	5.0	ug/L	102	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	SS-1,2-Dichlorobenzene-d4	524.2	N/A	---		10.4020	10.0	ug/L	104	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	SS-1,2-Dichloroethane-d4	524.2	N/A	---		9.7660	10.0	ug/L	98	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	SS-Toluene-d8	524.2	N/A	---		9.9280	10.0	ug/L	99	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Acetone	524.2	5.0	---		8.1890	10.0	ug/L	82	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Benzene	524.2	0.5	---		9.7660	10.0	ug/L	98	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Bromobenzene	524.2	0.5	---		10.0460	10.0	ug/L	100	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Bromochloromethane	524.2	0.5	---		10.0440	10.0	ug/L	100	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Bromodichloromethane	524.2	0.5	---		9.0050	10.0	ug/L	90	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Bromoform	524.2	0.5	---		9.6160	10.0	ug/L	96	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Bromomethane	524.2	0.5	---		8.3630	10.0	ug/L	84	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	2-Butanone (MEK)	524.2	5.0	---		10.4540	10.0	ug/L	105	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	n-Butylbenzene	524.2	0.5	---		9.8440	10.0	ug/L	98	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	sec-Butylbenzene	524.2	0.5	---		10.3050	10.0	ug/L	103	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	tert-Butylbenzene	524.2	0.5	---		9.9880	10.0	ug/L	100	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Carbon disulfide	524.2	5.0	---		9.5770	10.0	ug/L	96	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Carbon tetrachloride	524.2	0.5	---		9.7300	10.0	ug/L	97	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Chlorobenzene	524.2	0.5	---		9.8660	10.0	ug/L	99	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Chloroethane	524.2	0.5	---		10.3070	10.0	ug/L	103	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Chloroform	524.2	0.5	---		9.7620	10.0	ug/L	98	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Chloromethane	524.2	0.5	---		9.8750	10.0	ug/L	99	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	2-Chlorotoluene	524.2	0.5	---		10.2330	10.0	ug/L	102	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	4-Chlorotoluene	524.2	0.5	---		9.8780	10.0	ug/L	99	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Dibromochloromethane	524.2	0.5	---		8.6500	10.0	ug/L	86	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	1,2-Dibromo-3-chloropropane (DBCP)	524.2	0.2	---		10.4110	10.0	ug/L	104	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	1,2-Dibromoethane (EDB)	524.2	0.2	---		9.5900	10.0	ug/L	96	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Dibromomethane	524.2	0.5	---		9.6310	10.0	ug/L	96	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	1,2-Dichlorobenzene	524.2	0.5	---		9.8460	10.0	ug/L	98	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	1,3-Dichlorobenzene	524.2	0.5	---		9.9660	10.0	ug/L	100	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	1,4-Dichlorobenzene	524.2	0.5	---		9.8790	10.0	ug/L	99	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Dichlorodifluoromethane	524.2	0.5	---		9.0210	10.0	ug/L	90	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	1,1-Dichloroethane	524.2	0.5	---		10.0650	10.0	ug/L	101	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	1,2-Dichloroethane	524.2	0.5	---		9.6610	10.0	ug/L	97	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	1,1-Dichloroethylene	524.2	0.5	---		9.8990	10.0	ug/L	99	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	cis-1,2-Dichloroethylene	524.2	0.5	---		10.0380	10.0	ug/L	100	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	trans-1,2-Dichloroethylene	524.2	0.5	---		9.7050	10.0	ug/L	97	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	1,2-Dichloropropane	524.2	0.5	---		10.1800	10.0	ug/L	102	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	1,3-Dichloropropane	524.2	0.5	---		9.6540	10.0	ug/L	97	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	2,2-Dichloropropane	524.2	0.5	---		10.7800	10.0	ug/L	108	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	1,1-Dichloropropylene	524.2	0.5	---		9.5940	10.0	ug/L	96	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	cis-1,3-Dichloropropylene	524.2	0.5	---		9.7790	10.0	ug/L	98	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	trans-1,3-Dichloropropylene	524.2	0.5	---		9.4490	10.0	ug/L	94	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Ethylbenzene	524.2	0.5	---		10.0770	10.0	ug/L	101	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Hexachlorobutadiene	524.2	0.5	---		9.3440	10.0	ug/L	93	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	2-Hexanone	524.2	5.0	---		9.1880	10.0	ug/L	92	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Isopropylbenzene	524.2	0.5	---		10.0490	10.0	ug/L	100	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	4-Isopropyltoluene	524.2	0.5	---		9.9880	10.0	ug/L	100	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	4-Methyl-2-pentanone (MIBK)	524.2	2.0	---		9.2440	10.0	ug/L	92	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Methyl-t-butyl ether (MTBE)	524.2	0.5	---		9.6200	10.0	ug/L	96	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Naphthalene	524.2	0.5	---		8.9070	10.0	ug/L	89	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	n-Propylbenzene	524.2	0.5	---		10.0290	10.0	ug/L	100	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Styrene	524.2	0.5	---		10.1860	10.0	ug/L	102	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	1,1,1,2-Tetrachloroethane	524.2	0.5	---		9.9200	10.0	ug/L	99	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	1,1,2,2-Tetrachloroethane	524.2	0.5	---		10.9570	10.0	ug/L	110	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Tetrachloroethylene	524.2	0.5	---		9.6480	10.0	ug/L	96	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Tetrahydrofuran	524.2	5.0	---		9.2700	10.0	ug/L	93	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Toluene	524.2	0.5	---		9.9400	10.0	ug/L	99	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	1,2,3-Trichlorobenzene	524.2	0.5	---		9.6100	10.0	ug/L	96	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	1,2,4-Trichlorobenzene	524.2	0.5	---		9.8130	10.0	ug/L	98	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	1,1,1-Trichloroethane	524.2	0.5	---		10.0040	10.0	ug/L	100	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	1,1,2-Trichloroethane	524.2	0.5	---		9.5460	10.0	ug/L	95	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Trichloroethylene	524.2	0.5	---		10.0440	10.0	ug/L	100	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	Trichlorofluoromethane	524.2	0.5	---		9.6620	10.0	ug/L	97	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	1,2,3-Trichloropropane	524.2	0.5	---		8.9620	10.0	ug/L	90	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	1,1,2-Trichloro-1,2,2-trifluoroethane	524.2	0.5	---		9.7070	10.0	ug/L	97	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	1,2,3-Trimethylbenzene	524.2	0.5	---		10.0400	10.0	ug/L	100	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	1,2,4-Trimethylbenzene	524.2	0.5	---		10.1280	10.0	ug/L	101	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	1,3,5-Trimethylbenzene	524.2	0.5	---		10.0020	10.0	ug/L	100	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	Vinyl chloride	524.2	0.2	---		9.5970	10.0	ug/L	96	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	1,2-Xylene	524.2	0.5	---		10.0500	10.0	ug/L	100	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
CCC	1,3 + 1,4-Xylene	524.2	0.5	---		19.9440	20.0	ug/L	100	70 - 130	---	---	1.0	---	05/15/2019 20:52	4293301
LMB	IS-1,4-Difluorobenzene	524.2	N/A	---		245019	243271	ug/L	101	70 - 130	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	SS-Bromofluorobenzene	524.2	N/A	---		4.8180	5.0	ug/L	96	70 - 130	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	SS-1,2-Dichlorobenzene-d4	524.2	N/A	---		9.6440	10.0	ug/L	96	70 - 130	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	SS-1,2-Dichloroethane-d4	524.2	N/A	---		9.5900	10.0	ug/L	96	70 - 130	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	SS-Toluene-d8	524.2	N/A	---		9.9040	10.0	ug/L	99	70 - 130	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	Acetone	524.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	tert-Amyl Methyl ether	524.2	3.0	---	<	3.0		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	Benzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	Benzyl chloride	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	Bromobenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	Bromochloromethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	Bromodichloromethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	Bromomethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	Bromotoluene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	1,3-Butadiene	524.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	2-Butanone (MEK)	524.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	n-Butyl acrylate	524.2	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	n-Butylbenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	sec-Butylbenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	tert-Butylbenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	Carbon disulfide	524.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	Carbon tetrachloride	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	Chlorobenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	Chloroethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	Chloroform	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	Chloromethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	Chloroprene	524.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	2-Chlorotoluene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	4-Chlorotoluene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	Cyclohexanone	524.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	Dibromochloromethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	1,2-Dibromo-3-chloropropane (DBCP)	524.2	0.2	---	<	0.2		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	1,2-Dibromoethane (EDB)	524.2	0.2	---	<	0.2		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	Dibromomethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	1,2-Dichlorobenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	1,3-Dichlorobenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	1,4-Dichlorobenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953
LMB	Dichlorodifluoromethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	4293953

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LMB	1,1-Dichloroethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	1,2-Dichloroethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	1,1-Dichloroethylene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	cis-1,2-Dichloroethylene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	trans-1,2-Dichloroethylene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	1,2-Dichloropropane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	1,3-Dichloropropane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	2,2-Dichloropropane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	1,1-Dichloropropylene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	cis-1,3-Dichloropropylene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	trans-1,3-Dichloropropylene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	Diisopropyl ether	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	1,4-Dioxane	524.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	Epichlorohydrin	524.2	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	Ethyl acrylate	524.2	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	Ethylbenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	Ethyl tert-butyl ether	524.2	3.0	---	<	3.0		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	Hexachlorobutadiene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	2-Hexanone	524.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	Isopropylbenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	4-Isopropyltoluene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	4-Methyl-2-pentanone (MIBK)	524.2	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	tert-Butyl alcohol	524.2	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	Methyl-t-butyl ether (MTBE)	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	Naphthalene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	n-Propylbenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	Styrene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	1,1,1,2-Tetrachloroethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	1,1,2,2-Tetrachloroethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	Tetrachloroethylene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	Tetrahydrofuran	524.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	Toluene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	1,2,3-Trichlorobenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	1,2,4-Trichlorobenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	1,1,1-Trichloroethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	1,1,2-Trichloroethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	Trichloroethylene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	Trichlorofluoromethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	1,2,3-Trichloropropane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	1,1,2-Trichloro-1,2-trifluoroethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953
LMB	1,2,3-Trimethylbenzene	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/15/2019 21:55	42939953

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LMB	1,2,4-Trimethylbenzene	524.2	0.5	---	<	0.5		ug/L	----	----	----	----	1.0	----	05/15/2019 21:55	4293953
LMB	1,3,5-Trimethylbenzene	524.2	0.5	---	<	0.5		ug/L	----	----	----	----	1.0	----	05/15/2019 21:55	4293953
LMB	Vinyl acetate	524.2	5.0	---	<	5.0		ug/L	----	----	----	----	1.0	----	05/15/2019 21:55	4293953
LMB	Vinyl chloride	524.2	0.2	---	<	0.2		ug/L	----	----	----	----	1.0	----	05/15/2019 21:55	4293953
LMB	1,2-Xylene	524.2	0.5	---	<	0.5		ug/L	----	----	----	----	1.0	----	05/15/2019 21:55	4293953
LMB	1,3 + 1,4-Xylene	524.2	0.5	---	<	0.5		ug/L	----	----	----	----	1.0	----	05/15/2019 21:55	4293953



Eurofins Eaton Analytical

Run Log

Run ID: 259443 Method: 524.2

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
CCC	4295317		RW	B	05/16/2019 12:18	524 2-050319B-SC.mth
CCL	4295324		RW	B	05/16/2019 13:09	524 2-050319B-SC.mth
LMB	4295320		RW	B	05/16/2019 13:51	524 2-050319B-SC.mth
LTB	4291774	LTB-04/30/19	RW	B	05/16/2019 17:51	524 2-050319B-SC.mth
FS	4291763	Finished	DW	B	05/16/2019 18:24	524 2-050319B-SC.mth
CCC	4296949		RW	B	05/16/2019 21:11	524 2-050319B-SC.mth

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	Acrylonitrile	524.2	1.0	---		5.5970	5.0	ug/L	112	70 - 130	---	---	1.0	---	05/16/2019 12:18	4295317
CCC	Allyl chloride	524.2	5.0	---		5.9960	5.0	ug/L	120	70 - 130	---	---	1.0	---	05/16/2019 12:18	4295317
CCC	Chloroacetonitrile	524.2	5.0	---		5.7580	5.0	ug/L	115	70 - 130	---	---	1.0	---	05/16/2019 12:18	4295317
CCC	1-Chlorobutane	524.2	5.0	---		5.5400	5.0	ug/L	111	70 - 130	---	---	1.0	---	05/16/2019 12:18	4295317
CCC	trans-1,4-Dichloro-2-butylene	524.2	5.0	---		5.4280	5.0	ug/L	109	70 - 130	---	---	1.0	---	05/16/2019 12:18	4295317
CCC	Dichloromethane	524.2	0.5	---		4.6440	5.0	ug/L	93	70 - 130	---	---	1.0	---	05/16/2019 12:18	4295317
CCC	1,1-Dichloropropanone	524.2	5.0	---		5.1650	5.0	ug/L	103	70 - 130	---	---	1.0	---	05/16/2019 12:18	4295317
CCC	Ethyl Ether	524.2	2.0	---		6.1980	5.0	ug/L	124	70 - 130	---	---	1.0	---	05/16/2019 12:18	4295317
CCC	Ethyl methacrylate	524.2	1.0	---		5.7970	5.0	ug/L	116	70 - 130	---	---	1.0	---	05/16/2019 12:18	4295317
CCC	Hexachloroethane	524.2	2.0	---		4.9680	5.0	ug/L	99	70 - 130	---	---	1.0	---	05/16/2019 12:18	4295317
CCC	Methylacrylate	524.2	1.0	---		5.3510	5.0	ug/L	107	70 - 130	---	---	1.0	---	05/16/2019 12:18	4295317
CCC	Methyl iodide	524.2	2.0	---		4.8910	5.0	ug/L	98	70 - 130	---	---	1.0	---	05/16/2019 12:18	4295317
CCC	Methylmethacrylate	524.2	1.0	---		5.6000	5.0	ug/L	112	70 - 130	---	---	1.0	---	05/16/2019 12:18	4295317
CCC	Nitrobenzene	524.2	5.0	---		5.0610	5.0	ug/L	101	70 - 130	---	---	1.0	---	05/16/2019 12:18	4295317
CCC	2-Nitropropane	524.2	2.0	---		5.6350	5.0	ug/L	113	70 - 130	---	---	1.0	---	05/16/2019 12:18	4295317
CCC	Propionitrile	524.2	5.0	---		5.4740	5.0	ug/L	109	70 - 130	---	---	1.0	---	05/16/2019 12:18	4295317
CCL	Dichloromethane	524.2	0.5	---		0.3540	0.5	ug/L	71	50 - 150	---	---	1.0	---	05/16/2019 13:09	4295324
LMB	Acrylonitrile	524.2	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/16/2019 13:51	4295320
LMB	Allyl chloride	524.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/16/2019 13:51	4295320
LMB	Chloroacetonitrile	524.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/16/2019 13:51	4295320
LMB	1-Chlorobutane	524.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/16/2019 13:51	4295320
LMB	trans-1,4-Dichloro-2-butylene	524.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/16/2019 13:51	4295320
LMB	Dichloromethane	524.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/16/2019 13:51	4295320
LMB	1,1-Dichloropropanone	524.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/16/2019 13:51	4295320
LMB	Ethyl Ether	524.2	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/16/2019 13:51	4295320
LMB	Ethyl methacrylate	524.2	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/16/2019 13:51	4295320
LMB	Hexachloroethane	524.2	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/16/2019 13:51	4295320
LMB	Methylacrylate	524.2	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/16/2019 13:51	4295320
LMB	Methyl iodide	524.2	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/16/2019 13:51	4295320
LMB	Methylmethacrylate	524.2	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/16/2019 13:51	4295320
LMB	Nitrobenzene	524.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/16/2019 13:51	4295320
LMB	2-Nitropropane	524.2	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	---	05/16/2019 13:51	4295320
LMB	Propionitrile	524.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	---	05/16/2019 13:51	4295320
LTB	Acrylonitrile	524.2	1.0	LTB-04/30/19	<	1.0		ug/L	---	---	---	---	1.0	---	05/16/2019 17:51	4291774
LTB	Allyl chloride	524.2	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	---	05/16/2019 17:51	4291774
LTB	Chloroacetonitrile	524.2	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	---	05/16/2019 17:51	4291774
LTB	1-Chlorobutane	524.2	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	---	05/16/2019 17:51	4291774
LTB	trans-1,4-Dichloro-2-butylene	524.2	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	---	05/16/2019 17:51	4291774
LTB	Dichloromethane	524.2	0.5	LTB-04/30/19	<	0.5		ug/L	---	---	---	---	1.0	---	05/16/2019 17:51	4291774
LTB	1,1-Dichloropropanone	524.2	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	---	05/16/2019 17:51	4291774

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LTB	Ethyl Ether	524.2	2.0	LTB-04/30/19	<	2.0		ug/L	---	---	---	---	1.0	---	05/16/2019 17:51	4291774
LTB	Ethyl methacrylate	524.2	1.0	LTB-04/30/19	<	1.0		ug/L	---	---	---	---	1.0	---	05/16/2019 17:51	4291774
LTB	Hexachloroethane	524.2	2.0	LTB-04/30/19	<	2.0		ug/L	---	---	---	---	1.0	---	05/16/2019 17:51	4291774
LTB	Methylacrylate	524.2	1.0	LTB-04/30/19	<	1.0		ug/L	---	---	---	---	1.0	---	05/16/2019 17:51	4291774
LTB	Methyl iodide	524.2	2.0	LTB-04/30/19	<	2.0		ug/L	---	---	---	---	1.0	---	05/16/2019 17:51	4291774
LTB	Methylmethacrylate	524.2	1.0	LTB-04/30/19	<	1.0		ug/L	---	---	---	---	1.0	---	05/16/2019 17:51	4291774
LTB	Nitrobenzene	524.2	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	---	05/16/2019 17:51	4291774
LTB	2-Nitropropane	524.2	2.0	LTB-04/30/19	<	2.0		ug/L	---	---	---	---	1.0	---	05/16/2019 17:51	4291774
LTB	Propionitrile	524.2	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	---	05/16/2019 17:51	4291774
FS	Acrylonitrile	524.2	1.0	Finished	<	1.0		ug/L	---	---	---	---	1.0	---	05/16/2019 18:24	4291763
FS	Allyl chloride	524.2	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	---	05/16/2019 18:24	4291763
FS	Chloroacetonitrile	524.2	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	---	05/16/2019 18:24	4291763
FS	1-Chlorobutane	524.2	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	---	05/16/2019 18:24	4291763
FS	trans-1,4-Dichloro-2-butylene	524.2	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	---	05/16/2019 18:24	4291763
FS	Dichloromethane	524.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/16/2019 18:24	4291763
FS	1,1-Dichloropropanone	524.2	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	---	05/16/2019 18:24	4291763
FS	Ethyl Ether	524.2	2.0	Finished	<	2.0		ug/L	---	---	---	---	1.0	---	05/16/2019 18:24	4291763
FS	Ethyl methacrylate	524.2	1.0	Finished	<	1.0		ug/L	---	---	---	---	1.0	---	05/16/2019 18:24	4291763
FS	Hexachloroethane	524.2	2.0	Finished	<	2.0		ug/L	---	---	---	---	1.0	---	05/16/2019 18:24	4291763
FS	Methylacrylate	524.2	1.0	Finished	<	1.0		ug/L	---	---	---	---	1.0	---	05/16/2019 18:24	4291763
FS	Methyl iodide	524.2	2.0	Finished	<	2.0		ug/L	---	---	---	---	1.0	---	05/16/2019 18:24	4291763
FS	Methylmethacrylate	524.2	1.0	Finished	<	1.0		ug/L	---	---	---	---	1.0	---	05/16/2019 18:24	4291763
FS	Nitrobenzene	524.2	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	---	05/16/2019 18:24	4291763
FS	2-Nitropropane	524.2	2.0	Finished	<	2.0		ug/L	---	---	---	---	1.0	---	05/16/2019 18:24	4291763
FS	Propionitrile	524.2	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	---	05/16/2019 18:24	4291763
CCC	Acrylonitrile	524.2	1.0	---	<	8.2170	7.5	ug/L	110	70 - 130	---	---	1.0	---	05/16/2019 21:11	4296949
CCC	Allyl chloride	524.2	5.0	---	<	8.7790	7.5	ug/L	117	70 - 130	---	---	1.0	---	05/16/2019 21:11	4296949
CCC	Chloroacetonitrile	524.2	5.0	---	<	8.4970	7.5	ug/L	113	70 - 130	---	---	1.0	---	05/16/2019 21:11	4296949
CCC	1-Chlorobutane	524.2	5.0	---	<	7.9830	7.5	ug/L	106	70 - 130	---	---	1.0	---	05/16/2019 21:11	4296949
CCC	trans-1,4-Dichloro-2-butylene	524.2	5.0	---	<	7.7410	7.5	ug/L	103	70 - 130	---	---	1.0	---	05/16/2019 21:11	4296949
CCC	Dichloromethane	524.2	0.5	---	<	7.6070	7.5	ug/L	101	70 - 130	---	---	1.0	---	05/16/2019 21:11	4296949
CCC	1,1-Dichloropropanone	524.2	5.0	---	<	7.9380	7.5	ug/L	106	70 - 130	---	---	1.0	---	05/16/2019 21:11	4296949
CCC	Ethyl Ether	524.2	2.0	---	<	8.3650	7.5	ug/L	112	70 - 130	---	---	1.0	---	05/16/2019 21:11	4296949
CCC	Ethyl methacrylate	524.2	1.0	---	<	8.3290	7.5	ug/L	111	70 - 130	---	---	1.0	---	05/16/2019 21:11	4296949
CCC	Hexachloroethane	524.2	2.0	---	<	7.8630	7.5	ug/L	105	70 - 130	---	---	1.0	---	05/16/2019 21:11	4296949
CCC	Methylacrylate	524.2	1.0	---	<	8.0260	7.5	ug/L	107	70 - 130	---	---	1.0	---	05/16/2019 21:11	4296949
CCC	Methyl iodide	524.2	2.0	---	<	7.2370	7.5	ug/L	96	70 - 130	---	---	1.0	---	05/16/2019 21:11	4296949
CCC	Methylmethacrylate	524.2	1.0	---	<	7.8800	7.5	ug/L	105	70 - 130	---	---	1.0	---	05/16/2019 21:11	4296949
CCC	Nitrobenzene	524.2	5.0	---	<	7.5310	7.5	ug/L	100	70 - 130	---	---	1.0	---	05/16/2019 21:11	4296949
CCC	2-Nitropropane	524.2	2.0	---	<	8.2400	7.5	ug/L	110	70 - 130	---	---	1.0	---	05/16/2019 21:11	4296949
CCC	Propionitrile	524.2	5.0	---	<	8.6560	7.5	ug/L	115	70 - 130	---	---	1.0	---	05/16/2019 21:11	4296949

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD Limit	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #

Eurofins Eaton Analytical

Run Log

Run ID: **259793** Method: **525.2**

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
CCC	4301069		OS	DO	05/22/2019 16:40	525 2-DO-101018a-up2.mth
CCC	4301070		OS	DO	05/22/2019 17:22	525 2-DO-101018a-up2.mth
CCC	4301071		OS	DO	05/22/2019 18:04	525 2-DO-101018a-up2.mth
LFB	4301066		RW	DO	05/22/2019 18:46	525 2-DO-101018a-up2.mth
LFB	4301067		RW	DO	05/22/2019 19:28	525 2-DO-101018a-up2.mth
LFB	4301068		RW	DO	05/22/2019 20:10	525 2-DO-101018a-up2.mth
LMB	4301065		RW	DO	05/22/2019 22:16	525 2-DO-101018a-up2.mth
FS	4291759	Finished	DW	DO	05/22/2019 22:58	525 2-DO-101018a-up2.mth
CCC	4302646		OS	DO	05/23/2019 06:42	525 2-DO-101018a-up2.mth
CCC	4302647		OS	DO	05/23/2019 07:24	525 2-DO-101018a-up2.mth
CCC	4302648		OS	DO	05/23/2019 08:06	525 2-DO-101018a-up2.mth
LFB	4301067		RW	DO	05/23/2019 09:30	525 2-DO-101018a-up2.mth
LFB	4301068		RW	DO	05/23/2019 10:12	525 2-DO-101018a-up2.mth

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	IS-Chrysene-d12	525.2	N/A	--		1606000	1606000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	IS-Phenanthrene-d10	525.2	N/A	--		2173000	2173000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	IS-Pyrene-d10	525.2	N/A	--		1720000	1720000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	SS-4,4'-Dichlorobiphenyl	525.2	N/A	--		4.4570	5.0	ug/L	89	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	SS-2,4,5,6-Tetrachloro-m-xylene	525.2	N/A	--		5.5190	5.0	ug/L	110	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	SS-Triphenylphosphate	525.2	N/A	--		5.2880	5.0	ug/L	106	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Acenaphthene	525.2	0.1	--		5.0920	5.0	ug/L	102	65 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Acenaphthylene	525.2	0.1	--		5.4300	5.0	ug/L	109	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Anthracene	525.2	0.1	--		5.1620	5.0	ug/L	103	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Azinhpos-ethyl	525.2	0.5	--		5.8410	5.0	ug/L	117	70 - 145	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Bendiocarb	525.2	0.5	--		6.4820	5.0	ug/L	130	65 - 159	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Benzo(a)anthracene	525.2	0.1	--		4.7690	5.0	ug/L	95	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Benzo(b)fluoranthene	525.2	0.1	--		4.9070	5.0	ug/L	98	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Benzo(k)fluoranthene	525.2	0.1	--		5.0330	5.0	ug/L	101	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Benzo(g,h,i)perylene	525.2	0.1	--		4.8590	5.0	ug/L	97	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Butylbenzylphthalate	525.2	1.0	--		6.3570	5.0	ug/L	127	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Carbophenothion	525.2	0.5	--		5.3700	5.0	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Chlorfenvinphos	525.2	5.0	--		6.0350	4.55	ug/L	133	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	2-Chlorobiphenyl	525.2	0.1	--		5.2640	5.0	ug/L	105	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Chlorpyrifos methyl	525.2	0.5	--		6.6580	5.0	ug/L	133	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Chrysene	525.2	0.1	--		4.5620	5.0	ug/L	91	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Clopyralid	525.2	10	--		3.2700	10.0	ug/L	33	1 - 109	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Crotoxyphos	525.2	0.5	--		7.0690	5.0	ug/L	141	70 - 173	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Dibenzo(a,h)anthracene	525.2	0.1	--		5.0440	5.0	ug/L	101	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Di-n-butylphthalate	525.2	2.0	--		5.9530	5.0	ug/L	119	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	2,3-Dichlorobiphenyl	525.2	0.1	--		5.5010	5.0	ug/L	110	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Dicrotophos	525.2	0.5	--		7.9060	5.0	ug/L	158	70 - 191	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Diethylphthalate	525.2	1.0	--		5.5250	5.0	ug/L	110	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Dimethylphthalate	525.2	1.0	--		4.9420	5.0	ug/L	99	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Di-n-octylphthalate	525.2	2.0	--		6.6590	5.0	ug/L	133	48 - 173	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Dioxathion A	525.2	0.5	--		6.7890	2.8	ug/L	242	34 - 209	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Dioxathion B	525.2	0.5	--		4.0920	2.2	ug/L	186	14 - 498	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Disulfoton sulfoxide	525.2	10	--		10.4080	10.0	ug/L	104	69 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	EPN	525.2	0.5	--		7.1660	5.0	ug/L	143	70 - 147	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Erucylamide	525.2	5.0	--		19.4470	10.0	ug/L	194	43 - 251	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Esfenvalerate	525.2	0.5	--		5.7430	5.0	ug/L	115	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Fenarimol	525.2	1.0	--		6.7110	5.0	ug/L	134	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Fenitrothion	525.2	0.5	--		7.2370	5.0	ug/L	145	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Fenoxprop-ethyl	525.2	1.0	--		6.5660	5.0	ug/L	131	70 - 140	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Fensulfothion	525.2	0.5	--		7.2560	5.0	ug/L	145	70 - 191	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	Fluorene	525.2	0.1	---		5.3800	5.0	ug/L	108	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Fluridone	525.2	1.0	---		4.7880	5.0	ug/L	96	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	2,2',3',3',4',4',6-Heptachlorobiphenyl	525.2	0.5	---		3.9610	5.0	ug/L	79	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	2,2',4,4',5,6'-Hexachlorobiphenyl	525.2	0.1	---		5.1570	5.0	ug/L	103	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Indeno(1,2,3-cd)pyrene	525.2	0.1	---		5.0230	5.0	ug/L	100	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Iprodione	525.2	0.5	---		7.1640	5.0	ug/L	143	70 - 207	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Isofenphos	525.2	0.5	---		7.0560	5.0	ug/L	141	70 - 145	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Leptophos	525.2	0.5	---		5.5290	5.0	ug/L	111	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	MetaIaxyl	525.2	0.5	---		6.7630	5.0	ug/L	135	70 - 143	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Methyl parathion	525.2	0.5	---		9.0750	5.0	ug/L	182	70 - 206	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Metsulfuron-methyl	525.2	10	---		12.3990	10.0	ug/L	124	70 - 170	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Mirex	525.2	0.5	---		5.7990	5.0	ug/L	116	61 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Monocrotophos	525.2	0.5	---		5.9890	5.0	ug/L	120	70 - 185	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Naled	525.2	0.5	---		6.0960	5.0	ug/L	122	29 - 160	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	cis-Nonachlor	525.2	0.1	---		4.6370	5.0	ug/L	93	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Norflurazon	525.2	1.0	---		6.0430	5.0	ug/L	121	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	2,2',3',3',4',5',6',6'-Octachlorobiphenyl	525.2	0.5	---		5.4710	5.0	ug/L	109	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Oryzalin	525.2	10	---		12.2910	5.0	ug/L	246	1 - 389	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Oxychloridane	525.2	0.1	---		6.2900	5.0	ug/L	126	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Oxyfluorfen	525.2	0.5	---		5.5770	5.0	ug/L	112	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Parathion	525.2	0.5	---		6.6870	5.0	ug/L	134	63 - 145	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Pentachlorobenzene	525.2	0.5	---		5.8970	5.0	ug/L	118	61 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	2,2',3',4',6-Pentachlorobiphenyl	525.2	0.1	---		4.7900	5.0	ug/L	96	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Pentachlorophenol	525.2	1.0	---		7.1490	5.0	ug/L	143	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Phenanthrene	525.2	0.1	---		5.2280	5.0	ug/L	105	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Phosmet	525.2	0.5	---		4.9200	5.0	ug/L	98	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	E-Phosphamidon	525.2	0.5	---		2.2650	1.45	ug/L	156	70 - 302	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Z-Phosphamidon	525.2	0.5	---		5.4400	3.55	ug/L	153	70 - 189	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Propiconazole isomer a	525.2	5.0	---		2.0960	1.6	ug/L	131	1 - 278	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Propiconazole isomer b	525.2	5.0	---		3.8370	3.4	ug/L	113	61 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Prothiofos	525.2	0.5	---		6.5300	5.0	ug/L	131	70 - 135	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Pyrene	525.2	0.1	---		5.6920	5.0	ug/L	114	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Sulfotep	525.2	0.5	---		4.6040	5.0	ug/L	92	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Tebuthiuron	525.2	10	---		12.1150	10.0	ug/L	121	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	TEPP	525.2	1.0	---		9.5330	5.0	ug/L	191	1 - 267	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	1,2,4,5-Tetrachlorobenzene	525.2	0.5	---		5.5770	5.0	ug/L	112	57 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	2,2',4,4'-Tetrachlorobiphenyl	525.2	0.1	---		5.8190	5.0	ug/L	116	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Thiabendazole	525.2	10	---		7.7740	5.0	ug/L	155	6 - 190	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Thionazin	525.2	0.5	---		6.7090	5.0	ug/L	134	62 - 137	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Tribufos	525.2	0.1	---		6.7690	5.0	ug/L	135	70 - 136	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	2,4,5-Trichlorobiphenyl	525.2	0.1	---		5.5500	5.0	ug/L	111	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	Trichloronate	525.2	0.5	---		6.8460	5.0	ug/L	137	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Tricyclazole	525.2	1.0	---		5.5670	5.0	ug/L	111	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	Vinclozolin	525.2	0.5	---		6.4380	5.0	ug/L	129	70 - 137	---	---	1.0	05/21/2019 08:45	05/22/2019 16:40	4301069
CCC	IS-Chrysene-d12	525.2	N/A	---		1626000	1626000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	IS-Phenanthrene-d10	525.2	N/A	---		2568000	2568000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	IS-Pyrene-d10	525.2	N/A	---		1836000	1836000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	SS-4,4'-Dichlorobiphenyl	525.2	N/A	---		4.8170	5.0	ug/L	96	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	SS-2,4,5,6-Tetrachloro-m-xylene	525.2	N/A	---		4.8290	5.0	ug/L	97	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	SS-Triphenylphosphate	525.2	N/A	---		5.1470	5.0	ug/L	103	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Anilazine	525.2	1.0	---		3.5840	5.0	ug/L	72	41 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	alpha-Hexachlorocyclohexane	525.2	0.1	---		5.8080	5.0	ug/L	116	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	beta-BHC	525.2	0.1	---		5.2060	5.0	ug/L	104	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	delta-BHC	525.2	0.1	---		5.5070	5.0	ug/L	110	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Butylate	525.2	0.1	---		5.5340	5.0	ug/L	111	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Chlorobenzilate	525.2	0.1	---		6.0150	5.0	ug/L	120	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Chloroneb	525.2	0.1	---		6.1640	5.0	ug/L	123	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Cycloate	525.2	0.1	---		5.3350	5.0	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	DCPA	525.2	0.1	---		5.0870	5.0	ug/L	102	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	4,4'-DDD	525.2	0.1	---		5.0840	5.0	ug/L	102	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	4,4'-DDE	525.2	0.1	---		4.9560	5.0	ug/L	99	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Demeton O	525.2	0.5	---		2.5810	1.7	ug/L	152	70 - 174	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Demeton S	525.2	0.5	---		4.8200	3.3	ug/L	146	67 - 215	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Desethylatrazine	525.2	1.0	---		5.3290	5.0	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Diazinon	525.2	0.1	---		5.2480	5.0	ug/L	105	66 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Dichlofenthion	525.2	0.1	---		5.3040	5.0	ug/L	106	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Dichloran	525.2	0.5	---		5.7320	5.0	ug/L	115	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Dichlorvos	525.2	0.1	---		6.0480	5.0	ug/L	121	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Disulfoton	525.2	0.1	---		5.9090	5.0	ug/L	118	61 - 156	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Endosulfan I	525.2	0.1	---		5.3610	5.0	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Endosulfan II	525.2	0.1	---		5.3690	5.0	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Endosulfan sulfate	525.2	0.1	---		5.2950	5.0	ug/L	106	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Endrin aldehyde	525.2	0.5	---		5.0290	5.0	ug/L	101	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	EPTC	525.2	0.1	---		4.7840	5.0	ug/L	96	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Fenthion	525.2	0.1	---		4.6800	5.0	ug/L	94	63 - 137	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Fluchloralin	525.2	0.1	---		5.1250	5.0	ug/L	102	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Fluoranthene	525.2	0.1	---		5.0520	5.0	ug/L	101	65 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Fonofos	525.2	0.1	---		5.2590	5.0	ug/L	105	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	2-Methylnaphthalene	525.2	0.1	---		5.6890	5.0	ug/L	114	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	MKG 264 isomer a	525.2	0.1	---		3.0590	3.15	ug/L	97	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	MKG 264 isomer b	525.2	0.1	---		1.9710	1.85	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	MKG 326	525.2	0.1	---		5.4610	5.0	ug/L	109	70 - 131	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	Molinate	525.2	0.1	---		5.4820	5.0	ug/L	110	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Pebulate	525.2	0.1	---		4.8470	5.0	ug/L	97	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Pendimethalin	525.2	0.1	---		5.5530	5.0	ug/L	111	70 - 132	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	cis-Permethrin	525.2	0.1	---		2.7190	2.05	ug/L	133	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	trans-Permethrin	525.2	0.1	---		3.0850	2.95	ug/L	105	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Profluralin	525.2	0.1	---		5.0720	5.0	ug/L	101	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Prometryn	525.2	0.1	---		4.9540	5.0	ug/L	99	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Simetryn	525.2	0.1	---		5.5160	5.0	ug/L	110	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Terbutyn	525.2	0.1	---		4.9860	5.0	ug/L	100	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	Triadimefon	525.2	0.5	---		5.1250	5.0	ug/L	102	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 17:22	4301070
CCC	IS-Chrysene-d12	525.2	N/A	---		1575000	1575000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	IS-Phenanthrene-d10	525.2	N/A	---		2301000	2301000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	IS-Pyrene-d10	525.2	N/A	---		1749000	1749000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	SS-4,4'-Dichlorobiphenyl	525.2	N/A	---		5.6200	5.0	ug/L	112	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	SS-2,4,5,6-Tetrachloro-m-xylene	525.2	N/A	---		4.7840	5.0	ug/L	96	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	SS-Triphenylphosphate	525.2	N/A	---		5.4870	5.0	ug/L	110	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Acetochlor	525.2	0.1	---		6.1420	5.0	ug/L	123	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Alachlor	525.2	0.1	---		6.1680	5.0	ug/L	123	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Aldrin	525.2	0.1	---		6.3190	5.0	ug/L	126	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Ametryn	525.2	0.1	---		5.9000	5.0	ug/L	118	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Aspon	525.2	0.1	---		4.7860	5.0	ug/L	96	69 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Atrazine	525.2	0.1	---		5.7360	5.0	ug/L	115	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Atrazine	525.2	0.1	---		6.2280	5.0	ug/L	125	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Azinphos-methyl	525.2	0.5	---		7.2170	5.0	ug/L	144	70 - 151	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Benfluralin	525.2	0.1	---		5.7120	5.0	ug/L	114	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Benzo(a)pyrene	525.2	0.02	---		5.6190	5.0	ug/L	112	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	gamma-BHC (Lindane)	525.2	0.02	---		5.5070	5.0	ug/L	110	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Bolstar	525.2	0.1	---		5.1920	5.0	ug/L	104	70 - 132	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Bromacil	525.2	0.1	---		8.6590	5.0	ug/L	173	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Butachlor	525.2	0.1	---		5.8260	5.0	ug/L	117	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Carboxin	525.2	0.1	---		5.7950	5.0	ug/L	116	60 - 182	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	alpha-Chlordane	525.2	0.1	---		5.8050	5.0	ug/L	116	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	gamma-Chlordane	525.2	0.1	---		5.8760	5.0	ug/L	118	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Chloropropylate	525.2	0.1	---		5.6120	5.0	ug/L	112	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Chlorothalonil	525.2	0.1	---		5.7090	5.0	ug/L	114	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Chlorpropham	525.2	0.1	---		6.0440	5.0	ug/L	121	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Chlorpyrifos	525.2	0.1	---		6.1890	5.0	ug/L	124	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Clomazone	525.2	0.1	---		5.2260	5.0	ug/L	105	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Coumaphos	525.2	0.1	---		5.9970	5.0	ug/L	120	67 - 165	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Cyanazine	525.2	0.1	---		4.9940	5.0	ug/L	100	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	4,4'-DDT	525.2	0.1	---		5.5810	5.0	ug/L	112	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	Desisopropylatrazine	525.2	1.0	---		5.6360	5.0	ug/L	113	62 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Dichlobenil	525.2	0.1	---		5.9330	5.0	ug/L	119	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Dieldrin	525.2	0.1	---		5.5630	5.0	ug/L	111	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Di(2-ethylhexyl)adipate	525.2	0.6	---		5.9750	5.0	ug/L	120	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Di(2-ethylhexyl)phthalate	525.2	0.6	---		5.7550	5.0	ug/L	115	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Dimethoate	525.2	0.5	---		5.8820	5.0	ug/L	118	62 - 131	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	2,4-Dinitrotoluene	525.2	0.5	---		5.3460	5.0	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	2,6-Dinitrotoluene	525.2	0.5	---		5.7400	5.0	ug/L	115	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Diphenamid	525.2	0.1	---		5.8890	5.0	ug/L	118	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Disulfoton sulfone	525.2	0.1	---		6.2690	5.0	ug/L	125	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Endrin	525.2	0.01	---		6.1510	5.0	ug/L	123	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Ethalfuralin	525.2	0.1	---		6.4660	5.0	ug/L	129	70 - 136	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Ethion	525.2	5.0	---		5.3620	5.0	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Ethoflumesate	525.2	0.5	---		6.0480	5.0	ug/L	121	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Ethoprop	525.2	0.1	---		4.1840	5.0	ug/L	84	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Etridiazole	525.2	0.1	---		7.0360	5.0	ug/L	141	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Famphur	525.2	0.1	---		4.7760	5.0	ug/L	96	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Fenamiphos	525.2	0.1	---		7.4560	5.0	ug/L	149	61 - 234	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Fluazifop-butyl	525.2	0.1	---		4.7080	5.0	ug/L	94	66 - 140	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Fluometuron	525.2	0.5	---		8.3920	5.0	ug/L	168	42 - 159	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Heptachlor	525.2	0.04	---		6.3010	5.0	ug/L	126	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Heptachlor epoxide	525.2	0.02	---		5.5410	5.0	ug/L	111	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Hexachlorobenzene	525.2	0.1	---		4.2820	5.0	ug/L	86	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Hexachlorocyclopentadiene	525.2	0.1	---		4.4980	5.0	ug/L	90	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Hexazinone	525.2	0.1	---		5.6980	5.0	ug/L	114	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Isophorone	525.2	0.1	---		5.1190	5.0	ug/L	102	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Maliathion	525.2	0.1	---		5.5640	5.0	ug/L	111	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Methoxychlor	525.2	0.1	---		6.3660	5.0	ug/L	127	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	1-Methylnaphthalene	525.2	0.1	---		5.4520	5.0	ug/L	109	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Methyl paraoxon	525.2	0.5	---		7.0370	5.0	ug/L	141	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Metolachlor	525.2	0.1	---		5.9020	5.0	ug/L	118	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Metribuzin	525.2	0.1	---		6.3590	5.0	ug/L	127	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Mevinphos	525.2	0.1	---		6.1710	5.0	ug/L	123	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Naphthalene	525.2	0.1	---		5.0090	5.0	ug/L	100	68 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Napropamide	525.2	0.1	---		5.2530	5.0	ug/L	105	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	trans-Nonachlor	525.2	0.1	---		5.4700	5.0	ug/L	109	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Oxadiazon	525.2	0.1	---		4.4020	5.0	ug/L	88	58 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Pentachloronitrobenzene	525.2	0.5	---		5.7620	5.0	ug/L	115	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Phorate	525.2	0.1	---		5.9510	5.0	ug/L	119	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Prometon	525.2	1.0	---		6.2020	5.0	ug/L	124	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Pronamide	525.2	0.1	---		5.0520	5.0	ug/L	101	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	Propachlor	525.2	0.1	---		5.8920	5.0	ug/L	118	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Propanil	525.2	0.5	---		5.8420	5.0	ug/L	117	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Propazine	525.2	0.1	---		4.9050	5.0	ug/L	98	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Simazine	525.2	0.07	---		6.4970	5.0	ug/L	130	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Stirofos	525.2	0.1	---		6.7700	5.0	ug/L	135	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Terbacil	525.2	0.1	---		6.9640	5.0	ug/L	139	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Terbufos	525.2	0.5	---		5.6270	5.0	ug/L	113	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Thiobencarb	525.2	0.1	---		5.4610	5.0	ug/L	109	63 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Trifluralin	525.2	0.1	---		5.7900	5.0	ug/L	116	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
CCC	Vernolate	525.2	0.1	---		4.5920	5.0	ug/L	92	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:04	4301071
LFB	IS-Chrysene-d12	525.2	N/A	---		1678000	1575000	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	IS-Phenanthrene-d10	525.2	N/A	---		2578000	2301000	ug/L	112	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	IS-Pyrene-d10	525.2	N/A	---		1920000	1749000	ug/L	110	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	SS-4,4'-Dichlorobiphenyl	525.2	N/A	---		5.0690	5.0	ug/L	101	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	SS-2,4,5,6-Tetrachloro-m-xylene	525.2	N/A	---		4.3520	5.0	ug/L	87	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	SS-Triphenylphosphate	525.2	N/A	---		5.5250	5.0	ug/L	110	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Amtiazine	525.2	1.0	---		3.6820	5.0	ug/L	74	51 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	alpha-Hexachlorocyclohexane	525.2	0.1	---		5.6400	5.0	ug/L	113	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	beta-BHC	525.2	0.1	---		5.0520	5.0	ug/L	101	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	delta-BHC	525.2	0.1	---		5.5570	5.0	ug/L	111	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Butylate	525.2	0.1	---		5.3690	5.0	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Chlorobenzilate	525.2	0.1	---		5.9570	5.0	ug/L	119	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Chloroneb	525.2	0.1	---		6.0470	5.0	ug/L	121	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Cycoate	525.2	0.1	---		4.8050	5.0	ug/L	96	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	DCPA	525.2	0.1	---		4.9540	5.0	ug/L	99	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	4,4'-DDD	525.2	0.1	---		4.9710	5.0	ug/L	99	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	4,4'-DDE	525.2	0.1	---		4.9320	5.0	ug/L	99	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Demeton O	525.2	0.5	---		2.2960	1.7	ug/L	135	3 - 212	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Demeton S	525.2	0.5	---		4.6570	3.3	ug/L	141	1 - 242	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Desethylatrazine	525.2	1.0	---		2.0020	5.0	ug/L	40	18 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Diazinon	525.2	0.1	---		4.9990	5.0	ug/L	100	55 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Dichlofenthion	525.2	0.1	---		4.9950	5.0	ug/L	100	70 - 132	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Dichloran	525.2	0.5	---		5.5840	5.0	ug/L	112	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Dichlorvos	525.2	0.1	---		5.7670	5.0	ug/L	115	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Disulfoton	525.2	0.1	---		5.4280	5.0	ug/L	109	1 - 180	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Endosulfan I	525.2	0.1	---		5.3510	5.0	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Endosulfan II	525.2	0.1	---		5.4100	5.0	ug/L	108	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Endosulfan sulfate	525.2	0.1	---		5.3390	5.0	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Endrin aldehyde	525.2	0.5	---		5.0550	5.0	ug/L	101	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	EPTC	525.2	0.1	---		4.5600	5.0	ug/L	91	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Fenthion	525.2	0.1	---		4.5470	5.0	ug/L	91	67 - 137	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LFB	Fluchloralin	525.2	0.1	---		5.2760	5.0	ug/L	106	70 - 132	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Fluoranthene	525.2	0.1	---		4.9450	5.0	ug/L	99	64 - 139	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Fonofos	525.2	0.1	---		5.1190	5.0	ug/L	102	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	2-Methylnaphthalene	525.2	0.1	---		5.3740	5.0	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	MKG 264 isomer a	525.2	0.1	---		2.9270	3.15	ug/L	93	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	MKG 264 isomer b	525.2	0.1	---		1.8810	1.85	ug/L	102	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	MKG 326	525.2	0.1	---		5.4420	5.0	ug/L	109	70 - 136	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Molinate	525.2	0.1	---		5.3060	5.0	ug/L	106	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Pebulate	525.2	0.1	---		4.7240	5.0	ug/L	94	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Pendimethalin	525.2	0.1	---		5.7600	5.0	ug/L	115	70 - 141	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	cis-Permethrin	525.2	0.1	---		2.7500	2.05	ug/L	134	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	trans-Permethrin	525.2	0.1	---		3.0890	2.95	ug/L	105	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Profluralin	525.2	0.1	---		5.1060	5.0	ug/L	102	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Promethyn	525.2	0.1	---		4.9700	5.0	ug/L	99	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Simetyn	525.2	0.1	---		4.8630	5.0	ug/L	97	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Terbutyn	525.2	0.1	---		5.0050	5.0	ug/L	100	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	Triadimefon	525.2	0.5	---		5.6620	5.0	ug/L	113	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 18:46	4301066
LFB	IS-Chrysenes-d12	525.2	N/A	---		1543000	1461000	ug/L	98	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	IS-Phenanthrene-d10	525.2	N/A	---		2186000	2171000	ug/L	95	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	IS-Pyrene-d10	525.2	N/A	---		1811000	1588000	ug/L	104	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	SS-4,4'-Dichlorobiphenyl	525.2	N/A	---		5.1980	5.0	ug/L	104	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	SS-2,4,5,6-Tetrachloro-m-xylene	525.2	N/A	---		4.0130	5.0	ug/L	80	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	SS-Triphenylphosphate	525.2	N/A	---		5.1960	5.0	ug/L	104	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Acetochlor	525.2	0.1	---		6.1390	5.0	ug/L	123	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Alachlor	525.2	0.1	---		6.3710	5.0	ug/L	127	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Aldrin	525.2	0.1	---		6.4210	5.0	ug/L	128	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Amethyn	525.2	0.1	---		5.5220	5.0	ug/L	110	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Aspon	525.2	0.1	---		5.0040	5.0	ug/L	100	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Atraton	525.2	0.1	---		3.2270	5.0	ug/L	65	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Azinphos-methyl	525.2	0.5	---		7.6950	5.0	ug/L	154	70 - 166	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Benfluralin	525.2	0.1	---		5.7050	5.0	ug/L	114	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Benzo(a)pyrene	525.2	0.02	---		5.4350	5.0	ug/L	109	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	gamma-BHC (Lindane)	525.2	0.02	---		6.2450	5.0	ug/L	125	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Bolstar	525.2	0.1	---		5.1410	5.0	ug/L	103	55 - 140	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Bromacil	525.2	0.1	---		9.7520	5.0	ug/L	195	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Butachlor	525.2	0.1	---		5.8880	5.0	ug/L	118	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Carboxin	525.2	0.1	---		4.8320	5.0	ug/L	97	1 - 174	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	alpha-Chlordane	525.2	0.1	---		6.2700	5.0	ug/L	125	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	gamma-Chlordane	525.2	0.1	---		6.2430	5.0	ug/L	125	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Chloropropylate	525.2	0.1	---		5.5080	5.0	ug/L	110	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Chlorothalonil	525.2	0.1	---		5.9660	5.0	ug/L	119	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LFB	Chlorophoram	525.2	0.1	---		6.1660	5.0	ug/L	123	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Chlorpyrifos	525.2	0.1	---		6.3930	5.0	ug/L	128	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Clomazone	525.2	0.1	---		5.2410	5.0	ug/L	105	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Coumaphos	525.2	0.1	---		6.7140	5.0	ug/L	134	70 - 181	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Cyanazine	525.2	0.1	---		5.0640	5.0	ug/L	101	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	4,4'-DDT	525.2	0.1	---		5.5360	5.0	ug/L	111	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Desisopropylatrazine	525.2	1.0	---		1.4000	5.0	ug/L	28	3 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Dichlobenil	525.2	0.1	---		5.3510	5.0	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Dieldrin	525.2	0.1	---		5.7540	5.0	ug/L	115	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Di(2-ethylhexyl)adipate	525.2	0.6	---		6.0980	5.0	ug/L	122	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Di(2-ethylhexyl)phthalate	525.2	0.6	---		6.1930	5.0	ug/L	124	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Dimethoate	525.2	0.5	---		3.4500	5.0	ug/L	69	17 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	2,4-Dinitrotoluene	525.2	0.5	---		5.1740	5.0	ug/L	103	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	2,6-Dinitrotoluene	525.2	0.5	---		5.5540	5.0	ug/L	111	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Diphenamid	525.2	0.1	---		6.3110	5.0	ug/L	126	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Disulfoton sulfone	525.2	0.1	---		6.4790	5.0	ug/L	130	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Ethalfuralin	525.2	0.1	---		6.7960	5.0	ug/L	136	70 - 137	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Ethion	525.2	5.0	---		5.4000	5.0	ug/L	108	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Ethofumesate	525.2	0.5	---		6.1070	5.0	ug/L	122	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Ethoprop	525.2	0.1	---		4.4530	5.0	ug/L	89	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Etridiazole	525.2	0.1	---		6.0120	5.0	ug/L	120	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Famphur	525.2	0.1	---		4.9020	5.0	ug/L	98	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Fenamiphos	525.2	0.1	---		7.8490	5.0	ug/L	157	42 - 245	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Fluazifop-butyl	525.2	0.1	---		4.9230	5.0	ug/L	98	70 - 146	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Fluometuron	525.2	0.5	---		8.5180	5.0	ug/L	170	34 - 161	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Heptachlor	525.2	0.04	---		6.1680	5.0	ug/L	123	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Heptachlor epoxide	525.2	0.02	---		5.8570	5.0	ug/L	117	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Hexachlorobenzene	525.2	0.1	---		4.6490	5.0	ug/L	93	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Hexachlorocyclopentadiene	525.2	0.1	---		4.0250	5.0	ug/L	80	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Hexazinone	525.2	0.1	---		5.8400	5.0	ug/L	117	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Isophorone	525.2	0.1	---		5.4650	5.0	ug/L	109	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Malathion	525.2	0.1	---		5.9890	5.0	ug/L	120	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	1-Methylnaphthalene	525.2	0.1	---		5.1210	5.0	ug/L	102	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Methyl paraoxon	525.2	0.5	---		7.3390	5.0	ug/L	147	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Metolachlor	525.2	0.1	---		6.2780	5.0	ug/L	126	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Mevinphos	525.2	0.1	---		5.9930	5.0	ug/L	120	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Naphthalene	525.2	0.1	---		4.3140	5.0	ug/L	86	69 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Napropamide	525.2	0.1	---		5.2480	5.0	ug/L	105	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	trans-Nonachlor	525.2	0.1	---		5.9270	5.0	ug/L	119	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Oxadiazon	525.2	0.1	---		4.4890	5.0	ug/L	90	64 - 131	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Pentachloronitrobenzene	525.2	0.5	---		5.7100	5.0	ug/L	114	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LFB	Phorate	525.2	0.1	---		5.9440	5.0	ug/L	119	66 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Prometon	525.2	1.0	---		4.9050	5.0	ug/L	98	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Pronamide	525.2	0.1	---		5.3100	5.0	ug/L	106	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Propachlor	525.2	0.1	---		6.1560	5.0	ug/L	123	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Propanil	525.2	0.5	---		6.0670	5.0	ug/L	121	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Propazine	525.2	0.1	---		4.7510	5.0	ug/L	95	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Stirofos	525.2	0.1	---		6.5930	5.0	ug/L	132	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Terbacil	525.2	0.1	---		7.2730	5.0	ug/L	145	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Terbufos	525.2	0.5	---		5.4930	5.0	ug/L	110	69 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Thiobencarb	525.2	0.1	---		6.0100	5.0	ug/L	120	64 - 140	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Trifluralin	525.2	0.1	---		6.1490	5.0	ug/L	123	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	Vernolate	525.2	0.1	---		4.1340	5.0	ug/L	83	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 19:28	4301067
LFB	IS-Chrysene-d12	525.2	N/A	---		1670000	1461000	ug/L	106	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	IS-Phenanthrene-d10	525.2	N/A	---		2264000	2171000	ug/L	98	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	IS-Pyrene-d10	525.2	N/A	---		1905000	1588000	ug/L	109	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	SS-4,4'-Dichlorobiphenyl	525.2	N/A	---		4.0500	5.0	ug/L	81	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	SS-2,4,5,6-Tetrachloro-m-xylene	525.2	N/A	---		4.8790	5.0	ug/L	98	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	SS-Triphenylphosphate	525.2	N/A	---		5.2350	5.0	ug/L	105	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Acenaphthene	525.2	0.1	---		4.8450	5.0	ug/L	97	57 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Acenaphthylene	525.2	0.1	---		5.0300	5.0	ug/L	101	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Anthracene	525.2	0.1	---		4.8120	5.0	ug/L	96	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Azinhophen-ethyl	525.2	0.5	---		6.3470	5.0	ug/L	127	70 - 153	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Bendiocarb	525.2	0.5	---		6.6120	5.0	ug/L	132	70 - 170	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Benzo(a)anthracene	525.2	0.1	---		4.6750	5.0	ug/L	94	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Benzo(b)fluoranthene	525.2	0.1	---		5.0270	5.0	ug/L	101	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Benzo(k)fluoranthene	525.2	0.1	---		5.1350	5.0	ug/L	103	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Benzo(g,h,i)perylene	525.2	0.1	---		5.1270	5.0	ug/L	103	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Butylbenzylphthalate	525.2	1.0	---		6.5220	5.0	ug/L	130	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Carbophenothion	525.2	0.5	---		5.8050	5.0	ug/L	116	70 - 131	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Chlorfenvinphos	525.2	5.0	---		5.8820	4.55	ug/L	129	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	2-Chlorobiphenyl	525.2	0.1	---		4.9690	5.0	ug/L	99	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Chlorpyrifos methyl	525.2	0.5	---		6.8350	5.0	ug/L	137	70 - 135	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Chrysene	525.2	0.1	---		4.7650	5.0	ug/L	95	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Clopyralid	525.2	10	---		7.3010	10.0	ug/L	73	12 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Crotoxphos	525.2	0.5	---		7.8370	5.0	ug/L	157	70 - 193	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Dibenzo(a,h)anthracene	525.2	0.1	---		5.1500	5.0	ug/L	103	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Di-n-butylphthalate	525.2	2.0	---		6.0770	5.0	ug/L	122	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	2,3-Dichlorobiphenyl	525.2	0.1	---		5.4670	5.0	ug/L	109	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Dicrotophos	525.2	0.5	---		6.8560	5.0	ug/L	137	70 - 198	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Diethylphthalate	525.2	1.0	---		5.6460	5.0	ug/L	113	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Dimethylphthalate	525.2	1.0	---		5.1000	5.0	ug/L	102	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LFB	Di-n-octylphthalate	525.2	2.0	---		6.7910	5.0	ug/L	136	43 - 197	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Dioxathion A	525.2	0.5	---		6.9310	2.8	ug/L	248	29 - 222	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Dioxathion B	525.2	0.5	---		4.0900	2.2	ug/L	186	25 - 514	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Disulfoton sulfoxide	525.2	10	---		11.7620	10.0	ug/L	118	69 - 138	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	EPN	525.2	0.5	---		7.2880	5.0	ug/L	146	70 - 154	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Esfenvalerate	525.2	0.5	---		5.8430	5.0	ug/L	117	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Fenarimol	525.2	1.0	---		7.0360	5.0	ug/L	141	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Fenoxaprop-ethyl	525.2	1.0	---		5.0130	5.0	ug/L	100	66 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Fensulfothion	525.2	0.5	---		7.9950	5.0	ug/L	160	70 - 203	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Fluorene	525.2	0.1	---		5.1000	5.0	ug/L	102	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Fluridone	525.2	1.0	---		5.2610	5.0	ug/L	105	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	2,2',3',3',4,4',6-Heptachlorobiphenyl	525.2	0.5	---		4.0150	5.0	ug/L	80	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	2,2',4,4',5,6'-Hexachlorobiphenyl	525.2	0.1	---		5.1440	5.0	ug/L	103	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Indeno(1,2,3-cd)pyrene	525.2	0.1	---		5.1380	5.0	ug/L	103	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Iprodione	525.2	0.5	---		7.6760	5.0	ug/L	154	70 - 216	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Isofenphos	525.2	0.5	---		6.9700	5.0	ug/L	139	70 - 154	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Leptophos	525.2	0.5	---		5.9330	5.0	ug/L	119	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Metolaxyl	525.2	0.5	---		6.9080	5.0	ug/L	138	70 - 154	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Methyl parathion	525.2	0.5	---		9.2540	5.0	ug/L	185	70 - 213	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Metsulfuron-methyl	525.2	10	---		13.7550	10.0	ug/L	138	70 - 177	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Mirex	525.2	0.5	---		5.9590	5.0	ug/L	119	65 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Monocrotophos	525.2	0.5	---		2.9730	5.0	ug/L	59	28 - 159	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Naled	525.2	0.5	---		6.2750	5.0	ug/L	126	49 - 147	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	cis-Nonachlor	525.2	0.1	---		5.1350	5.0	ug/L	103	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Nonflurazon	525.2	1.0	---		6.3370	5.0	ug/L	127	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	2,2',3',3',4,5',6'-Octachlorobiphenyl	525.2	0.5	---		5.8400	5.0	ug/L	117	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Oryzalin	525.2	10	---		14.2930	5.0	ug/L	286	1 - 399	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Oxychloridane	525.2	0.1	---		5.9650	5.0	ug/L	119	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Oxyfluorene	525.2	0.5	---		5.9580	5.0	ug/L	119	70 - 142	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Parathion	525.2	0.5	---		7.0720	5.0	ug/L	141	68 - 164	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Pentachlorobenzene	525.2	0.5	---		5.5030	5.0	ug/L	110	59 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	2,2',3',4,6-Pentachlorobiphenyl	525.2	0.1	---		5.6600	5.0	ug/L	113	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Pentachlorophenol	525.2	1.0	---		8.2870	5.0	ug/L	166	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Phenanthrene	525.2	0.1	---		5.1640	5.0	ug/L	103	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Phosmet	525.2	0.5	---		5.2480	5.0	ug/L	105	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	E-Phosphamidon	525.2	0.5	---		2.4350	1.45	ug/L	168	70 - 301	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Z-Phosphamidon	525.2	0.5	---		5.6900	3.55	ug/L	160	70 - 193	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Propiconazole isomer a	525.2	5.0	---		2.2130	1.6	ug/L	138	1 - 276	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Propiconazole isomer b	525.2	5.0	---		3.9800	3.4	ug/L	117	64 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Prothiofos	525.2	0.5	---		6.9890	5.0	ug/L	140	70 - 144	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Pyrene	525.2	0.1	---		5.6960	5.0	ug/L	114	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LFB	Sulfotep	525.2	0.5	---		5.0580	5.0	ug/L	101	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Tebuthiuron	525.2	10	---		11.9060	10.0	ug/L	119	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	TEPP	525.2	1.0	---		7.9290	5.0	ug/L	159	1 - 222	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	1,2,4,5-Tetrachlorobenzene	525.2	0.5	---		5.0460	5.0	ug/L	101	56 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	2,2',4,4'-Tetrachlorobiphenyl	525.2	0.1	---		5.7140	5.0	ug/L	114	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Thiabenzazole	525.2	10	---		1.0320	5.0	ug/L	21	3 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Thionazin	525.2	0.5	---		7.0380	5.0	ug/L	141	60 - 147	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Tribufos	525.2	0.1	---		6.9400	5.0	ug/L	139	70 - 145	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	2,4,5-Trichlorobiphenyl	525.2	0.1	---		5.4140	5.0	ug/L	108	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Tincyclazole	525.2	1.0	---		5.9480	5.0	ug/L	119	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LFB	Vindozolin	525.2	0.5	---		7.0090	5.0	ug/L	140	70 - 145	---	---	1.0	05/21/2019 08:45	05/22/2019 20:10	4301068
LMB	IS-Chrysene-d12	525.2	N/A	---		1567000	1575000	ug/L	99	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	IS-Phenanthrene-d10	525.2	N/A	---		2292000	2301000	ug/L	100	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	IS-Pyrene-d10	525.2	N/A	---		1768000	1749000	ug/L	101	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	SS-4,4'-Dichlorobiphenyl	525.2	N/A	---		5.1530	5.0	ug/L	103	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	SS-2,4,5,6-Tetrachloro-m-xylene	525.2	N/A	---		4.9790	5.0	ug/L	100	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	SS-Triphenylphosphate	525.2	N/A	---		5.3610	5.0	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Acenaphthene	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Acenaphthylene	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Acetochlor	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Alachlor	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Aldrin	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Ametyrn	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Anilazine	525.2	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Anthracene	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Aspon	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Atraton	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Atrazine	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Azinphos-ethyl	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Azinphos-methyl	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Bendiocarb	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Benfluralin	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Benzo(a)anthracene	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Benzo(b)fluoranthene	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Benzo(k)fluoranthene	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Benzo(g,h,i)perylene	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Benzo(a)pyrene	525.2	0.02	---	<	0.02		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	alpha-Hexachlorocyclohexane	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	beta-BHC	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	delta-BHC	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	gamma-BHC (Lindane)	525.2	0.02	---	<	0.02		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LMB	Bolstar	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Bromacil	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Butachlor	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Butylate	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Butylbenzylphthalate	525.2	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Carbophenothion	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Carboxin	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	alpha-Chlordane	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	gamma-Chlordane	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Chlorfenvinphos	525.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Chlorobenzilate	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	2-Chlorobiphenyl	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Chloroneb	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Chloropropylate	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Chlorothalonil	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Chlorpropham	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Chlorpyrifos	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Chlorpyrifos methyl	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Chrysene	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Clomazone	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Clpyralid	525.2	10	---	<	10		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Coumaphos	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Crotoxyphos	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Cyanazine	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Cycoate	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	DCPA	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	4,4'-DDD	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	4,4'-DDE	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	4,4'-DDT	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Demeton O	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Demeton S	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Desethylatrazine	525.2	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Desisopropylatrazine	525.2	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Diazinon	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Dibenzo(a,h)anthracene	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Di-n-butylphthalate	525.2	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Dichlobenil	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Dichlofenthion	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Dichloran	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	2,3-Dichlorobiphenyl	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Dichlorvos	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LMB	Dicofophos	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Dieldrin	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Di(2-ethylhexyl)adipate	525.2	0.6	---	<	0.6		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Di(2-ethylhexyl)phthalate	525.2	0.6	---	<	0.6		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Diethylphthalate	525.2	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Dimethoate	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Dimethylphthalate	525.2	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	2,4-Dinitrotoluene	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	2,6-Dinitrotoluene	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Di-n-octylphthalate	525.2	2.0	---	<	2.0		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Dioxathion A	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Dioxathion B	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Diphenamid	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Disulfoton	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Disulfoton sulfone	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Disulfoton sulfoxide	525.2	10	---	<	10		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Endosulfan I	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Endosulfan II	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Endosulfan sulfate	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Endrin	525.2	0.01	---	<	0.01		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Endrin aldehyde	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	EPN	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	EPTC	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Erucylamide	525.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Esfenvalerate	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Ethalfuralin	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Ethion	525.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Ethofumesate	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Ethoprop	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Etridiazole	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Famphur	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Fenamiphos	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Fenarimol	525.2	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Fenitrothion	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Fenoxaprop-ethyl	525.2	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Fensulfothion	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Fenthion	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Fluazifop-butyl	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Fluchloralin	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Fluometuron	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Fluoranthene	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LMB	Fluorene	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Fluridone	525.2	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Fonofos	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Heptachlor	525.2	0.04	---	<	0.04		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Heptachlor epoxide	525.2	0.02	---	<	0.02		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	2,2',3',4',4',6'-Heptachlorobiphenyl	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Hexachlorobenzene	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	2,2',4',4',5',6'-Hexachlorobiphenyl	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Hexachlorocyclopentadiene	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Hexazinone	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Indeno(1,2,3-cd)pyrene	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Iprodione	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Isofenphos	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Isophorone	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Leptophos	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Malathion	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Metaxyl	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Methoxychlor	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	1-Methylnaphthalene	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	2-Methylnaphthalene	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Methyl paraoxon	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Methyl parathion	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Metolachlor	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Metribuzin	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Metsulfuron-methyl	525.2	10	---	<	10		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Mevinphos	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	MGK 264 isomer a	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	MGK 264 isomer b	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	MGK 326	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Mirex	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Molinate	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Monocrotophos	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Naled	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Naphthalene	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Napropamide	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	cis-Nonachlor	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	trans-Nonachlor	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Norflurazon	525.2	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	2,2',3',3',4',4',5',6',6'-Octachlorobiphenyl	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Oryzalin	525.2	10	---	<	10		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Oxadiazon	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LMB	Oxychloridane	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Oxyfluorfen	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Parathion	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Pebulate	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Pendimethalin	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Pentachlorobenzene	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	2,2',3',4',6-Pentachlorobiphenyl	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Pentachloronitrobenzene	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Pentachlorophenol	525.2	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	cis-Permethrin	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	trans-Permethrin	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Phenanthrene	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Phorate	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Phosmet	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	E-Phosphamidon	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Z-Phosphamidon	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Profluralin	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Prometon	525.2	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Prometryn	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Pronamide	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Propachlor	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Propanil	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Propazine	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Propiconazole isomer a	525.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Propiconazole isomer b	525.2	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Prothiofos	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Pyrene	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Simazine	525.2	0.07	---	<	0.07		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Simetryn	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Stirofos	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Sulfotep	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Tebuthiuron	525.2	10	---	<	10		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	TEPP	525.2	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Terbacil	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Terbufos	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Terbutryn	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	1,2,4,5-Tetrachlorobenzene	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	2,2',4,4'-Tetrachlorobiphenyl	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Thiabenzazole	525.2	10	---	<	10		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Thiobencarb	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Thionazin	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LMB	Triadimefon	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Tribufos	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	2,4,5-Trichlorobiphenyl	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Trichloronate	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Tricyclazole	525.2	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Trifluralin	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Vernolate	525.2	0.1	---	<	0.1		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
LMB	Vindozolin	525.2	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	05/21/2019 08:45	05/22/2019 22:16	4301065
FS	IS-Chrysene-d12	525.2	N/A	Finished		1607000	1575000	ug/L	102	70 - 130	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	IS-Phenanthrene-d10	525.2	N/A	Finished		2522000	2301000	ug/L	110	70 - 130	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	IS-Pyrene-d10	525.2	N/A	Finished		1853000	1749000	ug/L	106	70 - 130	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	SS-4,4'-Dichlorobiphenyl	525.2	N/A	Finished		5.0180	5.0	ug/L	102	70 - 130	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	SS-2,4,5,6-Tetrachloro-m-xylene	525.2	N/A	Finished		5.0380	5.0	ug/L	103	70 - 130	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	SS-Triphenylphosphate	525.2	N/A	Finished		5.8640	5.0	ug/L	120	70 - 130	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Acenaphthene	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Acenaphthylene	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Acetochlor	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Alachlor	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Aldrin	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Ametyrn	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Anilazine	525.2	1.0	Finished	<	1.0		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Anthracene	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Aspon	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Atlaton	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Atrazine	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Azinphos-ethyl	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Azinphos-methyl	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Bendiocarb	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Benfluralin	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Benzo(a)anthracene	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Benzo(b)fluoranthene	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Benzo(k)fluoranthene	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Benzo(g,h,i)perylene	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Benzo(a)pyrene	525.2	0.02	Finished	<	0.02		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	alpha-Hexachlorocyclohexane	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	beta-BHC	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	delta-BHC	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	gamma-BHC (Lindane)	525.2	0.02	Finished	<	0.02		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Bolstar	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Bromacil	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Butachlor	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FS	Butylate	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Butylbenzylphthalate	525.2	1.0	Finished	<	1.0		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Carbophenothion	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Carboxin	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	alpha-Chlordane	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	gamma-Chlordane	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Chlorfenvinphos	525.2	5.0	Finished	<	5.0		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Chlorobenzilate	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	2-Chlorobiphenyl	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Chloroneb	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Chloropropylate	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Chlorothalonil	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Chloroprotham	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Chlorpyrifos	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Chlorpyrifos methyl	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Chrysene	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Ciomezone	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Clopyralid	525.2	10	Finished	<	10		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Coumaphos	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Crotoxypfos	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Cyanazine	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Cydoate	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	DCPA	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	4,4'-DDD	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	4,4'-DDE	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	4,4'-DDT	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Demeton O	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Demeton S	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Desethylatrazine	525.2	1.0	Finished	<	1.0		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Desisopropylatrazine	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Diazinon	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Dibenzo(a,h)anthracene	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Di-n-butylphthalate	525.2	2.0	Finished	<	2.0		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Dichlobenil	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Dichlofenthion	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Dichloran	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	2,3-Dichlorobiphenyl	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Dichlorvos	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Dicrotophos	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Dieldrin	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Di(2-ethylhexyl)adipate	525.2	0.6	Finished	<	0.6		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FS	Di(2-ethylhexyl)phthalate	525.2	0.6	Finished	<	0.6		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Diethylphthalate	525.2	1.0	Finished	<	1.0		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Dimethoate	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Dimethylphthalate	525.2	1.0	Finished	<	1.0		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	2,4-Dinitrotoluene	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	2,6-Dinitrotoluene	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Di-n-octylphthalate	525.2	2.0	Finished	<	2.0		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Dioxathion A	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Dioxathion B	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Diphenamid	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Disulfoton	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Disulfoton sulfone	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Disulfoton sulfoxide	525.2	10	Finished	<	10		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Endosulfan I	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Endosulfan II	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Endosulfan sulfate	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Endrin	525.2	0.01	Finished	<	0.01		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Endrin aldehyde	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	EPN	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	EPTC	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Erucylamide	525.2	5.0	Finished	<	5.0		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Estenvalerate	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Ethalfuralin	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Ethion	525.2	5.0	Finished	<	5.0		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Ethofumesate	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Ethoprop	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Etridiazole	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Famphur	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Fenamiphos	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Fenarimol	525.2	1.0	Finished	<	1.0		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Fenitrothion	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Fenoxprop-ethyl	525.2	1.0	Finished	<	1.0		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Fensulfothion	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Fenthion	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Fluazifop-butyl	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Fluchloralin	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Fluometuron	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Fluoranthene	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Fluorene	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Fluridone	525.2	1.0	Finished	<	1.0		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Fonofos	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FS	Heptachlor	525.2	0.04	Finished	<	0.04		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Heptachlor epoxide	525.2	0.02	Finished	<	0.02		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	2,2',3,3',4,4',6-Heptachlorobiphenyl	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Hexachlorobenzene	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	2,2',4,4',5,6'-Hexachlorobiphenyl	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Hexachlorocyclopentadiene	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Hexazinone	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Indeno(1,2,3-cd)pyrene	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Iprodione	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Isofenphos	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Isophorone	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Leptophos	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Malathion	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Metaxyl	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Methoxychlor	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	1-Methylnaphthalene	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	2-Methylnaphthalene	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Methyl paraoxon	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Methyl parathion	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Metolachlor	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Metribuzin	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Metsulfuron-methyl	525.2	10	Finished	<	10		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Mevinphos	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	MGK 264 isomer a	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	MGK 264 isomer b	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	MGK 326	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Mirex	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Molinate	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Monocrotophos	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Naled	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Naphthalene	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Napropamide	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	cis-Nonachlor	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	trans-Nonachlor	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Norflurazon	525.2	1.0	Finished	<	1.0		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	2,2',3,3',4,5',6'-Octachlorobiphenyl	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Oryzalin	525.2	10	Finished	<	10		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Oxadiazon	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Oxychlorane	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Oxyfluorene	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Parathion	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FS	Pebulate	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Pendimethalin	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Pentachlorobenzene	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	2,2',3',4,6-Pentachlorobiphenyl	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Pentachloronitrobenzene	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Pentachlorophenol	525.2	1.0	Finished	<	1.0		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	cis-Permethrin	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	trans-Permethrin	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Permethrin, cis & trans	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Phenanthrene	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Phorate	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Phosmet	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	E-Phosphamidon	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Z-Phosphamidon	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Profluralin	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Prometon	525.2	1.0	Finished	<	1.0		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Prometryn	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Pronamide	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Propachlor	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Propanil	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Propazine	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Propiconazole isomer a	525.2	5.0	Finished	<	5.0		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Propiconazole isomer b	525.2	5.0	Finished	<	5.0		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Prothiofos	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Pyrene	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Simazine	525.2	0.07	Finished	<	0.07		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Simetryn	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Stirofos	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Sulfotep	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Tebuthiuron	525.2	10	Finished	<	10		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	TEPP	525.2	1.0	Finished	<	1.0		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Terbacil	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Terbufos	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Terbutryn	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	1,2,4,5-Tetrachlorobenzene	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	2,2',4,4'-Tetrachlorobiphenyl	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Thiabendazole	525.2	10	Finished	<	10		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Thiobencarb	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Thionazin	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Triadimefon	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Tribufos	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FS	2,4,5-Trichlorobiphenyl	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Trichloronate	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Tricyclazole	525.2	1.0	Finished	<	1.0		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Trifluralin	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Vernolate	525.2	0.1	Finished	<	0.1		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
FS	Vindozolin	525.2	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.98	05/21/2019 08:45	05/22/2019 22:58	4291759
CCC	IS-Chrysene-d12	525.2	N/A	---		1719000	1719000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	IS-Phenanthrene-d10	525.2	N/A	---		2498000	2498000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	IS-Pyrene-d10	525.2	N/A	---		1793000	1793000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	SS-4,4'-Dichlorobiphenyl	525.2	N/A	---		4.3800	5.0	ug/L	88	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	SS-2,4,5,6-Tetrachloro-m-xylene	525.2	N/A	---		5.3640	5.0	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	SS-Triphenylphosphate	525.2	N/A	---		4.9490	5.0	ug/L	99	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Acenaphthene	525.2	0.1	---		5.1750	5.0	ug/L	104	65 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Acenaphthylene	525.2	0.1	---		5.0430	5.0	ug/L	101	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Anthracene	525.2	0.1	---		5.1030	5.0	ug/L	102	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Aziphos-ethyl	525.2	0.5	---		5.6320	5.0	ug/L	113	70 - 145	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Bendiocarb	525.2	0.5	---		6.3630	5.0	ug/L	127	65 - 159	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Benzo(a)anthracene	525.2	0.1	---		4.6040	5.0	ug/L	92	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Benzo(b)fluoranthene	525.2	0.1	---		4.8460	5.0	ug/L	97	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Benzo(k)fluoranthene	525.2	0.1	---		4.9410	5.0	ug/L	99	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Benzo(g,h,i)perylene	525.2	0.1	---		4.7980	5.0	ug/L	96	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Butylbenzylphthalate	525.2	1.0	---		6.1080	5.0	ug/L	122	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Carbophenothion	525.2	0.5	---		5.4310	5.0	ug/L	109	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Chlorofeniphos	525.2	5.0	---		5.1820	4.55	ug/L	114	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	2-Chlorobiphenyl	525.2	0.1	---		5.1190	5.0	ug/L	102	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Chlorpyrifos methyl	525.2	0.5	---		6.6000	5.0	ug/L	132	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Chrysene	525.2	0.1	---		4.6440	5.0	ug/L	93	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Clopyralid	525.2	10	---		3.1960	10.0	ug/L	32	1 - 109	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Crotoxypfos	525.2	0.5	---		6.5730	5.0	ug/L	131	70 - 173	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Dibenzo(a,h)anthracene	525.2	0.1	---		5.1020	5.0	ug/L	102	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Di-n-butylphthalate	525.2	2.0	---		5.5350	5.0	ug/L	111	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	2,3-Dichlorobiphenyl	525.2	0.1	---		5.4020	5.0	ug/L	108	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Dicrotophos	525.2	0.5	---		6.4050	5.0	ug/L	128	70 - 191	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Diethylphthalate	525.2	1.0	---		5.4490	5.0	ug/L	109	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Dimethylphthalate	525.2	1.0	---		5.0890	5.0	ug/L	102	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Di-n-octylphthalate	525.2	2.0	---		6.4810	5.0	ug/L	130	48 - 173	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Dioxathion A	525.2	0.5	---		6.2230	2.8	ug/L	222	34 - 209	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Dioxathion B	525.2	0.5	---		3.6860	2.2	ug/L	168	14 - 498	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Disulfoton sulfoxide	525.2	10	---		9.7700	10.0	ug/L	98	69 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	EPN	525.2	0.5	---		6.7630	5.0	ug/L	135	70 - 147	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Erucylamide	525.2	5.0	---		18.6790	10.0	ug/L	187	43 - 251	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	Esfenvalerate	525.2	0.5	---		5.7100	5.0	ug/L	114	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Fenarimol	525.2	1.0	---		6.4170	5.0	ug/L	128	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Fenitrothion	525.2	0.5	---		6.7180	5.0	ug/L	134	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Fenoxaprop-ethyl	525.2	1.0	---		6.3280	5.0	ug/L	127	70 - 140	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Fensulfothion	525.2	0.5	---		6.4940	5.0	ug/L	130	70 - 191	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Fluorene	525.2	0.1	---		5.2590	5.0	ug/L	105	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Fluridone	525.2	1.0	---		4.6600	5.0	ug/L	93	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	2,2',3',3',4,4',6-Heptachlorobiphenyl	525.2	0.5	---		3.8240	5.0	ug/L	76	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	2,2',4,4',5,6'-Hexachlorobiphenyl	525.2	0.1	---		4.8650	5.0	ug/L	97	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Indeno(1,2,3-cd)pyrene	525.2	0.1	---		4.9900	5.0	ug/L	100	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Iprodione	525.2	0.5	---		6.3820	5.0	ug/L	128	70 - 207	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Isofenphos	525.2	0.5	---		6.1790	5.0	ug/L	124	70 - 145	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Leptophos	525.2	0.5	---		5.2840	5.0	ug/L	106	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Metaxalyl	525.2	0.5	---		6.3940	5.0	ug/L	128	70 - 143	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Methyl parathion	525.2	0.5	---		8.2850	5.0	ug/L	166	70 - 206	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Metsulfuron-methyl	525.2	10	---		11.3350	10.0	ug/L	113	70 - 170	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Mirex	525.2	0.5	---		5.6320	5.0	ug/L	113	61 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Monocrotophos	525.2	0.5	---		5.0540	5.0	ug/L	101	70 - 185	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Naled	525.2	0.5	---		6.0480	5.0	ug/L	121	29 - 160	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	cis-Nonachlor	525.2	0.1	---		4.4220	5.0	ug/L	88	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Nonflurazon	525.2	1.0	---		5.5870	5.0	ug/L	112	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	2,2',3',3',4,5',6',6'-Octachlorobiphenyl	525.2	0.5	---		5.2870	5.0	ug/L	106	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Oryzalin	525.2	10	---		11.1950	5.0	ug/L	224	1 - 389	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Oxychloridane	525.2	0.1	---		5.8470	5.0	ug/L	117	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Oxyfluorfen	525.2	0.5	---		5.4100	5.0	ug/L	108	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Parathion	525.2	0.5	---		6.1890	5.0	ug/L	124	63 - 145	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Pentachlorobenzene	525.2	0.5	---		5.7280	5.0	ug/L	115	61 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	2,2',3',4,6-Pentachlorobiphenyl	525.2	0.1	---		4.7150	5.0	ug/L	94	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Pentachlorophenol	525.2	1.0	---		6.9760	5.0	ug/L	140	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Phenanthrene	525.2	0.1	---		5.2430	5.0	ug/L	105	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Phosmet	525.2	0.5	---		4.5190	5.0	ug/L	90	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	E-Phosphamidon	525.2	0.5	---		2.2620	1.45	ug/L	156	70 - 302	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Z-Phosphamidon	525.2	0.5	---		5.1970	3.55	ug/L	146	70 - 189	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Propiconazole isomer a	525.2	5.0	---		2.0290	1.6	ug/L	127	1 - 278	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Propiconazole isomer b	525.2	5.0	---		3.6060	3.4	ug/L	106	61 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Prothiofos	525.2	0.5	---		6.0080	5.0	ug/L	120	70 - 135	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Pyrene	525.2	0.1	---		5.3320	5.0	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Sulfotep	525.2	0.5	---		4.7830	5.0	ug/L	96	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Tebuthiuron	525.2	10	---		10.5260	10.0	ug/L	105	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	TEPP	525.2	1.0	---		8.9040	5.0	ug/L	178	1 - 267	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	1,2,4,5-Tetrachlorobenzene	525.2	0.5	---		5.5830	5.0	ug/L	112	57 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	2,2',4,4'-Tetrachlorobiphenyl	525.2	0.1	---		5.6600	5.0	ug/L	113	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Thiabendazole	525.2	10	---		6.9430	5.0	ug/L	139	6 - 190	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Thionazin	525.2	0.5	---		6.2220	5.0	ug/L	124	62 - 137	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Tribufos	525.2	0.1	---		6.5630	5.0	ug/L	131	70 - 136	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	2,4,5-Trichlorobiphenyl	525.2	0.1	---		5.4100	5.0	ug/L	108	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Trichloronate	525.2	0.5	---		6.2680	5.0	ug/L	125	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Tricyclozole	525.2	1.0	---		5.3050	5.0	ug/L	106	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	Vindozolin	525.2	0.5	---		6.2080	5.0	ug/L	124	70 - 137	---	---	1.0	05/21/2019 08:45	05/23/2019 06:42	4302646
CCC	IS-Chrysene-d12	525.2	N/A	---		1467000	1467000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	IS-Phenanthrene-d10	525.2	N/A	---		2237000	2237000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	IS-Pyrene-d10	525.2	N/A	---		1658000	1658000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	SS-4,4'-Dichlorobiphenyl	525.2	N/A	---		5.2140	5.0	ug/L	104	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	SS-2,4,5,6-Tetrachloro-m-xylene	525.2	N/A	---		4.9480	5.0	ug/L	99	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	SS-Triphenylphosphate	525.2	N/A	---		5.3890	5.0	ug/L	108	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Anilazine	525.2	1.0	---		3.5130	5.0	ug/L	70	41 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	alpha-Hexachlorocyclohexane	525.2	0.1	---		5.9630	5.0	ug/L	119	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	beta-BHC	525.2	0.1	---		5.2850	5.0	ug/L	106	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	delta-BHC	525.2	0.1	---		5.7920	5.0	ug/L	116	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Butylate	525.2	0.1	---		5.6300	5.0	ug/L	113	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Chlorobenzilate	525.2	0.1	---		5.8720	5.0	ug/L	117	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Chloroneb	525.2	0.1	---		6.2370	5.0	ug/L	125	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Cydoate	525.2	0.1	---		5.5660	5.0	ug/L	111	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	DCPA	525.2	0.1	---		5.1740	5.0	ug/L	103	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	4,4'-DDD	525.2	0.1	---		5.1970	5.0	ug/L	104	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	4,4'-DDE	525.2	0.1	---		4.8880	5.0	ug/L	98	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Demeton O	525.2	0.5	---		2.5590	1.7	ug/L	151	70 - 174	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Demeton S	525.2	0.5	---		4.8460	3.3	ug/L	147	67 - 215	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Desethylatrazine	525.2	1.0	---		5.5410	5.0	ug/L	111	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Diazinon	525.2	0.1	---		5.3930	5.0	ug/L	108	66 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Dichlofenthion	525.2	0.1	---		5.3640	5.0	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Dichloran	525.2	0.5	---		5.9570	5.0	ug/L	119	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Dichlorvos	525.2	0.1	---		6.2620	5.0	ug/L	125	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Disulfoton	525.2	0.1	---		6.1300	5.0	ug/L	123	61 - 156	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Endosulfan I	525.2	0.1	---		5.7380	5.0	ug/L	115	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Endosulfan II	525.2	0.1	---		5.7270	5.0	ug/L	115	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Endosulfan sulfate	525.2	0.1	---		5.5440	5.0	ug/L	111	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Endrin aldehyde	525.2	0.5	---		4.9420	5.0	ug/L	99	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	EPTC	525.2	0.1	---		4.7920	5.0	ug/L	96	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Fenthion	525.2	0.1	---		4.9510	5.0	ug/L	99	63 - 137	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Fluchloralin	525.2	0.1	---		5.4180	5.0	ug/L	108	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Fluoranthene	525.2	0.1	---		5.2150	5.0	ug/L	104	65 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	Fonofos	525.2	0.1	---		5.3840	5.0	ug/L	108	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	2-Methylnaphthalene	525.2	0.1	---		5.9910	5.0	ug/L	120	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	MGK 264 isomer a	525.2	0.1	---		3.0500	3.15	ug/L	97	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	MGK 264 isomer b	525.2	0.1	---		2.0820	1.85	ug/L	113	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	MGK 326	525.2	0.1	---		5.7130	5.0	ug/L	114	70 - 131	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Molinate	525.2	0.1	---		5.3480	5.0	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Pebulate	525.2	0.1	---		5.0940	5.0	ug/L	102	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Pendimethalin	525.2	0.1	---		5.6590	5.0	ug/L	113	70 - 132	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	cis-Permethrin	525.2	0.1	---		2.7410	2.05	ug/L	134	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	trans-Permethrin	525.2	0.1	---		3.1890	2.95	ug/L	108	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Profluralin	525.2	0.1	---		5.3080	5.0	ug/L	106	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Promethyn	525.2	0.1	---		5.3270	5.0	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Simetyn	525.2	0.1	---		5.8940	5.0	ug/L	118	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Terbutyn	525.2	0.1	---		5.4590	5.0	ug/L	109	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	Triadimefon	525.2	0.5	---		5.5180	5.0	ug/L	110	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 07:24	4302647
CCC	IS-Chrysene-d12	525.2	N/A	---		1461000	1461000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	IS-Phenanthrene-d10	525.2	N/A	---		2171000	2171000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	IS-Pyrene-d10	525.2	N/A	---		1588000	1588000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	SS-4,4'-Dichlorobiphenyl	525.2	N/A	---		5.3660	5.0	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	SS-2,4,5,6-Tetrachloro-m-xylene	525.2	N/A	---		4.8000	5.0	ug/L	96	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	SS-Triphenylphosphate	525.2	N/A	---		5.5010	5.0	ug/L	110	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Acetochlor	525.2	0.1	---		6.0260	5.0	ug/L	121	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Alachlor	525.2	0.1	---		5.9800	5.0	ug/L	120	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Aldrin	525.2	0.1	---		6.5110	5.0	ug/L	130	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Amethyn	525.2	0.1	---		5.9720	5.0	ug/L	119	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Aspon	525.2	0.1	---		4.7670	5.0	ug/L	95	69 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Atraton	525.2	0.1	---		5.6570	5.0	ug/L	113	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Atrazine	525.2	0.1	---		6.1660	5.0	ug/L	123	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Azinphos-methyl	525.2	0.5	---		6.9510	5.0	ug/L	139	70 - 151	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Benfluralin	525.2	0.1	---		5.5860	5.0	ug/L	112	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Benzo(a)pyrene	525.2	0.02	---		5.6000	5.0	ug/L	112	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	gamma-BHC (Lindane)	525.2	0.02	---		5.7990	5.0	ug/L	116	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Bolstar	525.2	0.1	---		5.2100	5.0	ug/L	104	70 - 132	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Bromacil	525.2	0.1	---		8.8560	5.0	ug/L	177	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Butachlor	525.2	0.1	---		5.8250	5.0	ug/L	116	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Carboxin	525.2	0.1	---		6.0190	5.0	ug/L	120	60 - 182	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	alpha-Chlordane	525.2	0.1	---		5.9390	5.0	ug/L	119	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	gamma-Chlordane	525.2	0.1	---		5.9370	5.0	ug/L	119	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Chloropropylate	525.2	0.1	---		5.4810	5.0	ug/L	110	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Chlorothalonil	525.2	0.1	---		5.4360	5.0	ug/L	109	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Chlorpropham	525.2	0.1	---		6.0160	5.0	ug/L	120	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	Chlorpyrifos	525.2	0.1	---		5.9080	5.0	ug/L	118	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Clomazone	525.2	0.1	---		5.0500	5.0	ug/L	101	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Coumaphos	525.2	0.1	---		5.9210	5.0	ug/L	118	67 - 165	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Cyanazine	525.2	0.1	---		4.8960	5.0	ug/L	98	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	4,4'-DDT	525.2	0.1	---		5.9240	5.0	ug/L	118	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Desisopropylatrazine	525.2	1.0	---		5.3810	5.0	ug/L	108	62 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Dichlobenil	525.2	0.1	---		5.5920	5.0	ug/L	112	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Dieldrin	525.2	0.1	---		5.4970	5.0	ug/L	110	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Di(2-ethylhexyl)adipate	525.2	0.6	---		6.2190	5.0	ug/L	124	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Di(2-ethylhexyl)phthalate	525.2	0.6	---		5.9570	5.0	ug/L	119	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Dimethoate	525.2	0.5	---		6.2300	5.0	ug/L	125	62 - 131	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	2,4-Dinitrotoluene	525.2	0.5	---		5.2750	5.0	ug/L	106	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	2,6-Dinitrotoluene	525.2	0.5	---		5.5750	5.0	ug/L	112	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Diphenamid	525.2	0.1	---		6.0190	5.0	ug/L	120	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Disulfoton sulfone	525.2	0.1	---		6.0120	5.0	ug/L	120	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Endrin	525.2	0.01	---		6.1790	5.0	ug/L	124	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Ethalfuralin	525.2	0.1	---		5.9970	5.0	ug/L	120	70 - 136	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Ethion	525.2	5.0	---		5.4090	5.0	ug/L	108	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Ethofumesate	525.2	0.5	---		6.0390	5.0	ug/L	121	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Ethoprop	525.2	0.1	---		4.2970	5.0	ug/L	86	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Etridiazole	525.2	0.1	---		6.4150	5.0	ug/L	128	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Famphur	525.2	0.1	---		5.0180	5.0	ug/L	100	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Fenamiphos	525.2	0.1	---		7.6880	5.0	ug/L	154	61 - 234	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Fluzifop-butyl	525.2	0.1	---		5.0700	5.0	ug/L	101	66 - 140	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Fluometuron	525.2	0.5	---		9.7020	5.0	ug/L	194	42 - 159	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Heptachlor	525.2	0.04	---		6.1340	5.0	ug/L	123	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Heptachlor epoxide	525.2	0.02	---		5.7870	5.0	ug/L	116	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Hexachlorobenzene	525.2	0.1	---		3.9860	5.0	ug/L	80	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Hexachlorocyclopentadiene	525.2	0.1	---		4.0860	5.0	ug/L	82	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Hexazinone	525.2	0.1	---		5.7860	5.0	ug/L	116	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Isophorone	525.2	0.1	---		5.5790	5.0	ug/L	112	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Malathion	525.2	0.1	---		5.4740	5.0	ug/L	109	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Methoxychlor	525.2	0.1	---		6.1460	5.0	ug/L	123	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	1-Methylnaphthalene	525.2	0.1	---		5.5420	5.0	ug/L	111	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Methyl paraoxon	525.2	0.5	---		6.7380	5.0	ug/L	135	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Metolachlor	525.2	0.1	---		5.8640	5.0	ug/L	117	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Metribuzin	525.2	0.1	---		6.1660	5.0	ug/L	123	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Mevinphos	525.2	0.1	---		5.8560	5.0	ug/L	117	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Naphthalene	525.2	0.1	---		5.1140	5.0	ug/L	102	68 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Napropamide	525.2	0.1	---		5.4260	5.0	ug/L	109	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	trans-Nonachlor	525.2	0.1	---		5.4490	5.0	ug/L	109	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	Oxadiazon	525.2	0.1	---		4.3780	5.0	ug/L	88	58 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Pentachloronitrobenzene	525.2	0.5	---		5.5170	5.0	ug/L	110	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Phorate	525.2	0.1	---		5.9050	5.0	ug/L	118	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Prometon	525.2	1.0	---		6.2440	5.0	ug/L	125	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Pronamide	525.2	0.1	---		5.0220	5.0	ug/L	100	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Propachlor	525.2	0.1	---		5.7550	5.0	ug/L	115	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Propanil	525.2	0.5	---		5.5300	5.0	ug/L	111	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Propazine	525.2	0.1	---		4.5980	5.0	ug/L	92	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Simazine	525.2	0.07	---		6.1590	5.0	ug/L	123	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Stirofos	525.2	0.1	---		6.9500	5.0	ug/L	139	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Terbacil	525.2	0.1	---		6.9410	5.0	ug/L	139	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Terbufos	525.2	0.5	---		5.4420	5.0	ug/L	109	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Thiobencarb	525.2	0.1	---		5.4880	5.0	ug/L	110	63 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Trifluralin	525.2	0.1	---		5.7890	5.0	ug/L	116	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
CCC	Vernolate	525.2	0.1	---		4.4940	5.0	ug/L	90	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 08:06	4302648
LFB	Atrazine	525.2	0.1	---		6.2870	5.0	ug/L	126	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 09:30	4301067
LFB	Endrin	525.2	0.01	---		6.0910	5.0	ug/L	122	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 09:30	4301067
LFB	Methoxychlor	525.2	0.1	---		6.2320	5.0	ug/L	125	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 09:30	4301067
LFB	Metribuzin	525.2	0.1	---		6.1470	5.0	ug/L	123	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 09:30	4301067
LFB	Simazine	525.2	0.07	---		6.4450	5.0	ug/L	129	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 09:30	4301067
LFB	Erucylamide	525.2	5.0	---		23.2990	10.0	ug/L	233	70 - 234	---	---	1.0	05/21/2019 08:45	05/23/2019 10:12	4301068
LFB	Fenitrothion	525.2	0.5	---		6.9720	5.0	ug/L	139	70 - 143	---	---	1.0	05/21/2019 08:45	05/23/2019 10:12	4301068
LFB	Trichloronate	525.2	0.5	---		6.3540	5.0	ug/L	127	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 10:12	4301068



Eurofins Eaton Analytical

Run Log

Run ID: **259806** Method: **525.2**

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
CCC	4305008		OS	DO	05/23/2019 20:02	525 2-DO-101018a-up2.mth
CCC	4305009		OS	DO	05/23/2019 20:44	525 2-DO-101018a-up2.mth
CCC	4305010		OS	DO	05/23/2019 21:26	525 2-DO-101018a-up2.mth

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	IS-Chrysene-d12	525.2	N/A	--		1604000	1604000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	IS-Phenanthrene-d10	525.2	N/A	--		2330000	2330000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	IS-Pyrene-d10	525.2	N/A	--		1714000	1714000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	SS-4,4'-Dichlorobiphenyl	525.2	N/A	--		4.2670	5.0	ug/L	85	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	SS-2,4,5,6-Tetrachloro-m-xylene	525.2	N/A	--		5.3300	5.0	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	SS-Triphenylphosphate	525.2	N/A	--		5.3370	5.0	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Acenaphthene	525.2	0.1	--		5.0490	5.0	ug/L	101	65 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Acenaphthylene	525.2	0.1	--		4.9480	5.0	ug/L	99	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Anthracene	525.2	0.1	--		5.0550	5.0	ug/L	101	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Azinhpos-ethyl	525.2	0.5	--		5.8340	5.0	ug/L	117	70 - 145	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Bendiocarb	525.2	0.5	--		6.3530	5.0	ug/L	127	65 - 159	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Benzo(a)anthracene	525.2	0.1	--		4.7350	5.0	ug/L	95	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Benzo(b)fluoranthene	525.2	0.1	--		4.8990	5.0	ug/L	98	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Benzo(k)fluoranthene	525.2	0.1	--		5.0250	5.0	ug/L	100	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Benzo(g,h,i)perylene	525.2	0.1	--		4.6140	5.0	ug/L	92	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Butylbenzylphthalate	525.2	1.0	--		6.4860	5.0	ug/L	130	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Carbophenothion	525.2	0.5	--		5.5010	5.0	ug/L	110	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Chlorfenvinphos	525.2	5.0	--		5.5540	4.55	ug/L	122	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	2-Chlorobiphenyl	525.2	0.1	--		4.9210	5.0	ug/L	98	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Chlorpyrifos methyl	525.2	0.5	--		6.7940	5.0	ug/L	136	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Chrysene	525.2	0.1	--		4.7170	5.0	ug/L	94	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Clopyralid	525.2	10	--		3.8810	10.0	ug/L	39	1 - 109	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Crotoxphos	525.2	0.5	--		6.8250	5.0	ug/L	136	70 - 173	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Dibenzo(a,h)anthracene	525.2	0.1	--		4.7920	5.0	ug/L	96	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Di-n-butylphthalate	525.2	2.0	--		5.7860	5.0	ug/L	116	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	2,3-Dichlorobiphenyl	525.2	0.1	--		5.1820	5.0	ug/L	104	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Dicrotophos	525.2	0.5	--		7.2840	5.0	ug/L	146	70 - 191	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Diethylphthalate	525.2	1.0	--		5.2410	5.0	ug/L	105	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Dimethylphthalate	525.2	1.0	--		4.9990	5.0	ug/L	100	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Di-n-octylphthalate	525.2	2.0	--		6.6680	5.0	ug/L	133	48 - 173	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Dioxathion A	525.2	0.5	--		6.2740	2.8	ug/L	224	34 - 209	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Dioxathion B	525.2	0.5	--		3.9200	2.2	ug/L	178	14 - 498	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Disulfoton sulfoxide	525.2	10	--		10.4810	10.0	ug/L	105	69 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	EPN	525.2	0.5	--		6.8770	5.0	ug/L	138	70 - 147	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Erucylamide	525.2	5.0	--		19.6240	10.0	ug/L	196	43 - 251	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Esfenvalerate	525.2	0.5	--		5.7180	5.0	ug/L	114	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Fenarimol	525.2	1.0	--		6.8540	5.0	ug/L	137	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Fenitrothion	525.2	0.5	--		6.7560	5.0	ug/L	135	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Fenoxprop-ethyl	525.2	1.0	--		6.5960	5.0	ug/L	132	70 - 140	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Fensulfothion	525.2	0.5	--		7.3100	5.0	ug/L	146	70 - 191	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	Fluorene	525.2	0.1	---		5.1300	5.0	ug/L	103	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Fluridone	525.2	1.0	---		4.4130	5.0	ug/L	88	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	2,2',3',3',4',4',6-Heptachlorobiphenyl	525.2	0.5	---		4.0970	5.0	ug/L	82	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	2,2',4',4',5',6'-Hexachlorobiphenyl	525.2	0.1	---		4.8740	5.0	ug/L	97	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Indeno(1,2,3-cd)pyrene	525.2	0.1	---		4.7560	5.0	ug/L	95	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Iprodione	525.2	0.5	---		6.8380	5.0	ug/L	137	70 - 207	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Isofenphos	525.2	0.5	---		6.3700	5.0	ug/L	127	70 - 145	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Leptophos	525.2	0.5	---		5.7500	5.0	ug/L	115	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Metaalaxyl	525.2	0.5	---		6.4400	5.0	ug/L	129	70 - 143	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Methyl parathion	525.2	0.5	---		8.5310	5.0	ug/L	171	70 - 206	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Metsulfuron-methyl	525.2	10	---		11.8220	10.0	ug/L	118	70 - 170	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Mirex	525.2	0.5	---		5.9820	5.0	ug/L	120	61 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Monocrotophos	525.2	0.5	---		5.5290	5.0	ug/L	111	70 - 185	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Naled	525.2	0.5	---		5.9080	5.0	ug/L	118	29 - 160	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	cis-Nonachlor	525.2	0.1	---		4.7110	5.0	ug/L	94	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Norflurazon	525.2	1.0	---		5.9790	5.0	ug/L	120	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	2,2',3',3',4',5',6',6'-Octachlorobiphenyl	525.2	0.5	---		5.6870	5.0	ug/L	114	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Oryzalin	525.2	10	---		11.7200	5.0	ug/L	234	1 - 389	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Oxychloridane	525.2	0.1	---		5.8140	5.0	ug/L	116	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Oxyfluorfen	525.2	0.5	---		5.5510	5.0	ug/L	111	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Parathion	525.2	0.5	---		6.1760	5.0	ug/L	124	63 - 145	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Pentachlorobenzene	525.2	0.5	---		5.5020	5.0	ug/L	110	61 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	2,2',3',4',6-Pentachlorobiphenyl	525.2	0.1	---		4.9460	5.0	ug/L	99	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Pentachlorophenol	525.2	1.0	---		7.4770	5.0	ug/L	150	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Phenanthrene	525.2	0.1	---		5.1000	5.0	ug/L	102	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Phosmet	525.2	0.5	---		4.6410	5.0	ug/L	93	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	E-Phosphamidon	525.2	0.5	---		2.5720	1.45	ug/L	177	70 - 302	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Z-Phosphamidon	525.2	0.5	---		5.3310	3.55	ug/L	150	70 - 189	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Propiconazole isomer a	525.2	5.0	---		2.1110	1.6	ug/L	132	1 - 278	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Propiconazole isomer b	525.2	5.0	---		3.9230	3.4	ug/L	115	61 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Prothiofos	525.2	0.5	---		6.3390	5.0	ug/L	127	70 - 135	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Pyrene	525.2	0.1	---		5.4830	5.0	ug/L	110	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Sulfotep	525.2	0.5	---		4.6170	5.0	ug/L	92	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Tebuthiuron	525.2	10	---		11.3000	10.0	ug/L	113	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	TEPP	525.2	1.0	---		9.1840	5.0	ug/L	184	1 - 267	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	1,2,4,5-Tetrachlorobenzene	525.2	0.5	---		5.3640	5.0	ug/L	107	57 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	2,2',4,4'-Tetrachlorobiphenyl	525.2	0.1	---		5.5210	5.0	ug/L	110	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Thiabendazole	525.2	10	---		6.8370	5.0	ug/L	137	6 - 190	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Thionazin	525.2	0.5	---		6.2370	5.0	ug/L	125	62 - 137	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Tribufos	525.2	0.1	---		6.6940	5.0	ug/L	134	70 - 136	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	2,4,5-Trichlorobiphenyl	525.2	0.1	---		5.4080	5.0	ug/L	108	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	Trichloronate	525.2	0.5	---		6.3860	5.0	ug/L	128	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Tricyclazole	525.2	1.0	---		5.3940	5.0	ug/L	108	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	Vinclozolin	525.2	0.5	---		6.3200	5.0	ug/L	126	70 - 137	---	---	1.0	05/21/2019 08:45	05/23/2019 20:02	4305008
CCC	IS-Chrysene-d12	525.2	N/A	---		1492000	1492000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	IS-Phenanthrene-d10	525.2	N/A	---		2553000	2553000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	IS-Pyrene-d10	525.2	N/A	---		1776000	1776000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	SS-4,4'-Dichlorobiphenyl	525.2	N/A	---		4.8030	5.0	ug/L	96	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	SS-2,4,5,6-Tetrachloro-m-xylene	525.2	N/A	---		4.5980	5.0	ug/L	92	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	SS-Triphenylphosphate	525.2	N/A	---		5.6860	5.0	ug/L	114	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Anilazine	525.2	1.0	---		3.4020	5.0	ug/L	68	41 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	alpha-Hexachlorocyclohexane	525.2	0.1	---		5.6280	5.0	ug/L	113	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	beta-BHC	525.2	0.1	---		4.9540	5.0	ug/L	99	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	delta-BHC	525.2	0.1	---		5.5870	5.0	ug/L	112	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Butylate	525.2	0.1	---		5.2020	5.0	ug/L	104	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Chlorobenzilate	525.2	0.1	---		6.2750	5.0	ug/L	126	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Chloroneb	525.2	0.1	---		5.8300	5.0	ug/L	117	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Cycloate	525.2	0.1	---		5.5390	5.0	ug/L	111	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	DCPA	525.2	0.1	---		4.9530	5.0	ug/L	99	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	4,4'-DDD	525.2	0.1	---		5.6580	5.0	ug/L	113	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	4,4'-DDE	525.2	0.1	---		5.3550	5.0	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Demeton O	525.2	0.5	---		2.4410	1.7	ug/L	144	70 - 174	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Demeton S	525.2	0.5	---		4.7340	3.3	ug/L	143	67 - 215	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Desethylatrazine	525.2	1.0	---		5.1470	5.0	ug/L	103	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Diazinon	525.2	0.1	---		4.9990	5.0	ug/L	100	66 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Dichlofenthion	525.2	0.1	---		5.2560	5.0	ug/L	105	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Dichloran	525.2	0.5	---		5.6540	5.0	ug/L	113	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Dichlorvos	525.2	0.1	---		5.7590	5.0	ug/L	115	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Disulfoton	525.2	0.1	---		5.6950	5.0	ug/L	114	61 - 156	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Endosulfan I	525.2	0.1	---		5.4080	5.0	ug/L	108	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Endosulfan II	525.2	0.1	---		6.2200	5.0	ug/L	124	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Endosulfan sulfate	525.2	0.1	---		5.7890	5.0	ug/L	116	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Endrin aldehyde	525.2	0.5	---		5.3700	5.0	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	EPTC	525.2	0.1	---		4.9430	5.0	ug/L	99	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Fenthion	525.2	0.1	---		4.7810	5.0	ug/L	96	63 - 137	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Fluchloralin	525.2	0.1	---		5.0270	5.0	ug/L	101	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Fluoranthene	525.2	0.1	---		4.9930	5.0	ug/L	100	65 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Fonofos	525.2	0.1	---		5.0960	5.0	ug/L	102	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	2-Methylnaphthalene	525.2	0.1	---		5.8460	5.0	ug/L	117	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	MGK 264 isomer a	525.2	0.1	---		2.8950	3.15	ug/L	92	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	MGK 264 isomer b	525.2	0.1	---		1.9360	1.85	ug/L	105	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	MGK 326	525.2	0.1	---		5.3220	5.0	ug/L	106	70 - 131	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	Molinate	525.2	0.1	---		5.5900	5.0	ug/L	112	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Pebulate	525.2	0.1	---		4.9130	5.0	ug/L	98	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Pendimethalin	525.2	0.1	---		5.3090	5.0	ug/L	106	70 - 132	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	cis-Permethrin	525.2	0.1	---		2.8700	2.05	ug/L	140	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	trans-Permethrin	525.2	0.1	---		3.3140	2.95	ug/L	112	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Profluralin	525.2	0.1	---		4.9840	5.0	ug/L	100	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Prometryn	525.2	0.1	---		4.7490	5.0	ug/L	95	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Simetryn	525.2	0.1	---		5.0770	5.0	ug/L	110	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Terbutyn	525.2	0.1	---		5.0460	5.0	ug/L	101	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	Triadimefon	525.2	0.5	---		5.1030	5.0	ug/L	102	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 20:44	4305009
CCC	IS-Chrysene-d12	525.2	N/A	---		1573000	1573000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	IS-Phenanthrene-d10	525.2	N/A	---		2416000	2416000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	IS-Pyrene-d10	525.2	N/A	---		1688000	1688000	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	SS-4,4'-Dichlorobiphenyl	525.2	N/A	---		5.3970	5.0	ug/L	108	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	SS-2,4,5,6-Tetrachloro-m-xylene	525.2	N/A	---		4.9190	5.0	ug/L	98	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	SS-Triphenylphosphate	525.2	N/A	---		5.3830	5.0	ug/L	108	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Acetochlor	525.2	0.1	---		5.7400	5.0	ug/L	115	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Alachlor	525.2	0.1	---		5.4380	5.0	ug/L	109	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Aldrin	525.2	0.1	---		6.1780	5.0	ug/L	124	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Ametryn	525.2	0.1	---		5.4110	5.0	ug/L	108	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Aspon	525.2	0.1	---		4.5220	5.0	ug/L	90	69 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Atraton	525.2	0.1	---		5.7470	5.0	ug/L	115	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Atrazine	525.2	0.1	---		5.9960	5.0	ug/L	120	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Azinphos-methyl	525.2	0.5	---		6.1960	5.0	ug/L	124	70 - 151	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Benfluralin	525.2	0.1	---		5.3500	5.0	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Benzo(a)pyrene	525.2	0.02	---		5.4080	5.0	ug/L	108	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	gamma-BHC (Lindane)	525.2	0.02	---		5.4060	5.0	ug/L	108	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Bolstar	525.2	0.1	---		5.0450	5.0	ug/L	101	70 - 132	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Bromacil	525.2	0.1	---		8.0550	5.0	ug/L	161	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Butachlor	525.2	0.1	---		5.6660	5.0	ug/L	113	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Carboxin	525.2	0.1	---		5.5620	5.0	ug/L	111	60 - 182	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	alpha-Chlordane	525.2	0.1	---		5.5670	5.0	ug/L	111	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	gamma-Chlordane	525.2	0.1	---		5.5070	5.0	ug/L	110	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Chloropropylate	525.2	0.1	---		5.4620	5.0	ug/L	109	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Chlorothalonil	525.2	0.1	---		5.2690	5.0	ug/L	105	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Chlorpropham	525.2	0.1	---		5.8160	5.0	ug/L	116	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Chlorpyrifos	525.2	0.1	---		6.0690	5.0	ug/L	121	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Clomazone	525.2	0.1	---		5.0270	5.0	ug/L	101	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Coumaphos	525.2	0.1	---		5.7410	5.0	ug/L	115	67 - 165	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Cyanazine	525.2	0.1	---		4.4640	5.0	ug/L	89	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	4,4'-DDT	525.2	0.1	---		5.6480	5.0	ug/L	113	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	Desisopropylatrazine	525.2	1.0	---		5.5570	5.0	ug/L	111	62 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Dichlobenil	525.2	0.1	---		5.5770	5.0	ug/L	112	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Dieldrin	525.2	0.1	---		5.5230	5.0	ug/L	110	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Di(2-ethylhexyl)adipate	525.2	0.6	---		5.9250	5.0	ug/L	118	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Di(2-ethylhexyl)phthalate	525.2	0.6	---		5.7340	5.0	ug/L	115	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Dimethoate	525.2	0.5	---		5.5320	5.0	ug/L	111	62 - 131	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	2,4-Dinitrotoluene	525.2	0.5	---		5.2110	5.0	ug/L	104	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	2,6-Dinitrotoluene	525.2	0.5	---		5.8690	5.0	ug/L	117	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Diphenamid	525.2	0.1	---		5.5860	5.0	ug/L	112	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Disulfoton sulfone	525.2	0.1	---		6.7920	5.0	ug/L	136	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Endrin	525.2	0.01	---		5.9060	5.0	ug/L	118	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Ethalfuralin	525.2	0.1	---		5.6200	5.0	ug/L	112	70 - 136	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Ethion	525.2	5.0	---		5.3380	5.0	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Ethoflumesate	525.2	0.5	---		5.8600	5.0	ug/L	117	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Ethoprop	525.2	0.1	---		4.1240	5.0	ug/L	82	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Etridiazole	525.2	0.1	---		6.1160	5.0	ug/L	122	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Famphur	525.2	0.1	---		4.5170	5.0	ug/L	90	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Fenamiphos	525.2	0.1	---		7.2520	5.0	ug/L	145	61 - 234	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Fluazifop-butyl	525.2	0.1	---		4.9320	5.0	ug/L	99	66 - 140	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Fluometuron	525.2	0.5	---		9.0860	5.0	ug/L	182	42 - 159	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Heptachlor	525.2	0.04	---		5.8120	5.0	ug/L	116	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Heptachlor epoxide	525.2	0.02	---		5.0930	5.0	ug/L	102	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Hexachlorobenzene	525.2	0.1	---		4.1740	5.0	ug/L	83	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Hexachlorocyclopentadiene	525.2	0.1	---		4.4810	5.0	ug/L	90	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Hexazinone	525.2	0.1	---		5.6010	5.0	ug/L	112	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Isophorone	525.2	0.1	---		5.6570	5.0	ug/L	113	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Maliathion	525.2	0.1	---		5.3160	5.0	ug/L	106	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Methoxychlor	525.2	0.1	---		6.0540	5.0	ug/L	121	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	1-Methylnaphthalene	525.2	0.1	---		5.4580	5.0	ug/L	109	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Methyl paraoxon	525.2	0.5	---		6.4770	5.0	ug/L	130	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Metolachlor	525.2	0.1	---		5.6420	5.0	ug/L	113	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Metribuzin	525.2	0.1	---		5.7890	5.0	ug/L	116	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Mevinphos	525.2	0.1	---		5.7570	5.0	ug/L	115	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Naphthalene	525.2	0.1	---		5.1190	5.0	ug/L	102	68 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Napropamide	525.2	0.1	---		5.2830	5.0	ug/L	106	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	trans-Nonachlor	525.2	0.1	---		5.5620	5.0	ug/L	111	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Oxadiazon	525.2	0.1	---		4.4090	5.0	ug/L	88	58 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Pentachloronitrobenzene	525.2	0.5	---		5.7350	5.0	ug/L	115	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Phorate	525.2	0.1	---		5.8370	5.0	ug/L	117	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Prometon	525.2	1.0	---		6.0430	5.0	ug/L	121	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Pronamide	525.2	0.1	---		5.0850	5.0	ug/L	102	70 - 130	---	---	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	Propachlor	525.2	0.1	---		5.7350	5.0	ug/L	115	70 - 130	----	----	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Propanil	525.2	0.5	---		5.4570	5.0	ug/L	109	70 - 130	----	----	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Propazine	525.2	0.1	---		4.4860	5.0	ug/L	90	70 - 130	----	----	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Simazine	525.2	0.07	---		5.8730	5.0	ug/L	117	70 - 130	----	----	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Stirofos	525.2	0.1	---		6.7050	5.0	ug/L	134	70 - 130	----	----	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Terbacil	525.2	0.1	---		6.0670	5.0	ug/L	121	70 - 130	----	----	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Terbufos	525.2	0.5	---		5.4510	5.0	ug/L	109	70 - 130	----	----	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Thiobencarb	525.2	0.1	---		5.3010	5.0	ug/L	106	63 - 130	----	----	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Trifluralin	525.2	0.1	---		5.8700	5.0	ug/L	117	70 - 130	----	----	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010
CCC	Vernolate	525.2	0.1	---		4.0710	5.0	ug/L	81	70 - 130	----	----	1.0	05/21/2019 08:45	05/23/2019 21:26	4305010



Eurofins Eaton Analytical

Run Log

Run ID: 259716 Method: 527

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
CCL	4301406		OS	DN	05/23/2019 11:56	527-DN-012518.mth
LMB	4301401		RW	DN	05/23/2019 13:05	527-DN-012518.mth
FBH	4301402		RW	DN	05/23/2019 14:14	527-DN-012518.mth
FS	4292047	Finished	DW	DN	05/23/2019 15:23	527-DN-012518.mth
CCH	4301408		OS	DN	05/23/2019 16:33	527-DN-012518.mth

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCL	SS-1,3-Dimethyl-2-nitrobenzene	527	N/A	--		5.3170	5.0	ug/L	106	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	SS-Triphenylphosphate	527	N/A	--		4.5890	5.0	ug/L	92	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	Atrazine	527	0.5	--		0.5770	0.5	ug/L	115	50 - 150	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	Bifenthrin	527	0.5	--		0.5370	0.5	ug/L	107	50 - 150	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	Bromacil	527	0.5	--		0.5610	0.5	ug/L	112	50 - 150	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	Chlorpyrifos	527	0.5	--		0.5590	0.5	ug/L	112	50 - 150	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	Dimethoate	527	0.7	--		0.8700	0.7	ug/L	124	50 - 150	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	Esbiol	527	0.5	--		0.4850	0.5	ug/L	97	50 - 150	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	Esfenvalerate	527	0.5	--		0.7180	0.592	ug/L	121	50 - 150	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	Fenvalerate	527	0.5	--		0.4900	0.408	ug/L	120	50 - 150	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	2,2',4',4',5',5'-Hexabromobiphenyl (HBB)	527	0.7	--		0.8140	0.7	ug/L	116	50 - 150	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	2,2',4',4',5',5'-Hexabromodiphenyl ether (BDE-153)	527	0.8	--		0.9750	0.8	ug/L	122	50 - 150	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	Hexazinone	527	0.5	--		0.6000	0.5	ug/L	120	50 - 150	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	Kepone	527	0.5	--		0.6570	0.5	ug/L	131	50 - 150	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	Malathion	527	0.5	--		0.5090	0.5	ug/L	102	50 - 150	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	Mirex	527	0.5	--		0.5560	0.5	ug/L	111	50 - 150	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	Nitrofen	527	0.5	--		0.7230	0.5	ug/L	145	50 - 150	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	Norflurazon	527	0.5	--		0.5140	0.5	ug/L	103	50 - 150	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	Oxychlorodane	527	0.5	--		0.4420	0.5	ug/L	88	50 - 150	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	Parathion	527	0.5	--		0.6280	0.5	ug/L	126	50 - 150	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	2,2',4',4',5-Pentabromodiphenyl ether (BDE-99)	527	0.9	--		0.9040	0.9	ug/L	100	50 - 150	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	2,2',4',4',6-Pentabromodiphenyl ether (BDE-100)	527	0.5	--		0.5160	0.5	ug/L	103	50 - 150	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	Promethyn	527	0.5	--		0.5100	0.5	ug/L	102	50 - 150	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	Propazine	527	0.5	--		0.5520	0.5	ug/L	110	50 - 150	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	Terbufos-sulfone	527	0.4	--		0.5320	0.4	ug/L	133	50 - 150	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	2,2',4',4'-Tetrabromodiphenyl ether (BDE-47)	527	0.3	--		0.3140	0.3	ug/L	105	50 - 150	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	Thiobencarb	527	0.5	--		0.4790	0.5	ug/L	96	50 - 150	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
CCL	Vinclozolin	527	0.5	--		0.4890	0.5	ug/L	98	50 - 150	---	---	1.0	05/21/2019 08:55	05/23/2019 11:56	4301406
LMB	SS-1,3-Dimethyl-2-nitrobenzene	527	N/A	--		4.5870	5.0	ug/L	95	70 - 130	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
LMB	SS-Triphenylphosphate	527	N/A	--		4.3490	5.0	ug/L	90	70 - 130	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
LMB	Atrazine	527	0.5	--	<	0.5		ug/L	---	---	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
LMB	Bifenthrin	527	0.5	--	<	0.5		ug/L	---	---	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
LMB	Bromacil	527	0.5	--	<	0.5		ug/L	---	---	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
LMB	Chlorpyrifos	527	0.5	--	<	0.5		ug/L	---	---	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
LMB	Dimethoate	527	0.7	--	<	0.7		ug/L	---	---	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
LMB	Esbiol	527	0.5	--	<	0.5		ug/L	---	---	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
LMB	Esfenvalerate	527	0.5	--	<	0.5		ug/L	---	---	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
LMB	Fenvalerate	527	0.5	--	<	0.5		ug/L	---	---	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
LMB	2,2',4',4',5',5'-Hexabromobiphenyl (HBB)	527	0.7	--	<	0.7		ug/L	---	---	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
LMB	2,2',4',4',5',5'-Hexabromodiphenyl ether (BDE-153)	527	0.8	--	<	0.8		ug/L	---	---	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LMB	Hexazinone	527	0.5	---	<	0.5		ug/L	---	---	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
LMB	Keponone	527	0.5	---	<	0.5		ug/L	---	---	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
LMB	Malathion	527	0.5	---	<	0.5		ug/L	---	---	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
LMB	Mirex	527	0.5	---	<	0.5		ug/L	---	---	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
LMB	Nitrofen	527	0.5	---	<	0.5		ug/L	---	---	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
LMB	Norflurazon	527	0.5	---	<	0.5		ug/L	---	---	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
LMB	Oxychlorodane	527	0.5	---	<	0.5		ug/L	---	---	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
LMB	Parathion	527	0.5	---	<	0.5		ug/L	---	---	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
LMB	2,2',4',5'-Pentabromodiphenyl ether (BDE-99)	527	0.9	---	<	0.9		ug/L	---	---	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
LMB	2,2',4',4',6'-Pentabromodiphenyl ether (BDE-100)	527	0.5	---	<	0.5		ug/L	---	---	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
LMB	Prometyrn	527	0.5	---	<	0.5		ug/L	---	---	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
LMB	Propazine	527	0.5	---	<	0.5		ug/L	---	---	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
LMB	Terbufos-sulfone	527	0.4	---	<	0.4		ug/L	---	---	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
LMB	2,2',4',4'-Tetrabromodiphenyl ether (BDE-47)	527	0.3	---	<	0.3		ug/L	---	---	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
LMB	Thiobencarb	527	0.5	---	<	0.5		ug/L	---	---	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
LMB	Vinclozolin	527	0.5	---	<	0.5		ug/L	---	---	---	---	0.97	05/21/2019 08:55	05/23/2019 13:05	4301401
FBH	SS-1,3-Dimethyl-2-nitrobenzene	527	N/A	---		4.3810	5.0	ug/L	88	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FBH	SS-Triphenylphosphate	527	N/A	---		4.6820	5.0	ug/L	94	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FBH	Atrazine	527	0.5	---		4.4130	5.0	ug/L	88	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FBH	Bifenthrin	527	0.5	---		5.1160	5.0	ug/L	102	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FBH	Bromacil	527	0.5	---		5.1290	5.0	ug/L	103	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FBH	Chlorpyrifos	527	0.5	---		5.3440	5.0	ug/L	107	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FBH	Dimethoate	527	0.7	---		6.0210	7.0	ug/L	86	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FBH	Esbioi	527	0.5	---		4.1880	5.0	ug/L	84	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FBH	Esfenvalerate	527	0.5	---		6.2760	5.92	ug/L	106	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FBH	Fenvalerate	527	0.5	---		4.6290	4.08	ug/L	113	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FBH	2,2',4',4',5',5'-Hexabromobiphenyl (HBB)	527	0.7	---		8.0460	7.0	ug/L	115	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FBH	2,2',4',4',5',5'-Hexabromodiphenyl ether (BDE-153)	527	0.8	---		8.3220	8.0	ug/L	104	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FBH	Hexazinone	527	0.5	---		5.2060	5.0	ug/L	104	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FBH	Keponone	527	0.5	---		5.2410	5.0	ug/L	105	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FBH	Malathion	527	0.5	---		3.8640	5.0	ug/L	77	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FBH	Mirex	527	0.5	---		5.1490	5.0	ug/L	103	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FBH	Nitrofen	527	0.5	---		5.6650	5.0	ug/L	113	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FBH	Norflurazon	527	0.5	---		4.8670	5.0	ug/L	97	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FBH	Oxychlorodane	527	0.5	---		4.7510	5.0	ug/L	95	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FBH	Parathion	527	0.5	---		5.8020	5.0	ug/L	116	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FBH	2,2',4',4',5'-Pentabromodiphenyl ether (BDE-99)	527	0.9	---		7.9980	9.0	ug/L	89	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FBH	2,2',4',4',6'-Pentabromodiphenyl ether (BDE-100)	527	0.5	---		4.8710	5.0	ug/L	97	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FBH	Prometyrn	527	0.5	---		5.4360	5.0	ug/L	109	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FBH	Propazine	527	0.5	---		4.9700	5.0	ug/L	99	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FBH	Terbufos-sulfone	527	0.4	---		4.1650	4.0	ug/L	104	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FBH	2,2',4,4'-Tetrabromodiphenyl ether (BDE-47)	527	0.3	---		2.6750	3.0	ug/L	89	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FBH	Thiobencarb	527	0.5	---		4.4560	5.0	ug/L	89	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FBH	Vinclozolin	527	0.5	---		4.6310	5.0	ug/L	93	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 14:14	4301402
FS	SS-1,3-Dimethyl-2-nitrobenzene	527	N/A	Finished		4.4100	5.0	ug/L	94	70 - 130	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	SS-Triphenylphosphate	527	N/A	Finished		4.2230	5.0	ug/L	90	70 - 130	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	Arazine	527	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	Bifenthrin	527	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	Bromacil	527	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	Chlorpyrifos	527	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	Dimethoate	527	0.7	Finished	<	0.7		ug/L	---	---	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	Esbiol	527	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	Esfenvalerate	527	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	Fenvalerate	527	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	2,2',4,4',5,5'-Hexabromobiphenyl (HBB)	527	0.7	Finished	<	0.7		ug/L	---	---	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	2,2',4,4',5,5'-Hexabromodiphenyl ether (BDE-153)	527	0.8	Finished	<	0.8		ug/L	---	---	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	Hexazinone	527	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	Kepone	527	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	Malathion	527	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	Mirex	527	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	Nitrofen	527	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	Norflurazon	527	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	Oxychloridane	527	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	Parathion	527	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	2,2',4,4',5-Pentabromodiphenyl ether (BDE-99)	527	0.9	Finished	<	0.9		ug/L	---	---	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	2,2',4,4',6-Pentabromodiphenyl ether (BDE-100)	527	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	Promethyn	527	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	Propazine	527	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	Terbufos-sulfone	527	0.4	Finished	<	0.4		ug/L	---	---	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	2,2',4,4'-Tetrabromodiphenyl ether (BDE-47)	527	0.3	Finished	<	0.3		ug/L	---	---	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	Thiobencarb	527	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
FS	Vinclozolin	527	0.5	Finished	<	0.5		ug/L	---	---	---	---	0.94	05/21/2019 08:55	05/23/2019 15:23	4292047
CCH	SS-1,3-Dimethyl-2-nitrobenzene	527	N/A	---		4.7640	5.0	ug/L	95	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408
CCH	SS-Triphenylphosphate	527	N/A	---		4.9340	5.0	ug/L	99	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408
CCH	Arazine	527	0.5	---		4.2090	5.0	ug/L	84	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408
CCH	Bifenthrin	527	0.5	---		5.1040	5.0	ug/L	102	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408
CCH	Bromacil	527	0.5	---		5.2490	5.0	ug/L	105	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408
CCH	Chlorpyrifos	527	0.5	---		5.7340	5.0	ug/L	115	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408
CCH	Dimethoate	527	0.7	---		5.7190	7.0	ug/L	82	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408
CCH	Esbiol	527	0.5	---		4.6420	5.0	ug/L	93	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408
CCH	Esfenvalerate	527	0.5	---		6.6180	5.92	ug/L	112	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408
CCH	Fenvalerate	527	0.5	---		4.5830	4.08	ug/L	112	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCH	2,2',4,4',5,5'-Hexabromobiphenyl (HBB)	527	0.7	---		6.2050	7.0	ug/L	89	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408
CCH	2,2',4,4',5,5'-Hexabromodiphenyl ether (BDE-153)	527	0.8	---		8.5140	8.0	ug/L	106	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408
CCH	Hexazinone	527	0.5	---		5.4260	5.0	ug/L	109	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408
CCH	Kepone	527	0.5	---		5.6660	5.0	ug/L	113	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408
CCH	Malathion	527	0.5	---		4.8240	5.0	ug/L	96	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408
CCH	Mirex	527	0.5	---		5.4720	5.0	ug/L	109	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408
CCH	Nitrofen	527	0.5	---		6.2440	5.0	ug/L	125	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408
CCH	Norflurazon	527	0.5	---		4.9090	5.0	ug/L	98	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408
CCH	Oxychlorodane	527	0.5	---		5.8600	5.0	ug/L	117	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408
CCH	Parathion	527	0.5	---		6.3070	5.0	ug/L	126	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408
CCH	2,2',4,4',5-Pentabromodiphenyl ether (BDE-99)	527	0.9	---		9.0610	9.0	ug/L	101	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408
CCH	2,2',4,4',6-Pentabromodiphenyl ether (BDE-100)	527	0.5	---		5.3210	5.0	ug/L	106	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408
CCH	Promethyn	527	0.5	---		5.4760	5.0	ug/L	110	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408
CCH	Propazine	527	0.5	---		4.8950	5.0	ug/L	98	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408
CCH	Terbufos-sulfone	527	0.4	---		4.6430	4.0	ug/L	116	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408
CCH	2,2',4,4'-Tetrabromodiphenyl ether (BDE-47)	527	0.3	---		2.9590	3.0	ug/L	99	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408
CCH	Thiobencarb	527	0.5	---		4.5020	5.0	ug/L	90	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408
CCH	Vinclozolin	527	0.5	---		5.1910	5.0	ug/L	104	70 - 130	---	---	1.0	05/21/2019 08:55	05/23/2019 16:33	4301408



Eurofins Eaton Analytical

Run Log

Run ID: 259583 Method: 531.2 LL

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
CCL	4296514		RW	DR	05/21/2019 01:03	LL_531_2_051319DR
LMB	4296509		RW	DR	05/21/2019 01:34	LL_531_2_051319DR
FS	4291765	Finished	DW	DR	05/21/2019 02:05	LL_531_2_051319DR
MS	4296520	Finished	DW	DR	05/21/2019 02:36	LL_531_2_051319DR
MSD	4296521	Finished	DW	DR	05/21/2019 03:07	LL_531_2_051319DR
CCM	4296515		RW	DR	05/21/2019 06:13	LL_531_2_051319DR

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCL	SS-BDMC	531.2 LL	N/A	--		2.0130	2.0	ug/L	101	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 01:03	4296514
CCL	Aldicarb	531.2 LL	0.1	--		0.0720	0.1	ug/L	72	50 - 150	---	---	1.0	05/16/2019 08:20	05/21/2019 01:03	4296514
CCL	Aldicarb sulfone	531.2 LL	0.1	--		0.0960	0.1	ug/L	96	50 - 150	---	---	1.0	05/16/2019 08:20	05/21/2019 01:03	4296514
CCL	Aldicarb sulfoxide	531.2 LL	0.1	--		0.0960	0.1	ug/L	96	50 - 150	---	---	1.0	05/16/2019 08:20	05/21/2019 01:03	4296514
CCL	Baygon	531.2 LL	0.1	--		0.1200	0.1	ug/L	120	50 - 150	---	---	1.0	05/16/2019 08:20	05/21/2019 01:03	4296514
CCL	Carbaryl	531.2 LL	0.1	--		0.1090	0.1	ug/L	109	50 - 150	---	---	1.0	05/16/2019 08:20	05/21/2019 01:03	4296514
CCL	Carbofuran	531.2 LL	0.1	--		0.1210	0.1	ug/L	121	50 - 150	---	---	1.0	05/16/2019 08:20	05/21/2019 01:03	4296514
CCL	3-Hydroxycarbofuran	531.2 LL	0.1	--		0.0920	0.1	ug/L	92	50 - 150	---	---	1.0	05/16/2019 08:20	05/21/2019 01:03	4296514
CCL	Methiocarb	531.2 LL	1.0	--		0.1160	0.1	ug/L	116	50 - 150	---	---	1.0	05/16/2019 08:20	05/21/2019 01:03	4296514
CCL	Methomyl	531.2 LL	0.1	--		0.1090	0.1	ug/L	109	50 - 150	---	---	1.0	05/16/2019 08:20	05/21/2019 01:03	4296514
CCL	1-Naphthol	531.2 LL	1.0	--		0.1200	0.1	ug/L	120	50 - 150	---	---	1.0	05/16/2019 08:20	05/21/2019 01:03	4296514
CCL	Oxamyl	531.2 LL	0.1	--		0.1090	0.1	ug/L	109	50 - 150	---	---	1.0	05/16/2019 08:20	05/21/2019 01:03	4296514
LMB	SS-BDMC	531.2 LL	N/A	--		2.0070	2.0	ug/L	100	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 01:34	4296509
LMB	Aldicarb	531.2 LL	0.1	--	<	0.1		ug/L	---	---	---	---	1.0	05/16/2019 08:20	05/21/2019 01:34	4296509
LMB	Aldicarb sulfone	531.2 LL	0.1	--	<	0.1		ug/L	---	---	---	---	1.0	05/16/2019 08:20	05/21/2019 01:34	4296509
LMB	Aldicarb sulfoxide	531.2 LL	0.1	--	<	0.1		ug/L	---	---	---	---	1.0	05/16/2019 08:20	05/21/2019 01:34	4296509
LMB	Baygon	531.2 LL	0.1	--	<	0.1		ug/L	---	---	---	---	1.0	05/16/2019 08:20	05/21/2019 01:34	4296509
LMB	Carbaryl	531.2 LL	0.1	--	<	0.1		ug/L	---	---	---	---	1.0	05/16/2019 08:20	05/21/2019 01:34	4296509
LMB	Carbofuran	531.2 LL	0.1	--	<	0.1		ug/L	---	---	---	---	1.0	05/16/2019 08:20	05/21/2019 01:34	4296509
LMB	3-Hydroxycarbofuran	531.2 LL	0.1	--	<	0.1		ug/L	---	---	---	---	1.0	05/16/2019 08:20	05/21/2019 01:34	4296509
LMB	Methiocarb	531.2 LL	1.0	--	<	0.1		ug/L	---	---	---	---	1.0	05/16/2019 08:20	05/21/2019 01:34	4296509
LMB	Methomyl	531.2 LL	0.1	--	<	0.1		ug/L	---	---	---	---	1.0	05/16/2019 08:20	05/21/2019 01:34	4296509
LMB	1-Naphthol	531.2 LL	1.0	--	<	1.0		ug/L	---	---	---	---	1.0	05/16/2019 08:20	05/21/2019 01:34	4296509
LMB	Oxamyl	531.2 LL	0.1	--	<	0.1		ug/L	---	---	---	---	1.0	05/16/2019 08:20	05/21/2019 01:34	4296509
FS	SS-BDMC	531.2 LL	N/A	Finished		1.9340	2.0	ug/L	97	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 02:05	4291765
FS	Aldicarb	531.2 LL	0.1	Finished	<	0.1		ug/L	---	---	---	---	1.0	05/16/2019 08:20	05/21/2019 02:05	4291765
FS	Aldicarb sulfone	531.2 LL	0.1	Finished	<	0.1		ug/L	---	---	---	---	1.0	05/16/2019 08:20	05/21/2019 02:05	4291765
FS	Aldicarb sulfoxide	531.2 LL	0.1	Finished	<	0.1		ug/L	---	---	---	---	1.0	05/16/2019 08:20	05/21/2019 02:05	4291765
FS	Baygon	531.2 LL	0.1	Finished	<	0.1		ug/L	---	---	---	---	1.0	05/16/2019 08:20	05/21/2019 02:05	4291765
FS	Carbaryl	531.2 LL	0.1	Finished	<	0.1		ug/L	---	---	---	---	1.0	05/16/2019 08:20	05/21/2019 02:05	4291765
FS	Carbofuran	531.2 LL	0.1	Finished	<	0.1		ug/L	---	---	---	---	1.0	05/16/2019 08:20	05/21/2019 02:05	4291765
FS	3-Hydroxycarbofuran	531.2 LL	0.1	Finished	<	0.1		ug/L	---	---	---	---	1.0	05/16/2019 08:20	05/21/2019 02:05	4291765
FS	Methiocarb	531.2 LL	1.0	Finished	<	1.0		ug/L	---	---	---	---	1.0	05/16/2019 08:20	05/21/2019 02:05	4291765
FS	Methomyl	531.2 LL	0.1	Finished	<	0.1		ug/L	---	---	---	---	1.0	05/16/2019 08:20	05/21/2019 02:05	4291765
FS	1-Naphthol	531.2 LL	1.0	Finished	<	1.0		ug/L	---	---	---	---	1.0	05/16/2019 08:20	05/21/2019 02:05	4291765
FS	Oxamyl	531.2 LL	0.1	Finished	<	0.1		ug/L	---	---	---	---	1.0	05/16/2019 08:20	05/21/2019 02:05	4291765
MS	SS-BDMC	531.2 LL	N/A	Finished		2.2350	2.0	ug/L	112	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 02:36	4296520
MS	Aldicarb	531.2 LL	0.1	Finished		2.1350	2.0	ug/L	107	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 02:36	4296520
MS	Aldicarb sulfone	531.2 LL	0.1	Finished		2.1700	2.0	ug/L	108	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 02:36	4296520
MS	Aldicarb sulfoxide	531.2 LL	0.1	Finished		2.1210	2.0	ug/L	106	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 02:36	4296520

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
MS	Baygon	531.2 LL	0.1	Finished		2.1740	2.0	ug/L	109	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 02:36	4296520
MS	Carbaryl	531.2 LL	0.1	Finished		2.1430	2.0	ug/L	107	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 02:36	4296520
MS	Carbofuran	531.2 LL	0.1	Finished		2.2450	2.0	ug/L	112	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 02:36	4296520
MS	3-Hydroxycarbofuran	531.2 LL	0.1	Finished		2.1390	2.0	ug/L	107	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 02:36	4296520
MS	Methiocarb	531.2 LL	1.0	Finished		2.0850	2.0	ug/L	104	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 02:36	4296520
MS	Methomyl	531.2 LL	0.1	Finished		2.1420	2.0	ug/L	107	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 02:36	4296520
MS	1-Naphthol	531.2 LL	1.0	Finished		1.9740	2.0	ug/L	99	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 02:36	4296520
MS	Oxamyl	531.2 LL	0.1	Finished		2.1380	2.0	ug/L	107	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 02:36	4296520
MSD	SS-BDMC	531.2 LL	N/A	Finished		2.1730	2.0	ug/L	109	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 03:07	4296521
MSD	Aldicarb	531.2 LL	0.1	Finished		2.1940	2.0	ug/L	110	70 - 130	2.7	30	1.0	05/16/2019 08:20	05/21/2019 03:07	4296521
MSD	Aldicarb sulfone	531.2 LL	0.1	Finished		2.2320	2.0	ug/L	112	70 - 130	2.8	30	1.0	05/16/2019 08:20	05/21/2019 03:07	4296521
MSD	Aldicarb sulfoxide	531.2 LL	0.1	Finished		2.1820	2.0	ug/L	109	70 - 130	2.8	30	1.0	05/16/2019 08:20	05/21/2019 03:07	4296521
MSD	Baygon	531.2 LL	0.1	Finished		2.2510	2.0	ug/L	113	70 - 130	3.5	30	1.0	05/16/2019 08:20	05/21/2019 03:07	4296521
MSD	Carbaryl	531.2 LL	0.1	Finished		2.2090	2.0	ug/L	110	70 - 130	3.0	30	1.0	05/16/2019 08:20	05/21/2019 03:07	4296521
MSD	Carbofuran	531.2 LL	0.1	Finished		2.2580	2.0	ug/L	113	70 - 130	0.6	30	1.0	05/16/2019 08:20	05/21/2019 03:07	4296521
MSD	3-Hydroxycarbofuran	531.2 LL	0.1	Finished		2.1940	2.0	ug/L	110	70 - 130	2.5	30	1.0	05/16/2019 08:20	05/21/2019 03:07	4296521
MSD	Methiocarb	531.2 LL	1.0	Finished		2.2020	2.0	ug/L	110	70 - 130	5.5	30	1.0	05/16/2019 08:20	05/21/2019 03:07	4296521
MSD	Methomyl	531.2 LL	0.1	Finished		2.2040	2.0	ug/L	110	70 - 130	2.9	30	1.0	05/16/2019 08:20	05/21/2019 03:07	4296521
MSD	1-Naphthol	531.2 LL	1.0	Finished		2.1480	2.0	ug/L	107	70 - 130	8.4	30	1.0	05/16/2019 08:20	05/21/2019 03:07	4296521
MSD	Oxamyl	531.2 LL	0.1	Finished		2.2300	2.0	ug/L	112	70 - 130	4.2	30	1.0	05/16/2019 08:20	05/21/2019 03:07	4296521
CCM	SS-BDMC	531.2 LL	N/A	---		1.9680	2.0	ug/L	98	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 06:13	4296515
CCM	Aldicarb	531.2 LL	0.1	---		2.0420	2.0	ug/L	102	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 06:13	4296515
CCM	Aldicarb sulfone	531.2 LL	0.1	---		2.0000	2.0	ug/L	100	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 06:13	4296515
CCM	Aldicarb sulfoxide	531.2 LL	0.1	---		1.9910	2.0	ug/L	100	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 06:13	4296515
CCM	Baygon	531.2 LL	0.1	---		2.0080	2.0	ug/L	100	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 06:13	4296515
CCM	Carbaryl	531.2 LL	0.1	---		1.9950	2.0	ug/L	100	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 06:13	4296515
CCM	Carbofuran	531.2 LL	0.1	---		1.9730	2.0	ug/L	99	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 06:13	4296515
CCM	3-Hydroxycarbofuran	531.2 LL	0.1	---		2.0060	2.0	ug/L	100	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 06:13	4296515
CCM	Methiocarb	531.2 LL	1.0	---		1.9170	2.0	ug/L	96	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 06:13	4296515
CCM	Methomyl	531.2 LL	0.1	---		1.9900	2.0	ug/L	100	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 06:13	4296515
CCM	1-Naphthol	531.2 LL	1.0	---		1.8080	2.0	ug/L	90	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 06:13	4296515
CCM	Oxamyl	531.2 LL	0.1	---		1.9860	2.0	ug/L	99	70 - 130	---	---	1.0	05/16/2019 08:20	05/21/2019 06:13	4296515

Eurofins Eaton Analytical

Run Log

Run ID: **260136** Method: **544**

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
CCL	4315867		OS	FL	06/03/2019 18:54	060319M544a.mdb
LRB	4315854		RW	FL	06/03/2019 19:25	060319M544a.mdb
FBL	4315855		RW	FL	06/03/2019 19:41	060319M544a.mdb
FS	4292049	Finished	DW	FL	06/03/2019 20:57	060319M544a.mdb
CCM	4315868		OS	FL	06/03/2019 22:29	060319M544a.mdb

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCL	SS-C2D5-MC-LR	544	N/A	---		0.1031	0.1	ug/L	103	70 - 130	---	---	1.0	06/03/2019 15:00	06/03/2019 18:54	4315867
CCL	Microcystin-LA	544	0.008	---		0.0080	0.008	ug/L	101	50 - 150	---	---	1.0	06/03/2019 15:00	06/03/2019 18:54	4315867
CCL	Microcystin-LF	544	0.006	---		0.0063	0.006	ug/L	105	50 - 150	---	---	1.0	06/03/2019 15:00	06/03/2019 18:54	4315867
CCL	Microcystin-LR	544	0.020	---		0.0165	0.02	ug/L	82	50 - 150	---	---	1.0	06/03/2019 15:00	06/03/2019 18:54	4315867
CCL	Microcystin-LY	544	0.009	---		0.0082	0.009	ug/L	91	50 - 150	---	---	1.0	06/03/2019 15:00	06/03/2019 18:54	4315867
CCL	Microcystin-RR	544	0.006	---		0.0048	0.006	ug/L	79	50 - 150	---	---	1.0	06/03/2019 15:00	06/03/2019 18:54	4315867
CCL	Microcystin-YR	544	0.020	---		0.0158	0.02	ug/L	79	50 - 150	---	---	1.0	06/03/2019 15:00	06/03/2019 18:54	4315867
CCL	Nodularin	544	0.005	---		0.0042	0.005	ug/L	84	50 - 150	---	---	1.0	06/03/2019 15:00	06/03/2019 18:54	4315867
LRB	SS-C2D5-MC-LR	544	N/A	---		0.0842	0.1	ug/L	92	60 - 130	---	---	0.92	06/03/2019 08:30	06/03/2019 19:25	4315854
LRB	Microcystin-LA	544	0.008	---	<	0.008		ug/L	---	---	---	---	0.92	06/03/2019 08:30	06/03/2019 19:25	4315854
LRB	Microcystin-LF	544	0.006	---	<	0.006		ug/L	---	---	---	---	0.92	06/03/2019 08:30	06/03/2019 19:25	4315854
LRB	Microcystin-LR	544	0.020	---	<	0.020		ug/L	---	---	---	---	0.92	06/03/2019 08:30	06/03/2019 19:25	4315854
LRB	Microcystin-LY	544	0.009	---	<	0.009		ug/L	---	---	---	---	0.92	06/03/2019 08:30	06/03/2019 19:25	4315854
LRB	Microcystin-RR	544	0.006	---	<	0.006		ug/L	---	---	---	---	0.92	06/03/2019 08:30	06/03/2019 19:25	4315854
LRB	Microcystin-YR	544	0.020	---	<	0.020		ug/L	---	---	---	---	0.92	06/03/2019 08:30	06/03/2019 19:25	4315854
LRB	Nodularin	544	0.005	---	<	0.005		ug/L	---	---	---	---	0.92	06/03/2019 08:30	06/03/2019 19:25	4315854
FBL	SS-C2D5-MC-LR	544	N/A	---		0.0982	0.1	ug/L	98	60 - 130	---	---	1.0	06/03/2019 08:30	06/03/2019 19:41	4315855
FBL	Microcystin-LA	544	0.008	---		0.0084	0.008	ug/L	105	50 - 150	---	---	1.0	06/03/2019 08:30	06/03/2019 19:41	4315855
FBL	Microcystin-LF	544	0.006	---		0.0058	0.006	ug/L	97	50 - 150	---	---	1.0	06/03/2019 08:30	06/03/2019 19:41	4315855
FBL	Microcystin-LR	544	0.020	---		0.0200	0.02	ug/L	100	50 - 150	---	---	1.0	06/03/2019 08:30	06/03/2019 19:41	4315855
FBL	Microcystin-LY	544	0.009	---		0.0082	0.009	ug/L	91	50 - 150	---	---	1.0	06/03/2019 08:30	06/03/2019 19:41	4315855
FBL	Microcystin-RR	544	0.006	---		0.0057	0.006	ug/L	94	50 - 150	---	---	1.0	06/03/2019 08:30	06/03/2019 19:41	4315855
FBL	Microcystin-YR	544	0.020	---		0.0198	0.02	ug/L	99	50 - 150	---	---	1.0	06/03/2019 08:30	06/03/2019 19:41	4315855
FBL	Nodularin	544	0.005	---		0.0045	0.005	ug/L	90	50 - 150	---	---	1.0	06/03/2019 08:30	06/03/2019 19:41	4315855
FS	SS-C2D5-MC-LR	544	N/A	Finished		0.0774	0.1	ug/L	86	60 - 130	---	---	0.9	06/03/2019 08:30	06/03/2019 20:57	4292049
FS	Microcystin-LA	544	0.008	Finished	<	0.008		ug/L	---	---	---	---	0.9	06/03/2019 08:30	06/03/2019 20:57	4292049
FS	Microcystin-LF	544	0.006	Finished	<	0.006		ug/L	---	---	---	---	0.9	06/03/2019 08:30	06/03/2019 20:57	4292049
FS	Microcystin-LR	544	0.020	Finished	<	0.020		ug/L	---	---	---	---	0.9	06/03/2019 08:30	06/03/2019 20:57	4292049
FS	Microcystin-LY	544	0.009	Finished	<	0.009		ug/L	---	---	---	---	0.9	06/03/2019 08:30	06/03/2019 20:57	4292049
FS	Microcystin-RR	544	0.006	Finished	<	0.006		ug/L	---	---	---	---	0.9	06/03/2019 08:30	06/03/2019 20:57	4292049
FS	Microcystin-YR	544	0.020	Finished	<	0.020		ug/L	---	---	---	---	0.9	06/03/2019 08:30	06/03/2019 20:57	4292049
FS	Nodularin	544	0.005	Finished	<	0.005		ug/L	---	---	---	---	0.9	06/03/2019 08:30	06/03/2019 20:57	4292049
CCM	SS-C2D5-MC-LR	544	N/A	---		0.0806	0.1	ug/L	81	70 - 130	---	---	1.0	06/03/2019 15:00	06/03/2019 22:29	4315868
CCM	Microcystin-LA	544	0.008	---		0.0591	0.08	ug/L	74	70 - 130	---	---	1.0	06/03/2019 15:00	06/03/2019 22:29	4315868
CCM	Microcystin-LF	544	0.006	---		0.0450	0.06	ug/L	75	70 - 130	---	---	1.0	06/03/2019 15:00	06/03/2019 22:29	4315868
CCM	Microcystin-LR	544	0.020	---		0.1595	0.2	ug/L	79	70 - 130	---	---	1.0	06/03/2019 15:00	06/03/2019 22:29	4315868
CCM	Microcystin-LY	544	0.009	---		0.0724	0.09	ug/L	80	70 - 130	---	---	1.0	06/03/2019 15:00	06/03/2019 22:29	4315868
CCM	Microcystin-RR	544	0.006	---		0.0479	0.06	ug/L	80	70 - 130	---	---	1.0	06/03/2019 15:00	06/03/2019 22:29	4315868
CCM	Microcystin-YR	544	0.020	---		0.1663	0.2	ug/L	83	70 - 130	---	---	1.0	06/03/2019 15:00	06/03/2019 22:29	4315868
CCM	Nodularin	544	0.005	---		0.0426	0.05	ug/L	85	70 - 130	---	---	1.0	06/03/2019 15:00	06/03/2019 22:29	4315868

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD Limit	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #

Eurofins Eaton Analytical**Run Log**Run ID: **259519** Method: **545**

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
CCL	4296271		RW	CY	05/16/2019 17:40	051619M545a-CY.mdb
LRB	4296222		RW	CY	05/16/2019 18:09	051619M545a-CY.mdb
FS	4292048	Finished	DW	CY	05/16/2019 20:36	051619M545a-CY.mdb
CCM	4296272		RW	CY	05/16/2019 20:51	051619M545a-CY.mdb

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCL	IS-L-phenylalanine-d5	545	N/A	---		22264	22784	ug/L	98	50 - 150	---	---	1.0	05/13/2019 10:35	05/16/2019 17:40	4296271
CCL	IS-Uracil-d4	545	N/A	---		11842	11786	ug/L	100	50 - 150	---	---	1.0	05/13/2019 10:35	05/16/2019 17:40	4296271
CCL	Anatoxin-a	545	0.03	---		0.0226	0.03	ug/L	75	50 - 150	---	---	1.0	05/13/2019 10:35	05/16/2019 17:40	4296271
CCL	Cylindropermopsin	545	0.09	---		0.0818	0.09	ug/L	91	50 - 150	---	---	1.0	05/13/2019 10:35	05/16/2019 17:40	4296271
LRB	IS-L-phenylalanine-d5	545	N/A	---		22314	22784	ug/L	98	50 - 150	---	---	1.0	05/16/2019 13:50	05/16/2019 18:09	4296222
LRB	IS-Uracil-d4	545	N/A	---		11178	11786	ug/L	95	50 - 150	---	---	1.0	05/16/2019 13:50	05/16/2019 18:09	4296222
LRB	Anatoxin-a	545	0.03	---	<	0.03		ug/L	---	---	---	---	1.0	05/16/2019 13:50	05/16/2019 18:09	4296222
LRB	Cylindropermopsin	545	0.09	---	<	0.09		ug/L	---	---	---	---	1.0	05/16/2019 13:50	05/16/2019 18:09	4296222
FS	IS-L-phenylalanine-d5	545	N/A	Finished		22878	22784	ug/L	100	50 - 150	---	---	1.0	05/16/2019 13:50	05/16/2019 20:36	4292048
FS	IS-Uracil-d4	545	N/A	Finished		13390	11786	ug/L	114	50 - 150	---	---	1.0	05/16/2019 13:50	05/16/2019 20:36	4292048
FS	Anatoxin-a	545	0.03	Finished	<	0.03		ug/L	---	---	---	---	1.0	05/16/2019 13:50	05/16/2019 20:36	4292048
FS	Cylindropermopsin	545	0.09	Finished	<	0.09		ug/L	---	---	---	---	1.0	05/16/2019 13:50	05/16/2019 20:36	4292048
CCM	IS-L-phenylalanine-d5	545	N/A	---		23329	22784	ug/L	102	50 - 150	---	---	1.0	05/13/2019 10:35	05/16/2019 20:51	4296272
CCM	IS-Uracil-d4	545	N/A	---		12251	11786	ug/L	104	50 - 150	---	---	1.0	05/13/2019 10:35	05/16/2019 20:51	4296272
CCM	Anatoxin-a	545	0.03	---		0.5087	0.5	ug/L	102	70 - 130	---	---	1.0	05/13/2019 10:35	05/16/2019 20:51	4296272
CCM	Cylindropermopsin	545	0.09	---		1.4040	1.5	ug/L	94	70 - 130	---	---	1.0	05/13/2019 10:35	05/16/2019 20:51	4296272



Eurofins Eaton Analytical

Run Log

Run ID: 259359 Method: 549.2

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
CCC	4295399		RW	CL	05/16/2019 15:56	549_2-101518CL
LFB	4295398		RW	CL	05/16/2019 16:03	549_2-101518CL
LMB	4295395		RW	CL	05/16/2019 16:18	549_2-101518CL
FS	4291760	Finished	DW	CL	05/16/2019 17:51	549_2-101518CL
CCC	4295400		RW	CL	05/16/2019 18:20	549_2-101518CL

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCC	Diquat	549.2	0.4	---		4.8294	5.0	ug/L	97	80 - 120	---	---	1.0	04/26/2019 08:00	05/16/2019 15:56	4295399
CCC	Paraquat	549.2	0.4	---		4.6523	5.0	ug/L	93	80 - 120	---	---	1.0	04/26/2019 08:00	05/16/2019 15:56	4295399
LFB	Diquat	549.2	0.4	---		3.9134	5.0	ug/L	78	70 - 130	---	---	1.0	05/16/2019 08:20	05/16/2019 16:03	4295398
LFB	Paraquat	549.2	0.4	---		3.9766	5.0	ug/L	80	70 - 130	---	---	1.0	05/16/2019 08:20	05/16/2019 16:03	4295398
LMB	Diquat	549.2	0.4	---	<	0.4		ug/L	---	---	---	---	1.0	05/16/2019 08:20	05/16/2019 16:18	4295395
LMB	Paraquat	549.2	0.4	---	<	0.4		ug/L	---	---	---	---	1.0	05/16/2019 08:20	05/16/2019 16:18	4295395
FS	Diquat	549.2	0.4	Finished	<	0.4		ug/L	---	---	---	---	1.0	05/16/2019 08:20	05/16/2019 17:51	4291760
FS	Paraquat	549.2	0.4	Finished	<	0.4		ug/L	---	---	---	---	1.0	05/16/2019 08:20	05/16/2019 17:51	4291760
CCC	Diquat	549.2	0.4	---		1.8073	2.0	ug/L	90	80 - 120	---	---	1.0	04/26/2019 08:00	05/16/2019 18:20	4295400
CCC	Paraquat	549.2	0.4	---		1.6877	2.0	ug/L	84	80 - 120	---	---	1.0	04/26/2019 08:00	05/16/2019 18:20	4295400

Eurofins Eaton Analytical

Run Log

Run ID: **259805** Method: **556**

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
CCL	4297265		RW	AY	05/18/2019 01:16	m 556 051719AY
CCM	4297269		RW	AY	05/18/2019 03:26	m 556 051719AY
LMB	4297262		RW	AY	05/18/2019 04:31	m 556 051719AY
CCL	4297266		RW	AY	05/18/2019 16:24	m 556 051719AY
FS	4291768	Finished	DW	AY	05/19/2019 02:08	m 556 051719AY
LTB	4291773	LTB-04/30/19	RW	AY	05/19/2019 03:13	m 556 051719AY
CCM	4297270		RW	AY	05/19/2019 04:18	m 556 051719AY

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCL	Acetaldehyde	556	5.0	--		4.2784	5.0	ug/L	86	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 01:16	4297265
CCL	Benzaldehyde	556	5.0	--		4.6140	5.0	ug/L	92	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 01:16	4297265
CCL	Butanal	556	5.0	--		4.1616	5.0	ug/L	83	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 01:16	4297265
CCL	Crotonaldehyde	556	5.0	--		4.0908	5.0	ug/L	82	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 01:16	4297265
CCL	Cyclohexanone	556	5.0	--		3.9279	5.0	ug/L	79	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 01:16	4297265
CCL	Decanal	556	5.0	--		4.3856	5.0	ug/L	88	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 01:16	4297265
CCL	1,2-Dibromopropane	556	N/A	--		335	322	ug/L	104	70 - 130	---	---	1.0	05/17/2019 07:30	05/18/2019 01:16	4297265
CCL	Formaldehyde	556	5.0	--		4.5304	5.0	ug/L	91	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 01:16	4297265
CCL	Glyoxal	556	5.0	--		4.7438	5.0	ug/L	95	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 01:16	4297265
CCL	Heptanal	556	5.0	--		7.0810	5.0	ug/L	142	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 01:16	4297265
CCL	Hexanal	556	5.0	--		3.6602	5.0	ug/L	73	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 01:16	4297265
CCL	Methyl glyoxal	556	5.0	--		4.6791	5.0	ug/L	94	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 01:16	4297265
CCL	Nonanal	556	5.0	--		4.7128	5.0	ug/L	94	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 01:16	4297265
CCL	Octanal	556	5.0	--		4.3936	5.0	ug/L	88	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 01:16	4297265
CCL	Pentanal	556	5.0	--		4.5795	5.0	ug/L	92	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 01:16	4297265
CCL	Propanal	556	5.0	--		4.4685	5.0	ug/L	89	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 01:16	4297265
CCL	2,4,5-Trifluoroacetophenone	556	N/A	--		24.0515	25.0	ug/L	96	70 - 130	---	---	1.0	05/17/2019 07:30	05/18/2019 01:16	4297265
CCM	Acetaldehyde	556	5.0	--		21.0900	20.0	ug/L	105	70 - 130	---	---	1.0	05/17/2019 07:30	05/18/2019 03:26	4297269
CCM	Benzaldehyde	556	5.0	--		24.7755	20.0	ug/L	124	70 - 130	---	---	1.0	05/17/2019 07:30	05/18/2019 03:26	4297269
CCM	Butanal	556	5.0	--		21.3431	20.0	ug/L	107	70 - 130	---	---	1.0	05/17/2019 07:30	05/18/2019 03:26	4297269
CCM	Crotonaldehyde	556	5.0	--		20.9463	20.0	ug/L	105	70 - 130	---	---	1.0	05/17/2019 07:30	05/18/2019 03:26	4297269
CCM	Cyclohexanone	556	5.0	--		19.2659	20.0	ug/L	96	70 - 130	---	---	1.0	05/17/2019 07:30	05/18/2019 03:26	4297269
CCM	Decanal	556	5.0	--		21.6272	20.0	ug/L	108	70 - 130	---	---	1.0	05/17/2019 07:30	05/18/2019 03:26	4297269
CCM	1,2-Dibromopropane	556	N/A	--		321	322	ug/L	100	70 - 130	---	---	1.0	05/17/2019 07:30	05/18/2019 03:26	4297269
CCM	Formaldehyde	556	5.0	--		21.4976	20.0	ug/L	107	70 - 130	---	---	1.0	05/17/2019 07:30	05/18/2019 03:26	4297269
CCM	Glyoxal	556	5.0	--		19.7614	20.0	ug/L	99	70 - 130	---	---	1.0	05/17/2019 07:30	05/18/2019 03:26	4297269
CCM	Heptanal	556	5.0	--		22.9619	20.0	ug/L	115	70 - 130	---	---	1.0	05/17/2019 07:30	05/18/2019 03:26	4297269
CCM	Hexanal	556	5.0	--		20.4528	20.0	ug/L	102	70 - 130	---	---	1.0	05/17/2019 07:30	05/18/2019 03:26	4297269
CCM	Methyl glyoxal	556	5.0	--		21.1919	20.0	ug/L	106	70 - 130	---	---	1.0	05/17/2019 07:30	05/18/2019 03:26	4297269
CCM	Nonanal	556	5.0	--		21.4882	20.0	ug/L	107	70 - 130	---	---	1.0	05/17/2019 07:30	05/18/2019 03:26	4297269
CCM	Octanal	556	5.0	--		22.2303	20.0	ug/L	111	70 - 130	---	---	1.0	05/17/2019 07:30	05/18/2019 03:26	4297269
CCM	Pentanal	556	5.0	--		21.1291	20.0	ug/L	106	70 - 130	---	---	1.0	05/17/2019 07:30	05/18/2019 03:26	4297269
CCM	Propanal	556	5.0	--		21.0444	20.0	ug/L	105	70 - 130	---	---	1.0	05/17/2019 07:30	05/18/2019 03:26	4297269
CCM	2,4,5-Trifluoroacetophenone	556	N/A	--		25.9921	25.0	ug/L	104	70 - 130	---	---	1.0	05/17/2019 07:30	05/18/2019 03:26	4297269
LMB	Acetaldehyde	556	5.0	--	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/18/2019 04:31	4297262
LMB	Benzaldehyde	556	5.0	--	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/18/2019 04:31	4297262
LMB	Butanal	556	5.0	--	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/18/2019 04:31	4297262
LMB	Crotonaldehyde	556	5.0	--	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/18/2019 04:31	4297262
LMB	Cyclohexanone	556	5.0	--	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/18/2019 04:31	4297262
LMB	Decanal	556	5.0	--	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/18/2019 04:31	4297262
LMB	Decanal	556	5.0	--	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/18/2019 04:31	4297262

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LMB	1,2-Dibromopropane	556	N/A	---		329	322	ug/L	102	70 - 130	---	---	1.0	05/17/2019 07:30	05/18/2019 04:31	4297262
LMB	Formaldehyde	556	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/18/2019 04:31	4297262
LMB	Glyoxal	556	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/18/2019 04:31	4297262
LMB	Heptanal	556	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/18/2019 04:31	4297262
LMB	Hexanal	556	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/18/2019 04:31	4297262
LMB	Methyl glyoxal	556	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/18/2019 04:31	4297262
LMB	Nonanal	556	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/18/2019 04:31	4297262
LMB	Octanal	556	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/18/2019 04:31	4297262
LMB	Pentanal	556	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/18/2019 04:31	4297262
LMB	Propanal	556	5.0	---	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/18/2019 04:31	4297262
LMB	2,4,5-Trifluoroacetophenone	556	N/A	---		23.5546	25.0	ug/L	94	70 - 130	---	---	1.0	05/17/2019 07:30	05/18/2019 04:31	4297262
CCL	Acetaldehyde	556	5.0	---		3.8039	5.0	ug/L	76	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 16:24	4297266
CCL	Benzaldehyde	556	5.0	---		4.7671	5.0	ug/L	95	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 16:24	4297266
CCL	Butanal	556	5.0	---		4.2487	5.0	ug/L	85	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 16:24	4297266
CCL	Crotonaldehyde	556	5.0	---		4.3859	5.0	ug/L	88	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 16:24	4297266
CCL	Cyclohexanone	556	5.0	---		3.3161	5.0	ug/L	66	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 16:24	4297266
CCL	Decanal	556	5.0	---		4.0396	5.0	ug/L	81	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 16:24	4297266
CCL	1,2-Dibromopropane	556	N/A	---		313	322	ug/L	97	70 - 130	---	---	1.0	05/17/2019 07:30	05/18/2019 16:24	4297266
CCL	Formaldehyde	556	5.0	---		4.5479	5.0	ug/L	91	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 16:24	4297266
CCL	Glyoxal	556	5.0	---		4.7261	5.0	ug/L	95	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 16:24	4297266
CCL	Heptanal	556	5.0	---		5.9503	5.0	ug/L	119	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 16:24	4297266
CCL	Hexanal	556	5.0	---		3.0004	5.0	ug/L	60	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 16:24	4297266
CCL	Methyl glyoxal	556	5.0	---		4.4738	5.0	ug/L	89	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 16:24	4297266
CCL	Nonanal	556	5.0	---		5.0792	5.0	ug/L	102	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 16:24	4297266
CCL	Octanal	556	5.0	---		3.0202	5.0	ug/L	60	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 16:24	4297266
CCL	Pentanal	556	5.0	---		4.3686	5.0	ug/L	87	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 16:24	4297266
CCL	Propanal	556	5.0	---		5.2426	5.0	ug/L	105	50 - 150	---	---	1.0	05/17/2019 07:30	05/18/2019 16:24	4297266
CCL	2,4,5-Trifluoroacetophenone	556	N/A	---		24.8601	25.0	ug/L	99	70 - 130	---	---	1.0	05/17/2019 07:30	05/18/2019 16:24	4297266
FS	Acetaldehyde	556	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 02:08	4291768
FS	Benzaldehyde	556	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 02:08	4291768
FS	Butanal	556	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 02:08	4291768
FS	Crotonaldehyde	556	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 02:08	4291768
FS	Cyclohexanone	556	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 02:08	4291768
FS	Decanal	556	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 02:08	4291768
FS	1,2-Dibromopropane	556	N/A	Finished		325	322	ug/L	101	70 - 130	---	---	1.0	05/17/2019 07:30	05/19/2019 02:08	4291768
FS	Formaldehyde	556	5.0	Finished	<	7.1		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 02:08	4291768
FS	Glyoxal	556	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 02:08	4291768
FS	Heptanal	556	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 02:08	4291768
FS	Hexanal	556	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 02:08	4291768
FS	Methyl glyoxal	556	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 02:08	4291768
FS	Nonanal	556	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 02:08	4291768

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FS	Octanal	556	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 02:08	4291768
FS	Pentanal	556	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 02:08	4291768
FS	Propanal	556	5.0	Finished	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 02:08	4291768
FS	2,4,5-Trifluoroacetophenone	556	N/A	Finished		24.6892	25.0	ug/L	99	70 - 130	---	---	1.0	05/17/2019 07:30	05/19/2019 02:08	4291768
LTB	Acetaldehyde	556	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 03:13	4291773
LTB	Benzaldehyde	556	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 03:13	4291773
LTB	Butanal	556	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 03:13	4291773
LTB	Crotonaldehyde	556	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 03:13	4291773
LTB	Cyclohexanone	556	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 03:13	4291773
LTB	Decanal	556	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 03:13	4291773
LTB	1,2-Dibromopropane	556	N/A	LTB-04/30/19		345	322	ug/L	107	70 - 130	---	---	1.0	05/17/2019 07:30	05/19/2019 03:13	4291773
LTB	Formaldehyde	556	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 03:13	4291773
LTB	Glyoxal	556	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 03:13	4291773
LTB	Heptanal	556	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 03:13	4291773
LTB	Hexanal	556	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 03:13	4291773
LTB	Methyl glyoxal	556	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 03:13	4291773
LTB	Nonanal	556	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 03:13	4291773
LTB	Octanal	556	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 03:13	4291773
LTB	Pentanal	556	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 03:13	4291773
LTB	Propanal	556	5.0	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 03:13	4291773
LTB	2,4,5-Trifluoroacetophenone	556	N/A	LTB-04/30/19	<	5.0		ug/L	---	---	---	---	1.0	05/17/2019 07:30	05/19/2019 03:13	4291773
CCM	Acetaldehyde	556	5.0	---		24.8703	25.0	ug/L	99	70 - 130	---	---	1.0	05/17/2019 07:30	05/19/2019 03:13	4291773
CCM	Benzaldehyde	556	5.0	---		21.2172	20.0	ug/L	106	70 - 130	---	---	1.0	05/17/2019 07:30	05/19/2019 04:18	4297270
CCM	Butanal	556	5.0	---		25.1728	20.0	ug/L	126	70 - 130	---	---	1.0	05/17/2019 07:30	05/19/2019 04:18	4297270
CCM	Crotonaldehyde	556	5.0	---		21.0777	20.0	ug/L	105	70 - 130	---	---	1.0	05/17/2019 07:30	05/19/2019 04:18	4297270
CCM	Cyclohexanone	556	5.0	---		20.9611	20.0	ug/L	105	70 - 130	---	---	1.0	05/17/2019 07:30	05/19/2019 04:18	4297270
CCM	Decanal	556	5.0	---		20.7810	20.0	ug/L	104	70 - 130	---	---	1.0	05/17/2019 07:30	05/19/2019 04:18	4297270
CCM	1,2-Dibromopropane	556	5.0	---		21.3702	20.0	ug/L	107	70 - 130	---	---	1.0	05/17/2019 07:30	05/19/2019 04:18	4297270
CCM	Formaldehyde	556	N/A	---		323	322	ug/L	100	70 - 130	---	---	1.0	05/17/2019 07:30	05/19/2019 04:18	4297270
CCM	Glyoxal	556	5.0	---		20.6053	20.0	ug/L	103	70 - 130	---	---	1.0	05/17/2019 07:30	05/19/2019 04:18	4297270
CCM	Heptanal	556	5.0	---		19.7269	20.0	ug/L	99	70 - 130	---	---	1.0	05/17/2019 07:30	05/19/2019 04:18	4297270
CCM	Hexanal	556	5.0	---		20.4937	20.0	ug/L	102	70 - 130	---	---	1.0	05/17/2019 07:30	05/19/2019 04:18	4297270
CCM	Methyl glyoxal	556	5.0	---		20.4098	20.0	ug/L	102	70 - 130	---	---	1.0	05/17/2019 07:30	05/19/2019 04:18	4297270
CCM	Nonanal	556	5.0	---		21.1489	20.0	ug/L	106	70 - 130	---	---	1.0	05/17/2019 07:30	05/19/2019 04:18	4297270
CCM	Octanal	556	5.0	---		21.1689	20.0	ug/L	106	70 - 130	---	---	1.0	05/17/2019 07:30	05/19/2019 04:18	4297270
CCM	Pentanal	556	5.0	---		20.7692	20.0	ug/L	104	70 - 130	---	---	1.0	05/17/2019 07:30	05/19/2019 04:18	4297270
CCM	Propanal	556	5.0	---		21.3943	20.0	ug/L	107	70 - 130	---	---	1.0	05/17/2019 07:30	05/19/2019 04:18	4297270
CCM	2,4,5-Trifluoroacetophenone	556	5.0	---		20.6309	20.0	ug/L	103	70 - 130	---	---	1.0	05/17/2019 07:30	05/19/2019 04:18	4297270
CCM	2,4,5-Trifluoroacetophenone	556	N/A	---		26.2058	25.0	ug/L	105	70 - 130	---	---	1.0	05/17/2019 07:30	05/19/2019 04:18	4297270



Eaton Analytical

Eurofins Eaton Analytical

Run Log

Run ID: 259915 Method: 8330A

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
FS	4292046	Finished	DW	R L	05/24/2019 08:37	Ref Lab 8330A

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FS	2-Amino-4,6-dinitrotoluene	8330A	0.6	Finished	<	0.6		ug/L	---	---	---	---	1.0	---	05/24/2019 08:37	4292046
FS	4-Amino-2,6-dinitrotoluene	8330A	0.6	Finished	<	0.6		ug/L	---	---	---	---	1.0	---	05/24/2019 08:37	4292046
FS	1,3-Dinitrobenzene	8330A	0.6	Finished	<	0.6		ug/L	---	---	---	---	1.0	---	05/24/2019 08:37	4292046
FS	2,4-Dinitrotoluene	8330A	2.0	Finished	<	2.0		ug/L	---	---	---	---	1.0	---	05/24/2019 08:37	4292046
FS	2,6-Dinitrotoluene	8330A	1.3	Finished	<	1.3		ug/L	---	---	---	---	1.0	---	05/24/2019 08:37	4292046
FS	HMX (Octahydro-1,3,5,7-tetranitro-1,3,5,7-tetrazocin	8330A	2.0	Finished	<	2.0		ug/L	---	---	---	---	1.0	---	05/24/2019 08:37	4292046
FS	Nitrobenzene	8330A	0.6	Finished	<	0.6		ug/L	---	---	---	---	1.0	---	05/24/2019 08:37	4292046
FS	Nitroglycerin	8330A	15	Finished	<	15		ug/L	---	---	---	---	1.0	---	05/24/2019 08:37	4292046
FS	2-Nitrotoluene	8330A	0.75	Finished	<	0.75		ug/L	---	---	---	---	1.0	---	05/24/2019 08:37	4292046
FS	3-Nitrotoluene	8330A	1.2	Finished	<	1.2		ug/L	---	---	---	---	1.0	---	05/24/2019 08:37	4292046
FS	4-Nitrotoluene	8330A	1.2	Finished	<	1.2		ug/L	---	---	---	---	1.0	---	05/24/2019 08:37	4292046
FS	PETN (Pentaerythritol tetranitrate)	8330A	18	Finished	<	18		ug/L	---	---	---	---	1.0	---	05/24/2019 08:37	4292046
FS	RDX (Hexahydro-1,3,5-trinitro-1,3,5-triazine)	8330A	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/24/2019 08:37	4292046
FS	Tetryl (2,4,6-Trinitrophenylmethylnitramine)	8330A	0.8	Finished	<	0.8		ug/L	---	---	---	---	1.0	---	05/24/2019 08:37	4292046
FS	1,3,5-Trinitrobenzene	8330A	0.6	Finished	<	0.6		ug/L	---	---	---	---	1.0	---	05/24/2019 08:37	4292046
FS	2,4,6-Trinitrotoluene	8330A	0.6	Finished	<	0.6		ug/L	---	---	---	---	1.0	---	05/24/2019 08:37	4292046

Eurofins Eaton Analytical

Run Log

Run ID: 259809 Method: L306

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
ICL	4301881		RW	DQ	05/21/2019 16:24	052119L306a.mdb
ICS	4301882		RW	DQ	05/21/2019 16:48	052119L306a.mdb
ICS	4301883		RW	DQ	05/21/2019 17:13	052119L306a.mdb
ICS	4301884		RW	DQ	05/21/2019 17:37	052119L306a.mdb
ICS	4301885		RW	DQ	05/21/2019 18:01	052119L306a.mdb
ICS	4301886		RW	DQ	05/21/2019 18:26	052119L306a.mdb
LMB	4301887		RW	DQ	05/21/2019 19:14	052119L306a.mdb
FS	4291764	Finished	DW	DQ	05/21/2019 20:27	052119L306a.mdb
CCC	4301888		RW	DQ	05/21/2019 22:04	052119L306a.mdb

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
ICL	IS-Linuron-d6	L306	N/A	--		22869	22048	ug/L	104	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Diuron	L306	N/A	--		0.1038	0.1	ug/L	104	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Diflufenazuron	L306	N/A	--		0.1010	0.1	ug/L	101	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Methomyl	L306	N/A	--		0.1052	0.1	ug/L	105	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Siduron, Total	L306	N/A	--		0.1043	0.1	ug/L	104	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Neburon	L306	N/A	--		0.1032	0.1	ug/L	103	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	IS-Diuron-d6	L306	N/A	--		365865	360194	ug/L	102	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Tebuthiuron	L306	N/A	--		0.1044	0.1	ug/L	104	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Propanil	L306	N/A	--		0.0980	0.1	ug/L	98	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Bensulfide	L306	N/A	--		0.0993	0.1	ug/L	99	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Bendiocarb	L306	N/A	--		0.1036	0.1	ug/L	104	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Oxamyl	L306	N/A	--		1.1472	1.0	ug/L	115	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	3-Hydroxycarbofuran	L306	N/A	--		0.1055	0.1	ug/L	106	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Fenuron	L306	N/A	--		0.1008	0.1	ug/L	101	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Aldicarb sulfone	L306	N/A	--		1.1116	1.0	ug/L	111	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Baygon	L306	N/A	--		0.1022	0.1	ug/L	102	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Halofenozide	L306	N/A	--		0.1028	0.1	ug/L	103	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Paclobutrazol	L306	N/A	--		0.1189	0.1	ug/L	119	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Metalaxyl/Metalaxyl-M (Mefenoxam)	L306	N/A	--		0.1026	0.1	ug/L	103	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Oryzalin	L306	N/A	--		0.9474	1.0	ug/L	95	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Linuron	L306	N/A	--		0.9333	1.0	ug/L	93	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Propargite	L306	N/A	--		0.0991	0.1	ug/L	99	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Oxadiazon	L306	N/A	--		0.1001	0.1	ug/L	100	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Azoxystrobin	L306	N/A	--		1.0687	1.0	ug/L	107	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Triadimenol	L306	N/A	--		0.1009	0.1	ug/L	101	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Fluometuron	L306	N/A	--		0.0984	0.1	ug/L	98	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Aldicarb sulfoxide	L306	N/A	--		1.0660	1.0	ug/L	107	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Glomazone	L306	N/A	--		0.1094	0.1	ug/L	109	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Monuron	L306	N/A	--		0.1075	0.1	ug/L	108	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Halosulfuron methyl	L306	N/A	--		0.1034	0.1	ug/L	103	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Carbaryl	L306	N/A	--		0.1016	0.1	ug/L	102	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Thidiazuron	L306	N/A	--		0.0984	0.1	ug/L	98	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Carbofuran	L306	N/A	--		0.1028	0.1	ug/L	103	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Methiocarb	L306	N/A	--		0.1036	0.1	ug/L	104	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Imidacloprid	L306	N/A	--		0.1028	0.1	ug/L	103	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICL	Triadimefon	L306	N/A	--		0.1067	0.1	ug/L	107	50 - 150	---	---	1.0	---	05/21/2019 16:24	4301881
ICS	IS-Linuron-d6	L306	N/A	--		21404	22048	ug/L	97	50 - 150	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Diuron	L306	N/A	--		0.2126	0.2	ug/L	106	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Diflufenazuron	L306	N/A	--		0.2105	0.2	ug/L	105	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Methomyl	L306	N/A	--		0.2132	0.2	ug/L	107	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
ICS	Siduron, Total	L306	N/A	---		0.2155	0.2	ug/L	108	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Neburon	L306	N/A	---		0.2085	0.2	ug/L	104	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	IS-Diuron-d6	L306	N/A	---		364225	360194	ug/L	101	50 - 150	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Tebuthiuron	L306	N/A	---		0.2018	0.2	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Propanil	L306	N/A	---		0.2210	0.2	ug/L	110	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Bensulide	L306	N/A	---		0.2099	0.2	ug/L	105	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Bendiocarb	L306	N/A	---		0.2135	0.2	ug/L	107	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Oxamyl	L306	N/A	---		2.3271	2.0	ug/L	116	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	3-Hydroxycarbofuran	L306	N/A	---		0.2148	0.2	ug/L	107	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Fenuron	L306	N/A	---		0.2155	0.2	ug/L	108	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Aldicarb sulfone	L306	N/A	---		2.2687	2.0	ug/L	113	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Baygon	L306	N/A	---		0.2147	0.2	ug/L	107	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Halofenozide	L306	N/A	---		0.2224	0.2	ug/L	111	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Paclobutrazol	L306	N/A	---		0.2071	0.2	ug/L	104	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Metolaxyl/Metalaxyl-M (Mefenoxam)	L306	N/A	---		0.2110	0.2	ug/L	105	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Oryzalin	L306	N/A	---		2.1008	2.0	ug/L	105	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Linuron	L306	N/A	---		2.3037	2.0	ug/L	115	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Propargite	L306	N/A	---		0.2154	0.2	ug/L	108	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Oxadiazon	L306	N/A	---		0.2113	0.2	ug/L	106	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Azoxystrobin	L306	N/A	---		2.1864	2.0	ug/L	109	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Triadimenol	L306	N/A	---		0.2117	0.2	ug/L	106	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Fluometuron	L306	N/A	---		0.2052	0.2	ug/L	103	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Aldicarb sulfioxide	L306	N/A	---		2.2539	2.0	ug/L	113	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Clomazone	L306	N/A	---		0.2231	0.2	ug/L	112	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Monuron	L306	N/A	---		0.2201	0.2	ug/L	110	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Halosulfuron methyl	L306	N/A	---		0.1969	0.2	ug/L	98	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Carbaryl	L306	N/A	---		0.2104	0.2	ug/L	105	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Thidiazuron	L306	N/A	---		0.2121	0.2	ug/L	106	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Carbofuran	L306	N/A	---		0.2095	0.2	ug/L	105	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Methiocarb	L306	N/A	---		0.2108	0.2	ug/L	105	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Imidacloprid	L306	N/A	---		0.2161	0.2	ug/L	108	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	Triadimefon	L306	N/A	---		0.2059	0.2	ug/L	103	70 - 130	---	---	1.0	---	05/21/2019 16:48	4301882
ICS	IS-Linuron-d6	L306	N/A	---		22087	22048	ug/L	100	50 - 150	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Diuron	L306	N/A	---		0.4804	0.5	ug/L	96	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Diflufenuron	L306	N/A	---		0.4714	0.5	ug/L	94	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Methomyl	L306	N/A	---		0.4913	0.5	ug/L	98	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Siduron, Total	L306	N/A	---		0.4731	0.5	ug/L	95	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Neburon	L306	N/A	---		0.4759	0.5	ug/L	95	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	IS-Diuron-d6	L306	N/A	---		37047	360194	ug/L	103	50 - 150	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Tebuthiuron	L306	N/A	---		0.4736	0.5	ug/L	95	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Propanil	L306	N/A	---		0.4808	0.5	ug/L	96	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
ICS	Bensulide	L306	N/A	---		0.4497	0.5	ug/L	90	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Bendiocarb	L306	N/A	---		0.4847	0.5	ug/L	97	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Oxamyl	L306	N/A	---		5.1075	5.0	ug/L	102	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	3-Hydroxycarbofuran	L306	N/A	---		0.4800	0.5	ug/L	96	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Fenuron	L306	N/A	---		0.4787	0.5	ug/L	96	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Aldicarb sulfone	L306	N/A	---		4.9992	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Baygon	L306	N/A	---		0.4776	0.5	ug/L	96	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Halofenozide	L306	N/A	---		0.5103	0.5	ug/L	102	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Paclobutrazol	L306	N/A	---		0.4772	0.5	ug/L	95	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Metaxyl/Metaxyl-M (Mefenoxam)	L306	N/A	---		0.4831	0.5	ug/L	97	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Oryzalin	L306	N/A	---		4.9685	5.0	ug/L	99	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Linuron	L306	N/A	---		4.8739	5.0	ug/L	97	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Propargite	L306	N/A	---		0.4361	0.5	ug/L	87	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Oxadiazon	L306	N/A	---		0.4613	0.5	ug/L	92	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Azoxystrobin	L306	N/A	---		4.9408	5.0	ug/L	99	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Triadimenol	L306	N/A	---		0.4860	0.5	ug/L	97	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Fluometuron	L306	N/A	---		0.4878	0.5	ug/L	98	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Aldicarb sulfoxide	L306	N/A	---		4.8857	5.0	ug/L	98	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Glomazone	L306	N/A	---		0.4957	0.5	ug/L	99	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Monuron	L306	N/A	---		0.4865	0.5	ug/L	97	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Halosulfuron methyl	L306	N/A	---		0.4785	0.5	ug/L	96	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Carbaryl	L306	N/A	---		0.4749	0.5	ug/L	95	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Thidiazuron	L306	N/A	---		0.4800	0.5	ug/L	96	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Carbofuran	L306	N/A	---		0.4726	0.5	ug/L	95	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Methiocarb	L306	N/A	---		0.4842	0.5	ug/L	97	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Iridacloprid	L306	N/A	---		0.4714	0.5	ug/L	94	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	Triadimefon	L306	N/A	---		0.4801	0.5	ug/L	96	70 - 130	---	---	1.0	---	05/21/2019 17:13	4301883
ICS	IS-Linuron-d6	L306	N/A	---		22222	22048	ug/L	101	50 - 150	---	---	1.0	---	05/21/2019 17:37	4301884
ICS	Diuron	L306	N/A	---		1.0013	1.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301884
ICS	Diflufenuron	L306	N/A	---		1.0064	1.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301884
ICS	Methomyl	L306	N/A	---		1.0087	1.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301884
ICS	Siduron, Total	L306	N/A	---		1.0040	1.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301884
ICS	Neburon	L306	N/A	---		0.9997	1.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301884
ICS	IS-Diuron-d6	L306	N/A	---		361816	360194	ug/L	100	50 - 150	---	---	1.0	---	05/21/2019 17:37	4301884
ICS	Tebufluron	L306	N/A	---		1.0072	1.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301884
ICS	Propanil	L306	N/A	---		0.9838	1.0	ug/L	98	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301884
ICS	Bensulide	L306	N/A	---		1.0005	1.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301884
ICS	Bendiocarb	L306	N/A	---		1.0006	1.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301884
ICS	Oxamyl	L306	N/A	---		10.0747	10.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301884
ICS	3-Hydroxycarbofuran	L306	N/A	---		1.0074	1.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301884
ICS	Fenuron	L306	N/A	---		1.0070	1.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301884

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
ICS	Aldicarb sulfone	L306	N/A	---		10.0969	10.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301894
ICS	Baygon	L306	N/A	---		0.9994	1.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301894
ICS	Halofenozide	L306	N/A	---		0.9880	1.0	ug/L	99	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301894
ICS	Paclotrazol	L306	N/A	---		0.9628	1.0	ug/L	96	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301894
ICS	Metalaxyl/Metalaxyl-M (Mefenoxam)	L306	N/A	---		0.9978	1.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301894
ICS	Oryzalin	L306	N/A	---		9.9662	10.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301894
ICS	Linuron	L306	N/A	---		9.9900	10.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301894
ICS	Propargite	L306	N/A	---		0.9971	1.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301894
ICS	Oxadiazon	L306	N/A	---		0.9980	1.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301894
ICS	Azoxystrobin	L306	N/A	---		10.1097	10.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301894
ICS	Triadimenol	L306	N/A	---		1.0056	1.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301894
ICS	Fluometuron	L306	N/A	---		1.0097	1.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301894
ICS	Aldicarb sulfoxide	L306	N/A	---		9.5556	10.0	ug/L	96	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301894
ICS	Giomazone	L306	N/A	---		1.0000	1.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301894
ICS	Monuron	L306	N/A	---		1.0076	1.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301894
ICS	Halosulfuron methyl	L306	N/A	---		0.9979	1.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301894
ICS	Carbaryl	L306	N/A	---		1.0026	1.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301894
ICS	Thidiazuron	L306	N/A	---		0.9924	1.0	ug/L	99	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301894
ICS	Carbofuran	L306	N/A	---		1.0122	1.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301894
ICS	Methiocarb	L306	N/A	---		1.0076	1.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301894
ICS	Imidacloprid	L306	N/A	---		1.0108	1.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301894
ICS	Triadimefon	L306	N/A	---		0.9949	1.0	ug/L	99	70 - 130	---	---	1.0	---	05/21/2019 17:37	4301894
ICS	IS-Linuron-d6	L306	N/A	---		21891	22048	ug/L	99	50 - 150	---	---	1.0	---	05/21/2019 18:01	4301895
ICS	Diuron	L306	N/A	---		2.0009	2.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301895
ICS	Diflufenazuron	L306	N/A	---		2.0162	2.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301895
ICS	Methomyl	L306	N/A	---		1.9700	2.0	ug/L	99	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301895
ICS	Siduron, Total	L306	N/A	---		2.0042	2.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301895
ICS	Neburon	L306	N/A	---		2.0208	2.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301895
ICS	IS-Diuron-d6	L306	N/A	---		356992	360194	ug/L	100	50 - 150	---	---	1.0	---	05/21/2019 18:01	4301895
ICS	Tebuthiuron	L306	N/A	---		2.0217	2.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301895
ICS	Propanil	L306	N/A	---		2.0222	2.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301895
ICS	Bensulfide	L306	N/A	---		2.0614	2.0	ug/L	103	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301895
ICS	Bendiocarb	L306	N/A	---		1.9947	2.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301895
ICS	Oxamyl	L306	N/A	---		18.7691	20.0	ug/L	94	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301895
ICS	3-Hydroxycarbofuran	L306	N/A	---		1.9896	2.0	ug/L	99	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301895
ICS	Fenuron	L306	N/A	---		1.9970	2.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301895
ICS	Aldicarb sulfone	L306	N/A	---		19.2023	20.0	ug/L	96	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301895
ICS	Baygon	L306	N/A	---		2.0071	2.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301895
ICS	Halofenozide	L306	N/A	---		1.9710	2.0	ug/L	99	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301895
ICS	Paclotrazol	L306	N/A	---		2.0527	2.0	ug/L	103	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301895
ICS	Metalaxyl/Metalaxyl-M (Mefenoxam)	L306	N/A	---		2.0063	2.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301895

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
ICS	Oryzalin	L306	N/A	---		20.0304	20.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301885
ICS	Linuron	L306	N/A	---		19.8757	20.0	ug/L	99	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301885
ICS	Propargite	L306	N/A	---		2.0878	2.0	ug/L	104	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301885
ICS	Oxadiazon	L306	N/A	---		2.0427	2.0	ug/L	102	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301885
ICS	Azoxystrobin	L306	N/A	---		19.4763	20.0	ug/L	97	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301885
ICS	Triadimenol	L306	N/A	---		1.9924	2.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301885
ICS	Fluometuron	L306	N/A	---		1.9982	2.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301885
ICS	Aldicarb sulfioxide	L306	N/A	---		20.8858	20.0	ug/L	104	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301885
ICS	Clomazone	L306	N/A	---		1.9537	2.0	ug/L	98	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301885
ICS	Monuron	L306	N/A	---		1.9608	2.0	ug/L	98	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301885
ICS	Halosulfuron methyl	L306	N/A	---		2.0290	2.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301885
ICS	Carbaryl	L306	N/A	---		2.0119	2.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301885
ICS	Thidiazuron	L306	N/A	---		2.0206	2.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301885
ICS	Carbofuran	L306	N/A	---		2.0050	2.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301885
ICS	Methiocarb	L306	N/A	---		1.9916	2.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301885
ICS	Imidacloprid	L306	N/A	---		1.9992	2.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301885
ICS	Triadimefon	L306	N/A	---		2.0185	2.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 18:01	4301885
ICS	IS-Linuron-d6	L306	N/A	---		20685	22048	ug/L	94	50 - 150	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Diuron	L306	N/A	---		5.0008	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Diflufenzuron	L306	N/A	---		4.9950	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Methomyl	L306	N/A	---		5.0126	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Siduron, Total	L306	N/A	---		4.9995	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Neburon	L306	N/A	---		4.9935	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	IS-Diuron-d6	L306	N/A	---		340229	360194	ug/L	94	50 - 150	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Tebuflufen	L306	N/A	---		4.9928	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Propanil	L306	N/A	---		4.9956	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Bensulfide	L306	N/A	---		4.9795	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Bendiocarb	L306	N/A	---		5.0035	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Oxamyl	L306	N/A	---		50.7711	50.0	ug/L	102	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	3-Hydroxycarbofuran	L306	N/A	---		5.0045	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Fenuron	L306	N/A	---		5.0015	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Aldicarb sulfone	L306	N/A	---		50.3372	50.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Baygon	L306	N/A	---		4.9989	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Halofenozide	L306	N/A	---		5.0113	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Paclotrazol	L306	N/A	---		4.9890	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Metaxyl/Metalaxyl-M (Mefenoxam)	L306	N/A	---		4.9992	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Oryzalin	L306	N/A	---		49.9943	50.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Linuron	L306	N/A	---		50.0483	50.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Propargite	L306	N/A	---		4.9689	5.0	ug/L	99	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Oxadiazon	L306	N/A	---		4.9869	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Azoxystrobin	L306	N/A	---		50.2592	50.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
ICS	Triadimenol	L306	N/A	---		5.0026	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Fluometuron	L306	N/A	---		4.9998	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Clomazone	L306	N/A	---		5.0209	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Monuron	L306	N/A	---		5.0229	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Halosulfuron methyl	L306	N/A	---		4.9911	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Carbaryl	L306	N/A	---		4.9967	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Thidiazuron	L306	N/A	---		4.9951	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Carbofuran	L306	N/A	---		4.9978	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Methiocarb	L306	N/A	---		5.0029	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Imidacloprid	L306	N/A	---		5.0003	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
ICS	Triadimefon	L306	N/A	---		4.9956	5.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 18:26	4301886
LMB	IS-Linuron-d6	L306	N/A	---		20584	22048	ug/L	93	50 - 150	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Diuron	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Diflufenuron	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Methomyl	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Siduron, Total	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Neburon	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	IS-Diuron-d6	L306	N/A	---		372937	360194	ug/L	104	50 - 150	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Tebuthiuron	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Propanil	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Bensulide	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Bendiocarb	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Oxamyl	L306	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	3-Hydroxycarbofuran	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Fenuron	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Aldicarb sulfone	L306	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Baygon	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Halofenozide	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Paclobutrazol	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Metolaxyl/Metalaxyl-M (Mefenoxam)	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Oryzalin	L306	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Linuron	L306	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Propargite	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Oxadiazon	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Azoxystrobin	L306	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Triadimenol	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Fluometuron	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Aldicarb sulfoxide	L306	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Clomazone	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Monuron	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Halosulfuron methyl	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LMB	Carbaryl	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Thidiazuron	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Carbofuran	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Methiocarb	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Imidacloprid	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
LMB	Triadimefon	L306	0.05	---	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 19:14	4301887
FS	IS-Linuron-d6	L306	N/A	Finished	<	20821	22048	ug/L	94	50 - 150	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Diuron	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Diflufenuron	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Methomyl	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Siduron, Total	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Neburon	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	IS-Diuron-d6	L306	N/A	Finished	<	385424	360194	ug/L	107	50 - 150	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Tebuuthiuron	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Propanil	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Bensulfide	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Bendiocarb	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Oxamyl	L306	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	3-Hydroxycarbofuran	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Fenuron	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Aldicarb sulfone	L306	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Baygon	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Halofenozide	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Paclobutrazol	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Metolaxyl/Metolaxyl-M (Mefenoxam)	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Oryzalin	L306	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Linuron	L306	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Propargite	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Oxadiazon	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Azoxystrobin	L306	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Triadimenol	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Fluometuron	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Aldicarb sulfoxide	L306	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Clomazone	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Monuron	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Halosulfuron methyl	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Carbaryl	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Thidiazuron	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Carbofuran	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Methiocarb	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
FS	Imidacloprid	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FS	Triadimefon	L306	0.05	Finished	<	0.05		ug/L	---	---	---	---	1.0	---	05/21/2019 20:27	4291764
CCC	IS-Linuron-d6	L306	N/A	---		19851	22048	ug/L	90	50 - 150	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Diuron	L306	0.05	---		1.0293	1.0	ug/L	103	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Diflufenzuron	L306	0.05	---		0.9389	1.0	ug/L	94	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Methomyl	L306	0.05	---		0.9598	1.0	ug/L	96	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Siduron, Total	L306	0.05	---		0.9672	1.0	ug/L	97	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Neburon	L306	0.05	---		0.9591	1.0	ug/L	96	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	IS-Diuron-d6	L306	N/A	---		378110	360194	ug/L	105	50 - 150	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Tebuthiuron	L306	0.05	---		1.0351	1.0	ug/L	104	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Propanil	L306	0.05	---		1.0807	1.0	ug/L	108	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Bensulfide	L306	0.05	---		0.9513	1.0	ug/L	95	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Bendiocarb	L306	0.05	---		0.9805	1.0	ug/L	98	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Oxamyl	L306	0.5	---		9.6676	10.0	ug/L	97	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	3-Hydroxycarbofuran	L306	0.05	---		0.9540	1.0	ug/L	95	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Fenuron	L306	0.05	---		0.9971	1.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Aldicarb sulfone	L306	0.5	---		9.5358	10.0	ug/L	95	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Baygon	L306	0.05	---		0.9835	1.0	ug/L	98	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Halofenozide	L306	0.05	---		1.0252	1.0	ug/L	103	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Paclobutrazol	L306	0.05	---		0.9987	1.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Metaxyl/Metaxyl-M (Mefenoxam)	L306	0.05	---		0.9793	1.0	ug/L	98	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Oryzalin	L306	0.5	---		10.0370	10.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Linuron	L306	0.5	---		10.3307	10.0	ug/L	103	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Propargite	L306	0.05	---		1.0438	1.0	ug/L	104	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Oxadiazon	L306	0.05	---		1.0143	1.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Azoxystrobin	L306	0.5	---		9.5120	10.0	ug/L	95	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Triadimenol	L306	0.05	---		0.9917	1.0	ug/L	99	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Fluometuron	L306	0.05	---		1.0011	1.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Aldicarb sulfoxide	L306	0.5	---		8.9511	10.0	ug/L	90	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Clomazone	L306	0.05	---		1.0257	1.0	ug/L	103	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Monuron	L306	0.05	---		1.0076	1.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Halosulfuron methyl	L306	0.05	---		0.9284	1.0	ug/L	93	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Carbaryl	L306	0.05	---		0.9765	1.0	ug/L	98	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Thidiazuron	L306	0.05	---		0.9961	1.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Carbofuran	L306	0.05	---		1.0140	1.0	ug/L	101	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Methiocarb	L306	0.05	---		1.0008	1.0	ug/L	100	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Imidacloprid	L306	0.05	---		0.9694	1.0	ug/L	97	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888
CCC	Triadimefon	L306	0.05	---		0.9553	1.0	ug/L	96	70 - 130	---	---	1.0	---	05/21/2019 22:04	4301888

Eurofins Eaton Analytical

Run Log

Run ID: 260135 Method: L402

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
CCL	4310221		OS	FL	05/30/2019 22:16	3019L402a2-PFCQuant-NTA.m
LRB	4310196		RW	FL	05/30/2019 22:42	3019L402a2-PFCQuant-NTA.m
FBL	4310197		RW	FL	05/30/2019 23:08	3019L402a2-PFCQuant-NTA.m
FS	4291769	Finished	DW	FL	05/31/2019 01:20	3019L402a2-PFCQuant-NTA.m
CCM	4310222		OS	FL	05/31/2019 05:42	3019L402a2-PFCQuant-NTA.m

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCL	IS-4:2-FTS-13C2	L402	N/A	--		954540	994516	ng/L	96	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	IS-6:2-FTS-13C2	L402	N/A	--		262280	229135	ng/L	114	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	IS-8:2-FTS-13C2	L402	N/A	--		767884	673757	ng/L	114	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	IS-GenX-13C3	L402	N/A	--		33444	31003	ng/L	108	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	IS-NEFOSAA-d5	L402	N/A	--		786483	621684	ng/L	127	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	IS-NEIFOSE-d9	L402	N/A	--		129396	143081	ng/L	90	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	IS-NMeFOSAA-d3	L402	N/A	--		957568	906965	ng/L	106	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	IS-NMeFOSA-d3	L402	N/A	--		796826	893002	ng/L	89	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	IS-NMeFOSE-d7	L402	N/A	--		148400	149538	ng/L	99	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	IS-PFBA-13C4	L402	N/A	--		3556510	3380686	ng/L	105	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	IS-PFBS-13C3	L402	N/A	--		1128320	1055152	ng/L	107	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	IS-PFDA-13C2	L402	N/A	--		6206280	5022945	ng/L	124	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	IS-PFDoA-13C2	L402	N/A	--		3806160	3566740	ng/L	107	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	IS-PFHpa-13C4	L402	N/A	--		6590930	5972398	ng/L	110	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	IS-PFHXA-13C2	L402	N/A	--		5659140	5347586	ng/L	106	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	IS-PFNA-13C9	L402	N/A	--		2419550	2101506	ng/L	115	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	IS-PFOA-13C2	L402	N/A	--		7899940	6838625	ng/L	116	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	IS-PFOS-13C4	L402	N/A	--		854173	799486	ng/L	107	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	IS-PFPeA-13C5	L402	N/A	--		3443910	3340990	ng/L	103	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	IS-PFTeDA-13C2	L402	N/A	--		926949	1326521	ng/L	70	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	IS-PFUaA-13C7	L402	N/A	--		4802700	3866948	ng/L	124	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	10:2 Fluorotelomer sulfonic acid (10:2 FTS)	L402	2.0	--		1.5938	2.0	ng/L	80	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	4:2 Fluorotelomer sulfonic acid (4:2 FTS)	L402	2.0	--		2.1600	2.0	ng/L	108	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	6:2 Fluorotelomer sulfonic acid (6:2 FTS)	L402	2.0	--		1.8630	2.0	ng/L	93	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	8:2 Fluorotelomer sulfonic acid (8:2 FTS)	L402	2.0	--		2.1260	2.0	ng/L	106	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	ADONA	L402	2.0	--		2.1444	2.0	ng/L	107	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	9Cl-PF3ONS/F-53B Major	L402	2.0	--		1.7918	2.0	ng/L	90	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	11Cl-PF3OUs/F-53B Minor	L402	2.0	--		1.9013	2.0	ng/L	95	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	HFO-DA/GenX	L402	5.0	--		4.4626	5.0	ng/L	89	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	N-ethylperfluorooctane sulfonamide (NEFOSA)	L402	2.0	--		1.7923	2.0	ng/L	90	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	N-ethylperfluorooctane sulfonamideethanol	L402	2.0	--		1.9384	2.0	ng/L	97	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	N-methylperfluorooctane sulfonamide (NMeFOSA)	L402	2.0	--		1.8061	2.0	ng/L	90	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	N-methylperfluorooctane sulfonamideethanol	L402	2.0	--		1.7629	2.0	ng/L	88	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	Perfluorobutanesulfonic acid (PFBS)	L402	2.0	--		1.7626	2.0	ng/L	88	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	Perfluorobutanoic acid (PFBA)	L402	5.0	--		4.3606	5.0	ng/L	87	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	Perfluorodecanoic acid (PFDA)	L402	2.0	--		1.7825	2.0	ng/L	89	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	Perfluoroheptanoic acid (PFHpA)	L402	2.0	--		1.7983	2.0	ng/L	90	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCL	Perfluorohexanesulfonic acid (PFHxS)	L402	2.0	---		1.7689	2.0	ng/L	88	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	Perfluorohexanoic acid (PFHxA)	L402	2.0	---		1.8425	2.0	ng/L	92	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	Perfluorododecanoic acid (PFDoA)	L402	2.0	---		1.8303	2.0	ng/L	92	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	Perfluorononanoic acid (PFNA)	L402	2.0	---		1.8634	2.0	ng/L	93	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	Perfluorooctanesulfonic acid (PFOS)	L402	2.0	---		1.8492	2.0	ng/L	92	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	N-ethyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	---		1.9029	2.0	ng/L	95	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	N-methyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	---		1.8770	2.0	ng/L	94	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	Perfluorooctanoic acid (PFOA)	L402	2.0	---		1.8729	2.0	ng/L	94	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	Perfluorotridecanoic acid (PFTDA)	L402	2.0	---		2.2269	2.0	ng/L	111	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	Perfluoroundecanoic acid (PFUnA)	L402	2.0	---		1.8546	2.0	ng/L	93	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	Perfluorododecanesulfonic acid (PFDoS)	L402	2.0	---		2.2651	2.0	ng/L	113	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	Perfluorododecanesulfonic acid (PFDS)	L402	2.0	---		1.4414	2.0	ng/L	72	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	Perfluoroheptanesulfonic acid (PFHpS)	L402	2.0	---		1.7697	2.0	ng/L	88	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	Perfluorohexadecanoic acid (PFHxDA)	L402	2.0	---		2.4624	2.0	ng/L	123	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	Perfluoro-2-methoxyethoxyacetic acid	L402	5.0	---		5.4171	5.0	ng/L	108	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	Perfluoro-4-isopropoxybutanoic acid	L402	5.0	---		5.7127	5.0	ng/L	114	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	Perfluoro-4-methoxybutanoic acid (PFMOBA)	L402	5.0	---		4.9882	5.0	ng/L	100	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	Perfluorononanesulfonic acid (PFNS)	L402	2.0	---		1.7593	2.0	ng/L	88	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	Perfluorooctane sulfonamide (PFOSA)	L402	2.0	---		1.8277	2.0	ng/L	91	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	Perfluoropentanoic acid (PFPeA)	L402	2.0	---		1.8537	2.0	ng/L	93	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	Perfluoropentanesulfonic acid (PFPeS)	L402	2.0	---		1.6748	2.0	ng/L	84	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
CCL	Perfluorotetradecanoic acid (PFTeDA)	L402	2.0	---		2.0004	2.0	ng/L	100	50 - 150	---	---	1.0	05/30/2019 13:19	05/30/2019 22:16	4310221
LRB	10:2 Fluorotelomer sulfonic acid (10:2 FTS)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	4:2 Fluorotelomer sulfonic acid (4:2 FTS)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	6:2 Fluorotelomer sulfonic acid (6:2 FTS)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	8:2 Fluorotelomer sulfonic acid (8:2 FTS)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	ADONA	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	9Cl-PF3ONS/F-53B Major	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	11Cl-PF3OUdS/F-53B Minor	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	N-ethylperfluorooctane sulfonamide (NEFOSA)	L402	5.0	---	<	5.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	N-ethylperfluorooctane sulfonamidoethanol	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	N-methylperfluorooctane sulfonamide (MMeFOSA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	N-methylperfluorooctane sulfonamidoethanol	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	Perfluorobutanesulfonic acid (PFBS)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	Perfluorobutanoic acid (PFBA)	L402	5.0	---	<	5.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	Perfluorodecanoic acid (PFDA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	Perfluoroheptanoic acid (PFHpA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	Perfluorohexanesulfonic acid (PFHxS)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	Perfluorohexanoic acid (PFHxA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	Perfluorododecanoic acid (PFDoA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LRB	Perfluorononanoic acid (PFNA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	Perfluorooctanesulfonic acid (PFOS)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	N-ethyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	N-methyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	Perfluorododecanoic acid (PFDA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	Perfluorotridecanoic acid (PFTDA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	Perfluoroundecanoic acid (PFUNA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	Perfluorododecanesulfonic acid (PFDoS)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	Perfluorodecanesulfonic acid (PFDS)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	Perfluoroheptanesulfonic acid (PFHpS)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	Perfluorohexadecanoic acid (PFHxDA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	Perfluoro-2-methoxyethoxyacetic acid	L402	5.0	---	<	5.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	Perfluoro-4-isopropoxybutanoic acid	L402	5.0	---	<	5.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	Perfluoro-4-methoxybutanoic acid (PFMOBA)	L402	5.0	---	<	5.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	Perfluorononanesulfonic acid (PFNS)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	Perfluorooctane sulfonamide (PFOSA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	Perfluoropentanoic acid (PFPeA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	Perfluoropentanesulfonic acid (PFPeS)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	Perfluorotetradecanoic acid (PFTeDA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
LRB	10:2 Fluorotelomer sulfonic acid (10:2 FTS)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.87	05/30/2019 08:30	05/30/2019 22:42	4310196
FBL	4:2 Fluorotelomer sulfonic acid (4:2 FTS)	L402	2.0	---	<	1.4437	2.0	ng/L	72	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	6:2 Fluorotelomer sulfonic acid (6:2 FTS)	L402	2.0	---	<	2.0657	2.0	ng/L	103	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	8:2 Fluorotelomer sulfonic acid (8:2 FTS)	L402	2.0	---	<	2.0097	2.0	ng/L	100	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	ADONA	L402	2.0	---	<	2.2143	2.0	ng/L	111	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	9CI-PF3ONSIF-53B Major	L402	2.0	---	<	2.1919	2.0	ng/L	110	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	11CI-PF3OUdS/F-53B Minor	L402	2.0	---	<	1.8058	2.0	ng/L	90	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	HFPO-DA/GenX	L402	5.0	---	<	1.9994	2.0	ng/L	100	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	N-ethylperfluorooctane sulfonamide (NEFOSA)	L402	2.0	---	<	4.6055	5.0	ng/L	92	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	N-ethylperfluorooctane sulfonamidoethanol	L402	2.0	---	<	1.9548	2.0	ng/L	98	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	N-methylperfluorooctane sulfonamide (NMeFOSA)	L402	2.0	---	<	1.8410	2.0	ng/L	92	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	N-methylperfluorooctane sulfonamidoethanol	L402	2.0	---	<	1.9061	2.0	ng/L	95	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	Perfluorobutanesulfonic acid (PFBS)	L402	2.0	---	<	1.9849	2.0	ng/L	99	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	Perfluorobutanoic acid (PFBA)	L402	5.0	---	<	1.7515	2.0	ng/L	88	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	Perfluorodecanoic acid (PFDA)	L402	2.0	---	<	4.7924	5.0	ng/L	96	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	Perfluoroheptanoic acid (PFHpA)	L402	2.0	---	<	1.8278	2.0	ng/L	91	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	Perfluorohexanesulfonic acid (PFHxS)	L402	2.0	---	<	1.8040	2.0	ng/L	90	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	Perfluorohexanoic acid (PFHxA)	L402	2.0	---	<	1.7567	2.0	ng/L	88	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	Perfluorododecanoic acid (PFDoA)	L402	2.0	---	<	1.8798	2.0	ng/L	94	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	Perfluorononanoic acid (PFNA)	L402	2.0	---	<	1.8268	2.0	ng/L	91	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	Perfluorooctanesulfonic acid (PFOS)	L402	2.0	---	<	1.8559	2.0	ng/L	93	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	N-ethyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	---	<	1.9354	2.0	ng/L	97	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL		L402	2.0	---	<	1.9769	2.0	ng/L	99	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FBL	N-methyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	---		1.8060	2.0	ng/L	90	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	Perfluorooctanoic acid (PFOA)	L402	2.0	---		1.9508	2.0	ng/L	98	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	Perfluorotridecanoic acid (PFTDA)	L402	2.0	---		1.6870	2.0	ng/L	84	40 - 160	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	Perfluoroundecanoic acid (PFUnA)	L402	2.0	---		1.9400	2.0	ng/L	97	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	Perfluorododecanesulfonic acid (PFDoS)	L402	2.0	---		3.8974	2.0	ng/L	195	40 - 160	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	Perfluorodecanesulfonic acid (PFDS)	L402	2.0	---		1.1116	2.0	ng/L	56	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	Perfluoroheptanesulfonic acid (PFHps)	L402	2.0	---		1.7010	2.0	ng/L	85	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	Perfluorohexadecanoic acid (PFHxDA)	L402	2.0	---		2.2194	2.0	ng/L	111	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	Perfluoro-2-methoxyethoxyacetic acid	L402	5.0	---		5.8730	5.0	ng/L	117	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	Perfluoro-4-isopropoxybutanoic acid	L402	5.0	---		5.5349	5.0	ng/L	111	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	Perfluoro-4-methoxybutanoic acid (PFMOBA)	L402	5.0	---		4.3699	5.0	ng/L	87	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	Perfluoronanesulfonic acid (PFNS)	L402	2.0	---		1.7372	2.0	ng/L	87	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	Perfluorooctane sulfonamide (PFOSA)	L402	2.0	---		1.8959	2.0	ng/L	95	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	Perfluoropentanoic acid (PFPeA)	L402	2.0	---		1.8281	2.0	ng/L	91	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	Perfluoropentanesulfonic acid (PFPeS)	L402	2.0	---		1.5967	2.0	ng/L	80	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FBL	Perfluorotetradecanoic acid (PFTeDA)	L402	2.0	---		1.7860	2.0	ng/L	89	50 - 150	---	---	1.0	05/30/2019 08:30	05/30/2019 23:08	4310197
FS	10:2 Fluorotelomer sulfonic acid (10:2 FTS)	L402	2.0	Finished	<	2.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	4:2 Fluorotelomer sulfonic acid (4:2 FTS)	L402	2.0	Finished	<	2.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	6:2 Fluorotelomer sulfonic acid (6:2 FTS)	L402	2.0	Finished	<	2.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	8:2 Fluorotelomer sulfonic acid (8:2 FTS)	L402	2.0	Finished	<	2.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	ADONA	L402	2.0	Finished	<	2.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	9CI-PF3ONS/F-53B Major	L402	2.0	Finished	<	2.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	11CI-PF3OUdS/F-53B Minor	L402	2.0	Finished	<	2.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	HFPO-DA/GenX	L402	5.0	Finished	<	5.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	N-ethylperfluorooctane sulfonamide (NEFOSA)	L402	2.0	Finished	<	2.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	N-ethylperfluorooctane sulfonamidoethanol	L402	2.0	Finished	<	2.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	N-methylperfluorooctane sulfonamide (NMeFOSA)	L402	2.0	Finished	<	2.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	N-methylperfluorooctane sulfonamidoethanol	L402	2.0	Finished	<	2.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	Perfluorobutanesulfonic acid (PFBS)	L402	2.0	Finished	<	3.5		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	Perfluorobutanoic acid (PFBA)	L402	5.0	Finished	<	5.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	Perfluorodecanoic acid (PFDA)	L402	2.0	Finished	<	2.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	Perfluoroheptanoic acid (PFHpA)	L402	2.0	Finished	<	2.8		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	Perfluorohexanesulfonic acid (PFHxS)	L402	2.0	Finished	<	2.2		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	Perfluorohexanoic acid (PFHxA)	L402	2.0	Finished	<	5.6		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	Perfluorododecanoic acid (PFDoA)	L402	2.0	Finished	<	2.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	Perfluorononanoic acid (PFNA)	L402	2.0	Finished	<	2.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	Perfluorooctanesulfonic acid (PFOS)	L402	2.0	Finished	<	7.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	N-ethyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	Finished	<	2.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	N-methyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	Finished	<	2.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	Perfluorooctanoic acid (PFOA)	L402	2.0	Finished	<	5.3		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	Perfluorotridecanoic acid (PFTDA)	L402	2.0	Finished	<	2.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FS	Perfluoroundecanoic acid (PFUnA)	L402	2.0	Finished	<	2.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	Perfluorododecanesulfonic acid (PFDoS)	L402	2.0	Finished	<	2.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	Perfluorododecanesulfonic acid (PFDS)	L402	2.0	Finished	<	2.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	Perfluoroheptanesulfonic acid (PFHpS)	L402	2.0	Finished	<	2.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	Perfluorohexadecanoic acid (PFHxDA)	L402	2.0	Finished	<	2.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	Perfluoro-2-methoxyethoxyacetic acid	L402	5.0	Finished	<	5.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	Perfluoro-4-isopropoxybutanoic acid	L402	5.0	Finished	<	5.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	Perfluoro-4-methoxybutanoic acid (PFMOBA)	L402	5.0	Finished	<	5.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	Perfluorononanesulfonic acid (PFNS)	L402	2.0	Finished	<	2.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	Perfluorooctane sulfonamide (PFOSA)	L402	2.0	Finished	<	2.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	Perfluoropentanoic acid (PFPeA)	L402	2.0	Finished	<	5.8		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	Perfluoropentanesulfonic acid (PFPeS)	L402	2.0	Finished	<	2.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
FS	Perfluorotetradecanoic acid (PFTeDA)	L402	2.0	Finished	<	2.0		ng/L	---	---	---	---	0.9	05/30/2019 08:30	05/31/2019 01:20	4291769
CCM	IS-4-2-FTS-13C2	L402	N/A	---		500841	994516	ng/L	50	50 - 150	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	IS-6-2-FTS-13C2	L402	N/A	---		139064	229135	ng/L	61	50 - 150	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	IS-8-2-FTS-13C2	L402	N/A	---		456189	673757	ng/L	68	50 - 150	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	IS-GenX-13C3	L402	N/A	---		23926	31003	ng/L	77	50 - 150	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	IS-NEIFOSAA-d5	L402	N/A	---		546133	621684	ng/L	88	50 - 150	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	IS-NEIFOSE-d9	L402	N/A	---		127071	143081	ng/L	89	50 - 150	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	IS-NMeFOSAA-d3	L402	N/A	---		734395	906965	ng/L	81	50 - 150	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	IS-NMeFOSA-d3	L402	N/A	---		822746	893002	ng/L	92	50 - 150	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	IS-NMeFOSE-d7	L402	N/A	---		132270	149538	ng/L	88	50 - 150	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	IS-PFBA-13C4	L402	N/A	---		1735310	3380686	ng/L	51	50 - 150	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	IS-PFBS-13C3	L402	N/A	---		726064	1055152	ng/L	69	50 - 150	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	IS-PFDA-13C2	L402	N/A	---		4072510	5022945	ng/L	81	50 - 150	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	IS-PFDoA-13C2	L402	N/A	---		3242120	3566740	ng/L	91	50 - 150	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	IS-PFHpa-13C4	L402	N/A	---		4534000	5972398	ng/L	76	50 - 150	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	IS-PFHHA-13C2	L402	N/A	---		3586870	5347586	ng/L	67	50 - 150	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	IS-PFOA-13C8	L402	N/A	---		1826410	2101506	ng/L	87	50 - 150	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	IS-NEIFOSA-d5	L402	N/A	---		731902	767745	ng/L	95	50 - 150	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	IS-PFHXS-13C3	L402	N/A	---		699769	814946	ng/L	86	50 - 150	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	IS-PFNA-13C9	L402	N/A	---		6051470	7197693	ng/L	84	50 - 150	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	IS-PFOA-13C2	L402	N/A	---		5408980	6838625	ng/L	79	50 - 150	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	IS-PFOS-13C4	L402	N/A	---		722334	799486	ng/L	90	50 - 150	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	IS-PFPeA-13C5	L402	N/A	---		1991550	3340990	ng/L	60	50 - 150	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	IS-PFTeDA-13C2	L402	N/A	---		1409250	1326521	ng/L	106	50 - 150	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	IS-PFUnA-13C7	L402	N/A	---		3297380	3866948	ng/L	85	50 - 150	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	10:2 Fluorotelomer sulfonic acid (10:2 FTS)	L402	2.0	---		80.5275	100	ng/L	81	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	4:2 Fluorotelomer sulfonic acid (4:2 FTS)	L402	2.0	---		104.3490	100	ng/L	104	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	6:2 Fluorotelomer sulfonic acid (6:2 FTS)	L402	2.0	---		96.6068	100	ng/L	97	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	8:2 Fluorotelomer sulfonic acid (8:2 FTS)	L402	2.0	---		104.0960	100	ng/L	104	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCM	ADONA	L402	2.0	---		111.4940	100	ng/L	111	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	9CI-PF3ONS/F-53B Major	L402	2.0	---		102.6510	100	ng/L	103	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	11CI-PF3OUDS/F-53B Minor	L402	2.0	---		100.1140	100	ng/L	100	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	HFPO-DA/GenX	L402	5.0	---		246.7520	250	ng/L	99	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	N-ethylperfluorooctane sulfonamide (NEFOSA)	L402	2.0	---		97.4614	100	ng/L	97	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	N-ethylperfluorooctane sulfonamideethanol	L402	2.0	---		98.4860	100	ng/L	98	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	N-methylperfluorooctane sulfonamide (NMeFOSA)	L402	2.0	---		96.3009	100	ng/L	96	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	N-methylperfluorooctane sulfonamideethanol	L402	2.0	---		94.5296	100	ng/L	95	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	Perfluorobutanesulfonic acid (PFBS)	L402	2.0	---		97.2240	100	ng/L	97	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	Perfluorobutanoic acid (PFBA)	L402	5.0	---		234.8530	250	ng/L	94	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	Perfluorodecanoic acid (PFDA)	L402	2.0	---		98.2290	100	ng/L	98	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	Perfluoroheptanoic acid (PFHpA)	L402	2.0	---		94.2054	100	ng/L	94	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	Perfluorohexanesulfonic acid (PFHxS)	L402	2.0	---		93.7739	100	ng/L	94	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	Perfluorohexanoic acid (PFHxA)	L402	2.0	---		96.4471	100	ng/L	96	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	Perfluorododecanoic acid (PFDoA)	L402	2.0	---		96.9974	100	ng/L	97	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	Perfluorononanoic acid (PFNA)	L402	2.0	---		96.2101	100	ng/L	96	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	Perfluorooctanesulfonic acid (PFOS)	L402	2.0	---		91.0806	100	ng/L	91	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	N-ethyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	---		96.4327	100	ng/L	96	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	N-methyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	---		98.4550	100	ng/L	98	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	Perfluorooctanoic acid (PFOA)	L402	2.0	---		96.1204	100	ng/L	96	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	Perfluorotridecanoic acid (PFTDA)	L402	2.0	---		99.8868	100	ng/L	100	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	Perfluoroundecanoic acid (PFUnA)	L402	2.0	---		95.6777	100	ng/L	96	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	Perfluorododecanesulfonic acid (PFDoS)	L402	2.0	---		110.2860	100	ng/L	110	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	Perfluorodecanesulfonic acid (PFDS)	L402	2.0	---		96.4406	100	ng/L	96	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	Perfluoroheptanesulfonic acid (PFHpS)	L402	2.0	---		93.7576	100	ng/L	94	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	Perfluorohexadecanoic acid (PFHxDA)	L402	2.0	---		96.2029	100	ng/L	96	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	Perfluoro-2-methoxyethoxyacetic acid	L402	5.0	---		271.2550	250	ng/L	109	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	Perfluoro-4-isopropoxybutanoic acid	L402	5.0	---		244.8320	250	ng/L	98	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	Perfluoro-4-methoxybutanoic acid (PFMOBA)	L402	5.0	---		237.3930	250	ng/L	95	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	Perfluoronanesulfonic acid (PFNS)	L402	2.0	---		105.4790	100	ng/L	105	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	Perfluorooctane sulfonamide (PFOSA)	L402	2.0	---		95.6055	100	ng/L	96	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	Perfluoropentanoic acid (PFPeA)	L402	2.0	---		95.4972	100	ng/L	95	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	Perfluoropentanesulfonic acid (PFPeS)	L402	2.0	---		106.6990	100	ng/L	107	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222
CCM	Perfluorotetradecanoic acid (PFTeDA)	L402	2.0	---		95.0099	100	ng/L	95	70 - 130	---	---	1.0	05/30/2019 13:19	05/31/2019 05:42	4310222

Eurofins Eaton Analytical

Run Log

Run ID: 259695 Method: L520

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
ICL	4298354		RW	CY	05/17/2019 16:40	051719L520a.mdb
ICS	4298361		RW	CY	05/17/2019 16:53	051719L520a.mdb
ICS	4298362		RW	CY	05/17/2019 17:06	051719L520a.mdb
ICS	4298363		RW	CY	05/17/2019 17:19	051719L520a.mdb
ICS	4298364		RW	CY	05/17/2019 17:33	051719L520a.mdb
ICS	4298365		RW	CY	05/17/2019 17:46	051719L520a.mdb
LMB	4298367		RW	CY	05/17/2019 18:12	051719L520a.mdb
FS	4291767	Finished	DW	CY	05/17/2019 18:25	051719L520a.mdb
CCC	4298368		RW	CY	05/17/2019 19:57	051719L520a.mdb

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
ICL	Acrylamide	L520	N/A	---		0.5029	0.5	ug/L	101	50 - 150	---	---	1.0	---	05/17/2019 16:40	4298364
ICL	Aniline	L520	N/A	---		0.4733	0.5	ug/L	95	50 - 150	---	---	1.0	---	05/17/2019 16:40	4298364
ICL	Urethane	L520	N/A	---		0.9784	1.0	ug/L	98	50 - 150	---	---	1.0	---	05/17/2019 16:40	4298364
ICS	Acrylamide	L520	N/A	---		0.9758	1.0	ug/L	98	70 - 130	---	---	1.0	---	05/17/2019 16:53	4298361
ICS	Aniline	L520	N/A	---		1.0249	1.0	ug/L	102	70 - 130	---	---	1.0	---	05/17/2019 16:53	4298361
ICS	Urethane	L520	N/A	---		2.0367	2.0	ug/L	102	70 - 130	---	---	1.0	---	05/17/2019 16:53	4298361
ICS	Acrylamide	L520	N/A	---		2.0088	2.0	ug/L	100	70 - 130	---	---	1.0	---	05/17/2019 17:06	4298362
ICS	Aniline	L520	N/A	---		2.0831	2.0	ug/L	104	70 - 130	---	---	1.0	---	05/17/2019 17:06	4298362
ICS	Urethane	L520	N/A	---		4.0155	4.0	ug/L	100	70 - 130	---	---	1.0	---	05/17/2019 17:06	4298362
ICS	Acrylamide	L520	N/A	---		5.1594	5.0	ug/L	103	70 - 130	---	---	1.0	---	05/17/2019 17:19	4298363
ICS	Aniline	L520	N/A	---		4.9747	5.0	ug/L	99	70 - 130	---	---	1.0	---	05/17/2019 17:19	4298363
ICS	Urethane	L520	N/A	---		10.0419	10.0	ug/L	100	70 - 130	---	---	1.0	---	05/17/2019 17:19	4298363
ICS	Acrylamide	L520	N/A	---		9.7866	10.0	ug/L	98	70 - 130	---	---	1.0	---	05/17/2019 17:33	4298364
ICS	Aniline	L520	N/A	---		9.8950	10.0	ug/L	99	70 - 130	---	---	1.0	---	05/17/2019 17:33	4298364
ICS	Urethane	L520	N/A	---		19.8809	20.0	ug/L	99	70 - 130	---	---	1.0	---	05/17/2019 17:33	4298364
ICS	Acrylamide	L520	N/A	---		20.0644	20.0	ug/L	100	70 - 130	---	---	1.0	---	05/17/2019 17:46	4298365
ICS	Aniline	L520	N/A	---		20.0490	20.0	ug/L	100	70 - 130	---	---	1.0	---	05/17/2019 17:46	4298365
ICS	Urethane	L520	N/A	---		40.0467	40.0	ug/L	100	70 - 130	---	---	1.0	---	05/17/2019 17:46	4298365
LMB	Acrylamide	L520	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/17/2019 18:12	4298367
LMB	Aniline	L520	0.5	---	<	0.5		ug/L	---	---	---	---	1.0	---	05/17/2019 18:12	4298367
LMB	Urethane	L520	1.0	---	<	1.0		ug/L	---	---	---	---	1.0	---	05/17/2019 18:12	4298367
FS	Acrylamide	L520	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/17/2019 18:25	4291767
FS	Aniline	L520	0.5	Finished	<	0.5		ug/L	---	---	---	---	1.0	---	05/17/2019 18:25	4291767
FS	Urethane	L520	1.0	Finished	<	1.0		ug/L	---	---	---	---	1.0	---	05/17/2019 18:25	4291767
CCC	Acrylamide	L520	0.5	---		4.9741	5.0	ug/L	99	70 - 130	---	---	1.0	---	05/17/2019 19:57	4298368
CCC	Aniline	L520	0.5	---		5.0631	5.0	ug/L	101	70 - 130	---	---	1.0	---	05/17/2019 19:57	4298368
CCC	Urethane	L520	1.0	---		9.7975	10.0	ug/L	98	70 - 130	---	---	1.0	---	05/17/2019 19:57	4298368



Eaton Analytical

Eurofins Eaton Analytical

Run Log

Run ID: 260435 Method: LC/MS/MS

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
FS	4291766	Finished	DW	R L	05/23/2019 21:00	Ref_Lab

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FS	Acetaminophen	LC/MS/MS	20	Finished	<	88		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Albuterol	LC/MS/MS	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Amoxicillin	LC/MS/MS	20	Finished	<	20		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	4-Androstene-3,17-dione	LC/MS/MS	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Antipyrine	LC/MS/MS	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Atenolol	LC/MS/MS	5.0	Finished		5.8		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Atrazine	LC/MS/MS	5.0	Finished		16		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Azithromycin	LC/MS/MS	20	Finished	<	20		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Bendroflumethiazide	LC/MS/MS	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Bezafibrate	LC/MS/MS	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Bisphenol A	LC/MS/MS	10	Finished	<	10		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Bromacil	LC/MS/MS	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Butalbital	LC/MS/MS	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Butylparaben	LC/MS/MS	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Caffeine	LC/MS/MS	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Carbadox	LC/MS/MS	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Carbamazepine	LC/MS/MS	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Carisoprodol	LC/MS/MS	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Chloramphenicol	LC/MS/MS	10	Finished	<	10		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Chloridazon	LC/MS/MS	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Chlorotoluron	LC/MS/MS	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Cimetidine	LC/MS/MS	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Clofibric acid	LC/MS/MS	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Cotinine	LC/MS/MS	10	Finished	<	10		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Cyanazine	LC/MS/MS	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	2,4-D	LC/MS/MS	5.0	Finished		8.7		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	DEET	LC/MS/MS	10	Finished	<	10		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Dehydronifedipine	LC/MS/MS	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Desethylatrazine	LC/MS/MS	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Desisopropylatrazine	LC/MS/MS	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Diaminochlorotriazine	LC/MS/MS	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Diazepam	LC/MS/MS	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Diclofenac	LC/MS/MS	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Dilantin	LC/MS/MS	20	Finished	<	20		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Diltiazem	LC/MS/MS	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	1,7-Dimethylxanthine	LC/MS/MS	10	Finished	<	10		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Diuron	LC/MS/MS	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	Erythromycin	LC/MS/MS	10	Finished	<	10		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766
FS	17beta-Estradiol	LC/MS/MS	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/23/2019 21:00	4291766

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FS	Estrone	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Ethylparaben	LC/MS/MS	20	Finished	<	20		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	17alpha-Ethinyl estradiol	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Flumequine	LC/MS/MS	10	Finished	<	10		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Fluoxetine (Prozac)	LC/MS/MS	10	Finished	<	10		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Gemfibrozil	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Ibuprofen	LC/MS/MS	10	Finished	<	10		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Iohexal	LC/MS/MS	10	Finished	<	51		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Iopromide	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Isobutylparaben	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Isoproturon	LC/MS/MS	100	Finished	<	100		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Ketoprofen	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Ketorolac	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Lidocaine	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Lincomycin	LC/MS/MS	10	Finished	<	10		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Linuron	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Lopressor	LC/MS/MS	20	Finished	<	20		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Meclofenamic acid	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Meprobamate	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Metazochlor	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Methylparaben	LC/MS/MS	20	Finished	<	20		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Metolachlor	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Naproxen	LC/MS/MS	10	Finished	<	10		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Nifedipine	LC/MS/MS	20	Finished	<	20		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	4-Nonylphenol	LC/MS/MS	100	Finished	<	100		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Norethisterone	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	4-tert-Octylphenol	LC/MS/MS	50	Finished	<	50		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Oxolinic acid	LC/MS/MS	10	Finished	<	10		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Pentoxifylline	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Primidone	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Progesterone	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Propazine	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Propylparaben	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Quinoline	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Simazine	LC/MS/MS	5.0	Finished	<	16		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Sucralose	LC/MS/MS	100	Finished	<	580		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Sulfachloropyridazine	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Sulfadiazine	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Sulfadimethoxine	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Sulfamerazine	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Sulfamethazine	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FS	Sulfamethizole	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Sulfamethoxazole	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Sulfathiazole	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Testosterone	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Theobromine	LC/MS/MS	10	Finished	<	10		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Theophylline	LC/MS/MS	20	Finished	<	20		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Thiabendazole	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Triclocarban	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Triclosan	LC/MS/MS	10	Finished	<	10		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Trimethoprim	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Tris(2-carboxyethyl)phosphine hydrochloride	LC/MS/MS	10	Finished	<	10		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Tris(1,3-dichloro-2-propyl) phosphate	LC/MS/MS	100	Finished	<	100		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766
FS	Warfarin	LC/MS/MS	5.0	Finished	<	5.0		ng/L	***	***	***	***	1.0	***	05/23/2019 21:00	4291766

Eurofins Eaton Analytical

Run Log

Run ID: 259787 Method: SRL 524M-TCP

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
LFB	4306814		RW	ET	05/17/2019 15:09	CDHS-012819a-ET.m
LRB	4306816		RW	ET	05/17/2019 16:16	CDHS-012819a-ET.m
FS	4291762	Finished	DW	ET	05/17/2019 17:03	CDHS-012819a-ET.m
CCC	4297840		RW	ET	05/17/2019 21:17	CDHS-012819a-ET.m

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LFB	1,2,3-Trichloropropane	SRL 524M-TCP	5.0	---		10.7920	10.0	ng/L	108	80 - 120	---	---	1.0	---	05/17/2019 15:09	4306814
LRB	1,2,3-Trichloropropane	SRL 524M-TCP	5.0	---	<	5.0		ng/L	---	---	---	---	1.0	---	05/17/2019 16:16	4306816
FS	1,2,3-Trichloropropane	SRL 524M-TCP	5.0	Finished	<	5.0		ng/L	---	---	---	---	1.0	---	05/17/2019 17:03	4291762
CCC	1,2,3-Trichloropropane	SRL 524M-TCP	5.0	---		54.0460	50.0	ng/L	108	80 - 120	---	---	1.0	---	05/17/2019 21:17	4297840

Sample Type Key

<u>Type (Abbr.)</u>	<u>Sample Type</u>	<u>Type (Abbr.)</u>	<u>Sample Type</u>
CCV	Continuing Cali. Verification	SDL	Spike Duplicate Low
CCB	Continuing Calibration Blank		
CCC	Continuing Calibration Check		
CCH	Continuing Calibration High		
CCL	Continuing Calibration Low		
CCM	Continuing Calibration Mid		
FS	Field Sample		
FBH	Fortified Blank High		
FBL	Fortified Blank Low		
FBM	Fortified Blank Mid		
ICV	Initial Cali. Verification		
ICB	Initial Calibration Blank		
ICL	Initial Calibration Low		
ICS	Initial Calibration Sample		
IPC	Instrument Performance Check		
LFB	Laboratory Fortified Blank		
LMB	Laboratory Method Blank		
LRB	Laboratory Reagent Blank		
LTB	Laboratory Trip Blank		
MS	Matrix Spike		
MSD	Matrix Spike Duplicate		
MSL	Matrix Spike Low		
QCS	Quality Control Sample		

END OF REPORT